


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG080318\
 Data File : BG036088.D
 Acq On : 3 Aug 2018 11:05
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 31 12:30:06 2018



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.8	15657	PASS
68	69	0.00	2	1.5	288	PASS
69	198	0.00	100	30.7	19366	PASS
70	69	0.00	2	0.6	116	PASS
127	198	10	80	36.7	23213	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	63176	PASS
199	198	5	9	6.8	4295	PASS
275	198	10	60	27.0	17055	PASS
365	198	1	100	3.5	2199	PASS
441	198	0.01	100	9.3	5854	PASS
442	198	50	100	59.5	37570	PASS
443	442	15	24	19.3	7246	PASS