

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080319\
 Data File : BG041946.D
 Acq On : 4 Aug 2019 17:36
 Operator : HP/JU
 Sample : K3962-15
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BEP08

Quant Time: Aug 05 09:16:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 05 05:06:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	59346	20.00	ng/ul	0.00
18) Naphthalene-d8	10.86	136	242419	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.69	164	162157	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.45	188	344363	20.00	ng/ul	0.00
77) Chrysene-d12	21.73	240	338450	20.00	ng/ul	0.00
85) Perylene-d12	25.01	264	355090	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.42	96	5606	4.13	ng/uL	0.00
5) Phenol-d5	7.18	99	138699	29.78	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.35	67	80989	31.09	ng/ul	0.00
9) 2-Chlorophenol-d4	7.57	132	127701	32.07	ng/ul	0.00
13) 4-Methylphenol-d8	8.74	113	103778	26.54	ng/ul	0.00
19) Nitrobenzene-d5	9.20	128	65003	35.62	ng/ul	0.00
22) 2-Nitrophenol-d4	9.93	143	81085	38.27	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.47	165	148425	34.70	ng/ul	0.00
29) 4-Chloroaniline-d4	10.99	131	146558	30.88	ng/ul	0.00
43) Dimethylphthalate-d6	14.09	166	443821	35.29	ng/ul	0.00
46) Acenaphthylene-d8	14.39	160	508829	35.42	ng/ul	0.00
51) 4-Nitrophenol-d4	14.88	143	58387	27.65	ng/ul	0.00
57) Fluorene-d10	15.69	176	398493	35.62	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.80	200	59134	26.76	ng/ul	0.00
70) Anthracene-d10	17.55	188	618335	37.43	ng/ul	0.00
78) Pyrene-d10	19.83	212	689639	36.51	ng/ul	0.00
89) Benzo(a)pyrene-d12	24.79	264	736206	39.69	ng/ul	0.00

Target Compounds				Ovalue		
6) Phenol	7.21	94	7603	1.582	ng/ul	94
44) Dimethylphthalate	14.14	163	197549	15.820	ng/ul	98
76) Fluoranthene	19.50	202	26744	1.228	ng/ul	99
79) Pyrene	19.86	202	28556	1.254	ng/ul#	95
83) Bis(2-ethylhexyl)phthalate	21.63	149	44044	3.602	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080319\
Data File : BG041946.D
Acq On : 4 Aug 2019 17:36
Operator : HP/JU
Sample : K3962-15
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BEP08

Quant Time: Aug 05 09:16:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 05 05:06:29 2019
Response via : Initial Calibration

