

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080319\
 Data File : BG041933.D
 Acq On : 4 Aug 2019 8:39
 Operator : HP/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02080

Quant Time: Aug 04 09:48:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 04 03:32:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.474	0.444	6.3	96	0.00
3 S	1,4-Dioxane-d8	0.457	0.420	8.1	96	0.00
4	Benzaldehyde	0.742	0.744	-0.3	106	0.00
5 S	Phenol-d5	1.570	1.658	-5.6	115	0.00
6	Phenol	1.620	1.714	-5.8	114	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.878	0.889	-1.3	108	0.00
8	Bis(2-Chloroethyl)ether	1.233	1.244	-0.9	107	0.00
9 S	2-Chlorophenol-d4	1.342	1.422	-6.0	113	0.00
10	2-Chlorophenol	1.360	1.440	-5.9	113	0.00
11	2-Methylphenol	1.299	1.356	-4.4	114	0.00
12	2,2'-oxybis(1-Chloropropane	2.085	2.033	2.5	102	0.00
13 S	4-Methylphenol-d8	1.318	1.342	-1.8	111	0.00
14	Acetophenone	2.027	2.118	-4.5	112	0.00
15 P	N-Nitroso-di-n-propylamine	1.071	1.026	4.2	102	0.00
16	4-Methylphenol	1.403	1.459	-4.0	113	0.00
17	Hexachloroethane	0.563	0.479	14.9	91	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
19 S	Nitrobenzene-d5	0.151	0.168	-11.3	115	0.00
20	Nitrobenzene	0.343	0.368	-7.3	113	0.00
21	Isophorone	0.684	0.644	5.8	98	0.00
22 S	2-Nitrophenol-d4	0.175	0.206	-17.7	123	0.00
23 C	2-Nitrophenol	0.187	0.216	-15.5	118	0.00
24	2,4-Dimethylphenol	0.368	0.395	-7.3	110	0.00
25	Bis(2-Chloroethoxy)methane	0.406	0.418	-3.0	105	0.00
26 S	2,4-Dichlorophenol-d3	0.353	0.406	-15.0	118	0.00
27 C	2,4-Dichlorophenol	0.350	0.388	-10.9	114	0.00
28	Naphthalene	1.022	1.103	-7.9	112	0.00
29 S	4-Chloroaniline-d4	0.392	0.427	-8.9	109	0.00
30	4-Chloroaniline	0.394	0.431	-9.4	110	0.00
31 C	Hexachlorobutadiene	0.266	0.277	-4.1	107	0.00
32	Caprolactam	0.113	0.115	-1.8	107	-0.01
33 C	4-Chloro-3-methylphenol	0.345	0.368	-6.7	111	0.00
34	2-Methylnaphthalene	0.813	0.880	-8.2	112	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
36	1,2,4,5-Tetrachlorobenzene	0.715	0.838	-17.2	116	0.00
37	Hexachlorocyclopentadiene	0.451	0.388	14.0	87	0.00
38 C	2,4,6-Trichlorophenol	0.443	0.518	-16.9	115	0.00
39	2,4,5-Trichlorophenol	0.494	0.563	-14.0	112	0.00
40	1,1'-Biphenyl	1.454	1.614	-11.0	108	0.00
41	2-Chloronaphthalene	1.171	1.317	-12.5	110	0.00
42	2-Nitroaniline	0.306	0.336	-9.8	107	0.00
43 S	Dimethylphthalate-d6	1.551	1.613	-4.0	102	0.00
44	Dimethylphthalate	1.540	1.587	-3.1	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080319\
 Data File : BG041933.D
 Acq On : 4 Aug 2019 8:39
 Operator : HP/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02080

Quant Time: Aug 04 09:48:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 04 03:32:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.335	0.381	-13.7	111	0.00
46 S	Acenaphthylene-d8	1.772	1.920	-8.4	105	0.00
47	Acenaphthylene	1.788	1.923	-7.6	104	0.00
48	3-Nitroaniline	0.267	0.308	-15.4	118	0.00
49 C	Acenaphthene	1.250	1.343	-7.4	105	0.00
50	2,4-Dinitrophenol	0.183	0.221	-20.8	143	0.00
51 S	4-Nitrophenol-d4	0.260	0.277	-6.5	107	0.00
52	4-Nitrophenol	0.193	0.206	-6.7	106	0.00
53	Dibenzofuran	1.866	2.051	-9.9	106	0.00
54	2,4-Dinitrotoluene	0.487	0.522	-7.2	103	0.00
55	2,3,4,6-Tetrachlorophenol	0.468	0.519	-10.9	109	0.00
56	Diethylphthalate	1.532	1.514	1.2	96	0.00
57 S	Fluorene-d10	1.380	1.487	-7.8	104	0.00
58	Fluorene	1.503	1.591	-5.9	103	0.00
59	4-Chlorophenyl-phenylether	0.873	0.960	-10.0	108	0.00
60	4-Nitroaniline	0.298	0.337	-13.1	117	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.128	0.128	0.0	100	0.00
63	4,6-Dinitro-2-methylphenol	0.138	0.140	-1.4	104	0.00
64	N-Nitrosodiphenylamine	0.593	0.658	-11.0	103	0.00
65	4-Bromophenyl-phenylether	0.265	0.301	-13.6	106	0.00
66	Hexachlorobenzene	0.266	0.285	-7.1	100	0.00
67	Atrazine	0.249	0.257	-3.2	94	0.00
68 C	Pentachlorophenol	0.145	0.177	-22.1	134	0.00
69	Phenanthrene	1.095	1.183	-8.0	98	0.00
70 S	Anthracene-d10	0.960	1.043	-8.6	99	0.00
71	Anthracene	1.110	1.204	-8.5	98	0.00
72	1,2,3,4-Tetrachlorobenzene	0.319	0.382	-19.7	114	0.00
73	Pentachlorobenzene	0.322	0.376	-16.8	110	0.00
74	Carbazole	0.920	1.042	-13.3	95	0.00
75	Di-n-butylphthalate	1.134	1.121	1.1	88	0.00
76 C	Fluoranthene	1.265	1.485	-17.4	92	0.00
77 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
78 S	Pyrene-d10	1.116	1.221	-9.4	92	0.00
79	Pyrene	1.345	1.475	-9.7	91	0.00
80	Butylbenzylphthalate	0.530	0.525	0.9	85	0.00
81	3,3'-Dichlorobenzidine	0.455	0.501	-10.1	93	0.00
82	Benzo(a)anthracene	1.387	1.514	-9.2	92	0.00
83	Bis(2-ethylhexyl)phthalate	0.722	0.692	4.2	81	0.00
84	Chrysene	1.294	1.417	-9.5	93	0.00
85 I	Perylene-d12	1.000	1.000	0.0	81	-0.02
86	Di-n-octyl phthalate	0.990	1.172	-18.4	85	0.00
87	Benzo(b)fluoranthene	1.255	1.393	-11.0	87	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080319\
Data File : BG041933.D
Acq On : 4 Aug 2019 8:39
Operator : HP/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02080

Quant Time: Aug 04 09:48:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Aug 04 03:32:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.207	1.447	-19.9	94	-0.01
89 S	Benzo(a)pyrene-d12	1.045	1.215	-16.3	92	-0.01
90 C	Benzo(a)pyrene	1.196	1.374	-14.9	90	0.00
91	Indeno(1,2,3-cd)pyrene	1.395	1.376	1.4	78	-0.03
92	Dibenzo(a,h)anthracene	1.138	1.185	-4.1	82	-0.03
93	Benzo(a,h,i)perylene	1.165	1.000	14.2	68	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0