

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080320\
 Data File : BG046149.D
 Acq On : 3 Aug 2020 11:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 12:43:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 18:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.524	0.526	-0.4	89	0.00
3	Pyridine	1.442	1.529	-6.0	92	0.00
4	n-Nitrosodimethylamine	0.614	0.652	-6.2	90	0.00
5 S	2-Fluorophenol	1.151	1.220	-6.0	92	0.00
6	Aniline	1.866	1.961	-5.1	93	0.00
7 S	Phenol-d6	1.568	1.676	-6.9	93	0.00
8	2-Chlorophenol	1.272	1.313	-3.2	92	0.00
9	Benzaldehyde	0.930	0.990	-6.5	92	0.00
10 C	Phenol	1.575	1.657	-5.2	93	0.00
11	bis(2-Chloroethyl)ether	1.253	1.303	-4.0	93	0.00
12	1,3-Dichlorobenzene	1.541	1.572	-2.0	90	0.00
13 C	1,4-Dichlorobenzene	1.548	1.591	-2.8	92	0.00
14	1,2-Dichlorobenzene	1.460	1.488	-1.9	91	0.00
15	Benzyl Alcohol	1.083	1.209	-11.6	97	0.00
16	2,2'-oxybis(1-Chloropropane	2.054	2.174	-5.8	95	0.00
17	2-Methylphenol	1.064	1.131	-6.3	93	0.00
18	Hexachloroethane	0.518	0.544	-5.0	94	0.00
19 P	n-Nitroso-di-n-propylamine	1.020	1.092	-7.1	95	0.00
20	3+4-Methylphenols	1.456	1.583	-8.7	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.513	0.530	-3.3	93	0.00
23 S	Nitrobenzene-d5	0.369	0.391	-6.0	95	0.00
24	Nitrobenzene	0.394	0.416	-5.6	95	0.00
25	Isophorone	0.735	0.789	-7.3	95	0.00
26 C	2-Nitrophenol	0.178	0.201	-12.9	96	0.00
27	2,4-Dimethylphenol	0.290	0.305	-5.2	93	0.00
28	bis(2-Chloroethoxy)methane	0.400	0.414	-3.5	93	0.00
29 C	2,4-Dichlorophenol	0.342	0.364	-6.4	93	0.00
30	1,2,4-Trichlorobenzene	0.400	0.406	-1.5	92	0.00
31	Naphthalene	1.058	1.097	-3.7	93	0.00
32	Benzoic acid	0.178	0.203	-14.0	99	0.00
33	4-Chloroaniline	0.432	0.459	-6.3	94	0.00
34 C	Hexachlorobutadiene	0.266	0.275	-3.4	93	0.00
35	Caprolactam	0.112	0.125	-11.6	96	0.00
36 C	4-Chloro-3-methylphenol	0.335	0.365	-9.0	95	0.00
37	2-Methylnaphthalene	0.814	0.848	-4.2	94	0.00
38	1-Methylnaphthalene	0.771	0.790	-2.5	92	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.721	0.746	-3.5	93	0.00
41 P	Hexachlorocyclopentadiene	0.413	0.437	-5.8	91	0.00
42 S	2,4,6-Tribromophenol	0.307	0.328	-6.8	93	0.00
43 C	2,4,6-Trichlorophenol	0.460	0.497	-8.0	92	0.00
44	2,4,5-Trichlorophenol	0.502	0.545	-8.6	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080320\
 Data File : BG046149.D
 Acq On : 3 Aug 2020 11:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 03 12:43:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 18:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.377	1.434	-4.1	94	0.00
46	1,1'-Biphenyl	1.569	1.627	-3.7	93	0.00
47	2-Chloronaphthalene	1.251	1.282	-2.5	92	0.00
48	2-Nitroaniline	0.328	0.367	-11.9	96	0.00
49	Acenaphthylene	1.943	2.028	-4.4	93	0.00
50	Dimethylphthalate	1.535	1.599	-4.2	92	0.00
51	2,6-Dinitrotoluene	0.357	0.383	-7.3	94	0.00
52 C	Acenaphthene	1.283	1.344	-4.8	93	0.00
53	3-Nitroaniline	0.354	0.381	-7.6	91	0.00
54 P	2,4-Dinitrophenol	0.187	0.207	-10.7	97	0.00
55	Dibenzofuran	1.935	1.996	-3.2	93	0.00
56 P	4-Nitrophenol	0.307	0.326	-6.2	92	0.00
57	2,4-Dinitrotoluene	0.491	0.536	-9.2	93	0.00
58	Fluorene	1.567	1.618	-3.3	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.472	0.504	-6.8	93	0.00
60	Diethylphthalate	1.510	1.576	-4.4	93	0.00
61	4-Chlorophenyl-phenylether	0.819	0.850	-3.8	93	0.00
62	4-Nitroaniline	0.356	0.386	-8.4	94	0.00
63	Azobenzene	1.415	1.493	-5.5	94	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.132	0.138	-4.5	94	0.00
66 c	n-Nitrosodiphenylamine	0.605	0.628	-3.8	92	0.00
67	4-Bromophenyl-phenylether	0.247	0.250	-1.2	91	0.00
68	Hexachlorobenzene	0.277	0.281	-1.4	91	0.00
69	Atrazine	0.233	0.243	-4.3	91	0.00
70 C	Pentachlorophenol	0.180	0.187	-3.9	93	0.00
71	Phenanthrene	1.096	1.125	-2.6	93	0.00
72	Anthracene	1.090	1.119	-2.7	92	0.00
73	Carbazole	1.051	1.099	-4.6	94	0.00
74	Di-n-butylphthalate	1.098	1.178	-7.3	94	0.00
75 C	Fluoranthene	1.368	1.421	-3.9	94	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
77	Benzidine	0.567	0.606	-6.9	86	0.00
78	Pyrene	1.344	1.378	-2.5	93	0.00
79 S	Terphenyl-d14	0.982	0.964	1.8	95	0.00
80	Butylbenzylphthalate	0.487	0.541	-11.1	97	0.00
81	Benzo(a)anthracene	1.338	1.364	-1.9	94	0.00
82	3,3'-Dichlorobenzidine	0.556	0.577	-3.8	91	0.00
83	Chrysene	1.311	1.335	-1.8	94	0.00
84	Bis(2-ethylhexyl)phthalate	0.696	0.740	-6.3	96	0.00
85 c	Di-n-octyl phthalate	1.178	1.289	-9.4	96	0.00
86 I	Perylene-d12	1.000	1.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	1.525	1.549	-1.6	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080320\
Data File : BG046149.D
Acq On : 3 Aug 2020 11:44
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 03 12:43:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG073020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 30 18:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.340	1.372	-2.4	93	0.00
89	Benzo(k)fluoranthene	1.315	1.349	-2.6	93	0.00
90 C	Benzo(a)pyrene	1.247	1.296	-3.9	93	0.00
91	Dibenzo(a,h)anthracene	1.275	1.302	-2.1	91	-0.01
92	Benzo(g,h,i)perylene	1.288	1.308	-1.6	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0