

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080321\
 Data File : BG049631.D
 Acq On : 3 Aug 2021 19:53
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 02:23:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:15:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	40.819	-2.0	101	0.00
3	Pyridine	40.000	41.091	-2.7	97	0.00
4	n-Nitrosodimethylamine	40.000	43.413	-8.5	105	0.00
5 S	2-Fluorophenol	80.000	81.280	-1.6	96	0.00
6	Aniline	40.000	41.117	-2.8	98	0.00
7 S	Phenol-d6	80.000	82.181	-2.7	97	0.00
8	2-Chlorophenol	40.000	40.759	-1.9	100	0.00
9	Benzaldehyde	40.000	38.691	3.3	95	0.00
10 C	Phenol	40.000	40.484	-1.2	95	0.00
11	bis(2-Chloroethyl)ether	40.000	41.175	-2.9	99	0.00
12	1,3-Dichlorobenzene	40.000	40.413	-1.0	100	0.00
13 C	1,4-Dichlorobenzene	40.000	40.956	-2.4	100	0.00
14	1,2-Dichlorobenzene	40.000	39.791	0.5	95	0.00
15	Benzyl Alcohol	40.000	41.359	-3.4	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.793	0.5	95	0.00
17	2-Methylphenol	40.000	42.212	-5.5	99	0.00
18	Hexachloroethane	40.000	40.004	-0.0	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.392	-1.0	95	0.00
20	3+4-Methylphenols	40.000	41.097	-2.7	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	40.897	-2.2	96	0.00
23 S	Nitrobenzene-d5	80.000	84.385	-5.5	98	0.00
24	Nitrobenzene	40.000	42.356	-5.9	98	0.00
25	Isophorone	40.000	42.491	-6.2	99	0.00
26 C	2-Nitrophenol	40.000	43.289	-8.2	99	0.00
27	2,4-Dimethylphenol	40.000	43.107	-7.8	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.035	-5.1	98	0.00
29 C	2,4-Dichlorophenol	40.000	42.444	-6.1	99	0.00
30	1,2,4-Trichlorobenzene	40.000	41.323	-3.3	98	0.00
31	Naphthalene	40.000	41.403	-3.5	98	0.00
32	Benzoic acid	40.000	40.539	-1.3	95	0.00
33	4-Chloroaniline	40.000	42.695	-6.7	98	0.00
34 C	Hexachlorobutadiene	40.000	41.244	-3.1	98	0.00
35	Caprolactam	40.000	44.410	-11.0	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.714	-9.3	101	0.00
37	2-Methylnaphthalene	40.000	42.206	-5.5	97	0.00
38	1-Methylnaphthalene	40.000	41.982	-5.0	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.183	-0.5	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.250	-0.6	98	0.00
42 S	2,4,6-Tribromophenol	80.000	85.077	-6.3	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.686	-6.7	100	0.00
44	2,4,5-Trichlorophenol	40.000	41.840	-4.6	98	0.00
45 S	2-Fluorobiphenyl	80.000	81.887	-2.4	98	0.00
46	1,1'-Biphenyl	40.000	41.538	-3.8	100	0.00
47	2-Chloronaphthalene	40.000	41.416	-3.5	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080321\
 Data File : BG049631.D
 Acq On : 3 Aug 2021 19:53
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 02:23:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:15:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.417	-6.0	100	0.00
49	Acenaphthylene	40.000	41.887	-4.7	101	0.00
50	Dimethylphthalate	40.000	41.005	-2.5	100	0.00
51	2,6-Dinitrotoluene	40.000	40.393	-1.0	94	0.00
52 C	Acenaphthene	40.000	41.579	-3.9	100	0.00
53	3-Nitroaniline	40.000	42.363	-5.9	101	0.00
54 P	2,4-Dinitrophenol	40.000	42.888	-7.2	102	0.00
55	Dibenzofuran	40.000	41.667	-4.2	99	0.00
56 P	4-Nitrophenol	40.000	43.448	-8.6	98	0.00
57	2,4-Dinitrotoluene	40.000	41.714	-4.3	102	0.00
58	Fluorene	40.000	41.378	-3.4	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.124	-7.8	105	0.00
60	Diethylphthalate	40.000	41.770	-4.4	102	0.00
61	4-Chlorophenyl-phenylether	40.000	41.049	-2.6	99	0.00
62	4-Nitroaniline	40.000	42.718	-6.8	100	0.00
63	Azobenzene	40.000	41.208	-3.0	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.559	-3.9	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.558	-1.4	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.619	-1.5	101	0.00
68	Hexachlorobenzene	40.000	40.565	-1.4	101	0.00
69	Atrazine	40.000	41.578	-3.9	103	0.00
70 C	Pentachlorophenol	40.000	41.596	-4.0	102	0.00
71	Phenanthrene	40.000	40.471	-1.2	101	0.00
72	Anthracene	40.000	40.511	-1.3	102	0.00
73	Carbazole	40.000	41.040	-2.6	102	0.00
74	Di-n-butylphthalate	40.000	41.028	-2.6	103	0.00
75 C	Fluoranthene	40.000	41.387	-3.5	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	42.700	-6.8	94	0.00
78	Pyrene	40.000	40.142	-0.4	101	0.00
79 S	Terphenyl-d14	80.000	81.919	-2.4	103	0.00
80	Butylbenzylphthalate	40.000	41.179	-2.9	103	0.00
81	Benzo(a)anthracene	40.000	41.191	-3.0	103	0.00
82	3,3'-Dichlorobenzidine	40.000	41.347	-3.4	101	0.00
83	Chrysene	40.000	40.988	-2.5	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.579	-1.4	101	0.00
85 c	Di-n-octyl phthalate	40.000	41.919	-4.8	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.475	-3.7	105	0.00
88	Benzo(b)fluoranthene	40.000	41.183	-3.0	105	0.00
89	Benzo(k)fluoranthene	40.000	41.275	-3.2	103	0.00
90 C	Benzo(a)pyrene	40.000	41.181	-3.0	104	0.00
91	Dibenzo(a,h)anthracene	40.000	41.706	-4.3	104	0.00
92	Benzo(g,h,i)perylene	40.000	41.110	-2.8	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080321\
Data File : BG049631.D
Acq On : 3 Aug 2021 19:53
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 04 02:23:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:15:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0