


Data Path : Z:\HPCHEM1\BNA_G\Data\BG080415\
Data File : BG018254.D
Acq On : 4 Aug 2015 13:43
Operator : UM/TP
Sample : DFTPP25
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP25

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG073115.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 06 09:29:11 2015



AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.5	12470	PASS
68	69	0.00	2	1.1	175	PASS
69	198	0.00	100	35.1	15362	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.9	21370	PASS
197	198	0.00	2	0.3	129	PASS
198	198	100	100	100.0	43712	PASS
199	198	5	9	6.8	2953	PASS
275	198	10	60	31.5	13781	PASS
365	198	1	100	6.4	2806	PASS
441	442	0.01	24	16.1	6325	PASS
442	198	50	100	90.0	39325	PASS
443	442	15	24	19.4	7623	PASS

m/z	Abundance
37.10	183.0
38.00	224.0
38.90	463.0
39.90	170.0
43.00	274.0
50.00	3703.0
51.00	11656.0
52.00	414.0
55.10	250.0
55.90	230.0
57.00	1075.0
60.00	164.0
61.10	222.0
63.00	873.0
64.80	267.0
69.00	13338.0
72.90	617.0
74.00	1865.0
75.00	2848.0
77.00	16456.0
78.00	1288.0
79.00	1274.0
79.90	847.0
81.00	1971.0
81.90	358.0
82.90	688.0
84.90	204.0
85.90	497.0
86.80	526.0
87.60	200.0
90.90	334.0
91.90	550.0
93.00	2546.0
94.00	366.0
95.80	285.0
97.90	2257.0
99.00	1345.0
101.10	602.0
103.00	647.0
104.10	444.0
105.00	515.0
105.90	321.0
106.10	326.0
106.90	5959.0
107.90	760.0
110.00	9767.0
111.00	1359.0
112.10	248.0
115.80	349.0
117.00	5383.0
118.10	314.0
121.00	273.0
122.00	611.0
123.10	506.0
124.00	392.0
127.00	17640.0
128.10	2031.0
129.00	8681.0
130.00	835.0
131.00	174.0
132.70	214.0
135.00	763.0
135.90	496.0
137.00	341.0
141.00	1250.0
141.90	339.0
142.70	316.0
144.10	301.0
146.00	232.0
147.00	750.0
148.00	1601.0
149.00	424.0
150.80	250.0
151.50	167.0
153.10	545.0
153.80	214.0
155.00	544.0
155.90	904.0
157.10	217.0
158.00	178.0
159.00	290.0
160.00	397.0
161.00	755.0
164.90	458.0
165.90	592.0
167.00	2185.0
167.90	1340.0
171.90	421.0
172.90	395.0
173.90	679.0
175.00	902.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

176.20	204.0
176.90	506.0
178.90	1377.0
180.00	1192.0
180.90	650.0
183.90	434.0
184.90	953.0
186.00	5573.0
187.00	1560.0
187.90	153.0
189.00	472.0
190.00	186.0
191.10	407.0
192.00	1001.0
193.00	887.0
194.00	170.0
195.90	929.0
196.70	387.0
197.90	34264.0
198.90	2779.0
199.80	285.0
203.00	476.0
204.00	1389.0
205.00	2346.0
206.00	11024.0
207.00	1467.0
207.90	707.0
209.00	282.0
209.90	288.0
210.80	533.0
216.00	315.0
216.90	2714.0
217.90	445.0
219.10	197.0
221.00	2214.0
221.60	621.0
223.00	813.0
224.00	6021.0
225.00	1676.0
225.80	285.0
226.90	3795.0
227.90	920.0
229.00	385.0
229.80	203.0
231.10	379.0
234.00	381.0
235.00	361.0
237.20	175.0
238.90	175.0
240.00	262.0
241.00	392.0
242.10	345.0
243.10	455.0
244.00	4193.0
245.00	957.0
245.90	1448.0
247.00	414.0
253.10	202.0
253.80	189.0
254.90	22088.0
256.00	3660.0
256.90	218.0
257.20	186.0
258.00	1200.0
258.70	199.0
259.10	213.0
265.00	378.0
266.70	271.0
272.10	298.0
273.00	1037.0
273.90	2024.0
275.00	9809.0
276.00	1414.0
276.90	841.0
278.10	275.0
281.00	244.0
285.10	336.0
292.90	304.0
294.00	176.0
295.10	205.0
296.00	3312.0
297.00	378.0
302.90	531.0
304.00	156.0
308.80	243.0
315.00	318.0
323.00	949.0
323.90	267.0
327.10	358.0
328.20	210.0
332.10	356.0
332.90	262.0
333.90	878.0
334.90	247.0
346.10	292.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

351.90	546.0
353.10	263.0
354.00	575.0
365.00	2145.0
366.00	274.0
372.10	678.0
384.90	277.0
401.50	250.0
402.10	425.0
403.00	213.0
422.00	310.0
423.10	1573.0
424.10	347.0
441.00	4178.0
442.10	24400.0
443.10	4347.0
444.10	614.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

m/z	Abundance
39.00	1499.0
41.10	154.0
43.10	327.0
50.10	4586.0
51.00	15343.0
52.00	1134.0
55.10	212.0
55.90	516.0
56.90	1140.0
58.00	172.0
61.90	437.0
63.00	750.0
64.10	201.0
65.10	205.0
66.90	249.0
67.80	526.0
69.00	19048.0
70.90	239.0
73.00	163.0
74.00	2111.0
75.00	3240.0
75.90	1341.0
77.00	21960.0
78.00	1282.0
78.90	1793.0
79.90	1201.0
80.90	1973.0
81.90	441.0
82.90	626.0
84.00	226.0
85.00	448.0
86.00	460.0
86.70	195.0
91.10	527.0
92.00	623.0
93.00	2766.0
93.90	240.0
95.00	178.0
96.00	308.0
97.20	266.0
98.00	2650.0
99.00	2214.0
99.80	250.0
100.90	887.0
102.90	375.0
103.80	461.0
105.00	898.0
106.00	502.0
107.00	7909.0
108.10	1732.0
109.90	11616.0
111.00	1887.0
111.70	257.0
111.90	250.0
113.10	223.0
115.90	484.0
117.00	8089.0
118.00	497.0
120.00	281.0
121.70	671.0
123.00	1032.0
123.80	343.0
124.70	228.0
125.30	194.0
127.00	25288.0
128.00	2052.0
129.00	11806.0
129.90	1574.0
130.90	205.0
132.10	220.0
132.40	227.0
133.00	240.0
133.90	323.0
135.00	999.0
136.00	401.0
136.90	456.0
137.90	209.0
138.80	168.0
139.80	261.0
140.90	1963.0
142.00	399.0
143.00	289.0
143.90	230.0
145.80	234.0
147.00	1233.0
147.90	2078.0
149.00	364.0
150.00	167.0
150.90	214.0
152.10	181.0
152.90	589.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

154.00	677.0
154.90	729.0
156.00	1534.0
157.20	184.0
158.10	366.0
159.00	185.0
159.90	764.0
160.90	1111.0
162.00	244.0
162.80	179.0
164.50	208.0
165.00	791.0
165.90	778.0
167.00	3192.0
167.90	1509.0
168.90	317.0
171.10	225.0
171.80	353.0
172.90	311.0
174.00	731.0
174.90	975.0
175.90	362.0
176.90	1092.0
178.20	314.0
179.00	2506.0
179.90	1369.0
181.00	742.0
182.00	360.0
182.90	323.0
183.90	482.0
184.90	1426.0
186.00	9800.0
187.00	3131.0
188.00	401.0
188.90	860.0
190.90	357.0
192.00	797.0
193.00	923.0
194.80	185.0
195.90	1723.0
197.90	50872.0
198.90	3494.0
200.10	292.0
200.90	278.0
201.50	261.0
202.90	843.0
204.00	1789.0
205.00	4126.0
206.00	13795.0
207.00	2265.0
208.00	410.0
208.80	303.0
209.90	257.0
211.00	756.0
214.40	165.0
216.90	4691.0
217.90	908.0
221.00	2120.0
222.00	631.0
223.00	1002.0
224.00	8473.0
225.00	2759.0
226.00	353.0
227.00	4766.0
227.90	807.0
228.90	753.0
230.00	166.0
230.90	584.0
234.00	195.0
234.90	387.0
235.80	194.0
238.70	278.0
239.10	286.0
240.90	502.0
241.90	358.0
244.00	6368.0
245.10	900.0
245.90	1713.0
246.90	277.0
248.90	464.0
252.00	172.0
252.90	181.0
253.80	475.0
255.00	38112.0
256.00	5799.0
257.00	429.0
258.00	2515.0
259.00	339.0
264.00	181.0
265.00	1225.0
271.10	164.0
272.00	173.0
273.00	884.0
273.90	2804.0
275.00	16094.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

276.00	2013.0
277.10	2109.0
277.90	355.0
283.00	298.0
285.00	336.0
292.60	260.0
293.00	382.0
294.20	162.0
294.90	158.0
296.00	5451.0
296.90	1106.0
302.00	159.0
303.00	927.0
304.10	253.0
314.20	359.0
314.50	378.0
315.10	618.0
316.10	658.0
321.10	391.0
323.10	1570.0
324.00	462.0
327.00	719.0
328.10	177.0
333.20	277.0
334.10	1068.0
335.00	279.0
341.00	158.0
346.00	316.0
351.90	324.0
353.10	399.0
354.00	533.0
364.90	2938.0
365.80	227.0
372.10	766.0
372.90	280.0
383.10	328.0
402.00	428.0
403.10	418.0
404.10	234.0
421.10	523.0
422.00	583.0
423.10	3366.0
424.00	492.0
441.00	6621.0
442.00	47136.0
443.10	8993.0
444.10	752.0
523.50	193.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

m/z	Abundance
37.90	195.0
39.00	973.0
39.90	161.0
40.90	194.0
42.10	168.0
43.10	177.0
44.00	150.0
44.90	293.0
48.90	399.0
50.00	3376.0
51.00	10412.0
52.00	598.0
52.90	158.0
55.00	477.0
56.00	319.0
57.00	1056.0
61.00	294.0
62.00	266.0
63.00	669.0
65.00	177.0
69.00	13701.0
71.00	328.0
73.00	316.0
74.00	1564.0
75.00	2982.0
76.00	1053.0
77.00	15089.0
78.00	1134.0
79.00	1174.0
79.90	1083.0
81.00	1289.0
81.90	428.0
83.00	315.0
85.00	403.0
86.10	220.0
90.90	316.0
91.70	403.0
93.00	2364.0
94.80	336.0
97.20	340.0
98.00	2095.0
99.00	1315.0
101.00	556.0
103.00	364.0
104.00	671.0
104.90	965.0
107.00	7194.0
107.90	1047.0
109.90	10219.0
111.00	2466.0
112.10	372.0
117.00	6656.0
118.00	477.0
121.90	801.0
123.00	472.0
125.00	512.0
127.00	21184.0
128.00	1636.0
129.00	9677.0
130.00	825.0
130.80	156.0
133.90	559.0
134.90	845.0
136.00	403.0
136.90	473.0
139.90	224.0
141.00	1653.0
142.00	261.0
143.00	246.0
146.10	320.0
147.00	682.0
147.90	1401.0
152.90	227.0
153.90	259.0
154.90	652.0
156.00	820.0
156.90	254.0
158.10	161.0
159.00	322.0
160.00	395.0
161.00	501.0
161.90	257.0
163.00	168.0
165.10	965.0
165.90	461.0
167.00	2557.0
167.90	1100.0
168.90	438.0
170.00	200.0
172.80	346.0
174.00	289.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

174.90	1224.0
176.10	305.0
177.00	683.0
177.70	294.0
178.90	2703.0
180.00	1969.0
180.90	558.0
185.00	830.0
186.00	8264.0
187.00	2408.0
188.10	342.0
189.00	797.0
191.90	833.0
193.10	926.0
193.90	173.0
195.90	1286.0
197.90	46000.0
198.90	2587.0
199.90	173.0
201.20	540.0
203.00	536.0
203.90	2204.0
205.00	3677.0
206.00	13305.0
207.00	2304.0
207.90	610.0
209.00	529.0
209.90	399.0
210.90	376.0
215.00	282.0
215.70	179.0
217.00	3964.0
217.90	661.0
220.90	2376.0
221.60	918.0
223.10	620.0
224.00	8372.0
225.00	2107.0
226.00	537.0
227.00	3671.0
228.00	466.0
228.90	844.0
230.10	228.0
231.10	268.0
233.90	280.0
235.10	183.0
237.00	329.0
238.90	317.0
240.80	214.0
242.00	155.0
242.90	516.0
244.00	6220.0
245.00	774.0
246.00	1781.0
247.00	445.0
248.70	382.0
252.10	188.0
252.90	181.0
254.00	606.0
255.00	35528.0
256.00	4581.0
257.10	260.0
257.90	3005.0
259.00	277.0
260.00	187.0
261.20	231.0
264.90	986.0
272.00	161.0
273.00	1450.0
273.90	2841.0
275.00	15440.0
276.00	2491.0
276.90	1570.0
278.10	317.0
281.00	270.0
283.00	283.0
283.90	157.0
285.00	412.0
292.90	312.0
296.00	6234.0
297.00	783.0
302.00	221.0
303.00	590.0
314.00	293.0
315.00	638.0
315.90	386.0
321.00	267.0
323.00	1790.0
324.00	235.0
326.90	401.0
327.90	165.0
333.90	1064.0
335.10	234.0
341.00	225.0
346.00	360.0

Instrument :

BNA_G

ClientSampleId :

DFTPP25

352.10	585.0
352.90	528.0
353.90	308.0
356.10	187.0
365.00	3337.0
365.80	442.0
372.00	1256.0
373.10	176.0
382.90	280.0
402.00	756.0
403.00	718.0
420.90	317.0
422.00	488.0
423.00	3657.0
424.00	734.0
441.10	8177.0
442.00	46440.0
443.00	9530.0
444.10	938.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25

m/z	Abundance
41.00	217.0
43.00	214.0
43.80	173.0
44.90	179.0
55.00	161.0
67.00	289.0
81.20	236.0
91.00	231.0
94.90	165.0
103.00	155.0
123.10	155.0
149.00	218.0
167.00	226.0
187.10	165.0
206.90	241.0
218.10	298.0
367.40	203.0

Instrument :
BNA_G
ClientSampleId :
DFTPP25