


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080422\
Data File : BG054511.D
Acq On : 4 Aug 2022 9:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 28 04:30:50 2022






AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2512

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.8	13451	PASS
68	69	0.00	2	0.4	56	PASS
69	198	0.00	100	23.5	13287	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	32.2	18189	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	56512	PASS
199	198	5	9	7.3	4120	PASS
275	198	10	60	36.0	20331	PASS
365	198	1	100	4.6	2607	PASS
441	198	0.01	100	18.3	10329	PASS
442	442	50	100	100.0	63640	PASS
443	442	15	24	19.8	12588	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080422\
Data File : BG054511.D
Acq On : 4 Aug 2022 9:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Aug 04 15:45:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 28 04:30:50 2022
Response via : Initial Calibration



TIC: BG054511.D\data.ms

(70) Pentachlorophenol (C)

17.036min (-0.001) 19308.73 ng




response 40653

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	70.20	69.38
264.00	66.20	62.23
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG080422\
Data File : BG054511.D
Acq On : 4 Aug 2022 9:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Aug 04 15:45:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG071522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 28 04:30:50 2022
Response via : Initial Calibration



TIC: BG054511.D\data.ms

(77) Benzidine

19.598min (-0.001) 0.00 ng

response 355240

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.36
183.00	11.10	12.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/4/2022	BNA_G	BG054511.D
Compound Name	Response	Retention Time
DDT	216647	20.82
DDD	4817	20.38
DDE	661	19.87
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5478	222125	2.47

Instrument :
BNA_G
ClientSampleId :
DFTPP