

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
 Data File : BG023506.D
 Acq On : 6 Aug 2016 3:08
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02012

Quant Time: Aug 06 05:15:24 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 04 17:27:34 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.494	0.519	-5.1	110	0.00
3 S	1,4-Dioxane-d8	0.470	0.450	4.3	103	0.00
4	Benzaldehyde	1.228	1.367	-11.3	117	0.00
5 S	Phenol-d5	2.015	2.049	-1.7	114	0.00
6	Phenol	2.098	2.144	-2.2	116	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.291	1.328	-2.9	118	0.00
8	Bis(2-Chloroethyl)ether	1.523	1.608	-5.6	119	0.00
9 S	2-Chlorophenol-d4	1.374	1.409	-2.5	116	0.00
10	2-Chlorophenol	1.400	1.467	-4.8	119	0.00
11	2-Methylphenol	1.580	1.597	-1.1	116	0.00
12	2,2'-oxybis(1-Chloropropane	2.084	2.255	-8.2	122	0.00
13 S	4-Methylphenol-d8	1.573	1.597	-1.5	118	0.00
14	Acetophenone	2.523	2.675	-6.0	116	0.00
15 P	N-Nitroso-di-n-propylamine	1.518	1.533	-1.0	113	0.00
16	4-Methylphenol	1.771	1.746	1.4	112	0.00
17	Hexachloroethane	0.604	0.612	-1.3	118	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
19 S	Nitrobenzene-d5	0.158	0.168	-6.3	119	0.00
20	Nitrobenzene	0.462	0.480	-3.9	116	0.00
21	Isophorone	0.850	0.887	-4.4	113	0.00
22 S	2-Nitrophenol-d4	0.181	0.191	-5.5	116	0.00
23 C	2-Nitrophenol	0.195	0.203	-4.1	119	0.00
24	2,4-Dimethylphenol	0.429	0.452	-5.4	116	0.00
25	Bis(2-Chloroethoxy)methane	0.490	0.507	-3.5	116	0.00
26 S	2,4-Dichlorophenol-d3	0.342	0.359	-5.0	116	0.00
27 C	2,4-Dichlorophenol	0.344	0.360	-4.7	116	0.00
28	Naphthalene	1.015	1.072	-5.6	116	0.00
29 S	4-Chloroaniline-d4	0.401	0.457	-14.0	116	0.00
30	4-Chloroaniline	0.398	0.449	-12.8	117	0.00
31 C	Hexachlorobutadiene	0.229	0.245	-7.0	121	0.00
32	Caprolactam	0.143	0.139	2.8	106	0.00
33 C	4-Chloro-3-methylphenol	0.443	0.447	-0.9	113	0.00
34	2-Methylnaphthalene	0.823	0.863	-4.9	115	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
36	1,2,4,5-Tetrachlorobenzene	0.614	0.664	-8.1	112	0.00
37	Hexachlorocyclopentadiene	0.340	0.346	-1.8	115	0.00
38 C	2,4,6-Trichlorophenol	0.425	0.440	-3.5	106	0.00
39	2,4,5-Trichlorophenol	0.456	0.464	-1.8	108	0.00
40	1,1'-Biphenyl	1.470	1.606	-9.3	114	0.00
41	2-Chloronaphthalene	1.143	1.228	-7.4	113	0.00
42	2-Nitroaniline	0.467	0.499	-6.9	109	0.00
43 S	Dimethylphthalate-d6	1.638	1.713	-4.6	108	0.00
44	Dimethylphthalate	1.622	1.663	-2.5	105	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
 Data File : BG023506.D
 Acq On : 6 Aug 2016 3:08
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02012

Quant Time: Aug 06 05:15:24 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 04 17:27:34 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.352	0.358	-1.7	103	0.00
46 S	Acenaphthylene-d8	1.843	1.990	-8.0	112	0.00
47	Acenaphthylene	1.928	2.091	-8.5	112	0.00
48	3-Nitroaniline	0.331	0.340	-2.7	105	0.00
49 C	Acenaphthene	1.313	1.383	-5.3	109	0.00
50	2,4-Dinitrophenol	0.217	0.203	6.5	107	0.00
51 S	4-Nitrophenol-d4	0.280	0.284	-1.4	110	0.00
52	4-Nitrophenol	0.280	0.286	-2.1	106	0.00
53	Dibenzofuran	1.946	2.056	-5.7	110	0.00
54	2,4-Dinitrotoluene	0.533	0.557	-4.5	106	0.00
55	2,3,4,6-Tetrachlorophenol	0.468	0.459	1.9	101	0.00
56	Diethylphthalate	1.695	1.744	-2.9	105	0.00
57 S	Fluorene-d10	1.513	1.548	-2.3	106	0.00
58	Fluorene	1.621	1.650	-1.8	106	0.00
59	4-Chlorophenyl-phenylether	0.815	0.841	-3.2	107	0.00
60	4-Nitroaniline	0.383	0.409	-6.8	108	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.128	0.129	-0.8	103	0.00
63	4,6-Dinitro-2-methylphenol	0.135	0.142	-5.2	105	0.00
64	N-Nitrosodiphenylamine	0.559	0.612	-9.5	105	0.00
65	4-Bromophenyl-phenylether	0.229	0.233	-1.7	101	0.00
66	Hexachlorobenzene	0.244	0.254	-4.1	105	0.00
67	Atrazine	0.231	0.247	-6.9	100	0.00
68 C	Pentachlorophenol	0.135	0.127	5.9	98	0.00
69	Phenanthrene	1.036	1.117	-7.8	102	0.00
70 S	Anthracene-d10	0.901	0.950	-5.4	102	0.00
71	Anthracene	1.057	1.159	-9.6	104	0.00
72	Carbazole	0.852	0.997	-17.0	100	0.00
73	Di-n-butylphthalate	1.067	1.189	-11.4	103	0.00
74 C	Fluoranthene	1.060	1.305	-23.1	102	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
76 S	Pyrene-d10	1.012	1.048	-3.6	103	0.00
77	Pyrene	1.236	1.285	-4.0	101	0.00
78	Butylbenzylphthalate	0.474	0.515	-8.6	106	0.00
79	3,3'-Dichlorobenzidine	0.352	0.403	-14.5	110	0.00
80	Benzo(a)anthracene	1.126	1.203	-6.8	104	0.00
81	Bis(2-ethylhexyl)phthalate	0.631	0.709	-12.4	108	0.00
82	Chrysene	1.024	1.111	-8.5	107	0.00
83 I	Perylene-d12	1.000	1.000	0.0	107	0.00
84	Di-n-octyl phthalate	1.010	1.215	-20.3#	109	0.00
85	Benzo(b)fluoranthene	1.138	1.221	-7.3	104	0.00
86	Benzo(k)fluoranthene	1.120	1.210	-8.0	109	0.00
87 S	Benzo(a)pyrene-d12	0.905	0.960	-6.1	106	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
Data File : BG023506.D
Acq On : 6 Aug 2016 3:08
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02012

Quant Time: Aug 06 05:15:24 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 04 17:27:34 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.098	1.192	-8.6	107	0.00
89	Indeno(1,2,3-cd)pyrene	1.291	1.375	-6.5	108	0.02
90	Dibenzo(a,h)anthracene	1.101	1.174	-6.6	107	0.00
91	Benzo(g,h,i)perylene	1.070	1.125	-5.1	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0