

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
 Data File : BG023519.D
 Acq On : 6 Aug 2016 11:47
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02013

Quant Time: Aug 08 03:18:55 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 01 04:44:16 2004
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	0.00
2	1,4-Dioxane	0.494	0.477	3.4	106	0.00
3 S	1,4-Dioxane-d8	0.470	0.468	0.4	112	0.00
4	Benzaldehyde	1.228	1.327	-8.1	119	0.00
5 S	Phenol-d5	2.015	1.988	1.3	116	0.00
6	Phenol	2.098	2.135	-1.8	121	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.291	1.297	-0.5	121	0.00
8	Bis(2-Chloroethyl)ether	1.523	1.577	-3.5	122	0.00
9 S	2-Chlorophenol-d4	1.374	1.364	0.7	117	0.00
10	2-Chlorophenol	1.400	1.407	-0.5	119	0.00
11	2-Methylphenol	1.580	1.522	3.7	116	0.00
12	2,2'-oxybis(1-Chloropropane	2.084	2.192	-5.2	124	0.00
13 S	4-Methylphenol-d8	1.573	1.592	-1.2	123	0.00
14	Acetophenone	2.523	2.616	-3.7	119	0.00
15 P	N-Nitroso-di-n-propylamine	1.518	1.536	-1.2	119	0.00
16	4-Methylphenol	1.771	1.754	1.0	118	0.00
17	Hexachloroethane	0.604	0.601	0.5	121	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
19 S	Nitrobenzene-d5	0.158	0.166	-5.1	119	0.00
20	Nitrobenzene	0.462	0.479	-3.7	118	0.00
21	Isophorone	0.850	0.910	-7.1	118	0.00
22 S	2-Nitrophenol-d4	0.181	0.188	-3.9	116	0.00
23 C	2-Nitrophenol	0.195	0.205	-5.1	123	0.00
24	2,4-Dimethylphenol	0.429	0.451	-5.1	118	0.00
25	Bis(2-Chloroethoxy)methane	0.490	0.523	-6.7	121	0.00
26 S	2,4-Dichlorophenol-d3	0.342	0.362	-5.8	119	0.00
27 C	2,4-Dichlorophenol	0.344	0.358	-4.1	118	0.00
28	Naphthalene	1.015	1.066	-5.0	117	0.00
29 S	4-Chloroaniline-d4	0.401	0.469	-17.0	121	0.00
30	4-Chloroaniline	0.398	0.450	-13.1	120	0.00
31 C	Hexachlorobutadiene	0.229	0.233	-1.7	118	0.00
32	Caprolactam	0.143	0.136	4.9	105	0.00
33 C	4-Chloro-3-methylphenol	0.443	0.452	-2.0	116	0.00
34	2-Methylnaphthalene	0.823	0.865	-5.1	117	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
36	1,2,4,5-Tetrachlorobenzene	0.614	0.654	-6.5	116	0.00
37	Hexachlorocyclopentadiene	0.340	0.343	-0.9	119	0.00
38 C	2,4,6-Trichlorophenol	0.425	0.453	-6.6	115	0.00
39	2,4,5-Trichlorophenol	0.456	0.482	-5.7	118	0.00
40	1,1'-Biphenyl	1.470	1.569	-6.7	116	0.00
41	2-Chloronaphthalene	1.143	1.223	-7.0	117	0.00
42	2-Nitroaniline	0.467	0.499	-6.9	114	0.00
43 S	Dimethylphthalate-d6	1.638	1.701	-3.8	113	0.00
44	Dimethylphthalate	1.622	1.676	-3.3	111	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
 Data File : BG023519.D
 Acq On : 6 Aug 2016 11:47
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02013

Quant Time: Aug 08 03:18:55 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 01 04:44:16 2004
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.352	0.367	-4.3	111	0.00
46 S	Acenaphthylene-d8	1.843	1.974	-7.1	116	0.00
47	Acenaphthylene	1.928	2.054	-6.5	115	0.00
48	3-Nitroaniline	0.331	0.358	-8.2	116	0.00
49 C	Acenaphthene	1.313	1.375	-4.7	113	0.00
50	2,4-Dinitrophenol	0.217	0.203	6.5	112	0.00
51 S	4-Nitrophenol-d4	0.280	0.286	-2.1	116	0.00
52	4-Nitrophenol	0.280	0.281	-0.4	109	0.00
53	Dibenzofuran	1.946	2.036	-4.6	114	0.00
54	2,4-Dinitrotoluene	0.533	0.558	-4.7	111	0.00
55	2,3,4,6-Tetrachlorophenol	0.468	0.467	0.2	108	0.00
56	Diethylphthalate	1.695	1.764	-4.1	111	0.00
57 S	Fluorene-d10	1.513	1.560	-3.1	112	0.00
58	Fluorene	1.621	1.675	-3.3	112	0.00
59	4-Chlorophenyl-phenylether	0.815	0.851	-4.4	113	0.00
60	4-Nitroaniline	0.383	0.397	-3.7	109	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.128	0.133	-3.9	111	0.00
63	4,6-Dinitro-2-methylphenol	0.135	0.142	-5.2	111	0.00
64	N-Nitrosodiphenylamine	0.559	0.611	-9.3	110	0.00
65	4-Bromophenyl-phenylether	0.229	0.240	-4.8	109	0.00
66	Hexachlorobenzene	0.244	0.263	-7.8	114	0.00
67	Atrazine	0.231	0.247	-6.9	106	0.00
68 C	Pentachlorophenol	0.135	0.138	-2.2	112	0.00
69	Phenanthrene	1.036	1.109	-7.0	106	0.00
70 S	Anthracene-d10	0.901	0.967	-7.3	109	0.00
71	Anthracene	1.057	1.146	-8.4	108	0.00
72	Carbazole	0.852	1.016	-19.2	107	0.00
73	Di-n-butylphthalate	1.067	1.192	-11.7	109	0.00
74 C	Fluoranthene	1.060	1.311	-23.7	108	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
76 S	Pyrene-d10	1.012	1.064	-5.1	106	0.00
77	Pyrene	1.236	1.325	-7.2	105	0.00
78	Butylbenzylphthalate	0.474	0.528	-11.4	111	0.00
79	3,3'-Dichlorobenzidine	0.352	0.404	-14.8	112	0.00
80	Benzo(a)anthracene	1.126	1.208	-7.3	106	0.00
81	Bis(2-ethylhexyl)phthalate	0.631	0.731	-15.8	113	0.00
82	Chrysene	1.024	1.107	-8.1	109	0.00
83 I	Perylene-d12	1.000	1.000	0.0	111	0.00
84	Di-n-octyl phthalate	1.010	1.216	-20.4#	114	0.00
85	Benzo(b)fluoranthene	1.138	1.219	-7.1	108	0.00
86	Benzo(k)fluoranthene	1.120	1.166	-4.1	109	0.00
87 S	Benzo(a)pyrene-d12	0.905	0.954	-5.4	110	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
Data File : BG023519.D
Acq On : 6 Aug 2016 11:47
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02013

Quant Time: Aug 08 03:18:55 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 01 04:44:16 2004
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.098	1.161	-5.7	108	0.00
89	Indeno(1,2,3-cd)pyrene	1.291	1.363	-5.6	111	0.00
90	Dibenzo(a,h)anthracene	1.101	1.159	-5.3	110	0.00
91	Benzo(g,h,i)perylene	1.070	1.111	-3.8	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0