

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
 Data File : BG023529.D
 Acq On : 6 Aug 2016 19:06
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02014

Quant Time: Aug 08 02:13:25 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 01 04:44:16 2004
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.494	0.495	-0.2	108	0.00
3 S	1,4-Dioxane-d8	0.470	0.463	1.5	109	0.00
4	Benzaldehyde	1.228	1.493	-21.6#	131	0.00
5 S	Phenol-d5	2.015	2.134	-5.9	123	0.00
6	Phenol	2.098	2.244	-7.0	125	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.291	1.353	-4.8	124	0.00
8	Bis(2-Chloroethyl)ether	1.523	1.622	-6.5	123	0.00
9 S	2-Chlorophenol-d4	1.374	1.468	-6.8	124	0.00
10	2-Chlorophenol	1.400	1.478	-5.6	123	0.00
11	2-Methylphenol	1.580	1.652	-4.6	124	0.00
12	2,2'-oxybis(1-Chloropropane	2.084	2.281	-9.5	127	0.00
13 S	4-Methylphenol-d8	1.573	1.706	-8.5	129	0.00
14	Acetophenone	2.523	2.768	-9.7	124	0.00
15 P	N-Nitroso-di-n-propylamine	1.518	1.602	-5.5	122	0.00
16	4-Methylphenol	1.771	1.850	-4.5	122	0.00
17	Hexachloroethane	0.604	0.620	-2.6	123	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
19 S	Nitrobenzene-d5	0.158	0.163	-3.2	123	0.00
20	Nitrobenzene	0.462	0.465	-0.6	120	0.00
21	Isophorone	0.850	0.892	-4.9	122	0.00
22 S	2-Nitrophenol-d4	0.181	0.192	-6.1	124	0.00
23 C	2-Nitrophenol	0.195	0.205	-5.1	128	0.00
24	2,4-Dimethylphenol	0.429	0.453	-5.6	125	0.00
25	Bis(2-Chloroethoxy)methane	0.490	0.506	-3.3	123	0.00
26 S	2,4-Dichlorophenol-d3	0.342	0.360	-5.3	124	0.00
27 C	2,4-Dichlorophenol	0.344	0.354	-2.9	122	0.00
28	Naphthalene	1.015	1.065	-4.9	123	0.00
29 S	4-Chloroaniline-d4	0.401	0.455	-13.5	123	0.00
30	4-Chloroaniline	0.398	0.453	-13.8	126	0.00
31 C	Hexachlorobutadiene	0.229	0.236	-3.1	125	0.00
32	Caprolactam	0.143	0.144	-0.7	117	0.00
33 C	4-Chloro-3-methylphenol	0.443	0.460	-3.8	124	0.00
34	2-Methylnaphthalene	0.823	0.856	-4.0	122	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
36	1,2,4,5-Tetrachlorobenzene	0.614	0.662	-7.8	123	0.00
37	Hexachlorocyclopentadiene	0.340	0.316	7.1	115	0.00
38 C	2,4,6-Trichlorophenol	0.425	0.444	-4.5	118	0.00
39	2,4,5-Trichlorophenol	0.456	0.469	-2.9	121	0.00
40	1,1'-Biphenyl	1.470	1.570	-6.8	123	0.00
41	2-Chloronaphthalene	1.143	1.181	-3.3	120	0.00
42	2-Nitroaniline	0.467	0.504	-7.9	121	0.00
43 S	Dimethylphthalate-d6	1.638	1.653	-0.9	115	0.00
44	Dimethylphthalate	1.622	1.638	-1.0	114	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
 Data File : BG023529.D
 Acq On : 6 Aug 2016 19:06
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02014

Quant Time: Aug 08 02:13:25 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 01 04:44:16 2004
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.352	0.360	-2.3	114	0.00
46 S	Acenaphthylene-d8	1.843	1.940	-5.3	120	0.00
47	Acenaphthylene	1.928	2.045	-6.1	120	0.00
48	3-Nitroaniline	0.331	0.360	-8.8	123	0.00
49 C	Acenaphthene	1.313	1.351	-2.9	118	0.00
50	2,4-Dinitrophenol	0.217	0.184	15.2	107	0.00
51 S	4-Nitrophenol-d4	0.280	0.298	-6.4	128	0.00
52	4-Nitrophenol	0.280	0.292	-4.3	119	0.00
53	Dibenzofuran	1.946	1.984	-2.0	117	0.00
54	2,4-Dinitrotoluene	0.533	0.538	-0.9	113	0.00
55	2,3,4,6-Tetrachlorophenol	0.468	0.477	-1.9	116	0.00
56	Diethylphthalate	1.695	1.698	-0.2	113	0.00
57 S	Fluorene-d10	1.513	1.520	-0.5	115	0.00
58	Fluorene	1.621	1.620	0.1	114	0.00
59	4-Chlorophenyl-phenylether	0.815	0.832	-2.1	117	0.00
60	4-Nitroaniline	0.383	0.402	-5.0	117	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.128	0.133	-3.9	114	0.00
63	4,6-Dinitro-2-methylphenol	0.135	0.140	-3.7	112	0.00
64	N-Nitrosodiphenylamine	0.559	0.619	-10.7	114	0.00
65	4-Bromophenyl-phenylether	0.229	0.243	-6.1	113	0.00
66	Hexachlorobenzene	0.244	0.264	-8.2	118	0.00
67	Atrazine	0.231	0.253	-9.5	111	0.00
68 C	Pentachlorophenol	0.135	0.135	0.0	112	0.00
69	Phenanthrene	1.036	1.114	-7.5	109	0.00
70 S	Anthracene-d10	0.901	0.974	-8.1	112	0.00
71	Anthracene	1.057	1.157	-9.5	112	0.00
72	Carbazole	0.852	1.016	-19.2	110	0.00
73	Di-n-butylphthalate	1.067	1.211	-13.5	113	0.00
74 C	Fluoranthene	1.060	1.316	-24.2	111	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
76 S	Pyrene-d10	1.012	1.070	-5.7	110	0.00
77	Pyrene	1.236	1.309	-5.9	108	0.00
78	Butylbenzylphthalate	0.474	0.531	-12.0	115	0.00
79	3,3'-Dichlorobenzidine	0.352	0.415	-17.9	118	0.00
80	Benzo(a)anthracene	1.126	1.203	-6.8	109	0.00
81	Bis(2-ethylhexyl)phthalate	0.631	0.724	-14.7	115	0.00
82	Chrysene	1.024	1.090	-6.4	110	0.00
83 I	Perylene-d12	1.000	1.000	0.0	111	0.00
84	Di-n-octyl phthalate	1.010	1.255	-24.3#	118	0.00
85	Benzo(b)fluoranthene	1.138	1.205	-5.9	107	0.00
86	Benzo(k)fluoranthene	1.120	1.222	-9.1	115	0.00
87 S	Benzo(a)pyrene-d12	0.905	0.972	-7.4	112	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080516\
Data File : BG023529.D
Acq On : 6 Aug 2016 19:06
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02014

Quant Time: Aug 08 02:13:25 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 01 04:44:16 2004
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.098	1.182	-7.7	110	0.00
89	Indeno(1,2,3-cd)pyrene	1.291	1.384	-7.2	113	0.02
90	Dibenzo(a,h)anthracene	1.101	1.161	-5.4	110	0.01
91	Benzo(g,h,i)perylene	1.070	1.140	-6.5	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0