


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080715\
Data File : BG018309.D
Acq On : 7 Aug 2015 12:02
Operator : UM/TP
Sample : DFTPP28
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP28

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080515.M
Title : SVOA CALIBRATION
Last Update : Fri Aug 07 23:39:35 2015



AutoFind: Scans 2490, 2491, 2492; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	15796	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.2	16489	PASS
70	69	0.00	2	0.4	68	PASS
127	198	10	80	48.4	19888	PASS
197	198	0.00	2	0.9	364	PASS
198	198	100	100	100.0	41056	PASS
199	198	5	9	6.8	2781	PASS
275	198	10	60	33.5	13771	PASS
365	198	1	100	7.4	3055	PASS
441	442	0.01	24	15.2	4856	PASS
442	198	50	100	77.7	31917	PASS
443	442	15	24	20.9	6665	PASS

m/z	Abundance
38.00	236.0
39.00	923.0
49.00	384.0
50.00	4558.0
51.00	13432.0
51.90	652.0
52.70	286.0
53.10	152.0
55.90	361.0
56.90	981.0
61.80	396.0
63.00	677.0
64.90	213.0
69.00	17368.0
73.00	238.0
74.00	1955.0
75.00	2609.0
76.00	1135.0
77.00	19824.0
78.10	976.0
79.00	1749.0
79.90	1332.0
81.00	1313.0
81.80	479.0
83.00	233.0
84.90	270.0
86.00	631.0
86.90	232.0
91.10	348.0
92.10	265.0
93.00	4197.0
97.90	2080.0
98.90	1628.0
99.90	192.0
101.00	734.0
103.90	243.0
105.00	441.0
105.90	379.0
107.00	7069.0
108.00	1011.0
109.00	359.0
110.00	9213.0
111.00	1639.0
112.00	557.0
117.00	7335.0
117.90	806.0
122.00	352.0
123.00	410.0
123.80	248.0
125.00	157.0
127.00	17224.0
128.00	1653.0
129.00	8604.0
129.80	590.0
133.00	253.0
134.10	386.0
134.90	712.0
135.90	620.0
137.00	401.0
137.80	155.0
139.90	310.0
141.00	1309.0
141.70	187.0
142.20	417.0
142.90	470.0
143.90	273.0
145.80	336.0
146.90	605.0
147.90	2076.0
148.90	360.0
151.20	269.0
151.80	201.0
152.90	495.0
154.00	323.0
154.90	399.0
156.00	787.0
156.90	311.0
158.00	382.0
158.90	241.0
160.00	449.0
161.00	413.0
162.80	239.0
164.90	610.0
166.00	391.0
167.00	3552.0
168.00	1068.0
168.80	164.0
172.00	498.0
173.00	293.0
174.00	603.0
175.00	1055.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

176.10	472.0
177.10	645.0
177.90	238.0
179.00	1440.0
179.90	1056.0
180.90	352.0
182.00	218.0
184.00	196.0
184.90	827.0
186.00	5840.0
187.10	2116.0
188.00	242.0
189.00	743.0
191.10	429.0
191.90	709.0
193.00	842.0
193.90	223.0
195.90	1441.0
197.90	32720.0
198.90	2036.0
199.90	323.0
201.50	294.0
203.10	586.0
204.00	1775.0
205.00	2169.0
206.00	10960.0
207.00	1527.0
207.90	751.0
209.00	369.0
209.80	308.0
210.90	1057.0
211.80	175.0
214.80	249.0
216.10	479.0
216.90	3597.0
218.00	369.0
220.90	1659.0
222.00	474.0
222.90	936.0
224.00	5858.0
225.00	1837.0
226.00	318.0
227.00	3519.0
227.90	440.0
228.90	563.0
230.10	191.0
234.80	371.0
235.90	152.0
236.90	477.0
238.80	214.0
239.80	172.0
241.10	247.0
241.60	187.0
242.10	200.0
243.00	571.0
244.00	4609.0
245.10	635.0
246.00	1516.0
247.00	397.0
253.20	152.0
255.00	23840.0
255.90	4134.0
256.70	264.0
256.90	289.0
258.00	1859.0
259.00	508.0
260.90	207.0
265.00	1000.0
270.90	155.0
272.10	188.0
272.90	810.0
274.00	2235.0
275.00	10221.0
276.00	1933.0
276.90	1500.0
278.00	215.0
284.80	285.0
289.90	173.0
293.00	354.0
296.00	4402.0
297.00	545.0
303.20	311.0
314.00	206.0
315.20	332.0
315.80	206.0
321.10	273.0
323.00	668.0
327.00	472.0
334.00	563.0
335.10	231.0
342.10	243.0
345.80	301.0
352.00	680.0
353.00	292.0
353.90	351.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

364.90	2258.0
366.10	373.0
372.00	584.0
389.90	234.0
400.30	185.0
402.00	257.0
402.80	175.0
404.10	176.0
421.10	224.0
422.30	182.0
423.10	1578.0
423.90	266.0
441.00	3281.0
442.10	20632.0
443.10	4714.0
443.90	212.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

m/z	Abundance
36.90	184.0
38.20	320.0
39.10	856.0
39.90	161.0
41.00	212.0
43.00	179.0
48.70	485.0
50.00	5805.0
51.00	18848.0
52.00	1066.0
55.90	466.0
56.90	1308.0
60.90	151.0
61.80	178.0
63.00	974.0
63.80	255.0
65.00	202.0
68.90	19096.0
69.90	204.0
73.10	420.0
73.90	1779.0
74.90	3788.0
76.10	1709.0
77.00	23096.0
78.00	1383.0
79.00	2171.0
79.90	1509.0
81.00	2181.0
82.00	637.0
82.80	281.0
83.90	325.0
85.00	279.0
86.00	717.0
86.80	367.0
90.90	492.0
92.00	391.0
93.00	3915.0
94.10	433.0
94.70	376.0
97.10	355.0
98.00	3515.0
99.00	2237.0
100.90	929.0
103.00	404.0
104.00	603.0
104.90	470.0
107.00	9535.0
108.00	1069.0
110.00	12119.0
110.90	1728.0
111.80	338.0
112.80	220.0
115.90	517.0
117.00	11182.0
118.10	655.0
121.00	194.0
122.00	599.0
123.00	654.0
124.00	332.0
125.00	347.0
127.00	21592.0
127.90	1892.0
129.00	10700.0
130.00	1054.0
131.10	220.0
132.40	160.0
133.90	840.0
135.00	1106.0
136.00	649.0
137.00	552.0
139.60	177.0
140.90	2177.0
141.80	477.0
142.90	331.0
146.00	495.0
146.90	927.0
147.90	3413.0
149.00	624.0
150.00	241.0
150.80	164.0
151.20	198.0
153.00	596.0
153.70	263.0
155.00	1175.0
156.00	1281.0
157.00	451.0
158.00	166.0
159.00	274.0
159.80	783.0
160.80	440.0
162.00	203.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

165.00	340.0
166.00	582.0
166.90	4242.0
168.00	1563.0
169.90	176.0
171.10	299.0
171.90	260.0
172.90	473.0
174.00	1144.0
175.00	847.0
176.10	397.0
176.90	443.0
177.80	368.0
178.90	2309.0
180.00	1169.0
180.90	797.0
182.00	169.0
184.00	215.0
185.00	1106.0
186.00	7968.0
187.00	2427.0
188.00	727.0
188.90	688.0
192.00	953.0
193.00	720.0
194.10	218.0
194.90	330.0
196.00	1830.0
196.90	768.0
197.90	47560.0
198.90	2991.0
200.00	228.0
202.90	717.0
204.00	2949.0
205.00	4090.0
206.00	13402.0
207.00	2500.0
208.00	627.0
208.90	256.0
210.20	363.0
210.90	517.0
212.10	183.0
213.10	241.0
214.90	209.0
215.90	489.0
216.90	5083.0
217.80	517.0
220.90	2166.0
222.00	496.0
222.90	1523.0
224.00	10114.0
225.00	2663.0
227.00	4859.0
228.00	595.0
228.90	1104.0
231.00	581.0
231.80	203.0
233.80	308.0
234.80	261.0
236.10	265.0
237.00	327.0
239.00	360.0
240.00	229.0
240.90	348.0
242.00	508.0
243.00	580.0
244.00	7130.0
245.00	1225.0
246.00	1917.0
247.00	505.0
249.00	302.0
251.00	155.0
253.10	216.0
255.00	35632.0
255.90	5245.0
256.90	630.0
258.00	3088.0
258.90	452.0
263.00	199.0
264.10	197.0
265.00	1453.0
265.70	190.0
271.00	222.0
273.00	1165.0
274.00	3507.0
275.00	15311.0
276.00	2451.0
277.00	1684.0
278.00	523.0
284.90	401.0
290.00	159.0
292.10	199.0
294.00	280.0
296.00	6181.0
297.00	1061.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

302.90	736.0
304.10	183.0
314.10	299.0
315.00	894.0
316.10	410.0
321.20	312.0
323.10	1605.0
324.10	366.0
327.00	383.0
327.80	278.0
333.00	382.0
334.00	1125.0
334.70	195.0
339.20	194.0
341.00	413.0
345.90	377.0
352.10	584.0
353.00	645.0
353.90	763.0
355.00	256.0
357.30	170.0
358.80	221.0
365.00	3736.0
365.90	332.0
371.00	353.0
372.00	746.0
373.10	372.0
402.00	485.0
403.00	428.0
421.00	339.0
422.10	293.0
423.10	2680.0
424.10	415.0
441.00	5350.0
442.10	35912.0
443.10	8536.0
444.10	754.0
525.30	179.0

Instrument :

BNA_G

ClientSampleId :

DFTPP28

m/z	Abundance
38.10	201.0
39.00	921.0
49.00	513.0
50.00	3821.0
51.00	15109.0
52.10	947.0
55.00	416.0
55.90	424.0
57.00	871.0
63.00	765.0
64.10	237.0
67.20	303.0
68.90	13005.0
70.90	205.0
73.20	284.0
74.00	2052.0
75.00	2578.0
76.10	1169.0
77.00	17112.0
77.90	1098.0
79.00	1208.0
80.00	942.0
81.00	2273.0
81.80	183.0
83.00	441.0
83.80	233.0
85.10	584.0
85.90	501.0
87.00	660.0
91.00	499.0
92.00	684.0
93.00	2423.0
96.10	166.0
98.00	2347.0
99.00	1705.0
100.20	184.0
100.90	842.0
103.00	490.0
104.00	794.0
105.00	611.0
107.00	8349.0
108.00	1270.0
108.90	323.0
110.00	9681.0
111.00	1587.0
112.20	342.0
112.80	201.0
116.00	558.0
116.90	9295.0
118.00	451.0
120.00	210.0
121.90	395.0
122.90	1391.0
124.10	230.0
124.90	330.0
125.70	152.0
127.00	20848.0
128.00	2105.0
129.00	8833.0
130.00	861.0
130.90	159.0
132.00	156.0
133.90	465.0
135.00	922.0
135.80	367.0
137.00	420.0
139.90	165.0
140.90	1170.0
141.80	543.0
142.90	497.0
144.90	189.0
146.00	419.0
146.70	521.0
148.00	1765.0
149.00	296.0
151.00	249.0
153.00	468.0
154.00	307.0
154.90	665.0
156.00	1379.0
156.90	282.0
158.00	246.0
159.90	706.0
161.00	860.0
164.00	186.0
165.00	648.0
166.10	413.0
166.90	3090.0
167.80	1480.0
169.10	234.0
170.10	391.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

170.90	195.0
172.00	278.0
174.00	639.0
174.90	1226.0
175.80	379.0
176.90	796.0
177.90	185.0
178.90	2105.0
179.90	1270.0
180.90	550.0
184.10	174.0
184.90	1691.0
186.00	7465.0
186.90	1977.0
188.00	284.0
188.90	752.0
189.90	255.0
191.00	352.0
192.00	762.0
193.00	887.0
194.10	230.0
195.90	608.0
196.80	325.0
197.90	42888.0
198.90	3316.0
199.90	197.0
202.20	152.0
203.00	592.0
204.00	2810.0
205.00	3744.0
206.00	12853.0
206.90	1470.0
208.00	650.0
208.90	177.0
209.90	290.0
211.00	743.0
211.70	272.0
214.90	187.0
217.00	3673.0
218.00	466.0
218.90	151.0
221.00	2202.0
221.90	559.0
222.90	1091.0
224.00	8069.0
225.00	2019.0
226.10	269.0
226.90	5625.0
228.00	656.0
228.80	822.0
231.10	261.0
233.90	344.0
234.90	390.0
235.90	388.0
236.90	433.0
239.00	175.0
239.90	176.0
241.00	520.0
242.00	215.0
243.00	633.0
244.00	7078.0
245.10	843.0
245.90	2153.0
246.90	586.0
248.80	559.0
251.80	183.0
253.10	264.0
255.00	32728.0
256.00	5800.0
256.90	502.0
258.00	2560.0
265.00	1147.0
266.10	360.0
272.00	181.0
273.00	727.0
274.00	2918.0
275.00	15783.0
276.00	2011.0
277.00	2036.0
277.80	304.0
279.10	213.0
283.10	248.0
284.20	295.0
285.00	303.0
292.90	500.0
296.00	6869.0
297.00	912.0
298.20	159.0
303.00	606.0
304.00	156.0
307.90	173.0
308.90	265.0
309.90	165.0
314.00	377.0
315.00	1082.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

316.00	266.0
322.00	283.0
323.00	1413.0
323.90	443.0
326.90	173.0
328.10	402.0
334.10	1189.0
335.00	427.0
341.10	168.0
346.00	488.0
352.00	592.0
352.80	199.0
354.00	636.0
364.90	3173.0
365.80	310.0
372.00	1207.0
382.90	181.0
390.20	181.0
392.00	255.0
401.90	264.0
403.20	561.0
404.10	180.0
421.10	477.0
422.00	305.0
423.00	2816.0
424.00	529.0
431.60	197.0
438.60	239.0
441.00	5939.0
442.10	39208.0
443.10	6747.0
443.90	699.0
451.10	203.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28

Scan 2485 (17.289 min): BG018309.D
DFTPP28

m/z	Abundance
35.80	252.0
40.00	157.0
42.00	183.0
48.90	217.0
83.00	166.0
91.00	257.0
129.30	293.0
160.60	192.0
186.10	166.0
207.00	188.0
214.10	158.0
222.00	173.0
342.80	157.0
427.80	245.0

Instrument :
BNA_G
ClientSampleId :
DFTPP28