


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080717\
 Data File : BG028149.D
 Acq On : 7 Aug 2017 10:01
 Operator : SJ/JU
 Sample : DFTPP59
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP59

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Thu Aug 10 05:14:45 2017



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	25176	PASS
68	69	0.00	2	1.9	431	PASS
69	198	0.00	100	34.2	22337	PASS
70	69	0.00	2	1.0	234	PASS
127	198	10	80	40.9	26693	PASS
197	198	0.00	2	0.4	259	PASS
198	198	100	100	100.0	65317	PASS
199	198	5	9	7.1	4653	PASS
275	198	10	60	26.5	17332	PASS
365	198	1	100	3.8	2473	PASS
441	442	0.01	100	16.7	8377	PASS
442	198	50	100	77.0	50264	PASS
443	442	15	24	18.8	9443	PASS

m/z	Abundance
38.10	277.0
39.20	1026.0
40.00	170.0
41.10	243.0
43.10	311.0
44.10	940.0
45.00	192.0
46.90	151.0
49.00	232.0
50.10	4410.0
51.10	17016.0
52.00	774.0
55.10	266.0
56.10	900.0
57.00	1592.0
61.10	315.0
62.10	179.0
63.00	620.0
65.20	399.0
69.00	16052.0
70.20	281.0
71.00	190.0
73.10	312.0
74.00	2231.0
75.00	2771.0
76.10	1121.0
77.10	20096.0
78.10	1528.0
79.00	906.0
80.10	746.0
81.00	1517.0
82.10	540.0
83.10	310.0
84.10	288.0
85.00	343.0
86.10	504.0
87.10	169.0
91.00	457.0
92.00	364.0
93.10	1959.0
93.80	252.0
94.00	250.0
98.00	1804.0
99.00	938.0
100.00	182.0
101.00	829.0
104.00	572.0
105.00	652.0
107.00	6450.0
108.00	664.0
109.10	226.0
110.00	11108.0
111.00	1460.0
111.90	245.0
116.00	179.0
117.00	4040.0
118.20	261.0
121.90	353.0
123.00	786.0
124.00	485.0
125.00	254.0
127.00	19488.0
128.10	1326.0
129.10	8450.0
129.90	673.0
131.10	317.0
132.00	178.0
133.80	185.0
135.00	931.0
136.00	373.0
137.10	283.0
138.10	170.0
139.20	235.0
140.00	184.0
141.00	1078.0
142.00	658.0
142.90	290.0
145.20	291.0
147.00	904.0
148.10	961.0
149.10	408.0
153.00	311.0
153.80	327.0
155.00	747.0
156.10	1075.0
159.10	167.0
160.00	667.0
161.10	537.0
161.60	178.0
165.00	458.0
165.90	462.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

167.00	2737.0
168.10	1137.0
168.90	194.0
171.90	268.0
173.00	370.0
174.00	742.0
175.10	1134.0
176.00	226.0
176.30	191.0
177.00	329.0
179.00	2013.0
180.10	907.0
181.10	705.0
182.50	158.0
185.00	1068.0
186.00	7134.0
187.00	2001.0
188.20	272.0
189.00	372.0
191.10	237.0
192.00	654.0
193.10	626.0
194.10	230.0
195.00	199.0
196.00	1141.0
198.00	53328.0
199.00	3954.0
201.10	264.0
203.00	350.0
204.00	2515.0
205.00	2785.0
206.10	15509.0
207.00	1963.0
208.00	485.0
209.00	211.0
211.10	468.0
216.20	337.0
217.00	4449.0
217.90	614.0
221.00	2026.0
221.80	569.0
222.10	561.0
223.00	1093.0
224.00	8687.0
225.00	2009.0
226.00	204.0
227.00	3708.0
228.00	727.0
229.00	903.0
231.20	332.0
234.10	243.0
235.00	239.0
237.00	489.0
238.90	330.0
240.10	183.0
240.90	325.0
242.10	948.0
243.00	429.0
244.00	7021.0
245.10	1010.0
245.90	1358.0
247.00	262.0
249.00	312.0
253.10	183.0
255.00	38672.0
256.00	5177.0
257.00	705.0
258.00	1981.0
259.00	279.0
265.00	867.0
265.90	162.0
273.00	1118.0
274.00	2736.0
275.00	14941.0
276.00	2334.0
277.00	1131.0
278.10	335.0
284.00	180.0
284.90	268.0
287.90	163.0
292.90	384.0
293.90	312.0
296.00	4709.0
297.10	970.0
300.90	154.0
302.00	169.0
302.70	287.0
313.80	155.0
314.90	641.0
316.00	248.0
323.00	1381.0
324.00	261.0
327.00	354.0
328.10	173.0
332.00	159.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

333.00	253.0
334.00	565.0
335.00	248.0
341.10	162.0
345.80	196.0
352.00	418.0
353.00	494.0
354.00	648.0
354.90	218.0
365.00	2618.0
365.90	365.0
368.00	150.0
370.90	199.0
372.00	620.0
372.80	177.0
391.10	201.0
401.90	403.0
403.00	482.0
404.00	242.0
421.00	481.0
422.00	335.0
423.00	3027.0
423.90	527.0
441.00	8548.0
442.00	50920.0
443.00	9399.0
444.00	741.0
444.90	187.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

m/z	Abundance
38.10	337.0
39.10	2310.0
41.10	363.0
43.10	420.0
44.10	1459.0
45.00	239.0
47.10	155.0
49.10	340.0
50.10	6214.0
51.10	29192.0
52.10	1897.0
53.20	159.0
55.10	403.0
56.10	719.0
57.10	2653.0
58.00	262.0
61.00	362.0
62.00	174.0
63.10	1331.0
65.10	851.0
68.00	573.0
69.00	26728.0
70.00	265.0
73.10	393.0
74.00	2438.0
75.00	4460.0
76.10	1676.0
77.10	30608.0
78.10	2141.0
79.10	1998.0
80.00	1457.0
81.00	2657.0
82.10	661.0
83.10	612.0
83.90	268.0
85.00	629.0
86.00	446.0
87.10	474.0
88.00	172.0
91.00	595.0
92.00	730.0
93.00	3235.0
94.10	235.0
96.20	168.0
98.00	2129.0
99.00	1662.0
101.00	1305.0
103.00	402.0
104.00	871.0
105.10	1032.0
106.10	424.0
107.00	9236.0
108.00	1716.0
110.00	16648.0
111.00	2837.0
112.00	497.0
116.00	592.0
117.10	6043.0
117.90	464.0
120.10	174.0
122.10	508.0
123.10	1206.0
123.90	703.0
125.00	352.0
125.80	225.0
127.00	30128.0
128.00	2719.0
129.00	12219.0
130.00	893.0
131.20	256.0
134.00	411.0
135.10	1180.0
136.10	361.0
137.00	645.0
138.00	151.0
140.10	163.0
141.00	2228.0
141.90	474.0
143.10	281.0
143.90	197.0
145.90	187.0
147.00	1078.0
148.00	2255.0
149.10	756.0
150.00	330.0
151.10	343.0
151.60	165.0
153.00	383.0
154.10	279.0
155.10	601.0
156.00	1649.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

157.00	385.0
158.00	337.0
159.00	454.0
160.00	525.0
161.00	797.0
162.20	297.0
165.00	384.0
165.90	579.0
167.00	2428.0
168.00	1729.0
169.10	349.0
171.10	250.0
172.00	281.0
173.00	416.0
174.00	978.0
175.00	2012.0
176.00	435.0
177.00	708.0
178.00	318.0
178.90	2768.0
180.00	1948.0
181.00	1040.0
184.00	347.0
185.00	1593.0
186.10	12340.0
187.00	2928.0
188.00	322.0
188.90	869.0
190.20	261.0
191.90	1005.0
193.00	1457.0
194.20	188.0
195.00	294.0
196.00	1961.0
196.70	779.0
198.00	75856.0
199.00	5216.0
200.10	583.0
201.10	293.0
201.60	436.0
203.00	1011.0
204.00	3432.0
205.10	6005.0
206.10	21944.0
207.10	2501.0
208.00	956.0
209.10	521.0
210.00	503.0
211.00	991.0
214.00	187.0
215.00	344.0
215.90	593.0
217.00	5704.0
218.00	928.0
218.90	174.0
219.90	171.0
221.10	3197.0
222.00	669.0
223.00	1494.0
224.00	13295.0
225.10	3196.0
226.00	436.0
227.00	5285.0
228.00	931.0
229.00	1240.0
230.90	501.0
232.00	264.0
234.00	364.0
235.00	396.0
236.10	181.0
237.00	506.0
239.00	462.0
241.10	355.0
242.10	737.0
243.10	788.0
244.10	8468.0
245.10	938.0
246.00	1869.0
246.90	528.0
249.00	180.0
253.00	440.0
255.00	46960.0
256.00	7656.0
257.00	592.0
258.00	3838.0
259.00	609.0
265.00	1410.0
272.10	176.0
273.00	1189.0
274.00	3780.0
275.00	20360.0
276.00	2675.0
277.00	1511.0
278.10	383.0
283.00	370.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

284.10	193.0
285.10	573.0
293.00	352.0
295.10	191.0
296.00	6246.0
297.10	827.0
298.00	351.0
303.00	521.0
304.00	281.0
314.00	213.0
315.10	605.0
315.90	301.0
323.10	1484.0
324.00	394.0
328.00	298.0
332.00	164.0
333.10	251.0
334.10	1363.0
335.10	417.0
341.10	223.0
346.00	289.0
352.10	487.0
353.10	564.0
354.10	435.0
358.90	157.0
365.00	3031.0
365.90	492.0
370.90	245.0
372.10	1165.0
373.10	368.0
383.10	212.0
402.00	305.0
403.00	350.0
404.00	269.0
420.80	517.0
421.90	289.0
423.10	2388.0
424.00	621.0
433.80	229.0
441.00	10117.0
442.00	58144.0
443.00	11174.0
444.00	1056.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

m/z	Abundance
38.10	285.0
39.10	1810.0
41.00	451.0
42.00	174.0
43.10	571.0
44.00	938.0
45.10	373.0
49.20	475.0
50.10	7571.0
51.10	29320.0
52.10	1675.0
55.10	315.0
56.00	767.0
57.10	2476.0
60.10	242.0
61.00	340.0
62.10	266.0
63.10	1037.0
65.00	531.0
67.20	162.0
68.00	720.0
69.00	24232.0
70.10	156.0
71.10	178.0
73.00	671.0
74.10	3018.0
75.10	4239.0
76.10	1350.0
77.10	30320.0
78.10	1995.0
79.10	1959.0
80.00	1305.0
81.00	2796.0
81.90	462.0
83.00	761.0
84.10	169.0
85.10	374.0
86.10	805.0
87.10	225.0
88.10	222.0
91.00	677.0
92.00	474.0
93.00	2897.0
94.10	226.0
95.10	207.0
96.10	330.0
98.10	2156.0
99.00	1984.0
100.00	225.0
101.10	1499.0
103.00	373.0
104.00	736.0
105.10	1023.0
106.20	452.0
107.10	9135.0
108.00	1670.0
110.00	15922.0
111.10	3082.0
112.00	195.0
115.90	612.0
116.30	488.0
117.00	6764.0
118.00	638.0
122.10	318.0
123.10	1140.0
124.00	488.0
125.10	369.0
127.00	30464.0
128.00	2132.0
129.00	11399.0
130.10	907.0
134.00	549.0
135.10	1044.0
136.00	367.0
137.10	581.0
137.90	218.0
139.10	315.0
141.00	1343.0
142.00	747.0
143.00	393.0
146.00	546.0
147.00	548.0
148.00	1270.0
149.20	474.0
149.90	184.0
151.00	306.0
152.00	236.0
153.00	685.0
153.90	348.0
155.00	842.0
156.00	1463.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

157.10	481.0
158.00	186.0
160.00	595.0
161.00	777.0
162.10	400.0
165.10	606.0
165.90	533.0
167.00	2436.0
168.10	1744.0
169.00	255.0
169.70	215.0
170.90	194.0
172.00	230.0
173.00	488.0
173.90	815.0
175.10	1480.0
176.10	449.0
177.00	846.0
177.90	206.0
179.00	2679.0
180.00	2062.0
181.00	879.0
184.10	215.0
185.00	1927.0
186.10	9621.0
187.00	2805.0
188.00	439.0
189.00	552.0
191.00	497.0
192.00	681.0
193.10	1356.0
196.10	2241.0
198.00	66768.0
199.00	4791.0
200.10	532.0
201.40	340.0
202.00	185.0
203.10	552.0
204.00	3021.0
205.00	5279.0
206.10	20656.0
207.10	2503.0
208.00	837.0
209.10	346.0
209.90	345.0
211.00	905.0
211.90	529.0
216.10	210.0
217.00	4360.0
218.00	765.0
219.00	183.0
221.10	2346.0
221.80	580.0
223.00	926.0
224.00	10028.0
225.10	2714.0
226.10	454.0
227.00	4768.0
228.00	781.0
229.00	1534.0
229.90	240.0
231.00	494.0
234.00	329.0
235.10	213.0
236.00	314.0
237.10	545.0
240.10	160.0
241.00	347.0
242.10	659.0
243.10	581.0
244.10	7577.0
245.00	1158.0
246.00	1719.0
247.00	402.0
249.00	226.0
252.90	266.0
255.00	41392.0
256.00	6659.0
257.00	559.0
258.00	2685.0
259.10	592.0
265.10	689.0
272.00	172.0
273.00	1266.0
274.00	3406.0
275.00	16696.0
276.00	2384.0
277.00	1633.0
278.10	370.0
284.90	398.0
292.20	154.0
292.90	202.0
296.00	4625.0
297.00	859.0
298.20	233.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59

303.00	553.0
303.90	190.0
314.00	296.0
314.90	589.0
315.90	288.0
321.00	206.0
323.00	1592.0
324.00	179.0
327.00	367.0
327.90	163.0
332.10	191.0
334.00	1083.0
335.00	407.0
341.00	406.0
345.80	465.0
352.00	548.0
353.10	320.0
354.20	656.0
365.00	1771.0
370.10	210.0
372.00	730.0
373.10	336.0
390.10	151.0
402.10	249.0
403.00	651.0
420.80	273.0
421.00	275.0
422.00	324.0
423.00	2644.0
424.00	843.0
441.00	6467.0
442.10	41728.0
443.00	7758.0
444.00	784.0

Instrument :

BNA_G

ClientSampleId :

DFTPP59

m/z	Abundance
40.00	219.0
41.10	210.0
43.10	391.0
44.10	1050.0
45.10	251.0
55.10	401.0
56.00	181.0
57.00	262.0
59.80	186.0
71.10	182.0
84.00	158.0
91.10	270.0
105.10	180.0
119.10	186.0
134.20	188.0
149.00	232.0
192.80	152.0
207.00	392.0
474.40	179.0

Instrument :
BNA_G
ClientSampleId :
DFTPP59