

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080720\
 Data File : BG046188.D
 Acq On : 7 Aug 2020 11:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 07 13:39:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 18:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.524	0.513	2.1	97	0.00
3	Pyridine	1.442	1.417	1.7	95	0.00
4	n-Nitrosodimethylamine	0.614	0.562	8.5	87	0.00
5 S	2-Fluorophenol	1.151	1.162	-1.0	98	0.00
6	Aniline	1.866	1.874	-0.4	99	0.00
7 S	Phenol-d6	1.568	1.600	-2.0	100	0.00
8	2-Chlorophenol	1.272	1.269	0.2	99	0.00
9	Benzaldehyde	0.930	0.870	6.5	91	0.00
10 C	Phenol	1.575	1.586	-0.7	100	0.00
11	bis(2-Chloroethyl)ether	1.253	1.234	1.5	98	0.00
12	1,3-Dichlorobenzene	1.541	1.566	-1.6	101	0.00
13 C	1,4-Dichlorobenzene	1.548	1.547	0.1	99	0.00
14	1,2-Dichlorobenzene	1.460	1.447	0.9	99	0.00
15	Benzyl Alcohol	1.083	1.123	-3.7	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.054	1.869	9.0	91	0.00
17	2-Methylphenol	1.064	1.109	-4.2	102	0.00
18	Hexachloroethane	0.518	0.515	0.6	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.020	1.018	0.2	99	0.00
20	3+4-Methylphenols	1.456	1.505	-3.4	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.513	0.520	-1.4	100	0.00
23 S	Nitrobenzene-d5	0.369	0.372	-0.8	99	0.00
24	Nitrobenzene	0.394	0.385	2.3	96	0.00
25	Isophorone	0.735	0.766	-4.2	101	0.00
26 C	2-Nitrophenol	0.178	0.205	-15.2	107	0.00
27	2,4-Dimethylphenol	0.290	0.310	-6.9	104	0.00
28	bis(2-Chloroethoxy)methane	0.400	0.402	-0.5	99	0.00
29 C	2,4-Dichlorophenol	0.342	0.381	-11.4	106	0.00
30	1,2,4-Trichlorobenzene	0.400	0.414	-3.5	103	0.00
31	Naphthalene	1.058	1.080	-2.1	101	0.00
32	Benzoic acid	0.178	0.226	-27.0#	120	0.00
33	4-Chloroaniline	0.432	0.463	-7.2	104	0.00
34 C	Hexachlorobutadiene	0.266	0.279	-4.9	103	0.00
35	Caprolactam	0.112	0.129	-15.2	108	0.00
36 C	4-Chloro-3-methylphenol	0.335	0.363	-8.4	104	0.00
37	2-Methylnaphthalene	0.814	0.852	-4.7	104	0.00
38	1-Methylnaphthalene	0.771	0.811	-5.2	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	0.721	0.743	-3.1	106	0.00
41 P	Hexachlorocyclopentadiene	0.413	0.418	-1.2	101	0.00
42 S	2,4,6-Tribromophenol	0.307	0.361	-17.6	118	0.00
43 C	2,4,6-Trichlorophenol	0.460	0.508	-10.4	109	0.00
44	2,4,5-Trichlorophenol	0.502	0.548	-9.2	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080720\
 Data File : BG046188.D
 Acq On : 7 Aug 2020 11:07
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 07 13:39:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 18:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.377	1.401	-1.7	106	0.00
46	1,1'-Biphenyl	1.569	1.597	-1.8	106	0.00
47	2-Chloronaphthalene	1.251	1.255	-0.3	104	0.00
48	2-Nitroaniline	0.328	0.346	-5.5	105	0.00
49	Acenaphthylene	1.943	2.014	-3.7	107	0.00
50	Dimethylphthalate	1.535	1.611	-5.0	107	0.00
51	2,6-Dinitrotoluene	0.357	0.382	-7.0	108	0.00
52 C	Acenaphthene	1.283	1.352	-5.4	108	0.00
53	3-Nitroaniline	0.354	0.388	-9.6	107	0.00
54 P	2,4-Dinitrophenol	0.187	0.221	-18.2	120	0.00
55	Dibenzofuran	1.935	2.012	-4.0	108	0.00
56 P	4-Nitrophenol	0.307	0.338	-10.1	110	0.00
57	2,4-Dinitrotoluene	0.491	0.553	-12.6	111	0.00
58	Fluorene	1.567	1.632	-4.1	108	0.00
59	2,3,4,6-Tetrachlorophenol	0.472	0.529	-12.1	113	0.00
60	Diethylphthalate	1.510	1.560	-3.3	106	0.00
61	4-Chlorophenyl-phenylether	0.819	0.859	-4.9	109	0.00
62	4-Nitroaniline	0.356	0.403	-13.2	113	0.00
63	Azobenzene	1.415	1.401	1.0	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	0.132	0.146	-10.6	118	0.00
66 c	n-Nitrosodiphenylamine	0.605	0.623	-3.0	109	0.00
67	4-Bromophenyl-phenylether	0.247	0.262	-6.1	113	0.00
68	Hexachlorobenzene	0.277	0.292	-5.4	113	0.00
69	Atrazine	0.233	0.248	-6.4	111	0.00
70 C	Pentachlorophenol	0.180	0.197	-9.4	117	0.00
71	Phenanthrene	1.096	1.114	-1.6	110	0.00
72	Anthracene	1.090	1.119	-2.7	110	0.00
73	Carbazole	1.051	1.096	-4.3	112	0.00
74	Di-n-butylphthalate	1.098	1.140	-3.8	109	0.00
75 C	Fluoranthene	1.368	1.406	-2.8	112	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
77	Benzidine	0.567	0.563	0.7	100	0.00
78	Pyrene	1.344	1.326	1.3	111	0.00
79 S	Terphenyl-d14	0.982	0.902	8.1	111	0.00
80	Butylbenzylphthalate	0.487	0.509	-4.5	114	0.00
81	Benzo(a)anthracene	1.338	1.334	0.3	114	0.00
82	3,3'-Dichlorobenzidine	0.556	0.585	-5.2	115	0.00
83	Chrysene	1.311	1.287	1.8	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.696	0.691	0.7	112	0.00
85 c	Di-n-octyl phthalate	1.178	1.225	-4.0	114	0.00
86 I	Perylene-d12	1.000	1.000	0.0	117	0.00
87	Indeno(1,2,3-cd)pyrene	1.525	1.569	-2.9	116	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG080720\
Data File : BG046188.D
Acq On : 7 Aug 2020 11:07
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 07 13:39:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG073020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 30 18:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.340	1.366	-1.9	117	0.00
89	Benzo(k)fluoranthene	1.315	1.288	2.1	112	0.00
90 C	Benzo(a)pyrene	1.247	1.274	-2.2	115	0.00
91	Dibenzo(a,h)anthracene	1.275	1.303	-2.2	115	0.00
92	Benzo(q,h,i)perylene	1.288	1.302	-1.1	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0