

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080817\
 Data File : BG028212.D
 Acq On : 9 Aug 2017 5:37
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02054

Quant Time: Aug 09 08:18:23 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 09 02:30:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.451	0.509	-12.9	129	0.00
3 S	1,4-Dioxane-d8	0.429	0.445	-3.7	128	0.00
4	Benzaldehyde	1.269	1.294	-2.0	106	0.00
5 S	Phenol-d5	2.197	1.934	12.0	103	0.00
6	Phenol	2.183	1.943	11.0	103	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.391	1.300	6.5	104	0.00
8	Bis(2-Chloroethyl)ether	1.487	1.397	6.1	106	0.00
9 S	2-Chlorophenol-d4	1.386	1.329	4.1	104	0.00
10	2-Chlorophenol	1.350	1.306	3.3	108	0.00
11	2-Methylphenol	1.541	1.392	9.7	102	0.00
12	2,2'-oxybis(1-Chloropropane	3.610	3.337	7.6	100	0.00
13 S	4-Methylphenol-d8	1.836	1.641	10.6	99	0.00
14	Acetophenone	2.618	2.403	8.2	100	0.00
15 P	N-Nitroso-di-n-propylamine	1.488	1.443	3.0	103	0.00
16	4-Methylphenol	1.740	1.582	9.1	101	0.00
17	Hexachloroethane	0.560	0.555	0.9	109	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
19 S	Nitrobenzene-d5	0.156	0.157	-0.6	107	0.00
20	Nitrobenzene	0.453	0.446	1.5	105	0.00
21	Isophorone	0.860	0.872	-1.4	107	0.00
22 S	2-Nitrophenol-d4	0.171	0.182	-6.4	112	0.00
23 C	2-Nitrophenol	0.173	0.168	2.9	101	0.00
24	2,4-Dimethylphenol	0.430	0.433	-0.7	104	0.00
25	Bis(2-Chloroethoxy)methane	0.469	0.451	3.8	100	0.00
26 S	2,4-Dichlorophenol-d3	0.355	0.349	1.7	105	0.00
27 C	2,4-Dichlorophenol	0.323	0.316	2.2	105	0.00
28	Naphthalene	0.991	0.975	1.6	104	0.00
29 S	4-Chloroaniline-d4	0.395	0.442	-11.9	105	0.00
30	4-Chloroaniline	0.383	0.425	-11.0	105	0.00
31 C	Hexachlorobutadiene	0.238	0.241	-1.3	109	0.00
32	Caprolactam	0.134	0.155	-15.7	122	0.00
33 C	4-Chloro-3-methylphenol	0.412	0.432	-4.9	109	0.00
34	2-Methylnaphthalene	0.796	0.794	0.3	106	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
36	1,2,4,5-Tetrachlorobenzene	0.645	0.606	6.0	108	0.00
37	Hexachlorocyclopentadiene	0.356	0.303	14.9	107	0.00
38 C	2,4,6-Trichlorophenol	0.402	0.416	-3.5	116	0.00
39	2,4,5-Trichlorophenol	0.438	0.426	2.7	109	0.00
40	1,1'-Biphenyl	1.461	1.409	3.6	109	0.00
41	2-Chloronaphthalene	1.154	1.079	6.5	104	0.00
42	2-Nitroaniline	0.484	0.509	-5.2	112	0.00
43 S	Dimethylphthalate-d6	1.779	1.834	-3.1	113	0.00
44	Dimethylphthalate	1.638	1.672	-2.1	114	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080817\
 Data File : BG028212.D
 Acq On : 9 Aug 2017 5:37
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02054

Quant Time: Aug 09 08:18:23 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 09 02:30:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.338	0.357	-5.6	113	0.00
46 S	Acenaphthylene-d8	2.006	1.976	1.5	110	0.00
47	Acenaphthylene	1.917	1.862	2.9	108	0.00
48	3-Nitroaniline	0.305	0.335	-9.8	118	0.00
49 C	Acenaphthene	1.293	1.259	2.6	106	0.00
50	2,4-Dinitrophenol	0.191	0.194	-1.6	132	0.00
51 S	4-Nitrophenol-d4	0.283	0.298	-5.3	119	0.00
52	4-Nitrophenol	0.293	0.308	-5.1	115	0.00
53	Dibenzofuran	1.886	1.868	1.0	110	0.00
54	2,4-Dinitrotoluene	0.512	0.552	-7.8	118	0.00
55	2,3,4,6-Tetrachlorophenol	0.408	0.447	-9.6	122	0.00
56	Diethylphthalate	1.921	1.951	-1.6	114	0.00
57 S	Fluorene-d10	1.596	1.641	-2.8	115	0.00
58	Fluorene	1.669	1.675	-0.4	113	0.00
59	4-Chlorophenyl-phenylether	0.896	0.882	1.6	112	0.00
60	4-Nitroaniline	0.361	0.390	-8.0	122	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.130	0.127	2.3	128	0.00
63	4,6-Dinitro-2-methylphenol	0.132	0.126	4.5	124	0.00
64	N-Nitrosodiphenylamine	0.526	0.502	4.6	112	0.00
65	4-Bromophenyl-phenylether	0.215	0.210	2.3	116	0.00
66	Hexachlorobenzene	0.229	0.230	-0.4	121	0.00
67	Atrazine	0.229	0.230	-0.4	119	0.00
68 C	Pentachlorophenol	0.116	0.110	5.2	130	0.00
69	Phenanthrene	1.007	0.994	1.3	117	0.00
70 S	Anthracene-d10	0.959	0.953	0.6	120	0.00
71	Anthracene	1.029	1.024	0.5	119	0.00
72	Carbazole	0.895	0.896	-0.1	120	0.00
73	Di-n-butylphthalate	1.114	1.111	0.3	116	0.00
74 C	Fluoranthene	1.223	1.224	-0.1	117	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
76 S	Pyrene-d10	1.026	0.994	3.1	117	0.00
77	Pyrene	1.193	1.148	3.8	118	0.00
78	Butylbenzylphthalate	0.447	0.436	2.5	121	0.00
79	3,3'-Dichlorobenzidine	0.386	0.370	4.1	116	0.00
80	Benzo(a)anthracene	1.156	1.134	1.9	122	0.00
81	Bis(2-ethylhexyl)phthalate	0.636	0.618	2.8	123	0.00
82	Chrysene	1.041	1.003	3.7	119	0.00
83 I	Perylene-d12	1.000	1.000	0.0	124	0.00
84	Di-n-octyl phthalate	1.199	1.131	5.7	122	0.00
85	Benzo(b)fluoranthene	1.197	1.196	0.1	128	0.00
86	Benzo(k)fluoranthene	1.169	1.145	2.1	120	0.00
87 S	Benzo(a)pyrene-d12	0.936	0.925	1.2	125	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG080817\
Data File : BG028212.D
Acq On : 9 Aug 2017 5:37
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02054

Quant Time: Aug 09 08:18:23 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 09 02:30:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.159	1.133	2.2	123	0.00
89	Indeno(1,2,3-cd)pyrene	1.357	1.343	1.0	128	0.00
90	Dibenzo(a,h)anthracene	1.137	1.122	1.3	127	0.00
91	Benzo(g,h,i)perylene	1.116	1.112	0.4	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0