


Data Path : Z:\HPCHEM1\BNA G\DATA\BG080915\
 Data File : BG018326.D
 Acq On : 9 Aug 2015 12:55
 Operator : UM/TP
 Sample : DFTPP30
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP30

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 10 07:50:23 2015



AutoFind: Scans 2558, 2559, 2560; Background Corrected with Scan 2553

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.5	51872	PASS
68	69	0.00	2	0.8	379	PASS
69	198	0.00	100	42.6	49646	PASS
70	69	0.00	2	0.6	276	PASS
127	198	10	80	47.8	55618	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	116443	PASS
199	198	5	9	7.4	8563	PASS
275	198	10	60	26.5	30861	PASS
365	198	1	100	3.9	4536	PASS
441	442	0.01	24	16.1	14876	PASS
442	198	50	100	79.5	92589	PASS
443	442	15	24	20.0	18563	PASS

m/z	Abundance
35.90	238.0
39.10	241.0
40.00	1347.0
41.00	970.0
41.90	234.0
43.00	1096.0
44.00	2614.0
46.80	246.0
47.70	150.0
50.00	158.0
53.00	150.0
55.00	883.0
56.00	340.0
56.90	730.0
58.00	893.0
59.00	184.0
64.90	336.0
67.00	432.0
68.00	295.0
68.90	418.0
70.10	460.0
71.10	525.0
73.00	786.0
76.90	307.0
80.90	194.0
82.90	267.0
83.90	620.0
84.90	199.0
86.90	157.0
87.80	226.0
88.90	318.0
90.90	971.0
91.90	177.0
94.90	257.0
95.90	1049.0
97.00	451.0
98.00	289.0
102.90	303.0
104.70	167.0
109.90	272.0
114.80	236.0
119.00	187.0
120.80	166.0
129.10	181.0
129.80	150.0
131.90	225.0
132.90	316.0
134.00	1004.0
134.90	248.0
146.90	538.0
147.70	162.0
161.00	268.0
162.80	249.0
166.90	348.0
176.00	160.0
177.00	310.0
190.90	558.0
191.80	174.0
192.80	389.0
197.70	191.0
204.70	169.0
206.90	5414.0
207.90	629.0
208.80	906.0
218.90	173.0
253.00	159.0
281.00	585.0
281.90	268.0
317.40	152.0
324.50	227.0
504.20	175.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

m/z	Abundance
35.80	161.0
36.20	226.0
36.90	156.0
38.00	842.0
39.00	4430.0
40.00	1582.0
40.90	768.0
42.10	624.0
43.00	1187.0
44.00	2712.0
44.90	325.0
46.10	154.0
47.00	161.0
49.00	701.0
50.00	16063.0
51.00	57808.0
52.00	2478.0
52.90	440.0
54.00	235.0
55.00	1286.0
55.90	2129.0
57.00	4474.0
57.90	876.0
59.00	302.0
61.00	873.0
62.00	648.0
62.90	2480.0
64.00	251.0
64.90	1454.0
66.00	296.0
67.00	357.0
68.00	1226.0
68.90	53440.0
69.90	326.0
70.90	376.0
73.00	1306.0
74.00	4820.0
74.90	7818.0
76.00	3308.0
77.00	62800.0
78.00	3453.0
78.90	4520.0
79.90	3733.0
80.90	5265.0
82.00	1287.0
82.90	1392.0
83.90	262.0
85.00	1080.0
85.90	1303.0
86.90	568.0
88.90	375.0
91.00	2122.0
91.90	1310.0
92.90	7310.0
93.90	707.0
94.90	165.0
96.00	806.0
97.00	869.0
97.90	6558.0
98.90	4529.0
100.00	539.0
100.90	2752.0
101.70	231.0
102.90	787.0
103.80	1607.0
104.90	2065.0
106.00	738.0
106.90	18864.0
107.90	2855.0
108.90	681.0
109.90	28968.0
110.90	4874.0
111.90	373.0
112.70	162.0
113.10	308.0
114.90	376.0
115.80	1019.0
116.90	15939.0
118.00	999.0
119.00	522.0
119.90	264.0
120.90	246.0
121.90	989.0
122.90	1643.0
123.90	875.0
124.80	887.0
126.90	58560.0
127.90	4771.0
128.90	23888.0
129.90	1973.0
130.90	470.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

131.90	269.0
132.90	775.0
134.00	2068.0
134.90	2032.0
135.80	811.0
136.90	961.0
137.90	339.0
139.00	258.0
139.70	433.0
140.90	2769.0
141.90	1125.0
142.90	845.0
144.10	199.0
144.70	187.0
146.00	768.0
147.00	1274.0
147.90	4446.0
148.90	913.0
149.90	364.0
151.00	664.0
152.00	370.0
153.00	1392.0
153.90	871.0
155.00	1689.0
156.00	2542.0
156.90	710.0
157.60	271.0
158.80	494.0
159.90	909.0
161.00	1671.0
161.80	330.0
162.80	262.0
163.90	248.0
164.90	1083.0
165.90	1179.0
166.90	6615.0
167.90	3236.0
168.90	793.0
170.90	419.0
171.80	432.0
172.90	539.0
173.90	1517.0
175.00	2375.0
175.80	806.0
176.90	1477.0
177.90	573.0
178.90	4902.0
179.90	3453.0
181.00	1472.0
181.80	373.0
183.20	201.0
183.80	530.0
184.90	2005.0
185.90	17288.0
186.90	4982.0
187.90	699.0
189.00	1378.0
190.10	215.0
190.90	903.0
191.90	1683.0
192.90	2488.0
193.90	709.0
194.50	285.0
195.90	4215.0
196.60	1351.0
197.90	108440.0
198.90	7968.0
199.90	642.0
201.40	668.0
202.90	1280.0
203.90	5163.0
204.90	7035.0
205.90	30712.0
206.90	10433.0
207.90	2049.0
208.90	1044.0
209.80	557.0
210.30	273.0
210.90	946.0
211.90	221.0
214.90	654.0
215.90	734.0
216.80	8631.0
217.90	948.0
218.90	198.0
220.90	3802.0
221.60	1042.0
222.90	2462.0
223.90	17416.0
224.90	4270.0
225.90	650.0
226.90	7870.0
227.90	1458.0
228.90	1561.0
229.80	271.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

230.80	621.0
231.80	213.0
232.90	293.0
233.90	482.0
234.90	741.0
236.00	366.0
236.80	674.0
237.90	231.0
238.80	321.0
239.80	326.0
240.90	588.0
241.90	993.0
242.90	956.0
243.90	12746.0
245.00	1242.0
245.90	3477.0
246.80	399.0
247.80	262.0
248.90	472.0
249.80	287.0
250.80	328.0
252.20	211.0
252.80	576.0
253.80	741.0
254.90	66896.0
255.90	9838.0
256.90	655.0
257.90	4595.0
258.80	1100.0
260.80	231.0
264.90	2352.0
265.80	316.0
266.70	290.0
268.90	161.0
270.00	252.0
271.90	238.0
272.80	1944.0
273.90	5859.0
274.90	26288.0
275.80	4062.0
276.90	2972.0
277.80	441.0
280.80	680.0
282.90	409.0
284.00	431.0
284.80	376.0
289.90	166.0
292.90	362.0
293.90	165.0
294.90	257.0
295.90	9314.0
296.90	1919.0
297.90	173.0
300.80	210.0
301.90	288.0
302.80	796.0
303.80	217.0
309.00	228.0
309.80	190.0
313.90	517.0
314.90	1266.0
315.70	340.0
320.90	260.0
322.90	2195.0
323.80	416.0
326.80	603.0
328.10	274.0
332.10	442.0
333.90	1811.0
334.80	737.0
340.80	256.0
345.80	712.0
351.90	866.0
352.90	336.0
353.90	723.0
354.90	251.0
364.90	3499.0
365.70	432.0
371.90	1381.0
372.80	369.0
373.50	153.0
382.90	355.0
386.20	326.0
401.80	651.0
403.00	493.0
419.80	169.0
420.90	474.0
421.90	371.0
422.90	4019.0
423.90	886.0
424.80	263.0
440.90	11643.0
441.90	67768.0
442.90	15633.0
443.90	1536.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

m/z	Abundance
37.00	187.0
38.00	916.0
39.00	5576.0
39.90	1373.0
41.10	1168.0
42.00	529.0
43.00	1175.0
44.00	2580.0
44.80	513.0
49.10	371.0
50.00	15157.0
51.00	58064.0
52.00	3529.0
52.90	250.0
54.10	183.0
55.00	1165.0
55.90	1889.0
57.00	4310.0
58.00	1261.0
58.90	393.0
60.00	225.0
60.90	687.0
61.90	935.0
63.00	2078.0
64.00	424.0
65.00	1363.0
66.00	304.0
67.00	418.0
68.90	58504.0
69.90	503.0
71.10	338.0
73.00	1530.0
73.90	6762.0
75.00	8955.0
76.00	3385.0
77.00	66496.0
77.90	4189.0
78.90	3273.0
79.90	3468.0
80.90	4283.0
81.90	1965.0
83.00	1157.0
83.90	347.0
84.90	894.0
85.90	1470.0
86.90	878.0
87.90	348.0
91.00	2117.0
91.90	1168.0
92.90	7576.0
93.90	736.0
95.00	192.0
96.00	1134.0
97.00	484.0
97.90	5706.0
98.90	3906.0
99.90	332.0
100.90	2424.0
102.00	300.0
102.90	962.0
103.90	1874.0
104.90	1606.0
106.90	18808.0
107.90	3032.0
109.00	1064.0
109.90	35312.0
110.90	5345.0
111.90	1036.0
112.90	243.0
114.90	240.0
116.00	1466.0
116.90	16696.0
117.90	1405.0
120.00	152.0
120.90	218.0
121.90	1479.0
122.90	2230.0
123.90	1185.0
124.90	1088.0
126.90	64296.0
127.90	4581.0
128.90	26080.0
129.90	1977.0
130.90	515.0
132.00	162.0
132.90	436.0
133.90	1651.0
134.90	2276.0
136.00	1031.0
136.90	1376.0
138.00	388.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

139.90	180.0
140.90	3545.0
141.90	1200.0
142.80	555.0
143.90	190.0
144.80	281.0
145.90	533.0
146.90	1865.0
147.90	4429.0
148.90	952.0
150.00	176.0
151.00	230.0
151.30	305.0
152.00	186.0
152.80	1241.0
153.90	988.0
154.90	1950.0
155.90	2757.0
157.00	666.0
157.90	460.0
159.00	488.0
159.90	1037.0
161.00	1454.0
162.00	645.0
162.80	287.0
163.90	326.0
164.90	1662.0
165.90	1273.0
166.90	7691.0
167.90	3507.0
168.90	797.0
170.90	324.0
172.00	892.0
172.90	819.0
173.90	2139.0
174.90	2276.0
175.90	817.0
176.90	1674.0
177.80	856.0
178.90	5485.0
179.90	3647.0
180.90	1837.0
182.00	605.0
184.00	957.0
184.90	1917.0
185.90	22640.0
186.90	5723.0
187.80	895.0
188.90	1189.0
189.80	534.0
190.90	1893.0
191.80	1702.0
192.90	2373.0
193.90	548.0
194.90	634.0
195.90	3935.0
197.90	137984.0
198.90	10329.0
199.90	859.0
201.40	365.0
201.80	311.0
202.90	1570.0
203.90	6159.0
204.90	10113.0
205.90	37080.0
206.90	10744.0
207.90	2532.0
208.90	1240.0
209.90	693.0
210.90	2153.0
211.90	344.0
212.90	176.0
214.80	601.0
215.90	702.0
216.90	10617.0
217.90	1315.0
218.60	335.0
219.60	181.0
220.90	5295.0
221.70	1425.0
222.90	2883.0
223.90	23368.0
224.90	5601.0
225.80	719.0
226.90	9793.0
227.90	1171.0
228.90	2060.0
229.60	173.0
230.00	195.0
230.90	900.0
231.80	167.0
232.90	276.0
233.80	808.0
234.80	1124.0
235.90	692.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

236.90	874.0
238.90	342.0
239.70	208.0
240.30	199.0
240.90	539.0
242.00	1225.0
243.00	982.0
243.90	18000.0
244.90	2026.0
245.90	4388.0
247.00	766.0
248.80	1077.0
250.10	253.0
250.90	195.0
251.70	242.0
252.00	241.0
252.80	561.0
254.90	88240.0
255.90	13048.0
256.90	889.0
257.90	6185.0
258.80	1208.0
259.50	215.0
263.90	456.0
264.90	3123.0
265.70	248.0
266.60	215.0
268.00	158.0
270.90	221.0
271.80	228.0
272.90	2471.0
273.90	7092.0
274.90	39248.0
275.90	5758.0
276.90	4015.0
277.90	656.0
278.90	184.0
280.90	588.0
281.90	420.0
282.90	245.0
283.60	231.0
284.00	225.0
284.90	414.0
285.90	216.0
289.80	176.0
290.90	162.0
292.90	628.0
295.90	13724.0
296.90	1335.0
301.00	178.0
302.90	1522.0
308.80	216.0
309.80	361.0
313.90	416.0
314.80	1540.0
315.80	904.0
320.80	377.0
321.60	179.0
322.90	3592.0
324.00	743.0
326.80	545.0
328.00	272.0
332.00	470.0
332.70	515.0
333.90	2449.0
334.90	807.0
340.80	287.0
341.90	167.0
345.90	780.0
346.90	261.0
351.80	1214.0
352.90	1028.0
354.00	1250.0
354.60	241.0
364.90	5511.0
365.80	1151.0
371.90	1518.0
373.00	463.0
373.80	167.0
382.90	746.0
383.80	259.0
389.90	336.0
391.70	233.0
401.90	917.0
402.90	1121.0
403.90	311.0
414.70	157.0
420.90	979.0
422.00	844.0
422.90	5844.0
423.80	1001.0
424.80	179.0
431.80	194.0
434.20	162.0
439.10	192.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

439.80	164.0
440.90	17384.0
441.90	108352.0
442.90	20528.0
443.90	1878.0
445.00	184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

m/z	Abundance
35.90	202.0
36.90	363.0
38.00	769.0
39.00	3138.0
39.90	1103.0
41.00	775.0
41.90	197.0
43.00	909.0
44.00	1972.0
45.00	428.0
46.50	233.0
49.00	422.0
50.00	10345.0
51.00	39744.0
52.00	1451.0
53.00	233.0
53.80	164.0
55.00	1324.0
55.90	1044.0
56.90	3277.0
58.00	1256.0
59.00	190.0
61.00	269.0
61.90	478.0
62.90	2032.0
64.00	492.0
64.90	488.0
67.00	205.0
67.90	797.0
68.90	38248.0
70.10	526.0
70.90	319.0
73.00	703.0
74.00	3897.0
74.90	6575.0
75.90	2122.0
77.00	43608.0
77.90	3018.0
78.90	3099.0
80.00	1531.0
80.90	2466.0
82.00	932.0
83.00	592.0
83.90	788.0
84.90	971.0
85.90	916.0
87.00	546.0
87.70	405.0
88.80	254.0
90.90	1205.0
91.90	1046.0
92.90	4481.0
93.90	296.0
95.10	275.0
96.00	865.0
97.10	336.0
97.90	4246.0
98.90	3468.0
99.80	235.0
100.90	1797.0
102.90	782.0
103.90	1227.0
104.90	1130.0
105.80	664.0
106.90	13534.0
107.90	1949.0
108.90	669.0
109.90	24560.0
110.90	3701.0
111.90	502.0
113.00	213.0
114.90	317.0
116.00	941.0
116.90	12508.0
117.90	1014.0
118.80	165.0
119.90	256.0
121.90	751.0
122.90	1451.0
124.00	414.0
125.00	535.0
126.90	44000.0
127.90	3859.0
128.90	19424.0
129.90	1060.0
130.90	384.0
131.90	456.0
132.90	591.0
134.00	2038.0
134.90	1429.0
136.00	583.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

136.90	702.0
137.80	387.0
138.90	155.0
140.00	221.0
140.90	2758.0
141.80	927.0
142.90	594.0
144.80	353.0
145.30	223.0
145.50	219.0
145.90	524.0
146.90	1604.0
147.90	3085.0
149.00	779.0
149.90	229.0
150.90	527.0
151.90	229.0
152.90	1117.0
153.90	770.0
154.90	1601.0
156.00	1834.0
156.80	287.0
157.90	381.0
159.10	287.0
159.90	866.0
160.90	1461.0
162.80	165.0
163.80	186.0
164.80	1097.0
165.90	1150.0
166.90	4680.0
167.90	3346.0
168.90	502.0
169.80	197.0
171.00	392.0
171.90	239.0
172.90	519.0
173.90	1715.0
174.90	2262.0
175.90	764.0
176.90	1124.0
177.90	635.0
178.90	4156.0
179.90	2762.0
180.90	1203.0
181.70	346.0
182.80	169.0
183.60	292.0
184.90	2535.0
185.90	15228.0
186.90	4125.0
187.90	374.0
188.90	1357.0
189.80	162.0
190.90	1221.0
191.90	1610.0
192.90	2227.0
194.00	407.0
194.80	167.0
195.90	3231.0
197.90	103480.0
198.90	7392.0
199.80	817.0
201.30	507.0
201.90	263.0
202.80	1145.0
203.90	4650.0
204.90	6430.0
205.90	29528.0
206.90	9558.0
207.90	1958.0
209.00	1102.0
209.90	630.0
210.80	1178.0
211.80	434.0
214.90	546.0
215.90	730.0
216.90	8387.0
217.90	1038.0
218.80	196.0
220.10	409.0
220.90	4675.0
221.80	1388.0
222.90	1903.0
223.90	17344.0
225.00	4448.0
226.00	598.0
226.90	7982.0
227.90	968.0
228.90	1896.0
229.90	288.0
230.10	302.0
230.90	926.0
233.80	488.0
234.90	680.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30

236.00	552.0
236.80	671.0
237.90	320.0
238.90	153.0
239.90	438.0
240.90	640.0
241.90	874.0
243.00	1336.0
243.90	12906.0
244.90	2138.0
245.90	2981.0
246.80	663.0
248.80	810.0
249.90	340.0
250.90	230.0
251.80	209.0
252.90	219.0
254.90	71000.0
255.90	10362.0
256.90	756.0
257.90	4404.0
258.80	1034.0
260.70	210.0
263.60	237.0
264.90	1535.0
266.00	260.0
266.70	293.0
267.10	276.0
271.90	269.0
272.90	2639.0
273.90	5723.0
274.90	27048.0
275.90	3059.0
276.90	2512.0
277.80	830.0
280.90	1005.0
281.90	353.0
282.90	515.0
283.90	306.0
285.00	689.0
289.00	170.0
289.70	161.0
292.00	167.0
292.80	280.0
295.90	11043.0
296.90	1410.0
302.90	961.0
303.90	373.0
307.60	175.0
313.90	653.0
314.80	1031.0
315.90	491.0
318.90	168.0
320.70	186.0
322.90	3030.0
323.90	891.0
326.90	737.0
327.80	319.0
332.00	251.0
332.90	401.0
333.90	2121.0
334.90	446.0
340.10	173.0
340.90	389.0
345.80	721.0
351.90	1171.0
352.90	854.0
353.90	1189.0
354.70	188.0
364.00	241.0
364.80	4598.0
365.90	803.0
366.80	224.0
372.00	1398.0
372.90	532.0
382.90	736.0
401.80	844.0
402.80	977.0
403.80	289.0
414.80	151.0
417.10	150.0
420.80	861.0
421.80	783.0
422.90	5373.0
423.90	943.0
424.90	222.0
433.10	181.0
433.80	193.0
436.60	189.0
440.90	15601.0
441.90	101648.0
442.90	19528.0
443.90	1752.0

Instrument :
BNA_G
ClientSampleId :
DFTPP30