


Data Path : Z:\HPCHEM1\BNA_G\Data\BG080917\
 Data File : BG028218.D
 Acq On : 9 Aug 2017 9:50
 Operator : SJ/JU
 Sample : DFTPP61
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP61

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Thu Aug 10 07:27:12 2017



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	34296	PASS
68	69	0.00	2	0.8	272	PASS
69	198	0.00	100	36.2	32714	PASS
70	69	0.00	2	0.4	124	PASS
127	198	10	80	41.1	37064	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	90261	PASS
199	198	5	9	6.8	6107	PASS
275	198	10	60	27.6	24914	PASS
365	198	1	100	3.3	3002	PASS
441	442	0.01	100	16.4	12164	PASS
442	198	50	100	82.3	74266	PASS
443	442	15	24	20.1	14907	PASS

m/z	Abundance
38.20	367.0
39.10	2499.0
40.00	380.0
44.00	567.0
45.00	206.0
49.10	571.0
50.10	8007.0
51.10	35352.0
52.10	1584.0
55.10	334.0
56.00	934.0
57.10	2292.0
58.00	267.0
61.10	757.0
62.10	403.0
63.10	1812.0
65.10	841.0
68.00	816.0
69.00	34400.0
69.80	275.0
74.00	3539.0
75.10	5816.0
76.10	2234.0
77.10	36752.0
78.00	2102.0
79.00	2448.0
80.00	1632.0
81.00	3252.0
82.10	634.0
83.00	641.0
84.00	295.0
85.00	690.0
86.00	1091.0
87.10	408.0
88.80	172.0
91.00	606.0
92.00	1349.0
93.00	4521.0
94.00	196.0
96.00	151.0
98.10	2642.0
99.00	3093.0
100.10	429.0
101.00	1188.0
103.00	347.0
104.00	918.0
105.00	735.0
107.00	11123.0
108.00	1767.0
108.80	294.0
110.00	19592.0
111.00	3152.0
112.10	564.0
112.90	151.0
115.00	155.0
116.00	605.0
117.00	7962.0
118.00	404.0
119.80	252.0
122.00	697.0
122.90	771.0
123.90	525.0
125.10	452.0
127.00	36296.0
128.00	2750.0
129.00	13614.0
130.00	1282.0
132.10	187.0
133.90	370.0
135.00	1095.0
136.00	382.0
136.90	1108.0
141.00	2367.0
142.00	1000.0
143.00	541.0
143.80	269.0
145.10	283.0
146.00	605.0
147.00	2076.0
148.00	2448.0
149.00	637.0
150.00	219.0
151.10	342.0
152.00	164.0
152.90	474.0
154.10	330.0
155.00	1279.0
156.00	1639.0
157.00	545.0
158.00	421.0
158.80	258.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

160.10	960.0
161.00	1077.0
162.10	296.0
164.00	174.0
164.90	632.0
166.00	621.0
167.00	4432.0
168.00	2043.0
169.00	576.0
170.10	262.0
171.00	211.0
172.00	490.0
172.90	476.0
174.00	1055.0
175.10	1999.0
175.90	369.0
177.00	1608.0
178.10	341.0
179.00	3066.0
180.00	2469.0
181.00	1200.0
181.90	232.0
183.10	184.0
184.10	280.0
185.00	1612.0
186.00	12357.0
187.00	3377.0
188.10	630.0
189.00	591.0
190.00	210.0
190.80	256.0
192.00	1123.0
193.00	1065.0
194.90	216.0
196.00	2662.0
198.00	83176.0
199.00	6215.0
199.90	249.0
201.50	212.0
203.10	867.0
204.00	3471.0
205.00	6440.0
206.10	24632.0
207.00	3872.0
208.00	621.0
209.00	283.0
210.00	202.0
211.00	1073.0
212.00	164.0
214.90	466.0
216.10	917.0
217.00	6776.0
218.00	635.0
220.10	351.0
221.10	3269.0
222.00	1065.0
223.00	1247.0
224.00	13491.0
225.00	3711.0
227.00	6025.0
227.90	757.0
228.90	1241.0
230.20	216.0
231.00	414.0
234.10	384.0
235.00	447.0
235.90	416.0
236.90	533.0
239.00	607.0
241.00	296.0
242.00	689.0
243.10	520.0
244.00	8461.0
245.10	1003.0
246.00	1669.0
247.00	525.0
248.60	259.0
249.00	187.0
250.90	184.0
251.90	179.0
253.70	681.0
255.00	55328.0
256.00	8701.0
256.90	292.0
258.00	3067.0
258.90	634.0
260.00	300.0
265.00	1167.0
265.90	269.0
270.90	157.0
271.90	283.0
273.00	1555.0
274.00	3384.0
275.00	20328.0
276.10	3298.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

276.90	1856.0
277.90	250.0
280.00	166.0
283.00	255.0
284.00	314.0
285.00	179.0
290.90	200.0
292.00	246.0
293.00	476.0
294.10	246.0
296.00	6202.0
296.90	941.0
301.00	226.0
302.20	153.0
303.00	692.0
304.00	269.0
314.00	591.0
314.70	544.0
316.00	478.0
320.80	199.0
323.10	1703.0
324.00	709.0
327.10	265.0
333.00	153.0
334.00	1213.0
335.10	370.0
341.10	569.0
346.00	475.0
351.80	393.0
353.10	411.0
353.90	561.0
354.90	223.0
365.00	2828.0
365.90	429.0
372.00	951.0
372.90	286.0
383.00	232.0
401.90	189.0
402.90	318.0
404.00	385.0
420.90	644.0
422.10	530.0
423.00	2723.0
424.00	949.0
441.00	8687.0
442.00	50024.0
443.00	11155.0
444.00	732.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

m/z	Abundance
38.00	325.0
39.10	2233.0
44.10	543.0
47.00	217.0
49.00	454.0
50.10	9478.0
51.10	40976.0
52.10	2164.0
55.20	270.0
56.00	1378.0
57.10	3313.0
58.10	385.0
61.10	416.0
62.10	730.0
63.10	1718.0
64.10	256.0
65.10	1074.0
69.00	38560.0
70.10	374.0
73.10	398.0
74.10	4400.0
75.10	6853.0
76.00	1581.0
77.10	41384.0
78.10	2627.0
79.00	2795.0
80.00	2171.0
81.00	2581.0
82.10	525.0
83.00	617.0
84.10	381.0
85.00	654.0
86.10	1345.0
87.00	227.0
88.00	156.0
90.80	519.0
92.10	746.0
93.00	4712.0
94.00	420.0
94.80	167.0
98.00	3083.0
99.00	2989.0
100.00	404.0
101.00	1561.0
101.80	165.0
103.00	541.0
104.00	1432.0
105.00	801.0
106.00	461.0
107.00	12233.0
108.00	2119.0
109.00	631.0
110.00	23944.0
111.00	3630.0
111.70	270.0
112.00	268.0
113.00	263.0
115.20	164.0
116.00	851.0
117.00	8933.0
118.00	526.0
119.30	168.0
119.90	232.0
122.00	937.0
123.10	1185.0
124.10	524.0
124.90	606.0
127.00	44744.0
128.10	3891.0
129.00	17296.0
130.00	1414.0
130.90	378.0
133.90	468.0
135.00	1672.0
136.00	791.0
137.10	789.0
138.00	295.0
140.20	385.0
141.00	2574.0
141.90	517.0
142.90	887.0
145.90	410.0
147.00	1643.0
148.00	2185.0
149.00	826.0
150.00	193.0
150.90	251.0
152.90	627.0
154.00	573.0
155.10	1763.0
156.00	2637.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

157.00	442.0
157.80	288.0
159.10	485.0
160.00	1040.0
161.00	1090.0
161.90	408.0
162.80	296.0
164.00	222.0
165.00	1323.0
166.10	897.0
167.00	4267.0
168.00	1914.0
169.10	539.0
172.00	575.0
173.10	1039.0
174.00	1554.0
175.00	1537.0
176.30	436.0
177.00	1122.0
178.10	417.0
179.00	3712.0
180.00	2743.0
181.00	1621.0
182.10	293.0
182.70	154.0
184.00	467.0
185.00	1793.0
186.00	17432.0
187.00	4672.0
187.90	605.0
189.10	1061.0
191.00	238.0
192.00	1556.0
193.10	1885.0
194.00	252.0
195.00	350.0
196.10	2828.0
198.00	105672.0
199.00	6821.0
199.90	658.0
201.60	264.0
203.10	1272.0
204.00	4139.0
205.00	8266.0
206.00	30336.0
207.10	4354.0
208.00	1319.0
209.00	517.0
210.00	446.0
211.20	1003.0
212.00	318.0
215.00	433.0
216.20	825.0
217.00	8128.0
218.10	900.0
219.00	357.0
220.10	326.0
221.00	4184.0
222.00	1623.0
223.00	1534.0
224.00	16528.0
225.00	4345.0
226.10	722.0
227.00	8302.0
228.00	1004.0
229.00	1588.0
230.00	360.0
231.00	1212.0
233.00	177.0
233.90	823.0
234.90	500.0
235.90	352.0
236.90	950.0
238.90	186.0
239.90	276.0
240.90	253.0
241.20	237.0
242.00	1041.0
243.00	894.0
244.00	11705.0
245.00	1871.0
246.00	2894.0
247.00	529.0
247.90	240.0
249.10	391.0
249.80	158.0
251.00	249.0
252.00	356.0
253.10	432.0
255.00	71896.0
256.00	9716.0
257.00	1046.0
258.00	5239.0
259.00	820.0
259.90	183.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

264.10	212.0
265.00	1490.0
266.10	284.0
270.00	324.0
270.60	262.0
271.10	183.0
272.20	261.0
272.90	2870.0
274.00	5306.0
275.00	28032.0
276.00	4332.0
277.00	2751.0
278.10	512.0
280.90	170.0
282.90	552.0
284.00	353.0
284.90	453.0
290.90	240.0
293.00	550.0
293.70	188.0
296.00	8659.0
296.90	782.0
300.90	248.0
302.00	458.0
303.00	831.0
304.00	275.0
313.90	334.0
315.00	626.0
316.00	616.0
321.00	305.0
322.00	205.0
323.00	2804.0
324.00	459.0
326.80	415.0
328.00	320.0
332.10	230.0
333.10	231.0
334.00	1707.0
335.00	330.0
340.00	201.0
341.00	544.0
342.00	302.0
346.00	327.0
352.00	1026.0
353.10	899.0
354.00	794.0
355.10	216.0
365.00	3309.0
366.00	582.0
367.70	160.0
371.10	307.0
372.00	1329.0
373.00	549.0
382.90	358.0
401.00	203.0
402.00	582.0
403.00	631.0
403.90	265.0
421.00	815.0
422.00	651.0
423.00	4421.0
424.00	832.0
441.00	14757.0
442.00	89208.0
443.00	16752.0
444.00	1323.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

m/z	Abundance
38.00	270.0
39.10	1328.0
41.00	166.0
44.00	665.0
49.00	412.0
50.10	5755.0
51.10	26560.0
52.10	1302.0
53.10	153.0
55.00	306.0
56.00	716.0
57.10	2564.0
58.00	191.0
61.10	228.0
62.10	406.0
63.10	939.0
64.00	158.0
65.10	704.0
69.00	25184.0
70.30	230.0
73.00	166.0
74.10	2647.0
75.10	3958.0
76.20	1381.0
77.10	29216.0
78.10	2017.0
79.00	1935.0
80.00	1826.0
81.00	2378.0
82.00	375.0
82.90	313.0
83.80	196.0
85.10	338.0
86.00	730.0
87.00	422.0
88.00	225.0
91.10	361.0
92.00	437.0
93.10	3068.0
93.90	405.0
95.90	205.0
98.00	2984.0
99.00	2036.0
99.80	317.0
101.10	1214.0
103.10	444.0
104.00	777.0
105.00	625.0
106.20	389.0
107.00	10390.0
108.00	1373.0
109.00	191.0
110.00	15756.0
111.10	2678.0
112.10	181.0
116.20	555.0
117.00	6844.0
118.10	664.0
119.10	239.0
121.90	534.0
123.10	1275.0
124.00	414.0
125.00	475.0
127.00	30152.0
128.00	2430.0
129.00	13789.0
130.00	1169.0
130.90	323.0
134.00	500.0
134.90	1169.0
136.10	493.0
137.00	497.0
140.10	164.0
141.00	1865.0
142.10	657.0
142.90	555.0
146.00	395.0
147.00	1133.0
148.00	2073.0
149.10	429.0
150.10	182.0
151.00	251.0
152.10	315.0
153.00	938.0
154.00	423.0
155.10	953.0
156.00	1624.0
156.90	659.0
158.00	312.0
159.10	377.0
160.00	433.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

160.90	956.0
162.10	267.0
165.00	640.0
166.00	793.0
167.10	2866.0
168.00	1373.0
169.10	336.0
169.90	195.0
171.90	402.0
173.00	357.0
174.00	1189.0
175.00	1948.0
176.00	651.0
177.00	1259.0
178.00	401.0
179.00	2767.0
180.00	2089.0
181.00	1404.0
182.20	176.0
183.00	158.0
183.90	351.0
185.00	1266.0
186.00	11399.0
187.10	2697.0
187.90	402.0
189.00	923.0
191.10	351.0
192.10	722.0
193.10	931.0
193.90	305.0
195.10	164.0
196.00	1719.0
198.00	81936.0
199.00	5287.0
199.90	383.0
201.50	425.0
203.00	604.0
204.00	3804.0
205.00	4870.0
206.00	22656.0
207.00	2847.0
208.00	955.0
208.90	441.0
210.30	483.0
211.00	1014.0
212.00	187.0
214.70	359.0
215.90	399.0
217.00	5863.0
218.00	486.0
221.00	2527.0
222.00	556.0
222.90	1703.0
224.00	12589.0
225.00	3529.0
227.00	4919.0
228.00	1261.0
228.90	1298.0
230.90	600.0
233.10	271.0
234.00	504.0
235.00	545.0
236.00	285.0
237.00	438.0
238.90	367.0
240.10	158.0
241.10	368.0
242.00	637.0
243.00	825.0
244.00	9994.0
245.00	872.0
246.00	2305.0
246.90	213.0
249.00	370.0
249.90	207.0
251.90	269.0
253.00	330.0
253.50	391.0
255.00	59664.0
256.00	7444.0
257.00	1066.0
258.00	4017.0
259.10	529.0
263.90	237.0
265.00	1324.0
266.00	169.0
271.20	160.0
271.70	252.0
273.00	1951.0
274.00	4126.0
275.00	26384.0
276.00	2853.0
276.90	2614.0
278.00	301.0
281.20	172.0

Instrument :

BNA_G

ClientSampleId :

DFTPP61

283.00	304.0
283.90	382.0
285.00	331.0
293.00	484.0
295.10	525.0
296.00	7020.0
296.90	1053.0
303.00	1063.0
304.00	264.0
313.80	474.0
315.00	666.0
316.00	540.0
321.90	179.0
323.00	2656.0
323.90	450.0
326.90	415.0
327.90	204.0
331.90	551.0
332.80	270.0
334.00	1860.0
335.10	653.0
345.80	206.0
351.90	873.0
352.80	454.0
354.00	586.0
355.10	186.0
363.90	184.0
365.00	2871.0
365.90	486.0
372.00	1556.0
373.00	250.0
382.90	417.0
383.70	160.0
391.10	270.0
392.00	155.0
401.80	424.0
403.00	498.0
404.00	393.0
420.90	813.0
421.90	563.0
423.00	3914.0
423.90	960.0
441.00	13050.0
442.00	83568.0
443.00	16816.0
444.00	1269.0
445.00	256.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

Scan 2465 (18.070 min): BG028218.D
DFTPP61

m/z	Abundance
44.00	733.0
49.00	380.0
55.10	190.0
83.80	183.0
138.10	159.0
207.90	173.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61