

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081016\
 Data File : BG023619.D
 Acq On : 11 Aug 2016 8:16
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025

Quant Time: Aug 11 08:49:56 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 03:58:16 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.11	152	152272	20.00	ng/ul	0.00
18) Naphthalene-d8	10.95	136	700783	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.79	164	481131	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.55	188	1078177	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	999493	20.00	ng/ul	0.00
83) Perylene-d12	25.28	264	1007235	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.49	96	23265	6.50	ng/uL	0.00
5) Phenol-d5	7.27	99	299961	19.55	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.43	67	198547	20.20	ng/ul	0.00
9) 2-Chlorophenol-d4	7.64	132	216943	20.73	ng/ul	0.00
13) 4-Methylphenol-d8	8.83	113	244002	20.38	ng/ul	0.00
19) Nitrobenzene-d5	9.29	128	114182	20.67	ng/ul	0.00
22) 2-Nitrophenol-d4	10.02	143	135398	21.34	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.57	165	255450	21.35	ng/ul	0.00
29) 4-Chloroaniline-d4	11.09	131	325735	23.20	ng/ul	0.00
43) Dimethylphthalate-d6	14.18	166	784478	19.91	ng/ul	0.00
46) Acenaphthylene-d8	14.48	160	943148	21.27	ng/ul	0.00
51) 4-Nitrophenol-d4	15.00	143	133437	19.84	ng/ul	0.00
57) Fluorene-d10	15.78	176	709161	19.48	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.91	200	136193	19.69	ng/ul	0.00
70) Anthracene-d10	17.65	188	1045525	21.52	ng/ul	0.00
76) Pyrene-d10	19.94	212	1041477	20.59	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.04	264	944451	20.72	ng/ul	0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.52	88	27715	7.38	ng/uL#	67
6) Phenol	7.30	94	328384	20.56	ng/ul#	76
8) Bis(2-Chloroethyl)ether	7.52	93	231899	20.00	ng/ul	90
10) 2-Chlorophenol	7.67	128	218282	20.48	ng/ul#	86
11) 2-Methylphenol	8.56	108	241423	20.07	ng/ul	94
12) 2,2'-oxybis(1-Chloropropan	8.64	45	334302	21.07	ng/ul	99
14) Acetophenone	8.94	105	388740	20.24	ng/ul#	85
15) N-Nitroso-di-n-propylamine	8.92	70	226788	19.63	ng/ul#	86
16) 4-Methylphenol	8.89	108	267971	19.87	ng/ul	96
17) Hexachloroethane	9.20	117	92985	20.22	ng/ul	95
20) Nitrobenzene	9.34	77	327468	20.22	ng/ul#	83
21) Isophorone	9.85	82	609378	20.46	ng/ul#	91
23) 2-Nitrophenol	10.05	139	140069	20.45	ng/ul#	64
24) 2,4-Dimethylphenol	10.11	107	307131	20.44	ng/ul	86
25) Bis(2-Chloroethoxy)methane	10.34	93	344147	20.03	ng/ul	98
27) 2,4-Dichlorophenol	10.59	162	253211	20.98	ng/ul	100
28) Naphthalene	11.00	128	729134	20.49	ng/ul	97
30) 4-Chloroaniline	11.11	127	304829	21.88	ng/ul	99
31) Hexachlorobutadiene	11.27	225	171078	21.29	ng/ul	91
32) Caprolactam	11.87	113	77273	15.42	ng/ul#	42
33) 4-Chloro-3-methylphenol	12.24	107	305030	19.64	ng/ul#	83
34) 2-Methylnaphthalene	12.61	142	578424	20.05	ng/ul	96
36) 1,2,4,5-Tetrachlorobenzene	12.97	216	331574	22.45	ng/ul	98


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081016\
 Data File : BG023619.D
 Acq On : 11 Aug 2016 8:16
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025

Quant Time: Aug 11 08:49:56 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 03:58:16 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) Hexachlorocyclopentadiene	12.95	237	143104	17.52	ng/ul	97
38) 2,4,6-Trichlorophenol	13.22	196	221588	21.66	ng/ul	94
39) 2,4,5-Trichlorophenol	13.30	196	235194	21.44	ng/ul	99
40) 1,1'-Biphenyl	13.61	154	765615	21.65	ng/ul	97
41) 2-Chloronaphthalene	13.66	162	594143	21.61	ng/ul	98
42) 2-Nitroaniline	13.87	65	234643	20.90	ng/ul#	68
44) Dimethylphthalate	14.23	163	760751	19.50	ng/ul	98
45) 2,6-Dinitrotoluene	14.36	165	171146	20.24	ng/ul#	89
47) Acenaphthylene	14.51	152	976263	21.05	ng/ul	99
48) 3-Nitroaniline	14.69	138	170634	21.46	ng/ul#	69
49) Acenaphthene	14.85	153	662253	20.96	ng/ul	96
50) 2,4-Dinitrophenol	14.90	184	83259	15.95	ng/ul#	81
53) Dibenzofuran	15.19	168	937098	20.02	ng/ul#	88
54) 2,4-Dinitrotoluene	15.15	165	250362	19.54	ng/ul#	71
55) 2,3,4,6-Tetrachlorophenol	15.42	232	219786	19.51	ng/ul	96
56) Diethylphthalate	15.59	149	787116	19.31	ng/ul	95
58) Fluorene	15.84	166	763102	19.57	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.83	204	390388	19.90	ng/ul	99
60) 4-Nitroaniline	15.87	138	188624	20.46	ng/ul#	71
63) 4,6-Dinitro-2-methylphenol	15.92	198	140925	19.41	ng/ul#	88
64) N-Nitrosodiphenylamine	16.04	169	681470	22.62	ng/ul	99
65) 4-Bromophenyl-phenylether	16.73	248	267070	21.60	ng/ul	96
66) Hexachlorobenzene	16.85	284	286995	21.79	ng/ul	97
67) Atrazine	17.00	200	272756	21.92	ng/ul	96
68) Pentachlorophenol	17.20	266	149062	20.51	ng/ul	89
69) Phenanthrene	17.60	178	1185530	21.23	ng/ul	97
71) Anthracene	17.69	178	1237922	21.73	ng/ul	98
72) Carbazole	17.96	167	1059872	23.09	ng/ul	98
73) Di-n-butylphthalate	18.50	149	1274348	22.16	ng/ul#	95
74) Fluoranthene	19.60	202	1280553	22.40	ng/ul#	93
77) Pyrene	19.97	202	1290333	20.88	ng/ul#	93
78) Butylbenzylphthalate	20.84	149	546679	23.07	ng/ul#	85
79) 3,3'-Dichlorobenzidine	21.76	252	424255	24.09	ng/ul	98
80) Benzo(a)anthracene	21.85	228	1183997	21.03	ng/ul	97
81) Bis(2-ethylhexyl)phthalate	21.73	149	752670	23.88	ng/ul#	95
82) Chrysene	21.92	228	1076013	21.02	ng/ul	97
84) Di-n-octyl phthalate	23.00	149	1269586	24.96	ng/ul	100
85) Benzo(b)fluoranthene	24.19	252	1180408	20.60	ng/ul#	94
86) Benzo(k)fluoranthene	24.26	252	1140070	20.21	ng/ul#	90
88) Benzo(a)pyrene	25.11	252	1151207	20.81	ng/ul#	94
89) Indeno(1,2,3-cd)pyrene	29.17	276	1367197	21.04	ng/ul#	89
91) Benzo(g,h,i)perylene	30.39	276	1127487	20.92	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.38 ng/uL

RT: 3.52 min Scan# 116

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025





Tgt Ion: 88 Resp: 27715

Ion Ratio Lower Upper

88 100

43 32.9 16.6 24.8#

58 85.9 47.2 70.8#



#6

Phenol

Concen: 20.56 ng/uL

RT: 7.30 min Scan# 759

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16




Tgt Ion: 94 Resp: 328384


Ion Ratio Lower Upper

94 100

65 30.6 16.8 25.2#

66 41.9 22.6 33.8#





#8

Bis(2-Chloroethyl)ether

Concen: 20.00 ng/ul

RT: 7.52 min Scan# 797

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025





Tgt Ion: 93 Resp: 231899

Ion Ratio Lower Upper

93 100

63 79.3 56.0 84.0

95 30.4 27.7 41.5



#10

2-Chlorophenol

Concen: 20.48 ng/ul

RT: 7.67 min Scan# 823

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16




Tgt Ion: 128 Resp: 218282


Ion Ratio Lower Upper

128 100

64 57.9 34.2 51.4#

130 32.2 27.0 40.4









#11
 2-Methylphenol
 Concen: 20.07 ng/ul
 RT: 8.56 min Scan# 974
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion: 108 Resp: 241423
 Ion Ratio Lower Upper
 108 100
 107 92.6 78.6 117.8



#12
 2,2'-oxybis(1-Chloropropane)
 Concen: 21.07 ng/ul
 RT: 8.64 min Scan# 988
 Delta R.T. -0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion: 45 Resp: 334302
 Ion Ratio Lower Upper
 45 100
 77 16.2 13.4 20.2
 79 12.2 10.2 15.4





#14

Acetophenone

Concen: 20.24 ng/ul

RT: 8.94 min Scan# 1039

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion: 105 Resp: 388740

Ion Ratio Lower Upper

105 100

77 82.9 56.1 84.1




51 25.0 14.6 22.0#



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#15

N-Nitroso-di-n-propylamine

Concen: 19.63 ng/ul

RT: 8.92 min Scan# 1035

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Tgt Ion: 70 Resp: 226788


Ion Ratio Lower Upper

70 100

42 49.3 31.4 47.2#

101 8.9 8.6 13.0

130 17.2 19.0 28.4#






Abundance Ion 70.00 (69.70 to 70.70): BGC

Ion 42.00 (41.70 to 42.70): BGC

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E









#16
4-Methylphenol
Concen: 19.87 ng/ul
RT: 8.89 min Scan# 1030
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion:108 Resp: 267971
Ion Ratio Lower Upper
108 100
107 106.4 88.3 132.5



#17
Hexachloroethane
Concen: 20.22 ng/ul
RT: 9.20 min Scan# 1083
Delta R.T. -0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion:117 Resp: 92985
Ion Ratio Lower Upper
117 100
201 97.6 76.8 115.2
199 62.4 43.5 65.3









#20
 Nitrobenzene
 Concen: 20.22 ng/ul
 RT: 9.34 min Scan# 1106
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion: 77 Resp: 327468
 Ion Ratio Lower Upper
 77 100
 123 36.6 39.5 59.3#
 65 14.6 9.0 13.4#



#21
 Isophorone
 Concen: 20.46 ng/ul
 RT: 9.85 min Scan# 1194
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion: 82 Resp: 609378
 Ion Ratio Lower Upper
 82 100
 95 7.5 5.1 7.7
 138 13.7 15.0 22.6#









#23
2-Nitrophenol
Concen: 20.45 ng/ul
RT: 10.05 min Scan# 1228
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion	Ratio	Lower	Upper
139	100		
65	63.8	31.8	47.8#
109	35.9	17.1	25.7#



#24
2,4-Dimethylphenol
Concen: 20.44 ng/ul
RT: 10.11 min Scan# 1238
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion	Ratio	Lower	Upper
107	100		
121	44.1	43.7	65.5
122	75.6	70.3	105.5





#25

Bis(2-Chloroethoxy)methane

Concen: 20.03 ng/ul

RT: 10.34 min Scan# 1277

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025





Tgt Ion: 93 Resp: 344147

Ion Ratio Lower Upper

93 100

95 32.6 26.0 39.0

123 10.6 10.1 15.1



#27

2,4-Dichlorophenol

Concen: 20.98 ng/ul

RT: 10.59 min Scan# 1320

Delta R.T. -0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16




Tgt Ion: 162 Resp: 253211


Ion Ratio Lower Upper

162 100

164 64.6 51.6 77.4

98 35.8 28.3 42.5





#28

Naphthalene

Concen: 20.49 ng/ul

RT: 11.00 min Scan# 1389

Delta R.T. -0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:128 Resp: 729134

Ion Ratio Lower Upper

128 100

129 10.5 9.4 14.2




127 12.8 11.2 16.8



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#30

4-Chloroaniline

Concen: 21.88 ng/ul

RT: 11.11 min Scan# 1408

Delta R.T. 0.00 min

Lab File: BG023619.D


Acq: 11 Aug 2016 8:16

Tgt Ion:127 Resp: 304829

Ion Ratio Lower Upper



127 100


129 33.0 26.9 40.3



Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E





#31

Hexachlorobutadiene

Concen: 21.29 ng/ul

RT: 11.27 min Scan# 1436

Delta R.T. -0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion: 225 Resp: 171078

Ion Ratio Lower Upper

225 100

223 59.1 54.7 82.1




227 60.3 52.2 78.4



Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#32

Caprolactam

Concen: 15.42 ng/ul

RT: 11.87 min Scan# 1538

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion: 113 Resp: 77273

Ion Ratio Lower Upper

113 100

55 191.1 99.6 149.4#



56 134.2 66.4 99.6#




Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC









#33
 4-Chloro-3-methylphenol
 Concen: 19.64 ng/ul
 RT: 12.24 min Scan# 1600
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion:107 Resp: 305030
 Ion Ratio Lower Upper
 107 100
 144 21.8 22.7 34.1#
 142 72.1 71.6 107.4



#34
 2-Methylnaphthalene
 Concen: 20.05 ng/ul
 RT: 12.61 min Scan# 1663
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion:142 Resp: 578424
 Ion Ratio Lower Upper
 142 100
 141 90.0 75.2 112.8









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 22.45 ng/ul
 RT: 12.97 min Scan# 1725
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion:	216	Resp:	331574
Ion	Ratio	Lower	Upper
216	100		
214	79.4	61.7	92.5
179	22.3	17.0	25.4
108	16.8	14.2	21.2



#37
 Hexachlorocyclopentadiene
 Concen: 17.52 ng/ul
 RT: 12.95 min Scan# 1721
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion:	237	Resp:	143104
Ion	Ratio	Lower	Upper
237	100		
235	59.8	49.6	74.4
272	10.6	9.3	13.9









#38
 2,4,6-Trichlorophenol
 Concen: 21.66 ng/ul
 RT: 13.22 min Scan# 1767
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion	Ratio	Resp	Lower	Upper
196	100	221588		
198	104.6	77.4	116.0	
200	34.1	26.6	39.8	



#39
 2,4,5-Trichlorophenol
 Concen: 21.44 ng/ul
 RT: 13.30 min Scan# 1781
 Delta R.T. 0.01 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	235194		
198	99.5	79.8	119.8	
200	31.1	24.0	36.0	









#40
 1,1'-Biphenyl
 Concen: 21.65 ng/ul
 RT: 13.61 min Scan# 1834
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion:	154	Resp:	765615
Ion	Ratio	Lower	Upper
154	100		
153	41.0	34.6	52.0
76	14.5	11.9	17.9



#41
 2-Chloronaphthalene
 Concen: 21.61 ng/ul
 RT: 13.66 min Scan# 1842
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion:	162	Resp:	594143
Ion	Ratio	Lower	Upper
162	100		
164	32.9	25.4	38.2
127	38.9	30.2	45.4









#42
2-Nitroaniline
Concen: 20.90 ng/ul
RT: 13.87 min Scan# 1878
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion: 65 Resp: 234643
Ion Ratio Lower Upper
65 100
92 66.8 67.1 100.7#
138 89.7 109.6 164.4#



#44
Dimethylphthalate
Concen: 19.50 ng/ul
RT: 14.23 min Scan# 1939
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion: 163 Resp: 760751
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.1 8.8 13.2





#45

2,6-Dinitrotoluene

Concen: 20.24 ng/ul

RT: 14.36 min Scan# 1961

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:165 Resp: 171146

Ion Ratio Lower Upper

165 100

89 62.6 41.5 62.3#




121 20.6 17.0 25.4



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 89.00 (88.70 to 89.70): BG

Ion 121.00 (120.70 to 121.70): E



#47

Acenaphthylene

Concen: 21.05 ng/ul

RT: 14.51 min Scan# 1987

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:152 Resp: 976263

Ion Ratio Lower Upper

152 100

151 22.2 17.5 26.3



153 14.4 11.1 16.7




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E









#48
 3-Nitroaniline
 Concen: 21.46 ng/ul
 RT: 14.69 min Scan# 2018
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02025


Tgt Ion:138 Resp: 170634
 Ion Ratio Lower Upper
 138 100
 108 12.9 6.2 9.2#
 92 131.7 79.3 118.9#



#49
 Acenaphthene
 Concen: 20.96 ng/ul
 RT: 14.85 min Scan# 2045
 Delta R.T. 0.01 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion:153 Resp: 662253
 Ion Ratio Lower Upper
 153 100
 152 46.9 37.3 55.9
 154 90.5 68.2 102.2









#50
2,4-Dinitrophenol
Concen: 15.95 ng/ul
RT: 14.90 min Scan# 2054
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion:184 Resp: 83259
Ion Ratio Lower Upper
184 100
63 75.4 40.5 60.7#
154 64.4 48.5 72.7



#53
Dibenzofuran
Concen: 20.02 ng/ul
RT: 15.19 min Scan# 2102
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion:168 Resp: 937098
Ion Ratio Lower Upper
168 100
139 41.2 27.4 41.0#









#54
 2,4-Dinitrotoluene
 Concen: 19.54 ng/ul
 RT: 15.15 min Scan# 2096
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampleId :
 SST02025


Tgt Ion:165 Resp: 250362
 Ion Ratio Lower Upper
 165 100
 63 49.5 25.7 38.5#
 182 2.2 2.1 3.1



#55
 2,3,4,6-Tetrachlorophenol
 Concen: 19.51 ng/ul
 RT: 15.42 min Scan# 2141
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion:232 Resp: 219786
 Ion Ratio Lower Upper
 232 100
 131 46.8 34.8 52.2
 130 2.2 2.2 3.2
 166 30.7 23.8 35.6





#56

Diethylphthalate

Concen: 19.31 ng/ul

RT: 15.59 min Scan# 2171

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:149 Resp: 787116

Ion Ratio Lower Upper

149 100

177 24.2 17.4 26.2




150 14.0 9.6 14.4



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#58

Fluorene

Concen: 19.57 ng/ul

RT: 15.84 min Scan# 2213

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:166 Resp: 763102

Ion Ratio Lower Upper

166 100

165 99.9 80.3 120.5



167 14.1 10.3 15.5




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
4-Chlorophenyl-phenylether
Concen: 19.90 ng/ul
RT: 15.83 min Scan# 2211
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion	Ratio	Lower	Upper
204	100		
206	33.9	27.4	41.2
141	59.8	47.1	70.7



#60
4-Nitroaniline
Concen: 20.46 ng/ul
RT: 15.87 min Scan# 2218
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion	Ratio	Lower	Upper
138	100		
92	59.2	39.0	58.4#
108	91.4	49.2	73.8#









#63
4,6-Dinitro-2-methylphenol
Concen: 19.41 ng/ul
RT: 15.92 min Scan# 2227
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Instrument :
BNA_G
ClientSampleId :
SSTD02025

Tgt Ion:198 Resp: 140925
Ion Ratio Lower Upper
198 100
182 3.5 4.2 6.2#
77 27.1 16.7 25.1#




Abundance Ion 198.00 (197.70 to 198.70): E
Ion 182.00 (181.70 to 182.70): E
Ion 77.00 (76.70 to 77.70): BGC






#64
N-Nitrosodiphenylamine
Concen: 22.62 ng/ul
RT: 16.04 min Scan# 2248
Delta R.T. 0.01 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion:169 Resp: 681470
Ion Ratio Lower Upper
169 100
168 73.5 59.0 88.6
167 37.2 30.6 46.0



Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E









#65
4-Bromophenyl-phenylether
Concen: 21.60 ng/ul
RT: 16.73 min Scan# 2364
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion:248 Resp: 267070
Ion Ratio Lower Upper
248 100
250 95.2 76.6 114.8
141 69.5 61.8 92.6



#66
Hexachlorobenzene
Concen: 21.79 ng/ul
RT: 16.85 min Scan# 2385
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion:284 Resp: 286995
Ion Ratio Lower Upper
284 100
142 32.7 27.8 41.8
249 31.3 24.4 36.6









#67
 Atrazine
 Concen: 21.92 ng/ul
 RT: 17.00 min Scan# 2410
 Delta R.T. 0.01 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16




Instrument :
 BNA_G
 ClientSampled :
 SSTD02025


Tgt Ion	Ratio	Lower	Upper
200	100		
173	23.0	21.0	31.6
215	46.1	37.9	56.9



#68
 Pentachlorophenol
 Concen: 20.51 ng/ul
 RT: 17.20 min Scan# 2445
 Delta R.T. 0.00 min
 Lab File: BG023619.D
 Acq: 11 Aug 2016 8:16

Tgt Ion	Ratio	Lower	Upper
266	100		
264	58.9	54.7	82.1
268	64.0	57.8	86.8





#69

Phenanthrene

Concen: 21.23 ng/ul

RT: 17.60 min Scan# 2512

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:178 Resp: 1185530

Ion Ratio Lower Upper

178 100

179 16.3 12.9 19.3




176 22.1 15.8 23.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 21.73 ng/ul

RT: 17.69 min Scan# 2528

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:178 Resp: 1237922

Ion Ratio Lower Upper

178 100

179 16.6 12.3 18.5



176 20.8 16.3 24.5




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 23.09 ng/ul

RT: 17.96 min Scan# 2574

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:167 Resp: 1059872

Ion Ratio Lower Upper

167 100

166 23.3 19.2 28.8




139 13.8 10.2 15.4



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#73

Di-n-butylphthalate

Concen: 22.16 ng/ul

RT: 18.50 min Scan# 2666

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:149 Resp: 1274348

Ion Ratio Lower Upper

149 100

150 10.0 7.1 10.7



104 6.7 3.3 4.9#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 22.40 ng/ul

RT: 19.60 min Scan# 2854

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:202 Resp: 1280553

Ion Ratio Lower Upper

202 100

101 10.6 10.6 16.0#




100 8.5 8.7 13.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 20.88 ng/ul

RT: 19.97 min Scan# 2917

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:202 Resp: 1290333

Ion Ratio Lower Upper

202 100

101 11.6 12.5 18.7#



100 11.5 10.2 15.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B









#78
Butylbenzylphthalate
Concen: 23.07 ng/ul
RT: 20.84 min Scan# 3065
Delta R.T. -0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16




Instrument :
BNA_G
ClientSampleId :
SSTD02025


Tgt Ion	Ratio	Lower	Upper
149	100		
91	80.8	54.2	81.4
206	22.2	14.1	21.1



#79
3,3'-Dichlorobenzidine
Concen: 24.09 ng/ul
RT: 21.76 min Scan# 3221
Delta R.T. 0.00 min
Lab File: BG023619.D
Acq: 11 Aug 2016 8:16

Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.9	51.4	77.2
126	10.8	10.7	16.1





#80

Benzo(a)anthracene

Concen: 21.03 ng/ul

RT: 21.85 min Scan# 3237

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:228 Resp: 1183997

Ion	Ratio	Lower	Upper
228	100		
229	21.7	15.4	23.2
226	29.7	22.8	34.2

228 100

229 21.7 15.4 23.2




226 29.7 22.8 34.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



#81

Bis(2-ethylhexyl)phthalate

Concen: 23.88 ng/ul

RT: 21.73 min Scan# 3216

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:149 Resp: 752670

Ion	Ratio	Lower	Upper
149	100		
167	29.8	21.9	32.9
279	5.0	2.9	4.3#

149 100

167 29.8 21.9 32.9



279 5.0 2.9 4.3#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#82

Chrysene

Concen: 21.02 ng/ul

RT: 21.92 min Scan# 3248

Delta R.T. 0.00 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:228 Resp: 1076013

Ion Ratio Lower Upper

228 100

226 32.1 24.2 36.4




229 20.5 15.0 22.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Di-n-octyl phthalate

Concen: 24.96 ng/ul


RT: 23.00 min Scan# 3432

Delta R.T. 0.00 min

Lab File: BG023619.D



Acq: 11 Aug 2016 8:16


Tgt Ion:149 Resp: 1269586



Abundance Ion 149.00 (148.70 to 149.70): E

Time-->





#85

Benzo(b)fluoranthene

Concen: 20.60 ng/ul

RT: 24.19 min Scan# 3634

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:252 Resp: 1180408

Ion Ratio Lower Upper




252 100

253 23.5 18.2 27.2

125 8.5 10.8 16.2#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#86

Benzo(k)fluoranthene

Concen: 20.21 ng/ul

RT: 24.26 min Scan# 3646

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:252 Resp: 1140070

Ion Ratio Lower Upper



252 100


253 23.1 15.8 23.6

125 8.0 10.6 16.0#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#88

Benzo(a)pyrene

Concen: 20.81 ng/ul

RT: 25.11 min Scan# 3792

Delta R.T. 0.01 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025


Tgt Ion:252 Resp: 1151207

Ion Ratio Lower Upper

252 100

253 24.2 19.0 28.4




125 9.4 12.0 18.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Indeno(1,2,3-cd)pyrene

Concen: 21.04 ng/ul

RT: 29.17 min Scan# 4483

Delta R.T. 0.03 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16


Tgt Ion:276 Resp: 1367197

Ion Ratio Lower Upper

276 100

138 16.6 20.8 31.2#



277 23.1 17.6 26.4




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#91

Benzo(g,h,i)perylene

Concen: 20.92 ng/ul

RT: 30.39 min Scan# 4690

Delta R.T. 0.02 min

Lab File: BG023619.D

Acq: 11 Aug 2016 8:16

Instrument :

BNA_G

ClientSampleId :

SSTD02025

Tgt Ion:276 Resp: 1127487

Ion Ratio Lower Upper

276 100

138 18.0 22.1 33.1#

277 25.9 17.0 25.4#

