


Data Path : Z:\HPCHEM1\BNA_G\Data\BG081017\
 Data File : BG028255.D
 Acq On : 10 Aug 2017 12:37
 Operator : SJ/JU
 Sample : DFTPP62
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP62

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 11 04:08:50 2017



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	20752	PASS
68	69	0.00	2	1.6	317	PASS
69	198	0.00	100	34.0	19876	PASS
70	69	0.00	2	0.6	128	PASS
127	198	10	80	40.4	23621	PASS
197	198	0.00	2	0.4	254	PASS
198	198	100	100	100.0	58442	PASS
199	198	5	9	6.6	3884	PASS
275	198	10	60	28.9	16861	PASS
365	198	1	100	3.9	2305	PASS
441	442	0.01	100	15.4	7576	PASS
442	198	50	100	84.4	49304	PASS
443	442	15	24	20.4	10054	PASS

m/z	Abundance
37.10	274.0
38.10	256.0
39.10	1851.0
40.10	293.0
40.90	240.0
44.00	398.0
49.20	461.0
50.10	7228.0
51.10	24688.0
52.10	1287.0
56.00	582.0
57.10	1935.0
58.20	174.0
61.10	266.0
62.10	516.0
63.10	777.0
65.00	433.0
68.10	488.0
69.00	23720.0
70.20	218.0
73.10	273.0
74.10	2419.0
75.00	4066.0
76.10	1267.0
77.10	27256.0
78.10	1827.0
79.10	1344.0
80.00	1463.0
81.10	2252.0
82.00	496.0
83.10	381.0
85.00	456.0
86.00	459.0
87.10	441.0
91.00	435.0
92.00	438.0
93.00	2590.0
94.10	255.0
97.10	170.0
97.90	1813.0
99.00	1512.0
100.00	286.0
100.90	1307.0
103.10	250.0
104.00	488.0
105.00	324.0
106.10	366.0
107.00	7405.0
108.10	807.0
109.10	326.0
110.00	12317.0
111.00	2025.0
112.10	236.0
116.20	730.0
117.00	5879.0
118.00	571.0
120.00	294.0
122.00	412.0
122.90	619.0
123.90	487.0
125.10	265.0
125.90	280.0
127.00	26952.0
128.10	1884.0
129.00	9722.0
130.10	607.0
130.90	323.0
131.90	189.0
132.40	170.0
134.10	242.0
134.90	654.0
136.10	420.0
137.00	645.0
139.00	168.0
139.90	189.0
141.00	1302.0
142.00	888.0
146.00	206.0
147.00	860.0
148.00	1619.0
148.90	277.0
150.80	185.0
151.20	189.0
153.00	484.0
153.90	348.0
155.00	854.0
156.10	1335.0
157.00	336.0
158.00	240.0
159.00	234.0
160.00	517.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

161.00	713.0
165.00	641.0
167.00	2678.0
168.00	1826.0
168.90	301.0
172.00	223.0
172.80	500.0
173.90	653.0
175.00	973.0
176.00	184.0
177.00	461.0
178.10	256.0
179.00	1726.0
180.00	1728.0
181.10	671.0
183.90	244.0
185.10	1104.0
186.00	8571.0
187.00	2186.0
188.10	358.0
189.10	387.0
192.10	669.0
193.00	1185.0
194.10	236.0
195.00	341.0
196.00	2044.0
196.70	764.0
198.00	61256.0
199.00	4009.0
200.00	267.0
201.60	207.0
202.20	202.0
203.00	495.0
204.00	2678.0
205.10	4056.0
206.00	17512.0
207.00	2211.0
208.00	881.0
209.80	166.0
211.10	701.0
211.70	285.0
214.90	287.0
216.00	467.0
217.00	4463.0
218.00	449.0
219.00	220.0
219.90	205.0
221.00	2806.0
222.10	569.0
223.00	1162.0
224.00	10699.0
225.00	2058.0
227.00	4456.0
228.00	674.0
229.00	912.0
231.10	324.0
233.90	528.0
235.00	299.0
235.90	214.0
237.10	498.0
238.80	155.0
239.90	212.0
240.90	195.0
242.10	330.0
244.00	6754.0
245.10	926.0
246.00	1742.0
247.00	303.0
249.00	248.0
250.10	206.0
252.90	333.0
253.80	421.0
255.00	39560.0
256.00	6040.0
256.90	393.0
257.90	2801.0
258.90	765.0
265.00	1030.0
272.00	277.0
273.00	1062.0
274.00	3083.0
275.00	16656.0
276.00	1747.0
277.00	1267.0
277.90	277.0
284.90	179.0
286.10	255.0
291.00	163.0
293.00	611.0
294.90	189.0
296.00	3541.0
297.00	500.0
301.00	219.0
303.00	753.0
307.00	182.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

314.10	385.0
315.00	356.0
315.90	348.0
320.90	277.0
323.10	1224.0
324.00	481.0
327.00	322.0
328.00	238.0
333.20	255.0
334.00	677.0
335.00	250.0
341.00	206.0
346.00	200.0
347.10	160.0
352.00	443.0
353.00	340.0
353.90	518.0
355.00	151.0
365.00	2162.0
365.90	172.0
370.70	172.0
372.10	661.0
382.80	191.0
403.00	714.0
403.80	224.0
421.00	424.0
421.90	216.0
423.00	2486.0
423.90	561.0
425.00	188.0
441.00	7125.0
442.00	38232.0
443.00	8096.0
444.00	698.0

Instrument :

BNA_G

ClientSampleId :

DFTPP62

m/z	Abundance
39.10	1883.0
39.80	201.0
43.90	569.0
49.00	318.0
50.10	5527.0
51.10	25744.0
52.10	1623.0
53.20	180.0
56.00	752.0
57.00	1362.0
61.00	382.0
62.10	181.0
63.10	1084.0
65.10	581.0
68.00	464.0
69.00	23480.0
70.00	166.0
73.10	183.0
74.10	2308.0
75.00	3616.0
76.20	1394.0
77.10	26616.0
78.00	1261.0
79.00	1786.0
80.00	1147.0
81.00	1665.0
82.10	189.0
83.00	634.0
85.20	828.0
86.10	843.0
87.00	207.0
88.00	154.0
90.90	528.0
91.90	591.0
93.00	3106.0
94.00	182.0
98.00	1886.0
99.00	2095.0
99.90	311.0
101.00	1055.0
103.00	555.0
104.10	708.0
105.00	537.0
106.00	223.0
107.00	7947.0
108.00	1636.0
110.00	13035.0
111.00	2313.0
112.10	196.0
116.00	1181.0
117.10	6363.0
118.10	615.0
120.10	236.0
122.10	397.0
123.00	1337.0
124.00	310.0
125.00	420.0
127.00	27144.0
128.00	1749.0
129.00	11296.0
130.00	1003.0
131.10	324.0
134.00	425.0
135.10	993.0
135.90	680.0
137.10	357.0
141.00	1760.0
142.00	523.0
142.90	316.0
146.00	256.0
147.00	591.0
148.10	1731.0
149.00	329.0
150.90	229.0
153.10	881.0
154.00	255.0
155.10	928.0
156.00	1271.0
157.90	300.0
158.90	233.0
160.00	370.0
160.30	331.0
160.90	751.0
161.80	363.0
163.10	188.0
165.00	485.0
166.10	414.0
167.00	2638.0
168.00	1392.0
169.00	362.0
169.90	313.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

171.00	261.0
172.00	271.0
173.00	347.0
174.10	712.0
175.10	1482.0
175.90	474.0
177.00	511.0
178.10	158.0
179.00	2496.0
180.00	1473.0
180.90	827.0
184.00	255.0
185.00	1707.0
186.00	9224.0
187.00	2543.0
188.10	171.0
189.00	569.0
190.80	193.0
191.10	205.0
192.00	1109.0
193.00	1419.0
194.00	383.0
196.10	2417.0
198.00	70968.0
198.90	5057.0
199.90	527.0
201.40	304.0
202.90	453.0
204.00	2538.0
205.00	3725.0
206.00	21960.0
207.00	2628.0
207.90	473.0
210.90	822.0
212.10	157.0
214.80	224.0
215.90	280.0
216.90	5508.0
218.00	781.0
219.00	264.0
219.90	327.0
220.90	2877.0
222.00	784.0
222.80	988.0
224.00	11428.0
225.00	2833.0
227.00	5010.0
228.00	668.0
228.90	1035.0
230.90	545.0
234.00	212.0
234.90	655.0
236.90	759.0
238.90	278.0
240.10	155.0
240.80	187.0
242.10	526.0
243.20	426.0
244.00	7956.0
245.00	764.0
246.00	2315.0
247.00	331.0
249.00	191.0
253.20	286.0
255.00	48440.0
256.00	7128.0
257.00	557.0
258.00	2685.0
258.70	450.0
259.00	471.0
264.00	153.0
264.90	1217.0
271.80	152.0
272.80	850.0
274.00	3236.0
275.00	20584.0
276.10	2216.0
277.00	1840.0
278.00	373.0
283.00	269.0
285.00	254.0
285.90	208.0
293.10	433.0
294.00	177.0
295.00	214.0
296.00	5390.0
297.10	804.0
300.80	154.0
302.10	186.0
303.00	962.0
314.10	307.0
315.00	860.0
315.90	216.0
321.20	199.0
323.00	2217.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

323.90	284.0
326.20	172.0
326.80	262.0
328.10	234.0
329.30	165.0
333.10	247.0
334.00	1142.0
335.10	288.0
338.90	221.0
341.00	398.0
346.00	522.0
349.90	203.0
351.90	396.0
353.00	421.0
354.00	788.0
364.90	2673.0
365.90	470.0
372.00	895.0
373.00	637.0
382.80	183.0
389.70	176.0
401.00	153.0
401.90	496.0
403.00	711.0
421.00	532.0
422.00	633.0
423.00	2996.0
424.00	1174.0
425.00	263.0
428.10	168.0
435.80	153.0
441.00	8790.0
442.00	61592.0
443.00	11868.0
444.00	986.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

m/z	Abundance
39.10	750.0
44.10	492.0
49.10	297.0
50.10	2639.0
51.10	11824.0
52.10	726.0
56.00	615.0
57.00	1115.0
61.10	177.0
63.10	732.0
65.10	369.0
69.00	12429.0
73.00	192.0
74.00	1675.0
75.10	2186.0
76.10	1062.0
77.10	14929.0
78.10	1248.0
79.00	808.0
80.00	846.0
81.00	1235.0
82.10	350.0
83.20	153.0
85.10	279.0
86.10	426.0
87.00	154.0
91.00	190.0
92.10	547.0
93.00	1479.0
96.10	324.0
98.00	1237.0
99.00	1013.0
101.00	891.0
103.00	395.0
103.90	600.0
105.10	303.0
105.90	153.0
107.00	4797.0
108.10	816.0
110.00	8008.0
111.10	1609.0
115.90	502.0
117.00	3355.0
118.10	276.0
122.00	272.0
123.00	527.0
124.00	185.0
124.90	393.0
127.00	16768.0
127.90	1097.0
129.00	6009.0
130.00	518.0
131.00	193.0
134.00	266.0
135.00	348.0
136.00	170.0
141.00	1140.0
141.90	241.0
143.00	219.0
146.00	175.0
147.00	1121.0
147.90	762.0
153.00	294.0
153.30	254.0
154.00	233.0
155.00	343.0
156.10	737.0
158.10	217.0
160.00	291.0
160.90	318.0
165.10	540.0
166.00	321.0
167.00	1779.0
168.00	1194.0
170.40	150.0
170.90	158.0
172.00	233.0
173.10	305.0
174.00	356.0
175.00	644.0
176.10	387.0
177.00	365.0
177.80	155.0
178.90	1442.0
180.00	1116.0
181.00	706.0
182.00	170.0
183.00	314.0
185.00	697.0
186.00	6313.0
187.00	1443.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

188.90	445.0
190.90	303.0
191.90	400.0
193.00	367.0
193.80	165.0
195.90	932.0
198.00	43104.0
199.00	2586.0
201.40	380.0
203.00	503.0
203.90	1702.0
205.00	3233.0
206.00	13172.0
207.10	1340.0
208.00	371.0
208.70	162.0
209.80	222.0
211.00	544.0
215.80	397.0
217.00	3556.0
217.90	379.0
220.30	230.0
221.00	1131.0
222.00	472.0
222.90	923.0
224.00	7513.0
225.00	2216.0
226.10	344.0
227.00	2978.0
228.00	314.0
229.00	719.0
229.90	184.0
230.90	204.0
234.00	239.0
235.00	299.0
235.90	201.0
237.10	342.0
238.80	179.0
241.90	317.0
242.90	530.0
243.30	429.0
244.00	5097.0
245.00	917.0
246.00	1044.0
246.90	326.0
248.10	198.0
249.00	437.0
251.00	276.0
253.00	160.0
255.00	31360.0
256.00	4459.0
257.00	429.0
257.90	1394.0
259.00	245.0
264.90	386.0
273.00	1125.0
273.90	2402.0
275.00	13345.0
276.00	1929.0
276.90	1390.0
277.90	215.0
282.80	181.0
285.00	306.0
292.90	305.0
296.00	4510.0
297.00	732.0
303.00	815.0
314.00	263.0
314.90	346.0
315.90	340.0
323.10	1982.0
324.00	301.0
326.90	211.0
328.00	291.0
333.20	195.0
334.00	917.0
335.10	293.0
341.00	235.0
351.90	459.0
352.90	222.0
354.00	634.0
355.20	182.0
364.90	2080.0
365.80	313.0
370.90	223.0
372.00	698.0
372.90	204.0
383.10	190.0
401.90	383.0
403.10	345.0
403.80	291.0
421.00	466.0
422.00	261.0
423.00	2027.0
424.00	486.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

441.00	6813.0
442.00	48088.0
443.00	10200.0
444.00	830.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

Scan 2464 (18.064 min): BG028255.D
DFTPP62

m/z	Abundance
44.10	490.0
48.90	184.0
57.10	208.0
67.10	168.0
78.00	162.0
84.00	156.0
386.90	164.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62