


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081018\
 Data File : BG036268.D
 Acq On : 10 Aug 2018 14:48
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG071218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 31 12:30:06 2018



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.4	14877	PASS
68	69	0.00	2	2.0	370	PASS
69	198	0.00	100	31.4	18375	PASS
70	69	0.00	2	0.5	97	PASS
127	198	10	80	37.0	21637	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	58525	PASS
199	198	5	9	6.9	4038	PASS
275	198	10	60	30.3	17718	PASS
365	198	1	100	3.9	2299	PASS
441	198	0.01	100	10.6	6227	PASS
442	198	50	100	66.4	38866	PASS
443	442	15	24	19.0	7374	PASS