


Data Path : Z:\HPCHEM1\BNA G\DATA\BG081115\
Data File : BG018377.D
Acq On : 11 Aug 2015 11:41
Operator : UM/MA
Sample : DFTPP32
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP32

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 12 00:54:17 2015



AutoFind: Scans 2553, 2554, 2555; Background Corrected with Scan 2548

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.8	25179	PASS
68	69	0.00	2	1.7	493	PASS
69	198	0.00	100	41.6	29267	PASS
70	69	0.00	2	0.1	29	PASS
127	198	10	80	48.4	34056	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	70354	PASS
199	198	5	9	6.8	4794	PASS
275	198	10	60	28.0	19723	PASS
365	198	1	100	4.3	2998	PASS
441	442	0.01	24	16.5	9269	PASS
442	198	50	100	79.7	56064	PASS
443	442	15	24	19.3	10848	PASS

m/z	Abundance
39.00	274.0
39.80	789.0
41.00	671.0
43.00	803.0
43.90	1374.0
45.00	265.0
50.00	155.0
50.90	663.0
53.00	327.0
55.00	1161.0
55.90	263.0
57.00	770.0
58.00	552.0
67.00	415.0
68.00	172.0
69.10	581.0
69.90	261.0
70.90	331.0
72.80	583.0
76.90	312.0
79.00	230.0
80.90	387.0
83.00	446.0
84.90	311.0
85.90	157.0
90.90	925.0
91.90	166.0
93.20	185.0
95.10	645.0
95.90	401.0
97.00	324.0
104.80	266.0
109.00	571.0
110.10	303.0
112.00	162.0
112.80	158.0
115.10	385.0
115.70	268.0
117.00	191.0
117.90	153.0
118.90	303.0
121.10	233.0
121.70	170.0
123.00	178.0
128.80	270.0
129.80	219.0
132.90	322.0
133.90	1333.0
134.90	323.0
139.00	196.0
140.90	169.0
141.90	155.0
142.80	475.0
146.90	287.0
147.90	160.0
150.80	243.0
153.40	175.0
156.80	192.0
160.60	220.0
162.60	153.0
163.20	221.0
166.20	151.0
167.00	255.0
170.40	204.0
174.00	163.0
175.10	194.0
176.70	253.0
178.90	325.0
184.90	242.0
190.90	452.0
192.80	248.0
194.00	378.0
200.10	179.0
205.10	217.0
205.90	163.0
206.90	1374.0
209.00	270.0
209.90	154.0
217.90	195.0
221.10	169.0
236.30	155.0
269.00	340.0
281.10	194.0
281.90	167.0
321.80	196.0
381.30	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z	Abundance
38.00	400.0
39.00	2705.0
40.00	850.0
41.00	1081.0
41.90	334.0
43.00	739.0
44.00	866.0
48.90	295.0
50.00	8186.0
51.00	27440.0
51.90	1546.0
52.90	180.0
55.00	1102.0
55.90	1119.0
57.00	2200.0
57.90	816.0
59.20	167.0
60.90	618.0
61.90	224.0
62.90	1134.0
64.00	329.0
65.00	653.0
66.90	634.0
67.90	643.0
68.90	32728.0
69.90	408.0
71.00	314.0
72.90	771.0
74.00	2887.0
74.90	5106.0
75.90	1563.0
77.00	35520.0
78.00	2386.0
78.90	2693.0
79.90	1830.0
80.90	2147.0
82.00	767.0
83.10	1261.0
83.90	247.0
84.90	426.0
85.90	727.0
86.90	365.0
91.00	1500.0
92.00	493.0
92.90	3205.0
93.80	306.0
95.00	549.0
95.90	449.0
97.00	668.0
97.90	2999.0
98.90	2356.0
100.00	170.0
100.90	1426.0
101.60	179.0
102.90	517.0
103.90	1126.0
105.00	779.0
105.90	538.0
106.90	10784.0
108.00	1842.0
108.90	554.0
109.90	18232.0
110.90	3329.0
112.80	251.0
114.90	179.0
115.90	617.0
116.90	9088.0
117.90	880.0
119.00	419.0
119.90	254.0
121.00	329.0
121.80	530.0
122.90	1163.0
123.90	394.0
124.90	946.0
126.90	33736.0
127.90	2193.0
128.90	13074.0
129.90	1412.0
130.90	322.0
131.80	208.0
132.90	393.0
134.00	1429.0
135.00	1412.0
135.90	259.0
136.90	720.0
137.50	330.0
139.80	232.0
140.90	1734.0
141.90	371.0
142.90	355.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

143.90	184.0
145.80	554.0
146.90	1317.0
147.90	2155.0
148.90	676.0
150.00	172.0
151.30	410.0
152.90	743.0
153.80	477.0
154.90	1081.0
156.00	985.0
157.20	370.0
157.70	415.0
159.00	296.0
159.80	625.0
160.80	477.0
162.00	373.0
163.10	282.0
164.10	174.0
164.90	855.0
165.90	731.0
166.90	3068.0
167.90	2173.0
168.90	430.0
170.70	174.0
171.80	565.0
173.00	864.0
173.90	602.0
174.90	1442.0
175.90	330.0
177.00	819.0
177.90	350.0
178.80	2822.0
179.90	2030.0
181.00	794.0
181.80	204.0
182.60	172.0
183.80	458.0
185.00	1601.0
185.90	9484.0
186.90	3487.0
188.00	529.0
188.90	619.0
190.00	352.0
190.90	706.0
191.80	992.0
192.90	1168.0
193.90	295.0
194.90	634.0
195.90	2632.0
197.90	68176.0
198.90	4759.0
200.00	479.0
201.30	340.0
202.00	345.0
202.90	908.0
203.90	2601.0
204.90	4703.0
205.90	19896.0
206.90	4162.0
207.80	988.0
208.90	243.0
209.80	532.0
210.20	465.0
211.00	1076.0
212.90	257.0
214.90	423.0
216.00	277.0
216.80	5217.0
218.10	319.0
218.90	190.0
220.90	2676.0
221.90	565.0
222.80	1112.0
223.90	11048.0
224.90	2532.0
225.90	260.0
226.90	4921.0
227.90	761.0
228.70	738.0
229.80	198.0
230.90	597.0
232.70	346.0
233.70	275.0
234.80	487.0
235.80	331.0
236.80	534.0
239.00	285.0
240.90	415.0
241.80	688.0
242.80	526.0
243.90	7890.0
244.90	1519.0
245.80	1557.0
246.80	368.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

248.90	320.0
251.90	237.0
252.80	194.0
253.70	463.0
254.90	40232.0
255.90	6038.0
256.90	379.0
257.80	3068.0
258.80	632.0
259.80	243.0
264.80	1236.0
265.80	264.0
267.10	236.0
269.00	190.0
272.90	1263.0
273.80	3057.0
274.90	16250.0
275.90	2477.0
276.80	1364.0
280.80	201.0
282.00	179.0
284.90	668.0
287.50	151.0
288.80	153.0
292.90	649.0
294.00	214.0
295.90	4575.0
296.90	993.0
302.90	560.0
304.00	192.0
308.90	218.0
313.90	177.0
314.70	391.0
315.90	456.0
321.00	173.0
322.90	1763.0
323.90	228.0
326.70	223.0
327.80	226.0
332.80	196.0
333.90	1159.0
345.80	403.0
350.90	262.0
351.80	651.0
352.90	375.0
353.80	427.0
354.80	260.0
364.90	1897.0
365.90	435.0
368.80	164.0
370.90	150.0
371.80	619.0
372.70	205.0
382.80	220.0
389.70	198.0
395.10	180.0
401.80	370.0
402.90	596.0
403.80	328.0
420.80	252.0
421.30	193.0
421.90	407.0
422.90	1965.0
423.80	512.0
440.90	6827.0
441.90	42016.0
442.90	8400.0
443.90	776.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z	Abundance
37.90	413.0
39.00	2917.0
39.90	1072.0
40.90	686.0
42.00	494.0
43.00	1043.0
44.00	1174.0
44.90	253.0
48.00	153.0
48.90	343.0
50.00	6781.0
51.00	30864.0
51.90	1728.0
53.10	197.0
55.00	997.0
55.90	1281.0
57.00	2423.0
58.00	751.0
61.00	305.0
61.90	509.0
63.00	1677.0
65.00	456.0
67.00	242.0
67.90	765.0
68.90	35880.0
70.00	253.0
71.00	324.0
73.00	452.0
73.90	3508.0
74.90	6106.0
76.00	1999.0
77.00	38424.0
78.00	3198.0
78.90	2129.0
79.90	1668.0
80.90	3411.0
81.90	667.0
83.00	1384.0
84.00	597.0
85.00	576.0
85.90	699.0
86.90	414.0
87.70	218.0
90.90	1420.0
92.00	637.0
92.90	3968.0
93.90	375.0
95.00	334.0
95.80	268.0
97.10	412.0
97.90	3855.0
98.90	2440.0
99.80	298.0
100.90	1172.0
102.90	1000.0
103.90	1271.0
104.90	1506.0
105.90	405.0
106.90	14011.0
107.90	2117.0
109.00	594.0
109.90	20456.0
110.90	3493.0
112.00	673.0
115.10	341.0
115.90	909.0
116.90	9851.0
117.90	622.0
118.90	309.0
119.80	270.0
121.10	295.0
121.90	1164.0
122.80	1079.0
123.80	308.0
125.00	668.0
125.90	363.0
126.90	41224.0
127.90	2894.0
128.90	16067.0
129.90	1582.0
131.00	644.0
131.90	184.0
132.90	608.0
134.00	1473.0
135.00	1922.0
135.70	409.0
136.90	933.0
137.80	249.0
139.90	314.0
140.90	2408.0
141.90	689.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

142.80	509.0
143.90	189.0
145.90	311.0
146.90	1327.0
147.90	2338.0
149.00	833.0
150.10	174.0
150.90	369.0
152.10	248.0
152.90	484.0
154.00	514.0
155.00	1245.0
155.90	1798.0
156.80	415.0
157.90	652.0
158.80	295.0
159.90	272.0
160.90	1232.0
162.00	535.0
163.00	307.0
164.90	901.0
165.90	847.0
166.90	6307.0
167.90	2807.0
169.10	573.0
170.80	539.0
171.90	383.0
172.90	530.0
173.80	925.0
174.80	1528.0
175.90	746.0
176.90	1102.0
177.90	491.0
178.90	3183.0
179.90	2029.0
180.80	704.0
181.80	156.0
182.10	174.0
182.50	175.0
182.80	193.0
183.90	385.0
184.90	2043.0
185.90	13220.0
186.90	4661.0
187.90	376.0
188.90	1239.0
189.90	264.0
190.90	663.0
192.00	1612.0
192.90	1559.0
193.90	304.0
195.90	2650.0
197.80	83992.0
198.80	5883.0
199.90	477.0
200.80	237.0
201.20	263.0
202.20	248.0
202.80	739.0
203.90	3694.0
204.90	4689.0
205.90	25096.0
206.90	4316.0
207.90	922.0
208.90	839.0
210.00	517.0
210.90	1106.0
211.70	215.0
212.90	267.0
214.90	378.0
215.80	416.0
216.90	6525.0
217.80	862.0
220.90	3635.0
221.80	735.0
222.80	1202.0
223.90	12722.0
224.90	2981.0
225.90	507.0
226.90	6241.0
227.90	1135.0
228.90	1629.0
230.90	627.0
232.00	160.0
233.90	625.0
234.80	499.0
235.90	585.0
236.90	593.0
238.80	342.0
239.90	364.0
240.80	500.0
242.00	870.0
242.90	960.0
243.90	10232.0
244.90	1839.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

245.90	1902.0
246.90	446.0
248.90	564.0
250.80	256.0
251.80	306.0
252.90	320.0
253.80	466.0
254.90	54216.0
255.90	7893.0
256.80	994.0
257.90	3510.0
258.90	532.0
263.80	170.0
264.90	1458.0
266.00	305.0
270.60	317.0
272.80	938.0
273.80	3918.0
274.90	23784.0
275.90	2995.0
276.90	2394.0
277.90	581.0
278.80	227.0
282.90	234.0
283.90	213.0
284.90	527.0
290.00	179.0
292.90	301.0
294.70	226.0
295.90	7425.0
296.90	1294.0
300.90	177.0
302.00	266.0
302.80	835.0
303.90	168.0
309.80	182.0
313.00	252.0
313.90	398.0
314.90	879.0
315.50	313.0
316.70	193.0
319.80	193.0
321.70	252.0
322.90	3211.0
324.00	489.0
326.70	303.0
327.80	440.0
332.80	191.0
333.90	1738.0
334.80	320.0
339.10	255.0
339.90	179.0
340.90	566.0
345.90	461.0
346.70	180.0
351.90	614.0
352.70	410.0
354.00	508.0
364.80	3237.0
365.80	506.0
370.90	219.0
371.80	882.0
372.90	376.0
379.10	152.0
382.70	323.0
383.80	275.0
389.90	164.0
391.00	376.0
401.70	413.0
402.80	554.0
403.70	465.0
420.90	473.0
421.90	1048.0
422.90	3901.0
423.70	754.0
434.50	216.0
439.10	182.0
440.90	10345.0
441.90	63432.0
442.90	12884.0
443.90	1241.0
460.60	171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

m/z	Abundance
36.00	234.0
38.00	459.0
39.00	1427.0
39.90	1215.0
41.00	518.0
42.00	391.0
43.00	696.0
44.00	1321.0
46.80	190.0
49.00	430.0
50.00	4834.0
51.00	19224.0
52.00	1067.0
53.80	151.0
54.90	657.0
55.90	634.0
57.00	1906.0
58.00	1075.0
61.00	345.0
62.00	499.0
63.00	1039.0
64.00	397.0
64.90	633.0
67.00	382.0
68.00	588.0
68.90	20936.0
70.10	209.0
71.00	261.0
73.00	694.0
73.90	2916.0
74.90	3441.0
76.00	1305.0
77.00	25280.0
78.00	1449.0
79.00	1797.0
79.90	1367.0
80.90	2043.0
82.00	592.0
82.90	819.0
83.90	315.0
84.90	779.0
86.00	778.0
86.90	456.0
90.90	1536.0
91.90	504.0
92.90	3085.0
94.00	363.0
95.00	273.0
96.00	860.0
96.90	536.0
97.90	2699.0
98.80	1474.0
100.80	1307.0
102.90	498.0
103.80	1063.0
104.80	1057.0
105.60	162.0
106.90	7987.0
107.90	1587.0
109.00	522.0
109.90	13750.0
110.90	2695.0
111.80	359.0
115.00	195.0
115.90	802.0
116.90	7510.0
117.90	415.0
118.90	622.0
119.90	182.0
121.10	361.0
121.90	941.0
122.80	805.0
123.90	368.0
124.80	472.0
126.90	27208.0
127.90	2507.0
128.90	10099.0
129.90	586.0
130.90	160.0
132.80	253.0
134.00	1524.0
134.90	801.0
135.90	264.0
137.00	663.0
140.90	1533.0
141.90	336.0
142.80	261.0
143.80	381.0
145.10	294.0
145.90	358.0
146.80	1278.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

147.90	2321.0
148.90	559.0
149.70	233.0
150.80	279.0
151.30	189.0
152.10	261.0
152.90	414.0
153.90	659.0
154.80	1147.0
155.90	1321.0
157.00	272.0
157.80	231.0
158.80	582.0
159.90	541.0
160.90	469.0
161.90	345.0
163.00	452.0
164.90	698.0
165.80	303.0
166.90	3458.0
167.90	1550.0
168.90	286.0
171.10	216.0
171.80	637.0
172.90	430.0
173.90	770.0
174.90	991.0
175.90	356.0
176.80	356.0
177.80	358.0
178.90	2463.0
179.90	1902.0
180.90	898.0
181.70	301.0
182.90	179.0
183.90	228.0
184.90	962.0
185.90	9293.0
186.90	2668.0
187.90	308.0
188.80	1091.0
190.00	237.0
190.90	682.0
191.80	1017.0
192.90	993.0
193.80	182.0
195.90	1162.0
197.80	58896.0
198.90	3740.0
199.80	464.0
201.50	324.0
202.90	444.0
203.90	2360.0
204.90	4137.0
205.90	18792.0
206.90	4177.0
207.90	1060.0
208.80	546.0
209.80	435.0
210.20	352.0
210.90	523.0
212.80	330.0
214.20	188.0
215.00	498.0
215.80	537.0
216.80	4633.0
218.00	649.0
219.10	225.0
220.80	2695.0
221.70	776.0
222.90	993.0
223.90	10862.0
224.90	2367.0
225.80	226.0
226.90	3855.0
227.80	509.0
228.80	1098.0
229.90	197.0
231.00	330.0
233.90	219.0
234.90	629.0
235.70	259.0
237.00	284.0
238.00	262.0
238.90	195.0
240.90	730.0
241.90	662.0
242.80	875.0
243.90	8629.0
244.90	584.0
245.80	1835.0
246.90	338.0
248.70	559.0
251.20	151.0
252.00	202.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32

254.90	41280.0
255.90	6623.0
256.80	885.0
257.80	3086.0
258.90	571.0
259.80	256.0
261.00	280.0
264.80	1034.0
266.00	263.0
268.90	210.0
272.90	1519.0
273.80	3155.0
274.90	19136.0
275.80	2285.0
276.90	1794.0
277.90	377.0
280.70	271.0
283.00	191.0
284.90	281.0
290.00	193.0
290.80	244.0
292.90	559.0
294.90	162.0
295.80	6779.0
296.80	914.0
302.80	727.0
303.80	350.0
313.70	265.0
314.90	897.0
315.70	472.0
322.90	2120.0
324.00	498.0
327.00	393.0
328.70	203.0
332.90	253.0
333.80	1292.0
334.80	402.0
340.90	157.0
345.80	295.0
351.80	866.0
352.90	481.0
353.80	452.0
364.80	3862.0
366.00	253.0
371.90	968.0
372.90	253.0
383.00	520.0
390.40	247.0
391.00	225.0
391.80	171.0
401.90	391.0
402.80	425.0
403.90	276.0
414.90	199.0
420.90	649.0
421.80	545.0
422.90	2862.0
423.90	548.0
439.50	176.0
440.90	10637.0
441.90	62744.0
442.90	11261.0
443.80	701.0

Instrument :
BNA_G
ClientSampleId :
DFTPP32