

Data Path : Z:\HPCHEM1\BNA_G\Data\BG081116\
 Data File : BG023621.D
 Acq On : 11 Aug 2016 9:31
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 SSTD02026

Quant Time: Aug 12 03:39:31 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 03:58:16 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	150739	20.00	ng/ul	0.00
18) Naphthalene-d8	10.95	136	702531	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.79	164	478945	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.55	188	1120989	20.00	ng/ul	0.00
75) Chrysene-d12	21.87	240	1032715	20.00	ng/ul	0.00
83) Perylene-d12	25.27	264	1040469	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.48	96	25367	7.15	ng/uL	0.00
5) Phenol-d5	7.26	99	307034	20.22	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.43	67	201810	20.74	ng/ul	0.00
9) 2-Chlorophenol-d4	7.64	132	218817	21.12	ng/ul	0.00
13) 4-Methylphenol-d8	8.83	113	243806	20.57	ng/ul	0.00
19) Nitrobenzene-d5	9.29	128	116788	21.09	ng/ul	0.00
22) 2-Nitrophenol-d4	10.02	143	139686	21.96	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.57	165	258353	21.54	ng/ul	0.00
29) 4-Chloroaniline-d4	11.09	131	316102	22.46	ng/ul	0.00
43) Dimethylphthalate-d6	14.18	166	791029	20.17	ng/ul	0.00
46) Acenaphthylene-d8	14.48	160	940415	21.31	ng/ul	0.00
51) 4-Nitrophenol-d4	15.00	143	129324	19.32	ng/ul	0.00
57) Fluorene-d10	15.78	176	725495	20.02	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.91	200	134727	18.73	ng/ul	0.00
70) Anthracene-d10	17.65	188	1052610	20.84	ng/ul	0.00
76) Pyrene-d10	19.94	212	1095733	20.96	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.04	264	956453	20.31	ng/ul	0.01

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.52	88	27630	7.43 ng/uL#	80
4) Benzaldehyde	7.24	77	215806	23.32 ng/ul	85
6) Phenol	7.29	94	318340	20.13 ng/ul#	73
8) Bis(2-Chloroethyl)ether	7.52	93	238005	20.73 ng/ul	90
10) 2-Chlorophenol	7.67	128	216815	20.55 ng/ul#	86
11) 2-Methylphenol	8.56	108	241598	20.29 ng/ul	93
12) 2,2'-oxybis(1-Chloropropan	8.65	45	331794	21.12 ng/ul	98
14) Acetophenone	8.94	105	383522	20.17 ng/ul#	78
15) N-Nitroso-di-n-propylamine	8.92	70	233168	20.38 ng/ul#	86
16) 4-Methylphenol	8.89	108	264639	19.83 ng/ul	100
17) Hexachloroethane	9.21	117	92863	20.40 ng/ul#	87
20) Nitrobenzene	9.33	77	329911	20.32 ng/ul#	85
21) Isophorone	9.85	82	614965	20.60 ng/ul	93
23) 2-Nitrophenol	10.05	139	143610	20.91 ng/ul#	59
24) 2,4-Dimethylphenol	10.11	107	303347	20.14 ng/ul	92
25) Bis(2-Chloroethoxy)methane	10.34	93	352271	20.46 ng/ul	98
27) 2,4-Dichlorophenol	10.59	162	250453	20.70 ng/ul	95
28) Naphthalene	11.00	128	752809	21.11 ng/ul	98
30) 4-Chloroaniline	11.11	127	307066	21.98 ng/ul	98
31) Hexachlorobutadiene	11.28	225	175108	21.73 ng/ul	96
32) Caprolactam	11.87	113	91435	18.20 ng/ul#	52
33) 4-Chloro-3-methylphenol	12.24	107	307490	19.75 ng/ul#	83
34) 2-Methylnaphthalene	12.61	142	589453	20.38 ng/ul	97


Data Path : Z:\HPCHEM1\BNA_G\Data\BG081116\
 Data File : BG023621.D
 Acq On : 11 Aug 2016 9:31
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02026

Quant Time: Aug 12 03:39:31 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 03:58:16 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	12.97	216	338081	22.99	ng/ul	98
37) Hexachlorocyclopentadiene	12.95	237	138456	17.02	ng/ul	96
38) 2,4,6-Trichlorophenol	13.22	196	220574	21.66	ng/ul	97
39) 2,4,5-Trichlorophenol	13.30	196	233168	21.36	ng/ul	99
40) 1,1'-Biphenyl	13.61	154	775491	22.03	ng/ul	98
41) 2-Chloronaphthalene	13.66	162	599811	21.92	ng/ul	97
42) 2-Nitroaniline	13.87	65	237812	21.28	ng/ul#	65
44) Dimethylphthalate	14.23	163	771751	19.87	ng/ul	99
45) 2,6-Dinitrotoluene	14.36	165	173158	20.57	ng/ul#	85
47) Acenaphthylene	14.51	152	985226	21.34	ng/ul	99
48) 3-Nitroaniline	14.69	138	176211	22.26	ng/ul#	64
49) Acenaphthene	14.85	153	656673	20.88	ng/ul	98
50) 2,4-Dinitrophenol	14.91	184	80629	15.52	ng/ul#	87
53) Dibenzofuran	15.19	168	961668	20.64	ng/ul	89
54) 2,4-Dinitrotoluene	15.15	165	256343	20.10	ng/ul#	70
55) 2,3,4,6-Tetrachlorophenol	15.42	232	218395	19.47	ng/ul	99
56) Diethylphthalate	15.59	149	788308	19.42	ng/ul	96
58) Fluorene	15.84	166	779609	20.08	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.82	204	398534	20.41	ng/ul	99
63) 4,6-Dinitro-2-methylphenol	15.92	198	142088	18.82	ng/ul#	90
64) N-Nitrosodiphenylamine	16.04	169	701849	22.40	ng/ul	95
65) 4-Bromophenyl-phenylether	16.73	248	278887	21.69	ng/ul	97
66) Hexachlorobenzene	16.85	284	294562	21.51	ng/ul	94
67) Atrazine	16.99	200	276004	21.33	ng/ul	98
68) Pentachlorophenol	17.20	266	141827	18.77	ng/ul	92
69) Phenanthrene	17.60	178	1224609	21.09	ng/ul	98
71) Anthracene	17.68	178	1261813	21.30	ng/ul	98
72) Carbazole	17.96	167	1084226	22.71	ng/ul	98
73) Di-n-butylphthalate	18.50	149	1291573	21.60	ng/ul#	95
74) Fluoranthene	19.61	202	1332807	22.42	ng/ul#	92
77) Pyrene	19.97	202	1342351	21.03	ng/ul#	94
78) Butylbenzylphthalate	20.85	149	568421	23.21	ng/ul#	93
79) 3,3'-Dichlorobenzidine	21.77	252	444747	24.44	ng/ul#	97
80) Benzo(a)anthracene	21.86	228	1236525	21.26	ng/ul	96
81) Bis(2-ethylhexyl)phthalate	21.73	149	769249	23.63	ng/ul#	97
82) Chrysene	21.92	228	1121649	21.21	ng/ul	97
84) Di-n-octyl phthalate	23.00	149	1282564	24.41	ng/ul	100
85) Benzo(b)fluoranthene	24.19	252	1219466	20.61	ng/ul#	93
86) Benzo(k)fluoranthene	24.26	252	1186270	20.36	ng/ul#	92
88) Benzo(a)pyrene	25.11	252	1172245	20.52	ng/ul#	95
89) Indeno(1,2,3-cd)pyrene	29.17	276	1388935	20.69	ng/ul#	86
90) Dibenzo(a,h)anthracene	29.25	278	1185117	20.70	ng/ul#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.43 ng/uL

RT: 3.52 min Scan# 115

Delta R.T. -0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion: 88 Resp: 27630

Ion Ratio Lower Upper

88 100

43 39.0 16.6 24.8#




58 69.1 47.2 70.8



Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 43.00 (42.70 to 43.70): BGC

Ion 58.00 (57.70 to 58.70): BGC



#4

Benzaldehyde

Concen: 23.32 ng/uL

RT: 7.24 min Scan# 749

Delta R.T. -0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion: 77 Resp: 215806

Ion Ratio Lower Upper

77 100

105 92.4 83.8 125.8



106 79.5 77.9 116.9




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC





#6

Phenol

Concen: 20.13 ng/ul

RT: 7.29 min Scan# 758

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion: 94 Resp: 318340

Ion Ratio Lower Upper

94 100

65 33.5 16.8 25.2#




66 42.8 22.6 33.8#



Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 20.73 ng/ul

RT: 7.52 min Scan# 797

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion: 93 Resp: 238005

Ion Ratio Lower Upper

93 100

63 80.2 56.0 84.0



95 31.5 27.7 41.5




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 63.00 (62.70 to 63.70): BGC

Ion 95.00 (94.70 to 95.70): BGC





#10

2-Chlorophenol

Concen: 20.55 ng/ul

RT: 7.67 min Scan# 823

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:128 Resp: 216815

Ion Ratio Lower Upper

128 100

64 57.4 34.2 51.4#




130 35.1 27.0 40.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BG

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 20.29 ng/ul

RT: 8.56 min Scan# 974

Delta R.T. 0.00 min

Lab File: BG023621.D


Acq: 11 Aug 2016 9:31

Tgt Ion:108 Resp: 241598

Ion Ratio Lower Upper



108 100


107 91.2 78.6 117.8



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
2,2'-oxybis(1-Chloropropane)
Concen: 21.12 ng/ul
RT: 8.65 min Scan# 989
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion	Ratio	Lower	Upper
45	100		
77	16.2	13.4	20.2
79	13.7	10.2	15.4



#14
Acetophenone
Concen: 20.17 ng/ul
RT: 8.94 min Scan# 1039
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion	Ratio	Lower	Upper
105	100		
77	87.5	56.1	84.1#
51	28.5	14.6	22.0#









#15
N-Nitroso-di-n-propylamine
Concen: 20.38 ng/ul
RT: 8.92 min Scan# 1035
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion	Ratio	Lower	Upper
70	100		
42	48.5	31.4	47.2#
101	9.0	8.6	13.0
130	15.9	19.0	28.4#



#16
4-Methylphenol
Concen: 19.83 ng/ul
RT: 8.89 min Scan# 1030
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion	Ratio	Lower	Upper
108	100		
107	110.7	88.3	132.5





#17

Hexachloroethane

Concen: 20.40 ng/ul

RT: 9.21 min Scan# 1084

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion: 117 Resp: 92863

Ion Ratio Lower Upper

117 100

201 105.3 76.8 115.2




199 68.3 43.5 65.3#



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 20.32 ng/ul

RT: 9.33 min Scan# 1105

Delta R.T. -0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion: 77 Resp: 329911

Ion Ratio Lower Upper

77 100

123 38.0 39.5 59.3#



65 14.4 9.0 13.4#




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BGC





#21

Isophorone

Concen: 20.60 ng/ul

RT: 9.85 min Scan# 1194

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion: 82 Resp: 614965

Ion Ratio Lower Upper

82 100

95 7.4 5.1 7.7




138 15.2 15.0 22.6



Abundance Ion 82.00 (81.70 to 82.70): BGC

Ion 95.00 (94.70 to 95.70): BGC

Ion 138.00 (137.70 to 138.70): BGC



#23

2-Nitrophenol

Concen: 20.91 ng/ul

RT: 10.05 min Scan# 1228

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion: 139 Resp: 143610

Ion Ratio Lower Upper

139 100

65 67.1 31.8 47.8#



109 38.1 17.1 25.7#




Abundance Ion 139.00 (138.70 to 139.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 109.00 (108.70 to 109.70): BGC









#24
2,4-Dimethylphenol
Concen: 20.14 ng/ul
RT: 10.11 min Scan# 1237
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion: 107 Resp: 303347
Ion Ratio Lower Upper
107 100
121 49.1 43.7 65.5
122 79.9 70.3 105.5



#25
Bis(2-Chloroethoxy)methane
Concen: 20.46 ng/ul
RT: 10.34 min Scan# 1277
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion: 93 Resp: 352271
Ion Ratio Lower Upper
93 100
95 32.6 26.0 39.0
123 10.4 10.1 15.1









#27
 2,4-Dichlorophenol
 Concen: 20.70 ng/ul
 RT: 10.59 min Scan# 1320
 Delta R.T. -0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31




Instrument :
 BNA_G
 ClientSampleId :
 SST02026


Tgt Ion	Ratio	Lower	Upper
162	100		
164	68.5	51.6	77.4
98	37.9	28.3	42.5



#28
 Naphthalene
 Concen: 21.11 ng/ul
 RT: 11.00 min Scan# 1389
 Delta R.T. -0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.4	14.2
127	14.8	11.2	16.8









#30
4-Chloroaniline
Concen: 21.98 ng/ul
RT: 11.11 min Scan# 1408
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion:127 Resp: 307066
Ion Ratio Lower Upper
127 100
129 32.3 26.9 40.3



#31
Hexachlorobutadiene
Concen: 21.73 ng/ul
RT: 11.28 min Scan# 1436
Delta R.T. -0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion:225 Resp: 175108
Ion Ratio Lower Upper
225 100
223 65.9 54.7 82.1
227 68.8 52.2 78.4









#32
 Caprolactam
 Concen: 18.20 ng/ul
 RT: 11.87 min Scan# 1538
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02026


Tgt Ion:113 Resp: 91435
 Ion Ratio Lower Upper
 113 100
 55 175.9 99.6 149.4#
 56 131.0 66.4 99.6#



#33
 4-Chloro-3-methylphenol
 Concen: 19.75 ng/ul
 RT: 12.24 min Scan# 1600
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31

Tgt Ion:107 Resp: 307490
 Ion Ratio Lower Upper
 107 100
 144 22.4 22.7 34.1#
 142 71.8 71.6 107.4





#34

2-Methylnaphthalene

Concen: 20.38 ng/ul

RT: 12.61 min Scan# 1663

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :





SSTD02026

Tgt Ion:142 Resp: 589453

Ion Ratio Lower Upper

142 100

141 91.0 75.2 112.8



#36

1,2,4,5-Tetrachlorobenzene

Concen: 22.99 ng/ul

RT: 12.97 min Scan# 1725

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Tgt Ion:216 Resp: 338081




Ion Ratio Lower Upper


216 100

214 76.1 61.7 92.5

179 19.1 17.0 25.4

108 15.7 14.2 21.2





#37

Hexachlorocyclopentadiene

Concen: 17.02 ng/ul

RT: 12.95 min Scan# 1721

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:237 Resp: 138456

Ion Ratio Lower Upper

237 100

235 64.8 49.6 74.4




272 10.2 9.3 13.9



Abundance Ion 237.00 (236.70 to 237.70): E

Ion 235.00 (234.70 to 235.70): E

Ion 272.00 (271.70 to 272.70): E



#38

2,4,6-Trichlorophenol

Concen: 21.66 ng/ul

RT: 13.22 min Scan# 1767

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:196 Resp: 220574

Ion Ratio Lower Upper

196 100

198 101.0 77.4 116.0



200 33.1 26.6 39.8




Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#39
 2,4,5-Trichlorophenol
 Concen: 21.36 ng/ul
 RT: 13.30 min Scan# 1780
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02026


Tgt Ion	Resp	Lower	Upper
196	233168		
196	100		
198	100.5	79.8	119.8
200	31.3	24.0	36.0



#40
 1,1'-Biphenyl
 Concen: 22.03 ng/ul
 RT: 13.61 min Scan# 1834
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31

Tgt Ion	Resp	Lower	Upper
154	775491		
154	100		
153	41.5	34.6	52.0
76	14.9	11.9	17.9









#41
 2-Chloronaphthalene
 Concen: 21.92 ng/ul
 RT: 13.66 min Scan# 1842
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02026


Tgt Ion	Ratio	Lower	Upper
162	100		
164	33.8	25.4	38.2
127	39.3	30.2	45.4



#42
 2-Nitroaniline
 Concen: 21.28 ng/ul
 RT: 13.87 min Scan# 1877
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31

Tgt Ion	Ratio	Lower	Upper
65	100		
92	63.3	67.1	100.7#
138	85.4	109.6	164.4#





#44

Dimethylphthalate

Concen: 19.87 ng/ul

RT: 14.23 min Scan# 1939

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:163 Resp: 771751

Ion Ratio Lower Upper

163 100

194 4.3 3.4 5.2




164 10.5 8.8 13.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#45

2,6-Dinitrotoluene

Concen: 20.57 ng/ul

RT: 14.36 min Scan# 1961

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:165 Resp: 173158

Ion Ratio Lower Upper

165 100

89 65.5 41.5 62.3#



121 23.2 17.0 25.4




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 89.00 (88.70 to 89.70): BG

Ion 121.00 (120.70 to 121.70): E





#47

Acenaphthylene

Concen: 21.34 ng/ul

RT: 14.51 min Scan# 1986

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:152 Resp: 985226

Ion Ratio Lower Upper

152 100

151 21.7 17.5 26.3




153 14.8 11.1 16.7



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#48

3-Nitroaniline

Concen: 22.26 ng/ul

RT: 14.69 min Scan# 1818

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:138 Resp: 176211

Ion Ratio Lower Upper

138 100

108 11.7 6.2 9.2#



92 136.8 79.3 118.9#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BG





#49

Acenaphthene

Concen: 20.88 ng/ul

RT: 14.85 min Scan# 2044

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampled :

SSTD02026


Tgt Ion:153 Resp: 656673

Ion Ratio Lower Upper

153 100

152 44.9 37.3 55.9




154 86.1 68.2 102.2



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#50

2,4-Dinitrophenol

Concen: 15.52 ng/ul

RT: 14.91 min Scan# 2055

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:184 Resp: 80629

Ion Ratio Lower Upper

184 100

63 67.9 40.5 60.7#



154 63.0 48.5 72.7




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 154.00 (153.70 to 154.70): E





#53

Dibenzenofuran

Concen: 20.64 ng/ul

RT: 15.19 min Scan# 2102

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :


SSTD02026

Tgt Ion:168 Resp: 961668

Ion Ratio Lower Upper




168 100

139 40.7 27.4 41.0



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E



#54

2,4-Dinitrotoluene

Concen: 20.10 ng/ul

RT: 15.15 min Scan# 2096

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:165 Resp: 256343

Ion Ratio Lower Upper

165 100

63 50.4 25.7 38.5#



182 2.4 2.1 3.1




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG023621.D (-2044) (-)

Ion 182.00 (181.70 to 182.70): E





#55

2,3,4,6-Tetrachlorophenol

Concen: 19.47 ng/ul

RT: 15.42 min Scan# 2141

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026

Tgt Ion:232 Resp: 218395


Ion Ratio Lower Upper

232 100

131 44.2 34.8 52.2

130 2.4 2.2 3.2

166 29.3 23.8 35.6






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#56

Diethylphthalate

Concen: 19.42 ng/ul

RT: 15.59 min Scan# 2171

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:149 Resp: 788308

Ion Ratio Lower Upper

149 100

177 23.8 17.4 26.2


150 13.0 9.6 14.4




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



Time-->





#58

Fluorene

Concen: 20.08 ng/ul

RT: 15.84 min Scan# 2213

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:166 Resp: 779609

Ion Ratio Lower Upper

166 100

165 100.9 80.3 120.5




167 15.0 10.3 15.5



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#59

4-Chlorophenyl-phenylether

Concen: 20.41 ng/ul

RT: 15.82 min Scan# 2210

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:204 Resp: 398534

Ion Ratio Lower Upper

204 100

206 33.6 27.4 41.2



141 59.2 47.1 70.7




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#63
 4,6-Dinitro-2-methylphenol
 Concen: 18.82 ng/ul
 RT: 15.92 min Scan# 2227
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02026


Tgt Ion:198 Resp: 142088
 Ion Ratio Lower Upper
 198 100
 182 4.9 4.2 6.2
 77 26.9 16.7 25.1#



#64
 N-Nitrosodiphenylamine
 Concen: 22.40 ng/ul
 RT: 16.04 min Scan# 2247
 Delta R.T. 0.00 min
 Lab File: BG023621.D
 Acq: 11 Aug 2016 9:31

Tgt Ion:169 Resp: 701849
 Ion Ratio Lower Upper
 169 100
 168 68.8 59.0 88.6
 167 36.4 30.6 46.0









#65
4-Bromophenyl-phenylether
Concen: 21.69 ng/ul
RT: 16.73 min Scan# 2364
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion:248 Resp: 278887
Ion Ratio Lower Upper
248 100
250 99.1 76.6 114.8
141 75.0 61.8 92.6



#66
Hexachlorobenzene
Concen: 21.51 ng/ul
RT: 16.85 min Scan# 2385
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion:284 Resp: 294562
Ion Ratio Lower Upper
284 100
142 31.4 27.8 41.8
249 33.6 24.4 36.6









#67
Atrazine
Concen: 21.33 ng/ul
RT: 16.99 min Scan# 2409
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.3	21.0	31.6
215	46.2	37.9	56.9



#68
Pentachlorophenol
Concen: 18.77 ng/ul
RT: 17.20 min Scan# 2445
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion	Ratio	Lower	Upper
266	100		
264	66.2	54.7	82.1
268	61.8	57.8	86.8





#69

Phenanthrene

Concen: 21.09 ng/ul

RT: 17.60 min Scan# 2512

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampled :

SSTD02026


Tgt Ion:178 Resp: 1224609

Ion Ratio Lower Upper




178 100

179 16.3 12.9 19.3

176 21.0 15.8 23.8



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 21.30 ng/ul

RT: 17.68 min Scan# 2527

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:178 Resp: 1261813

Ion Ratio Lower Upper



178 100


179 17.1 12.3 18.5

176 20.6 16.3 24.5



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 22.71 ng/ul

RT: 17.96 min Scan# 2574

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:167 Resp: 1084226

Ion Ratio Lower Upper

167 100

166 23.4 19.2 28.8




139 14.0 10.2 15.4



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#73

Di-n-butylphthalate

Concen: 21.60 ng/ul

RT: 18.50 min Scan# 2666

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:149 Resp: 1291573

Ion Ratio Lower Upper

149 100

150 10.3 7.1 10.7



104 6.3 3.3 4.9#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 22.42 ng/ul

RT: 19.61 min Scan# 2855

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:202 Resp: 1332807

Ion Ratio Lower Upper

202 100

101 9.6 10.6 16.0#




100 8.6 8.7 13.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 21.03 ng/ul

RT: 19.97 min Scan# 2916

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:202 Resp: 1342351

Ion Ratio Lower Upper

202 100

101 12.3 12.5 18.7#



100 11.6 10.2 15.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B









#78
Butylbenzylphthalate
Concen: 23.21 ng/ul
RT: 20.85 min Scan# 3066
Delta R.T. 0.00 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31




Instrument :
BNA_G
ClientSampleId :
SSTD02026


Tgt Ion	Ratio	Lower	Upper
149	100		
91	71.8	54.2	81.4
206	24.0	14.1	21.1#



#79
3,3'-Dichlorobenzidine
Concen: 24.44 ng/ul
RT: 21.77 min Scan# 3222
Delta R.T. 0.01 min
Lab File: BG023621.D
Acq: 11 Aug 2016 9:31

Tgt Ion	Ratio	Lower	Upper
252	100		
254	62.8	51.4	77.2
126	9.3	10.7	16.1#





#80

Benzo(a)anthracene

Concen: 21.26 ng/ul

RT: 21.86 min Scan# 3237

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:228 Resp: 1236525

Ion Ratio Lower Upper

228 100

229 22.2 15.4 23.2




226 29.5 22.8 34.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



#81

Bis(2-ethylhexyl)phthalate

Concen: 23.63 ng/ul

RT: 21.73 min Scan# 3216

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:149 Resp: 769249

Ion Ratio Lower Upper

149 100

167 28.8 21.9 32.9



279 4.5 2.9 4.3#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#82

Chrysene

Concen: 21.21 ng/ul

RT: 21.92 min Scan# 3248

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026





Tgt Ion:228 Resp: 1121649

Ion Ratio Lower Upper

228 100

226 30.9 24.2 36.4

229 21.7 15.0 22.6



#84

Di-n-octyl phthalate

Concen: 24.41 ng/ul




RT: 23.00 min Scan# 3432


Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Tgt Ion:149 Resp: 1282564





#85

Benzo(b)fluoranthene

Concen: 20.61 ng/ul

RT: 24.19 min Scan# 3634

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:252 Resp: 1219466

Ion Ratio Lower Upper

252 100

253 20.9 18.2 27.2




125 8.7 10.8 16.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#86

Benzo(k)fluoranthene

Concen: 20.36 ng/ul

RT: 24.26 min Scan# 3646

Delta R.T. 0.01 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31


Tgt Ion:252 Resp: 1186270

Ion Ratio Lower Upper

252 100

253 22.6 15.8 23.6



125 8.8 10.6 16.0#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#88

Benzo(a)pyrene

Concen: 20.52 ng/ul

RT: 25.11 min Scan# 3791

Delta R.T. 0.00 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026





Tgt Ion:252 Resp: 1172245

Ion Ratio Lower Upper

252 100

253 23.7 19.0 28.4

125 9.5 12.0 18.0#



#89

Indeno(1,2,3-cd)pyrene

Concen: 20.69 ng/ul

RT: 29.17 min Scan# 4483

Delta R.T. 0.03 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31




Tgt Ion:276 Resp: 1388935


Ion Ratio Lower Upper

276 100

138 15.3 20.8 31.2#

277 24.6 17.6 26.4





#90

Dibenzo(a,h)anthracene

Concen: 20.70 ng/ul

RT: 29.25 min Scan# 4495

Delta R.T. 0.03 min

Lab File: BG023621.D

Acq: 11 Aug 2016 9:31

Instrument :

BNA_G

ClientSampleId :

SSTD02026


Tgt Ion:278 Resp: 1185117

Ion Ratio Lower Upper

278 100

139 13.5 16.3 24.5#

279 26.3 19.1 28.7



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

