

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081116\
 Data File : BG023645.D
 Acq On : 12 Aug 2016 6:42
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 SSTD02028

Quant Time: Aug 12 07:25:20 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 04:34:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	162102	20.00	ng/ul	0.00
18) Naphthalene-d8	10.97	136	677622	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.80	164	463390	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.56	188	1108544	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	1144931	20.00	ng/ul	0.00
83) Perylene-d12	25.28	264	1158004	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.49	96	28419	7.45	ng/uL	0.00
5) Phenol-d5	7.29	99	317148	19.42	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.32	67	12431	1.19	ng/ul	-0.13
9) 2-Chlorophenol-d4	7.66	132	226227	20.31	ng/ul	0.00
13) 4-Methylphenol-d8	8.84	113	249460	19.57	ng/ul	0.00
19) Nitrobenzene-d5	9.31	128	115506	21.63	ng/ul	0.00
22) 2-Nitrophenol-d4	10.04	143	139016	22.66	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.59	165	247109	21.36	ng/ul	0.00
29) 4-Chloroaniline-d4	11.10	131	301577	22.21	ng/ul	0.00
43) Dimethylphthalate-d6	14.19	166	745941	19.66	ng/ul	0.00
46) Acenaphthylene-d8	14.49	160	903354	21.16	ng/ul	0.00
51) 4-Nitrophenol-d4	15.01	143	130767	20.19	ng/ul	0.00
57) Fluorene-d10	15.79	176	695620	19.84	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.92	200	138507	19.48	ng/ul	0.00
70) Anthracene-d10	17.66	188	1047918	20.98	ng/ul	0.00
76) Pyrene-d10	19.95	212	1117618	19.28	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.05	264	1083552	20.67	ng/ul	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.52	88	30889	7.72	ng/uL#	72
4) Benzaldehyde	7.26	77	210329	21.13	ng/ul	90
6) Phenol	7.32	94	337201	19.83	ng/ul#	77
8) Bis(2-Chloroethyl)ether	7.54	93	251104	20.34	ng/ul	91
10) 2-Chlorophenol	7.69	128	229449	20.22	ng/ul#	85
11) 2-Methylphenol	8.58	108	245394	19.16	ng/ul	99
12) 2,2'-oxybis(1-Chloropropan	8.67	45	351402	20.80	ng/ul	98
14) Acetophenone	8.96	105	389112	19.03	ng/ul#	80
15) N-Nitroso-di-n-propylamine	8.94	70	232150	18.87	ng/ul#	88
16) 4-Methylphenol	8.91	108	265848	18.52	ng/ul	99
17) Hexachloroethane	9.22	117	97598	19.93	ng/ul	89
20) Nitrobenzene	9.35	77	329876	21.07	ng/ul#	82
21) Isophorone	9.87	82	599815	20.83	ng/ul	94
23) 2-Nitrophenol	10.07	139	143509	21.67	ng/ul#	78
24) 2,4-Dimethylphenol	10.13	107	302912	20.85	ng/ul	85
25) Bis(2-Chloroethoxy)methane	10.35	93	346978	20.89	ng/ul#	98
27) 2,4-Dichlorophenol	10.61	162	248615	21.31	ng/ul	94
28) Naphthalene	11.02	128	717078	20.84	ng/ul	99
30) 4-Chloroaniline	11.13	127	286413	21.26	ng/ul	100
31) Hexachlorobutadiene	11.29	225	171202	22.03	ng/ul	91
32) Caprolactam	11.89	113	82801	17.09	ng/ul#	49
33) 4-Chloro-3-methylphenol	12.26	107	299605	19.95	ng/ul	87
34) 2-Methylnaphthalene	12.62	142	551227	19.76	ng/ul	96


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081116\
 Data File : BG023645.D
 Acq On : 12 Aug 2016 6:42
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02028

Quant Time: Aug 12 07:25:20 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 04:34:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	12.99	216	315808	22.20	ng/ul	95
37) Hexachlorocyclopentadiene	12.96	237	131086	16.66	ng/ul	95
38) 2,4,6-Trichlorophenol	13.23	196	217322	22.05	ng/ul	96
39) 2,4,5-Trichlorophenol	13.31	196	232579	22.02	ng/ul	98
40) 1,1'-Biphenyl	13.62	154	718504	21.10	ng/ul	99
41) 2-Chloronaphthalene	13.67	162	563392	21.28	ng/ul	95
42) 2-Nitroaniline	13.88	65	232327	21.49	ng/ul#	67
44) Dimethylphthalate	14.24	163	734860	19.56	ng/ul	99
45) 2,6-Dinitrotoluene	14.37	165	168067	20.63	ng/ul#	84
47) Acenaphthylene	14.52	152	922876	20.66	ng/ul	99
48) 3-Nitroaniline	14.71	138	176364	23.03	ng/ul#	70
49) Acenaphthene	14.86	153	624089	20.51	ng/ul	97
50) 2,4-Dinitrophenol	14.92	184	73836	14.69	ng/ul#	88
52) 4-Nitrophenol	15.02	109	127659	19.68	ng/ul#	76
53) Dibenzofuran	15.19	168	912876	20.25	ng/ul	92
54) 2,4-Dinitrotoluene	15.17	165	244134	19.78	ng/ul#	78
55) 2,3,4,6-Tetrachlorophenol	15.43	232	221840	20.44	ng/ul#	98
56) Diethylphthalate	15.60	149	762425	19.42	ng/ul	95
58) Fluorene	15.85	166	734466	19.55	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.84	204	370608	19.62	ng/ul	99
60) 4-Nitroaniline	15.88	138	183687	20.68	ng/ul#	76
63) 4,6-Dinitro-2-methylphenol	15.94	198	144181	19.31	ng/ul#	92
64) N-Nitrosodiphenylamine	16.05	169	681930	22.01	ng/ul	97
65) 4-Bromophenyl-phenylether	16.73	248	265773	20.91	ng/ul	95
66) Hexachlorobenzene	16.86	284	281386	20.78	ng/ul	95
67) Atrazine	17.00	200	279697	21.86	ng/ul	98
68) Pentachlorophenol	17.21	266	151199	20.23	ng/ul	88
69) Phenanthrene	17.60	178	1211261	21.10	ng/ul	97
71) Anthracene	17.69	178	1242033	21.20	ng/ul	99
72) Carbazole	17.97	167	1089497	23.08	ng/ul	98
73) Di-n-butylphthalate	18.51	149	1355698	22.93	ng/ul#	94
74) Fluoranthene	19.61	202	1385742	23.58	ng/ul#	93
77) Pyrene	19.98	202	1387513	19.60	ng/ul	95
78) Butylbenzylphthalate	20.85	149	610602	22.49	ng/ul#	85
79) 3,3'-Dichlorobenzidine	21.77	252	503985	24.98	ng/ul#	97
80) Benzo(a)anthracene	21.86	228	1368148	21.22	ng/ul	97
81) Bis(2-ethylhexyl)phthalate	21.73	149	878740	24.34	ng/ul#	96
82) Chrysene	21.93	228	1238010	21.12	ng/ul	94
84) Di-n-octyl phthalate	23.01	149	1483129	25.36	ng/ul	100
85) Benzo(b)fluoranthene	24.20	252	1364925	20.72	ng/ul#	94
86) Benzo(k)fluoranthene	24.27	252	1280196	19.74	ng/ul#	90
88) Benzo(a)pyrene	25.12	252	1312605	20.64	ng/ul#	95
89) Indeno(1,2,3-cd)pyrene	29.18	276	1549141	20.73	ng/ul#	86
90) Dibenzo(a,h)anthracene	29.25	278	1328221	20.84	ng/ul#	91
91) Benzo(g,h,i)perylene	30.42	276	1288962	20.81	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.72 ng/uL

RT: 3.52 min Scan# 116

Delta R.T. -0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42

Instrument :

BNA_G

ClientSampleId :

SSTD02028


Tgt Ion: 88 Resp: 30889

Ion Ratio Lower Upper

88 100

43 32.8 16.6 24.8#




58 80.3 47.2 70.8#



Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 43.00 (42.70 to 43.70): BGC

Ion 58.00 (57.70 to 58.70): BGC



#4

Benzaldehyde

Concen: 21.13 ng/uL

RT: 7.26 min Scan# 752

Delta R.T. 0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42


Tgt Ion: 77 Resp: 210329

Ion Ratio Lower Upper

77 100

105 94.8 83.8 125.8



106 88.2 77.9 116.9




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC





#6

Phenol

Concen: 19.83 ng/ul

RT: 7.32 min Scan# 762

Delta R.T. 0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42

Instrument :

BNA_G

ClientSampleId :

SSTD02028


Tgt Ion: 94 Resp: 337201

Ion Ratio Lower Upper




94 100

65 31.2 16.8 25.2#

66 41.1 22.6 33.8#



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 20.34 ng/ul

RT: 7.54 min Scan# 800

Delta R.T. 0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42


Tgt Ion: 93 Resp: 251104

Ion Ratio Lower Upper



93 100


63 77.5 56.0 84.0

95 30.4 27.7 41.5



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC





#10

2-Chlorophenol

Concen: 20.22 ng/ul

RT: 7.69 min Scan# 826

Delta R.T. 0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42

Instrument :

BNA_G

ClientSampleId :

SSTD02028


Tgt Ion:128 Resp: 229449

Ion Ratio Lower Upper

128 100

64 59.4 34.2 51.4#




130 33.6 27.0 40.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BG

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 19.16 ng/ul

RT: 8.58 min Scan# 977

Delta R.T. 0.00 min

Lab File: BG023645.D


Acq: 12 Aug 2016 6:42

Tgt Ion:108 Resp: 245394

Ion Ratio Lower Upper



108 100


107 97.5 78.6 117.8



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
 2,2'-oxybis(1-chloropropane)
 Concen: 20.80 ng/ul
 RT: 8.67 min Scan# 992
 Delta R.T. 0.01 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02028


Tgt Ion: 45 Resp: 351402
 Ion Ratio Lower Upper
 45 100
 77 15.7 13.4 20.2
 79 12.5 10.2 15.4



#14
 Acetophenone
 Concen: 19.03 ng/ul
 RT: 8.96 min Scan# 1042
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42

Tgt Ion: 105 Resp: 389112
 Ion Ratio Lower Upper
 105 100
 77 87.2 56.1 84.1#
 51 25.7 14.6 22.0#









#15
 N-Nitroso-di-n-propylamine
 Concen: 18.87 ng/ul
 RT: 8.94 min Scan# 1038
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02028


Tgt Ion	Ratio	Lower	Upper
70	100		
42	46.9	31.4	47.2
101	9.3	8.6	13.0
130	16.8	19.0	28.4



#16
 4-Methylphenol
 Concen: 18.52 ng/ul
 RT: 8.91 min Scan# 1033
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42

Tgt Ion	Ratio	Lower	Upper
108	100		
107	111.6	88.3	132.5





#17

Hexachloroethane

Concen: 19.93 ng/ul

RT: 9.22 min Scan# 1086

Delta R.T. 0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42

Instrument :

BNA_G

ClientSampleId :

SSTD02028


Tgt Ion: 117 Resp: 97598

Ion Ratio Lower Upper

117 100

201 104.0 76.8 115.2




199 65.3 43.5 65.3



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 21.07 ng/ul

RT: 9.35 min Scan# 1108

Delta R.T. 0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42


Tgt Ion: 77 Resp: 329876

Ion Ratio Lower Upper

77 100

123 35.3 39.5 59.3#



65 13.3 9.0 13.4




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): BGC

Ion 65.00 (64.70 to 65.70): BGC









#21
Isophorone
Concen: 20.83 ng/ul
RT: 9.87 min Scan# 1197
Delta R.T. 0.00 min
Lab File: BG023645.D
Acq: 12 Aug 2016 6:42




Instrument :
BNA_G
ClientSampleId :
SSTD02028


Tgt Ion	Ratio	Lower	Upper
82	100		
95	7.5	5.1	7.7
138	15.7	15.0	22.6



#23
2-Nitrophenol
Concen: 21.67 ng/ul
RT: 10.07 min Scan# 1231
Delta R.T. 0.00 min
Lab File: BG023645.D
Acq: 12 Aug 2016 6:42

Tgt Ion	Ratio	Lower	Upper
139	100		
65	55.9	31.8	47.8#
109	28.5	17.1	25.7#









#24
 2,4-Dimethylphenol
 Concen: 20.85 ng/ul
 RT: 10.13 min Scan# 1240
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02028


Tgt Ion: 107 Resp: 302912
 Ion Ratio Lower Upper
 107 100
 121 43.9 43.7 65.5
 122 73.5 70.3 105.5



#25
 Bis(2-Chloroethoxy)methane
 Concen: 20.89 ng/ul
 RT: 10.35 min Scan# 1279
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42

Tgt Ion: 93 Resp: 346978
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.0 39.0
 123 9.9 10.1 15.1#









#27
2,4-Dichlorophenol
Concen: 21.31 ng/ul
RT: 10.61 min Scan# 1323
Delta R.T. 0.00 min
Lab File: BG023645.D
Acq: 12 Aug 2016 6:42




Instrument :
BNA_G
ClientSampleId :
SSTD02028


Tgt Ion	Ratio	Lower	Upper
162	100		
164	68.2	51.6	77.4
98	39.7	28.3	42.5



#28
Naphthalene
Concen: 20.84 ng/ul
RT: 11.02 min Scan# 1392
Delta R.T. 0.00 min
Lab File: BG023645.D
Acq: 12 Aug 2016 6:42

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.4	14.2
127	13.9	11.2	16.8









#30
 4-Chloroaniline
 Concen: 21.26 ng/ul
 RT: 11.13 min Scan# 1411
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02028


Tgt Ion:127 Resp: 286413
 Ion Ratio Lower Upper
 127 100
 129 33.4 26.9 40.3



#31
 Hexachlorobutadiene
 Concen: 22.03 ng/ul
 RT: 11.29 min Scan# 1438
 Delta R.T. -0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42

Tgt Ion:225 Resp: 171202
 Ion Ratio Lower Upper
 225 100
 223 59.4 54.7 82.1
 227 59.3 52.2 78.4





#32

Caprolactam

Concen: 17.09 ng/ul

RT: 11.89 min Scan# 1540

Delta R.T. -0.00 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42

Instrument :

BNA_G

ClientSampleId :

SSTD02028


Tgt Ion:113 Resp: 82801

Ion Ratio Lower Upper

113 100

55 175.3 99.6 149.4#




56 136.5 66.4 99.6#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC



#33

4-Chloro-3-methylphenol

Concen: 19.95 ng/ul

RT: 12.26 min Scan# 1603

Delta R.T. 0.01 min

Lab File: BG023645.D

Acq: 12 Aug 2016 6:42


Tgt Ion:107 Resp: 299605

Ion Ratio Lower Upper

107 100

144 24.0 22.7 34.1



142 75.7 71.6 107.4




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#34
 2-Methylnaphthalene
 Concen: 19.76 ng/ul
 RT: 12.62 min Scan# 1665
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02028


Tgt Ion:142 Resp: 551227
 Ion Ratio Lower Upper
 142 100
 141 89.9 75.2 112.8



#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 22.20 ng/ul
 RT: 12.99 min Scan# 1727
 Delta R.T. 0.00 min
 Lab File: BG023645.D
 Acq: 12 Aug 2016 6:42

Tgt Ion:216 Resp: 315808
 Ion Ratio Lower Upper
 216 100
 214 82.5 61.7 92.5
 179 20.8 17.0 25.4
 108 16.7 14.2 21.2









#37
Hexachlorocyclopentadiene
Concen: 16.66 ng/ul
RT: 12.96 min Scan# 1723
Delta R.T. 0.00 min
Lab File: BG023645.D
Acq: 12 Aug 2016 6:42

Instrument :
BNA_G
ClientSampleId :
SSTD02028

Tgt Ion: 237 Resp: 131086
Ion Ratio Lower Upper
237 100
235 65.7 49.6 74.4
272 10.0 9.3 13.9



#38
2,4,6-Trichlorophenol
Concen: 22.05 ng/ul
RT: 13.23 min Scan# 1769
Delta R.T. 0.01 min
Lab File: BG023645.D
Acq: 12 Aug 2016 6:42

Tgt Ion: 196 Resp: 217322
Ion Ratio Lower Upper
196 100
198 91.6 77.4 116.0
200 33.1 26.6 39.8

