


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081117\
 Data File : BG028292.D
 Acq On : 11 Aug 2017 16:03
 Operator : SJ/JU
 Sample : DFTPP63
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP63

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 14 17:53:11 2017



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	25200	PASS
68	69	0.00	2	0.9	194	PASS
69	198	0.00	100	33.7	21240	PASS
70	69	0.00	2	0.6	118	PASS
127	198	10	80	43.6	27485	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	63104	PASS
199	198	5	9	6.4	4028	PASS
275	198	10	60	28.0	17691	PASS
365	198	1	100	3.1	1971	PASS
441	442	0.01	100	15.5	7832	PASS
442	198	50	100	80.3	50650	PASS
443	442	15	24	20.1	10188	PASS

m/z	Abundance
38.10	285.0
39.10	1887.0
40.10	182.0
41.10	170.0
44.00	634.0
44.70	163.0
49.20	398.0
50.10	7677.0
51.10	25904.0
52.10	793.0
55.10	294.0
56.00	1005.0
57.10	1814.0
61.10	797.0
61.90	207.0
63.00	1133.0
65.00	426.0
69.00	22592.0
70.10	173.0
73.00	224.0
74.00	1971.0
75.10	4053.0
76.10	1486.0
77.10	27088.0
78.10	2146.0
79.00	1456.0
80.00	1191.0
81.00	2337.0
82.10	373.0
83.10	646.0
84.00	249.0
85.10	416.0
86.00	444.0
87.10	289.0
91.00	508.0
91.90	472.0
93.00	3003.0
94.20	360.0
95.90	154.0
98.00	2156.0
99.00	2142.0
99.90	218.0
101.00	1511.0
103.00	359.0
104.00	782.0
104.90	663.0
106.20	299.0
107.00	8811.0
108.00	1215.0
110.00	14798.0
111.00	2594.0
112.10	374.0
116.00	689.0
117.00	5787.0
118.00	201.0
121.40	212.0
122.00	377.0
122.90	984.0
124.00	621.0
125.20	281.0
127.00	28568.0
128.10	1739.0
129.00	10902.0
130.00	772.0
131.10	231.0
132.90	290.0
134.00	273.0
134.20	262.0
135.00	774.0
136.30	351.0
137.00	506.0
138.90	152.0
139.90	201.0
141.00	1618.0
142.00	372.0
142.80	371.0
143.80	210.0
144.10	200.0
147.00	702.0
148.00	1367.0
149.10	323.0
151.90	193.0
152.70	290.0
152.90	292.0
153.80	248.0
155.10	570.0
156.10	1447.0
157.00	273.0
157.80	252.0
158.80	281.0
159.90	389.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

161.00	768.0
162.00	199.0
164.00	278.0
164.70	288.0
166.00	279.0
167.00	2956.0
168.00	1488.0
168.90	309.0
169.90	180.0
171.00	358.0
171.90	404.0
172.90	277.0
173.90	739.0
175.00	851.0
176.00	480.0
176.90	584.0
177.80	223.0
179.00	1951.0
180.00	1911.0
181.10	1026.0
182.00	176.0
185.10	1671.0
186.00	9431.0
187.00	2476.0
187.90	382.0
188.90	426.0
190.90	477.0
192.10	490.0
193.00	613.0
194.20	212.0
195.00	218.0
196.00	1377.0
198.00	59264.0
199.00	4273.0
199.90	462.0
203.10	482.0
204.00	2967.0
205.00	4855.0
206.00	17016.0
207.00	2436.0
207.90	778.0
208.90	213.0
210.20	181.0
211.00	1205.0
214.90	214.0
216.00	377.0
216.90	5327.0
218.00	446.0
221.00	2014.0
222.10	496.0
223.10	829.0
224.00	9469.0
225.00	2310.0
226.10	443.0
227.00	3624.0
228.00	695.0
228.90	694.0
231.20	200.0
234.00	403.0
234.90	183.0
236.90	307.0
239.00	185.0
240.10	284.0
241.90	566.0
243.20	620.0
244.00	8554.0
245.00	1272.0
246.00	1523.0
247.00	243.0
248.90	261.0
252.00	261.0
255.00	38984.0
256.00	4869.0
257.10	376.0
258.00	2047.0
264.00	179.0
265.00	1089.0
265.80	176.0
272.20	217.0
273.00	1349.0
274.00	2789.0
275.00	14483.0
275.90	2288.0
277.00	1329.0
277.90	404.0
280.80	180.0
285.00	376.0
291.70	182.0
293.00	428.0
294.00	415.0
296.00	4785.0
297.00	1111.0
303.00	281.0
304.10	189.0
309.80	152.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

314.90	420.0
316.00	316.0
320.80	167.0
323.00	1478.0
324.00	377.0
326.80	358.0
333.20	214.0
334.00	942.0
334.90	259.0
335.10	265.0
346.00	472.0
352.00	476.0
353.10	286.0
353.90	686.0
364.90	1181.0
366.00	263.0
371.20	301.0
372.00	813.0
373.00	232.0
383.00	228.0
402.10	195.0
402.90	270.0
404.00	202.0
421.00	422.0
422.00	426.0
423.00	1955.0
424.00	696.0
440.10	159.0
441.00	6532.0
442.00	35560.0
443.00	6775.0
444.00	799.0
445.10	229.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

m/z	Abundance
37.10	178.0
38.10	421.0
39.10	1585.0
41.00	152.0
43.10	175.0
44.10	657.0
49.10	334.0
50.10	7734.0
51.10	29816.0
52.10	1239.0
55.10	406.0
56.10	1074.0
57.10	2174.0
61.10	399.0
62.10	550.0
63.10	1376.0
64.00	228.0
65.00	597.0
68.10	584.0
69.00	23984.0
70.00	182.0
73.00	505.0
74.10	2371.0
75.00	4136.0
76.10	1398.0
77.10	28272.0
78.10	2339.0
79.10	1614.0
80.00	1369.0
81.00	2047.0
82.00	636.0
82.90	512.0
84.00	232.0
85.00	874.0
86.00	822.0
87.00	200.0
87.90	245.0
88.70	203.0
91.00	459.0
92.00	504.0
93.10	3776.0
94.00	299.0
96.00	265.0
98.10	2987.0
99.00	1946.0
99.80	155.0
101.00	1408.0
103.00	380.0
104.00	630.0
105.00	682.0
106.10	597.0
107.00	10033.0
108.00	1016.0
109.00	487.0
110.00	17168.0
111.00	2854.0
111.90	318.0
115.90	366.0
117.00	6996.0
117.80	398.0
122.00	614.0
123.10	992.0
123.90	775.0
125.00	306.0
127.00	31328.0
128.00	2845.0
129.00	13777.0
130.00	1410.0
130.80	215.0
134.00	659.0
135.00	1089.0
136.10	184.0
137.00	244.0
140.00	292.0
141.00	1623.0
142.00	700.0
143.10	337.0
144.00	190.0
147.00	844.0
148.00	1414.0
149.00	536.0
153.00	613.0
154.10	659.0
155.00	1210.0
156.00	1258.0
157.00	498.0
157.90	370.0
159.00	293.0
160.00	996.0
161.00	711.0
161.90	355.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

162.80	191.0
165.00	1248.0
166.00	699.0
167.00	3314.0
168.00	1633.0
168.90	215.0
172.00	406.0
173.20	199.0
174.00	1113.0
175.00	1529.0
175.90	430.0
177.00	620.0
177.80	237.0
179.00	2382.0
180.00	2019.0
181.00	1065.0
181.80	225.0
184.10	425.0
185.10	1408.0
186.00	10813.0
187.00	3247.0
188.00	345.0
189.00	742.0
190.00	312.0
191.00	325.0
192.00	1615.0
193.00	819.0
193.70	256.0
195.00	385.0
196.00	1665.0
198.00	74536.0
198.90	3997.0
199.90	289.0
201.60	341.0
202.90	757.0
204.00	3297.0
205.00	4738.0
206.00	24568.0
207.00	2997.0
208.00	1174.0
209.00	192.0
210.20	279.0
211.00	709.0
214.90	436.0
215.90	440.0
216.90	5669.0
217.90	853.0
221.00	2107.0
222.10	878.0
223.00	1622.0
224.00	11383.0
225.10	3034.0
227.00	5097.0
228.00	457.0
228.90	1112.0
230.00	362.0
230.90	431.0
233.90	304.0
235.00	459.0
237.10	425.0
239.40	233.0
240.00	229.0
241.00	422.0
241.90	342.0
243.10	944.0
244.00	7539.0
245.00	1170.0
246.00	1499.0
247.00	552.0
249.00	309.0
250.80	163.0
252.90	343.0
255.00	51880.0
256.00	6955.0
256.90	761.0
258.00	3141.0
259.00	363.0
260.00	176.0
260.90	227.0
265.00	1068.0
265.90	234.0
269.90	176.0
271.80	190.0
273.00	1795.0
274.00	3290.0
275.00	20824.0
276.00	2583.0
276.90	2204.0
278.00	269.0
281.90	270.0
284.00	225.0
285.10	472.0
288.00	180.0
291.20	203.0
292.90	471.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

295.90	6409.0
297.00	784.0
301.00	212.0
303.00	1160.0
303.90	226.0
307.70	206.0
314.10	206.0
314.90	623.0
316.00	418.0
317.20	187.0
321.00	211.0
321.80	153.0
323.00	1873.0
324.00	347.0
326.90	351.0
327.90	328.0
332.00	198.0
333.10	233.0
334.00	1230.0
335.00	204.0
340.90	187.0
346.00	453.0
352.00	263.0
353.00	393.0
354.10	667.0
365.00	2470.0
366.00	458.0
372.10	1082.0
373.00	582.0
390.00	182.0
402.00	335.0
402.90	425.0
403.90	376.0
421.10	362.0
422.00	614.0
423.00	2774.0
424.00	969.0
441.00	8312.0
442.00	59888.0
443.00	12244.0
444.00	1481.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

m/z	Abundance
39.10	1148.0
39.90	181.0
41.00	176.0
44.00	1220.0
49.00	425.0
50.10	4191.0
51.10	19880.0
52.10	971.0
54.10	187.0
55.20	286.0
56.00	634.0
57.00	1176.0
61.10	277.0
62.10	711.0
63.00	917.0
65.00	413.0
69.00	17144.0
73.00	327.0
74.00	2189.0
75.10	3016.0
77.10	19192.0
78.00	1778.0
79.10	1339.0
80.00	1708.0
81.00	1842.0
82.00	165.0
83.10	443.0
84.00	432.0
85.10	404.0
85.90	497.0
86.90	251.0
88.10	174.0
89.10	164.0
91.00	475.0
92.00	388.0
93.00	2320.0
98.00	1826.0
99.10	1496.0
101.00	827.0
103.00	345.0
103.80	258.0
105.00	484.0
106.00	231.0
107.00	6542.0
108.00	1254.0
110.00	10759.0
111.00	1735.0
112.10	318.0
115.90	352.0
117.00	4223.0
118.00	311.0
120.00	160.0
121.00	152.0
122.10	264.0
123.00	702.0
124.10	582.0
125.00	302.0
125.60	170.0
127.00	22560.0
128.00	1629.0
129.00	7710.0
129.90	1098.0
131.00	251.0
134.00	247.0
135.00	781.0
136.00	282.0
137.00	762.0
140.10	180.0
141.00	1658.0
142.00	324.0
142.90	314.0
145.00	265.0
145.90	492.0
147.00	975.0
148.00	1709.0
149.10	173.0
150.00	255.0
150.90	417.0
152.90	384.0
154.00	506.0
155.00	608.0
156.00	972.0
157.00	249.0
157.60	295.0
160.00	334.0
161.10	544.0
162.00	431.0
162.90	232.0
165.00	638.0
165.90	450.0
167.00	2729.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

168.00	1706.0
169.10	348.0
171.20	182.0
173.00	188.0
173.90	328.0
175.10	1117.0
176.10	230.0
176.90	516.0
178.00	300.0
178.90	1871.0
180.10	1718.0
181.00	614.0
185.00	1147.0
186.00	7402.0
187.10	2070.0
188.10	271.0
188.90	336.0
191.10	158.0
191.90	661.0
193.00	761.0
195.20	331.0
196.00	1658.0
198.00	55512.0
198.90	3816.0
200.00	505.0
201.50	251.0
202.10	259.0
203.00	355.0
204.00	2287.0
205.00	4562.0
206.00	17168.0
207.00	2714.0
208.00	358.0
209.90	175.0
211.00	854.0
215.90	234.0
217.00	3874.0
218.00	436.0
220.20	322.0
221.00	2223.0
221.90	513.0
222.90	928.0
224.00	9599.0
225.00	2505.0
226.00	376.0
227.00	4591.0
227.90	801.0
229.00	842.0
230.90	297.0
233.80	289.0
235.10	444.0
236.00	151.0
236.90	306.0
240.80	207.0
242.00	400.0
243.10	712.0
244.10	7001.0
245.00	660.0
245.90	1457.0
247.20	229.0
249.00	527.0
253.10	301.0
255.00	37992.0
256.00	5400.0
257.00	522.0
258.00	1871.0
258.90	346.0
262.90	286.0
265.00	968.0
270.10	237.0
271.10	178.0
273.00	1343.0
274.00	3116.0
275.00	17768.0
276.00	2198.0
277.00	2041.0
278.00	261.0
282.90	296.0
284.80	278.0
293.00	304.0
294.90	167.0
296.00	4933.0
297.00	405.0
303.00	694.0
308.10	242.0
313.10	168.0
314.00	271.0
314.90	663.0
315.90	311.0
321.20	166.0
323.00	1461.0
324.10	258.0
327.10	170.0
327.80	200.0
333.00	184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

334.10	1030.0
335.00	373.0
345.70	212.0
352.00	876.0
353.00	326.0
354.00	483.0
365.00	2264.0
365.90	195.0
370.80	176.0
371.90	761.0
373.10	197.0
383.00	366.0
390.00	303.0
391.20	230.0
401.90	159.0
403.00	426.0
403.80	150.0
421.00	361.0
422.00	331.0
423.00	3938.0
424.00	838.0
441.00	8652.0
442.00	56504.0
443.00	11545.0
444.00	845.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63

Scan 2463 (18.062 min): BG028292.D
DFTPP63

m/z	Abundance
44.00	806.0
49.00	291.0
135.20	160.0
314.90	177.0

Instrument :
BNA_G
ClientSampleId :
DFTPP63