


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081324\
Data File : BG062377.D
Acq On : 13 Aug 2024 9:18
Operator : MA/JU
Sample : DFTPP411
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP411

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG080524.MA.M
Title : SVOA CALIBRATION
Last Update : Fri Aug 09 01:00:03 2024



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2457

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	41.8	133611	PASS
68	69	0.00	2	0.7	743	PASS
69	198	0.00	100	32.4	103517	PASS
70	69	0.00	2	0.5	547	PASS
127	198	10	80	38.7	123704	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	319680	PASS
199	198	5	9	7.0	22256	PASS
275	198	10	60	26.8	85773	PASS
365	198	1	100	4.2	13275	PASS
441	198	0.01	100	11.3	36040	PASS
442	198	50	100	67.3	215253	PASS
443	442	15	24	19.5	41971	PASS

m/z	Abundance
41.25	189.0
43.05	302.0
44.05	1447.0
55.05	195.0
57.15	183.0
69.25	163.0
83.95	178.0
167.15	185.0
207.15	222.0

Instrument :

BNA_G

ClientSampleId :

DFTPP411

m/z	Abundance
37.15	377.0
38.15	1469.0
39.15	8026.0
40.15	542.0
41.25	309.0
43.15	186.0
44.15	1793.0
45.15	293.0
50.15	28296.0
51.15	118408.0
52.15	6206.0
53.25	382.0
55.05	666.0
56.15	3500.0
57.15	7350.0
58.05	259.0
61.15	1283.0
62.15	1422.0
63.15	4421.0
64.05	573.0
65.15	2175.0
65.95	254.0
69.05	90264.0
70.15	481.0
71.25	150.0
73.15	829.0
74.15	9366.0
75.05	14664.0
76.15	5477.0
77.15	108536.0
78.15	7147.0
79.05	6783.0
80.15	5385.0
81.05	6890.0
82.05	1883.0
83.15	1858.0
84.05	339.0
85.05	1504.0
86.05	1803.0
87.15	963.0
87.95	394.0
89.05	157.0
91.15	1471.0
92.05	1622.0
93.05	12350.0
94.05	703.0
95.05	299.0
96.15	428.0
98.15	8854.0
99.15	7088.0
99.95	429.0
101.15	4178.0
102.05	266.0
103.05	1051.0
104.05	2914.0
105.05	2381.0
107.05	32472.0
108.15	4800.0
110.05	56408.0
111.15	9313.0
112.15	1303.0
113.15	281.0
116.15	1510.0
117.05	26832.0
118.05	2097.0
119.05	284.0
120.15	449.0
122.05	1760.0
123.15	3212.0
124.05	1217.0
125.05	1467.0
127.15	100464.0
128.15	8458.0
129.05	39488.0
130.05	3408.0
131.05	585.0
132.15	409.0
134.05	1082.0
135.05	3544.0
136.05	1239.0
137.15	1509.0
137.75	362.0
139.05	170.0
140.05	618.0
141.15	4968.0
142.05	1508.0
143.05	1196.0
144.05	367.0
145.05	368.0
146.05	1103.0
147.05	2976.0
148.05	7274.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

149.05	1270.0
150.15	473.0
151.15	639.0
151.75	445.0
153.05	1407.0
154.15	1370.0
155.15	3304.0
156.15	4716.0
157.05	827.0
158.05	930.0
159.05	853.0
160.05	1573.0
161.15	2498.0
162.05	938.0
162.95	292.0
164.05	387.0
165.15	2029.0
166.05	2098.0
167.15	11440.0
168.05	5292.0
169.15	1051.0
169.95	455.0
171.05	599.0
171.95	940.0
173.05	1150.0
174.05	2516.0
175.15	4741.0
176.15	1181.0
177.05	2141.0
178.15	748.0
179.05	8658.0
180.05	5847.0
181.15	2530.0
182.15	450.0
183.15	261.0
184.05	762.0
185.15	4750.0
186.15	33368.0
187.05	9800.0
188.05	1138.0
189.05	1953.0
190.15	341.0
191.05	1313.0
192.05	2465.0
193.15	3885.0
194.15	860.0
194.95	427.0
196.05	6742.0
198.05	250240.0
199.05	18304.0
200.05	1254.0
201.55	1114.0
202.15	348.0
203.05	1679.0
204.15	9858.0
205.05	17144.0
206.15	69008.0
207.05	9453.0
208.05	2650.0
209.05	742.0
210.05	1006.0
211.05	2751.0
212.95	226.0
215.05	968.0
216.05	1593.0
217.05	18200.0
218.05	2275.0
219.05	234.0
221.05	8413.0
221.95	1955.0
223.05	4317.0
224.05	39104.0
225.05	9271.0
226.05	1200.0
227.05	18288.0
228.05	2711.0
229.05	3394.0
229.95	420.0
231.05	1665.0
231.95	256.0
233.05	334.0
234.05	838.0
235.05	1110.0
236.05	771.0
237.05	1293.0
238.05	277.0
239.05	524.0
240.15	498.0
241.05	853.0
242.05	2164.0
243.05	1936.0
244.05	29288.0
245.15	5012.0
246.05	6425.0
247.05	1252.0
248.05	303.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

249.05	987.0
249.85	279.0
251.05	214.0
251.95	421.0
252.95	712.0
255.05	151808.0
256.05	23504.0
257.05	1872.0
258.05	10491.0
259.15	1664.0
260.15	285.0
261.05	237.0
264.05	445.0
265.05	3977.0
266.05	536.0
270.15	206.0
271.05	342.0
272.05	424.0
273.05	4630.0
274.05	12040.0
275.05	62744.0
276.05	9022.0
277.05	6456.0
278.05	1026.0
279.05	238.0
282.95	748.0
284.05	295.0
284.95	982.0
285.95	178.0
288.95	215.0
291.05	332.0
292.05	304.0
293.05	1154.0
293.95	397.0
296.05	18600.0
297.05	3013.0
298.05	311.0
300.95	173.0
302.05	360.0
303.05	2133.0
304.05	456.0
308.05	278.0
309.05	239.0
309.95	230.0
313.05	278.0
314.05	775.0
315.05	2378.0
316.05	1349.0
316.85	294.0
321.15	757.0
322.05	381.0
323.05	6034.0
324.15	1213.0
326.95	1079.0
328.15	576.0
328.95	211.0
331.95	564.0
333.05	594.0
334.05	3607.0
335.05	984.0
340.15	152.0
341.05	491.0
342.05	170.0
346.05	1240.0
347.05	246.0
350.95	159.0
352.05	2042.0
353.05	1158.0
354.05	2010.0
355.05	389.0
364.95	9404.0
365.95	1520.0
366.85	175.0
371.15	486.0
372.05	2671.0
373.05	909.0
383.05	784.0
383.95	223.0
384.95	215.0
389.95	451.0
390.95	236.0
392.05	300.0
400.95	198.0
402.05	1317.0
403.05	1607.0
403.95	516.0
420.95	1359.0
421.95	1213.0
423.05	8171.0
424.05	1640.0
441.05	22864.0
442.05	139520.0
443.05	29704.0
444.05	2983.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

m/z	Abundance
36.05	216.0
37.15	600.0
38.15	1870.0
39.15	10359.0
40.15	688.0
41.15	441.0
43.15	353.0
44.05	1405.0
45.25	350.0
50.15	38528.0
51.15	156352.0
52.15	7806.0
53.15	261.0
55.15	661.0
56.15	4494.0
57.15	10033.0
58.05	421.0
60.05	166.0
61.05	1897.0
62.15	2290.0
63.15	6257.0
64.15	813.0
65.05	2664.0
66.15	238.0
67.15	287.0
68.15	2228.0
69.05	118840.0
70.15	603.0
71.05	472.0
73.05	752.0
74.15	14094.0
75.15	22264.0
76.15	7397.0
77.15	151616.0
78.15	9772.0
79.15	9154.0
80.15	7159.0
81.15	10476.0
82.15	2671.0
83.15	2331.0
84.15	190.0
85.15	2043.0
86.05	3185.0
87.15	1362.0
88.15	539.0
88.95	197.0
91.05	2301.0
92.05	2614.0
93.05	15757.0
94.15	1202.0
95.05	282.0
96.05	768.0
98.15	13626.0
99.15	10095.0
100.15	968.0
101.05	5943.0
102.15	265.0
103.15	1917.0
104.15	3684.0
105.05	3476.0
107.15	42232.0
108.15	7186.0
110.05	76520.0
111.15	12236.0
112.15	1559.0
113.15	509.0
116.15	2293.0
117.05	38648.0
118.05	2864.0
119.05	355.0
120.15	548.0
121.05	342.0
122.15	3026.0
123.05	4688.0
124.15	1695.0
125.15	2053.0
127.15	142848.0
128.15	11074.0
129.05	56384.0
130.15	4663.0
131.05	692.0
132.15	609.0
133.05	177.0
134.05	1807.0
135.05	4534.0
136.05	2174.0
137.15	2191.0
137.95	737.0
138.95	293.0
140.15	811.0
141.15	7942.0
142.05	2453.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

143.05	1717.0
144.05	539.0
145.05	378.0
146.05	1655.0
147.15	4030.0
148.05	10136.0
149.15	1956.0
150.05	516.0
151.15	1027.0
151.85	738.0
153.05	3122.0
154.05	1878.0
155.05	4889.0
156.05	7139.0
157.15	1047.0
158.05	1501.0
159.05	1168.0
160.05	2417.0
161.05	3924.0
162.05	917.0
163.25	379.0
164.15	589.0
165.05	3226.0
166.15	2737.0
167.05	16109.0
168.05	8315.0
169.05	1540.0
170.05	616.0
171.05	1206.0
172.05	1833.0
173.05	1912.0
174.05	3541.0
175.15	6898.0
176.05	1721.0
177.05	3602.0
178.15	1339.0
179.05	12690.0
180.05	8624.0
181.05	4189.0
182.15	716.0
182.95	465.0
184.05	787.0
185.05	6259.0
186.15	49072.0
187.15	14658.0
188.15	1527.0
189.05	2743.0
190.05	554.0
191.05	1298.0
192.05	4654.0
193.05	4759.0
194.15	1226.0
195.15	858.0
196.15	10634.0
198.05	367360.0
199.05	25496.0
200.15	1725.0
201.55	1342.0
203.15	3248.0
204.15	12910.0
205.15	23960.0
206.15	100824.0
207.15	13122.0
208.05	3660.0
209.15	1142.0
211.05	3705.0
213.05	284.0
215.05	1648.0
216.15	1888.0
217.05	28432.0
218.05	3497.0
218.95	415.0
221.15	12716.0
221.95	3949.0
223.05	6075.0
224.05	60320.0
225.15	15208.0
226.15	1607.0
227.05	26088.0
228.05	3692.0
229.05	5490.0
230.05	677.0
231.05	2314.0
232.15	426.0
233.15	293.0
234.15	1998.0
235.05	1794.0
236.05	1156.0
237.05	1884.0
238.15	315.0
239.05	1206.0
240.05	892.0
240.95	1236.0
242.05	2807.0
243.05	3094.0
244.05	42992.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

245.05	5737.0
246.05	9654.0
247.05	1627.0
248.05	403.0
249.05	1667.0
250.15	368.0
250.95	361.0
252.15	691.0
253.05	921.0
255.05	228608.0
256.05	32840.0
257.05	3001.0
258.05	15078.0
259.05	2249.0
259.95	289.0
260.95	435.0
264.05	214.0
265.05	6367.0
266.05	1116.0
266.85	355.0
270.05	383.0
271.15	487.0
272.15	970.0
273.05	6881.0
274.05	18696.0
275.05	97152.0
276.05	13416.0
277.05	9477.0
278.05	1491.0
279.15	244.0
281.95	212.0
283.05	834.0
284.15	735.0
285.15	1474.0
286.05	441.0
286.95	171.0
288.25	178.0
288.85	344.0
290.05	357.0
290.85	366.0
292.15	644.0
293.05	2141.0
294.25	484.0
296.05	30696.0
297.05	4533.0
298.05	420.0
300.95	482.0
302.15	518.0
303.05	3691.0
304.05	1290.0
305.15	157.0
308.05	471.0
309.05	312.0
310.15	589.0
312.95	258.0
314.15	1434.0
315.05	3898.0
316.05	1985.0
317.15	287.0
321.05	1051.0
322.15	594.0
323.05	9549.0
324.15	1963.0
326.05	279.0
326.95	1763.0
328.15	1060.0
332.05	561.0
333.15	930.0
334.05	6739.0
335.15	1548.0
336.15	206.0
338.85	186.0
340.15	274.0
341.05	891.0
342.15	252.0
345.95	2027.0
347.05	491.0
349.95	152.0
350.95	170.0
352.05	2522.0
353.05	2094.0
354.15	3390.0
355.05	443.0
359.05	259.0
365.05	14150.0
366.05	2201.0
367.05	179.0
370.05	249.0
371.05	753.0
372.05	5547.0
373.05	1266.0
376.95	176.0
383.05	1508.0
384.15	301.0
389.95	590.0
391.15	466.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

392.05	225.0
401.05	295.0
402.05	1843.0
403.05	2785.0
404.05	994.0
405.15	184.0
420.95	2006.0
422.05	2333.0
423.05	13762.0
424.05	3008.0
425.05	276.0
430.15	188.0
441.05	39616.0
442.05	234816.0
443.05	44248.0
444.05	4249.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

m/z	Abundance
36.05	154.0
37.05	432.0
38.15	1174.0
39.15	8765.0
40.15	483.0
41.15	408.0
43.15	455.0
44.15	1524.0
45.15	291.0
49.15	886.0
50.15	31088.0
51.15	126072.0
52.15	6001.0
53.15	186.0
54.95	659.0
56.05	3371.0
57.05	8249.0
58.05	291.0
59.15	184.0
61.15	1260.0
62.15	1696.0
63.15	4276.0
64.15	743.0
65.05	1962.0
66.15	174.0
69.05	101936.0
70.15	558.0
71.05	182.0
73.15	1002.0
74.15	10835.0
75.15	16225.0
76.15	5955.0
77.15	123224.0
78.15	7710.0
79.05	7678.0
80.15	5840.0
81.05	9130.0
82.05	2394.0
83.05	1903.0
84.15	310.0
85.15	1613.0
86.05	2124.0
87.05	1145.0
88.25	319.0
89.05	226.0
91.05	1805.0
92.05	2177.0
93.05	13092.0
94.05	1079.0
95.15	263.0
96.15	716.0
96.95	266.0
98.05	11102.0
99.05	7758.0
100.05	603.0
101.05	4639.0
102.25	270.0
103.15	1685.0
104.05	2837.0
105.15	3330.0
107.15	37752.0
108.15	5045.0
110.05	68144.0
111.15	9821.0
112.15	979.0
113.15	534.0
116.05	2095.0
117.05	33464.0
118.05	2448.0
119.05	271.0
120.15	456.0
122.05	2354.0
123.15	3734.0
124.05	1829.0
125.05	1778.0
127.15	127800.0
128.15	9967.0
129.05	48248.0
130.15	3910.0
131.05	942.0
132.15	292.0
133.95	1245.0
135.15	4084.0
136.15	1513.0
137.15	1805.0
138.15	599.0
139.05	340.0
140.05	647.0
141.05	6505.0
142.05	2321.0
143.05	1708.0
143.95	542.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

145.25	441.0
146.05	1378.0
147.05	4710.0
148.05	8587.0
149.05	1949.0
150.05	498.0
151.15	881.0
151.95	460.0
153.05	2642.0
154.05	1710.0
155.15	4794.0
156.15	6012.0
157.15	1269.0
157.95	1572.0
159.05	1125.0
160.15	1956.0
161.05	3731.0
162.15	980.0
163.05	175.0
164.25	514.0
165.05	2990.0
166.15	2302.0
167.05	15372.0
168.05	7003.0
169.05	1508.0
170.05	548.0
171.05	631.0
172.05	1310.0
173.05	1643.0
174.15	3711.0
175.15	6472.0
176.15	1605.0
177.05	2906.0
178.05	1171.0
179.05	12065.0
180.15	8061.0
181.05	3579.0
182.15	630.0
182.95	320.0
184.05	1087.0
185.15	6088.0
186.15	47112.0
187.15	12917.0
188.05	1207.0
189.05	3023.0
189.95	434.0
191.05	1582.0
192.05	4406.0
193.05	4472.0
194.15	1106.0
195.05	805.0
196.15	9619.0
198.05	341440.0
199.05	22968.0
200.05	1629.0
201.75	1403.0
203.15	2758.0
204.15	12691.0
205.15	23168.0
206.05	95832.0
207.05	12093.0
208.05	3042.0
209.05	1114.0
210.25	1683.0
211.15	3852.0
213.05	364.0
214.05	165.0
215.05	1177.0
216.05	2358.0
217.05	25568.0
218.05	3413.0
219.15	346.0
219.75	331.0
221.05	12257.0
222.05	3890.0
223.05	6167.0
224.05	54840.0
225.15	13885.0
226.15	1417.0
227.05	25400.0
228.15	3559.0
229.05	5155.0
230.15	494.0
231.15	1897.0
232.15	396.0
233.05	560.0
234.05	1776.0
235.05	1522.0
236.05	1195.0
237.15	1691.0
238.15	293.0
239.05	846.0
240.05	744.0
241.05	1643.0
242.05	3073.0
243.15	3095.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

244.05	40480.0
245.05	5818.0
246.05	8995.0
247.05	1914.0
248.15	604.0
249.05	1704.0
250.15	188.0
251.05	325.0
252.15	500.0
253.05	1045.0
255.05	227712.0
256.05	33160.0
257.05	2461.0
258.05	14827.0
259.15	2326.0
260.05	566.0
261.05	486.0
264.15	451.0
265.05	6407.0
266.05	966.0
269.55	193.0
270.15	274.0
271.15	552.0
271.95	930.0
273.05	6534.0
274.05	18192.0
275.05	97424.0
276.05	12888.0
277.05	9434.0
278.05	1695.0
278.95	299.0
282.05	225.0
282.95	1089.0
284.05	779.0
285.05	1813.0
285.95	279.0
289.15	324.0
290.05	388.0
291.15	283.0
292.05	437.0
293.15	2021.0
294.05	577.0
296.05	32608.0
297.05	4719.0
297.95	284.0
301.15	543.0
302.05	567.0
303.05	3805.0
304.15	1056.0
307.95	518.0
309.05	166.0
310.15	284.0
312.95	270.0
314.05	1605.0
315.05	3196.0
316.15	1780.0
317.25	305.0
320.15	170.0
320.95	836.0
322.15	406.0
323.05	9631.0
324.15	1776.0
325.15	187.0
325.95	209.0
326.95	2002.0
328.15	830.0
332.15	811.0
333.05	925.0
334.05	7158.0
335.15	1586.0
336.15	211.0
341.05	1088.0
342.05	517.0
346.05	2427.0
346.85	393.0
351.15	217.0
352.05	3236.0
353.05	2500.0
354.05	3662.0
355.05	604.0
359.15	322.0
365.05	16270.0
366.05	2126.0
366.95	213.0
369.85	173.0
371.05	777.0
372.05	4825.0
373.15	1076.0
373.95	288.0
377.05	211.0
383.05	1814.0
384.05	513.0
385.15	173.0
389.95	654.0
390.95	353.0
392.15	511.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411

401.05	430.0
401.95	2480.0
403.05	2776.0
404.15	846.0
404.95	307.0
421.05	2342.0
422.05	2428.0
423.05	16696.0
424.05	3227.0
425.05	279.0
441.05	45640.0
442.05	271424.0
443.05	51960.0
444.05	5313.0
445.05	258.0

Instrument :
BNA_G
ClientSampleId :
DFTPP411