

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081415\
 Data File : BG018442.D
 Acq On : 14 Aug 2015 23:44
 Operator : UM/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Aug 17 03:22:41 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 17 03:15:20 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.485	0.472	2.7	110	0.00
3 S	1,4-Dioxane-d8	0.453	0.445	1.8	109	0.00
4	Benzaldehyde	1.065	1.243	-16.7	114	0.00
5 S	Phenol-d5	1.815	1.842	-1.5	113	0.00
6	Phenol	1.852	1.961	-5.9	118	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.125	1.128	-0.3	111	0.00
8	Bis(2-Chloroethyl)ether	1.424	1.474	-3.5	112	0.00
9 S	2-Chlorophenol-d4	1.390	1.370	1.4	112	0.00
10	2-Chlorophenol	1.417	1.432	-1.1	111	0.00
11	2-Methylphenol	1.436	1.494	-4.0	114	0.00
12	2,2'-oxybis(1-Chloropropane	2.286	1.993	12.8	95	0.00
13 S	4-Methylphenol-d8	1.525	1.559	-2.2	114	0.00
14	Acetophenone	2.203	2.353	-6.8	114	0.00
15 P	N-Nitroso-di-n-propylamine	1.264	1.290	-2.1	113	0.00
16	4-Methylphenol	1.567	1.626	-3.8	117	0.00
17	Hexachloroethane	0.611	0.590	3.4	110	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
19 S	Nitrobenzene-d5	0.156	0.164	-5.1	122	0.00
20	Nitrobenzene	0.393	0.387	1.5	112	0.00
21	Isophorone	0.757	0.762	-0.7	115	0.00
22 S	2-Nitrophenol-d4	0.159	0.169	-6.3	120	0.00
23 C	2-Nitrophenol	0.176	0.183	-4.0	119	0.00
24	2,4-Dimethylphenol	0.395	0.388	1.8	115	0.00
25	Bis(2-Chloroethoxy)methane	0.472	0.470	0.4	113	0.00
26 S	2,4-Dichlorophenol-d3	0.337	0.344	-2.1	119	0.00
27 C	2,4-Dichlorophenol	0.316	0.320	-1.3	116	0.00
28	Naphthalene	1.057	1.057	0.0	113	0.00
29 S	4-Chloroaniline-d4	0.381	0.456	-19.7	116	0.00
30	4-Chloroaniline	0.387	0.461	-19.1	115	0.00
31 C	Hexachlorobutadiene	0.199	0.192	3.5	111	0.00
32	Caprolactam	0.127	0.140	-10.2	125	0.00
33 C	4-Chloro-3-methylphenol	0.366	0.383	-4.6	122	0.00
34	2-Methylnaphthalene	0.766	0.791	-3.3	118	0.00
35	1-Methylnaphthalene	0.748	0.769	-2.8	116	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
37	1,2,4,5-Tetrachlorobenzene	0.641	0.642	-0.2	118	0.00
38	Hexachlorocyclopentadiene	0.382	0.360	5.8	108	0.00
39 C	2,4,6-Trichlorophenol	0.437	0.441	-0.9	118	0.00
40	2,4,5-Trichlorophenol	0.470	0.452	3.8	113	0.00
41	1,1'-Biphenyl	1.669	1.693	-1.4	118	0.00
42	2-Chloronaphthalene	1.297	1.297	0.0	118	0.00
43	2-Nitroaniline	0.419	0.429	-2.4	122	0.00
44 S	Dimethylphthalate-d6	1.683	1.707	-1.4	120	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081415\
 Data File : BG018442.D
 Acq On : 14 Aug 2015 23:44
 Operator : UM/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Aug 17 03:22:41 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 17 03:15:20 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.660	1.683	-1.4	119	0.00
46	2,6-Dinitrotoluene	0.331	0.367	-10.9	129	0.00
47 S	Acenaphthylene-d8	2.119	2.175	-2.6	119	0.00
48	Acenaphthylene	2.125	2.215	-4.2	121	0.00
49	3-Nitroaniline	0.337	0.412	-22.3#	135	0.00
50 C	Acenaphthene	1.491	1.509	-1.2	119	0.00
51	2,4-Dinitrophenol	0.161	0.164	-1.9	132	0.00
52 S	4-Nitrophenol-d4	0.303	0.328	-8.3	132	0.00
53	4-Nitrophenol	0.263	0.262	0.4	115	0.00
54	Dibenzofuran	1.948	1.995	-2.4	118	0.00
55	2,4-Dinitrotoluene	0.472	0.526	-11.4	131	0.00
56	2,3,4,6-Tetrachlorophenol	0.409	0.407	0.5	115	0.00
57	Diethylphthalate	1.764	1.800	-2.0	122	0.00
58 S	Fluorene-d10	1.466	1.513	-3.2	122	0.00
59	Fluorene	1.676	1.700	-1.4	121	0.00
60	4-Chlorophenyl-phenylether	0.793	0.814	-2.6	119	0.00
61	4-Nitroaniline	0.372	0.420	-12.9	132	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.098	0.105	-7.1	133	0.00
64	4,6-Dinitro-2-methylphenol	0.105	0.111	-5.7	135	0.00
65	N-Nitrosodiphenylamine	0.602	0.599	0.5	123	0.00
66	4-Bromophenyl-phenylether	0.216	0.215	0.5	122	0.00
67	Hexachlorobenzene	0.229	0.226	1.3	121	0.00
68	Atrazine	0.225	0.245	-8.9	126	0.00
69 C	Pentachlorophenol	0.153	0.136	11.1	110	0.00
70	Phenanthrene	1.085	1.100	-1.4	125	0.00
71 S	Anthracene-d10	0.957	0.965	-0.8	122	0.00
72	Anthracene	1.106	1.133	-2.4	124	0.00
73	1,2,3,4-Tetrachlorobenzene	0.277	0.268	3.2	118	0.00
74	Pentachlorobenzene	0.264	0.256	3.0	120	0.00
75	Carbazole	0.970	1.085	-11.9	127	0.00
76	Di-n-butylphthalate	1.266	1.326	-4.7	126	0.00
77 C	Fluoranthene	1.159	1.327	-14.5	125	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
79 S	Pyrene-d10	1.038	1.067	-2.8	126	0.00
80	Pyrene	1.352	1.362	-0.7	123	0.00
81	Butylbenzylphthalate	0.572	0.602	-5.2	131	0.00
82	3,3'-Dichlorobenzidine	0.415	0.470	-13.3	128	0.00
83	Benzo(a)anthracene	1.240	1.259	-1.5	126	0.00
84	Bis(2-ethylhexyl)phthalate	0.808	0.850	-5.2	131	0.00
85	Chrysene	1.159	1.171	-1.0	126	0.00
86 I	Perylene-d12	1.000	1.000	0.0	131	0.00
87	Di-n-octyl phthalate	1.290	1.459	-13.1	130	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081415\
Data File : BG018442.D
Acq On : 14 Aug 2015 23:44
Operator : UM/MA
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Aug 17 03:22:41 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 17 03:15:20 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.257	1.248	0.7	126	0.00
89	Benzo(k)fluoranthene	1.210	1.218	-0.7	129	0.00
90 S	Benzo(a)pyrene-d12	1.062	1.050	1.1	125	-0.01
91 C	Benzo(a)pyrene	1.195	1.185	0.8	124	-0.01
92	Indeno(1,2,3-cd)pyrene	1.383	1.374	0.7	127	0.00
93	Dibenzo(a,h)anthracene	1.148	1.132	1.4	127	-0.01
94	Benzo(g,h,i)perylene	1.161	1.138	2.0	126	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0