

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081424\
 Data File : BG062399.D
 Acq On : 14 Aug 2024 17:39
 Operator : MA/JU
 Sample : P3555-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DCXT2

Quant Time: Aug 14 18:20:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 15:31:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.960	152	43891	20.000	ng/u1	0.00
20) Naphthalene-d8	10.751	136	205820	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	174383	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.318	188	434210	20.000	ng/u1	0.00
79) Chrysene-d12	21.560	240	458024	20.000	ng/u1	0.00
88) Perylene-d12	24.650	264	527767	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.419	96	5225	4.520	ng/uL	0.00
4) Pyridine-d5	3.831	84	13206	3.749	ng/u1	0.00
7) Phenol-d5	7.114	99	34733	7.955	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.285	67	83468	30.348	ng/u1	0.00
11) 2-Chlorophenol-d4	7.490	132	84751	28.415	ng/u1	0.00
15) 4-Methylphenol-d8	8.654	113	65438	17.879	ng/u1	0.00
21) Nitrobenzene-d5	9.106	128	49994	37.876	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	51358	39.115	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	103619	30.984	ng/u1	0.00
31) 4-Chloroaniline-d4	10.880	131	34090	6.428	ng/u1	0.00
46) Dimethylphthalate-d6	13.982	166	444137	32.743	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	454805	30.282	ng/u1	0.00
54) 4-Nitrophenol-d4	14.757	143	17341	8.014	ng/u1	0.00
60) Fluorene-d10	15.568	176	396626	32.420	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.673	200	62392	31.625	ng/u1	-0.01
73) Anthracene-d10	17.418	188	714897	34.273	ng/u1	0.00
81) Pyrene-d10	19.697	212	933115	34.787	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.444	264	983606	35.118	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081424\
Data File : BG062399.D
Acq On : 14 Aug 2024 17:39
Operator : MA/JU
Sample : P3555-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DCXT2

Quant Time: Aug 14 18:20:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 14 15:31:21 2024
Response via : Initial Calibration

