


Data Path : Z:\HPCHEM1\BNA_G\Data\BG081516\
 Data File : BG023717.D
 Acq On : 15 Aug 2016 9:25
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\Methods\8270-BG080116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 01 19:22:14 2016



AutoFind: Scans 2575, 2576, 2577; Background Corrected with Scan 2568

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.9	27906	PASS
68	69	0.00	2	1.2	397	PASS
69	198	0.00	100	40.7	33489	PASS
70	69	0.00	2	0.6	190	PASS
127	198	10	80	41.7	34293	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	82322	PASS
199	198	5	9	7.0	5791	PASS
275	198	10	60	27.1	22336	PASS
365	198	1	100	4.0	3333	PASS
441	442	0.01	24	16.3	9361	PASS
442	198	50	100	69.6	57314	PASS
443	442	15	24	19.9	11382	PASS