


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG081618\
 Data File : BG036304.D
 Acq On : 16 Aug 2018 9:35
 Operator : JU/SJ
 Sample : DFTPP02
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP02

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG081618MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Aug 16 14:12:03 2018



AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2480

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.5	55928	PASS
68	69	0.00	2	1.3	830	PASS
69	198	0.00	100	48.4	62112	PASS
70	69	0.00	2	0.7	427	PASS
127	198	10	80	47.0	60365	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	128434	PASS
199	198	5	9	7.0	9009	PASS
275	198	10	60	28.4	36498	PASS
365	198	1	100	3.6	4677	PASS
441	198	0.01	100	9.9	12687	PASS
442	198	50	100	65.2	83768	PASS
443	442	15	24	19.6	16388	PASS

m/z	Abundance
37.00	233.0
38.10	886.0
39.10	2862.0
40.00	468.0
44.00	1008.0
49.00	485.0
50.10	13029.0
51.10	45712.0
52.10	2996.0
55.00	302.0
56.00	1182.0
57.10	3986.0
57.90	257.0
61.00	999.0
62.10	610.0
63.10	2019.0
64.00	383.0
65.00	1190.0
68.00	1150.0
69.00	50624.0
70.00	201.0
73.00	786.0
74.10	6267.0
75.10	8184.0
76.00	2684.0
77.10	57056.0
78.20	3147.0
79.00	4396.0
80.00	3254.0
81.00	3712.0
82.10	860.0
83.00	739.0
84.10	230.0
85.00	760.0
86.00	1384.0
87.00	642.0
88.00	252.0
91.10	822.0
92.00	1094.0
93.10	7952.0
94.00	392.0
94.80	153.0
96.10	311.0
98.00	7227.0
99.00	4113.0
100.00	393.0
101.00	2674.0
103.00	584.0
104.10	1284.0
105.10	1466.0
105.80	440.0
107.00	19176.0
108.00	2103.0
110.00	23560.0
111.00	4743.0
111.80	294.0
112.10	259.0
113.10	212.0
114.90	194.0
116.10	1144.0
117.00	14653.0
118.10	949.0
119.10	216.0
119.90	251.0
122.10	1383.0
123.00	2634.0
124.10	720.0
125.00	580.0
125.90	282.0
127.00	45288.0
128.00	2956.0
129.00	21328.0
130.00	2115.0
130.90	359.0
132.00	288.0
133.90	764.0
135.00	2066.0
136.00	813.0
137.10	926.0
137.90	203.0
141.00	3314.0
141.90	918.0
142.60	261.0
143.10	461.0
143.90	226.0
145.90	447.0
147.10	1165.0
148.00	3378.0
149.10	699.0
151.00	253.0
153.00	972.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

154.00	863.0
155.10	2136.0
156.00	2347.0
156.90	831.0
158.10	1031.0
159.00	569.0
160.00	764.0
161.00	1400.0
162.10	542.0
165.00	1784.0
166.00	647.0
167.00	6365.0
168.00	2512.0
169.00	599.0
170.10	275.0
171.00	258.0
172.00	485.0
172.90	819.0
174.00	1217.0
175.00	1917.0
176.10	488.0
176.90	711.0
177.60	280.0
179.00	3496.0
180.00	3314.0
181.00	959.0
183.90	289.0
185.00	1927.0
186.00	14701.0
187.00	3548.0
188.00	645.0
189.10	998.0
189.80	167.0
191.00	626.0
192.10	1463.0
193.00	1601.0
194.20	370.0
195.00	329.0
196.00	2963.0
198.00	95960.0
199.00	6791.0
200.00	903.0
200.80	163.0
201.70	211.0
201.90	218.0
203.00	1050.0
204.00	4571.0
205.00	6949.0
206.00	31184.0
207.00	3815.0
208.00	1699.0
209.00	613.0
210.00	596.0
210.70	766.0
211.10	701.0
213.10	153.0
215.00	439.0
216.00	824.0
217.00	9349.0
218.00	1195.0
219.80	156.0
221.00	2409.0
222.00	1107.0
223.00	2199.0
224.00	18264.0
225.00	3855.0
226.10	612.0
227.00	9363.0
228.00	1181.0
229.00	1619.0
231.00	443.0
232.80	187.0
234.00	666.0
234.90	179.0
236.00	448.0
237.00	411.0
238.90	430.0
239.90	239.0
241.00	454.0
242.00	535.0
243.10	1833.0
244.10	9927.0
245.00	1093.0
246.00	2881.0
246.90	292.0
247.80	372.0
249.10	625.0
249.70	172.0
252.90	470.0
255.00	63048.0
256.00	11289.0
257.00	619.0
258.00	4656.0
258.90	976.0
264.00	220.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

265.00	1735.0
266.00	376.0
269.90	178.0
271.10	210.0
272.20	267.0
273.00	1487.0
274.00	4180.0
275.00	25272.0
276.00	3369.0
277.00	3095.0
277.90	532.0
278.90	185.0
281.10	192.0
281.80	205.0
283.00	390.0
285.10	625.0
292.90	1018.0
293.70	214.0
296.00	8569.0
297.00	1257.0
301.00	202.0
301.60	235.0
302.00	209.0
303.00	827.0
304.10	238.0
308.00	266.0
312.90	328.0
314.00	482.0
314.90	916.0
316.00	334.0
321.30	217.0
323.10	1692.0
324.20	166.0
326.90	206.0
328.00	254.0
330.30	235.0
332.00	157.0
333.10	152.0
333.90	1269.0
335.00	304.0
336.20	184.0
345.90	422.0
347.80	179.0
352.00	569.0
352.90	514.0
354.00	860.0
355.00	194.0
365.00	2643.0
365.90	159.0
372.00	1133.0
373.00	232.0
382.90	200.0
390.00	251.0
390.80	165.0
402.00	202.0
403.00	476.0
404.00	273.0
421.00	329.0
421.80	266.0
423.00	1776.0
424.00	988.0
441.00	5958.0
442.00	47624.0
443.00	8684.0
444.00	997.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

m/z	Abundance
37.00	192.0
38.20	754.0
39.10	4446.0
40.10	421.0
41.00	234.0
44.10	975.0
44.80	161.0
47.90	238.0
49.20	360.0
50.10	16568.0
51.10	65360.0
52.10	3336.0
56.10	1821.0
57.00	4074.0
61.00	982.0
62.10	976.0
63.10	3712.0
64.00	552.0
65.10	919.0
66.20	237.0
69.00	75600.0
70.00	476.0
73.30	623.0
74.10	8020.0
75.00	12813.0
76.10	4055.0
77.10	77160.0
78.00	5016.0
79.10	4961.0
80.00	5030.0
81.00	5978.0
82.10	1097.0
83.00	1693.0
84.00	304.0
85.00	765.0
86.00	1195.0
87.00	912.0
87.90	264.0
88.40	159.0
89.00	179.0
91.10	1261.0
92.00	2344.0
93.00	9650.0
94.10	352.0
96.10	165.0
97.20	421.0
98.00	8037.0
99.00	5254.0
99.90	469.0
101.10	3282.0
103.00	546.0
104.00	2399.0
105.00	2036.0
106.10	1098.0
107.00	25608.0
108.00	4410.0
110.00	33896.0
111.10	6983.0
112.10	604.0
113.20	157.0
114.10	154.0
114.60	286.0
116.00	1535.0
117.00	24280.0
118.10	1653.0
120.00	353.0
122.00	1104.0
123.00	2455.0
124.00	1068.0
125.00	1101.0
127.00	69376.0
128.10	5978.0
129.00	31704.0
130.00	3024.0
131.10	230.0
132.10	309.0
132.90	308.0
134.00	682.0
135.00	3269.0
135.90	950.0
137.00	1054.0
138.10	196.0
140.10	473.0
141.00	3649.0
142.10	1328.0
143.00	810.0
144.00	217.0
146.00	740.0
147.00	2451.0
148.00	6093.0
149.00	1081.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

149.80	289.0
150.90	551.0
151.80	240.0
152.10	296.0
153.00	1797.0
154.00	1415.0
155.10	3024.0
156.00	3510.0
156.90	466.0
158.10	910.0
158.90	917.0
160.00	1720.0
161.10	2853.0
162.10	613.0
162.80	206.0
163.10	235.0
163.80	161.0
165.00	2563.0
166.00	1345.0
167.00	7863.0
168.10	3736.0
169.00	704.0
169.90	408.0
170.90	564.0
172.00	725.0
173.00	1228.0
174.00	1558.0
175.00	3203.0
176.00	885.0
177.00	1446.0
178.10	563.0
179.00	8303.0
180.00	5089.0
181.00	1961.0
181.90	430.0
184.00	1151.0
185.00	3565.0
186.00	23024.0
187.00	6051.0
188.00	1118.0
189.00	1415.0
190.10	468.0
191.00	1072.0
192.00	1894.0
193.10	3035.0
194.00	708.0
196.00	2779.0
198.00	145920.0
198.90	10486.0
200.00	608.0
201.40	560.0
201.80	358.0
202.00	349.0
203.10	1338.0
204.00	6184.0
205.00	11802.0
206.10	47224.0
207.00	6982.0
208.00	1834.0
209.00	452.0
210.00	513.0
211.10	1593.0
212.00	406.0
213.00	184.0
215.00	729.0
216.10	1561.0
217.00	13385.0
218.00	1668.0
218.70	171.0
220.00	375.0
221.00	6469.0
222.00	1819.0
223.00	2624.0
224.00	25328.0
225.00	7500.0
227.00	14483.0
228.00	2621.0
229.00	3292.0
230.00	320.0
231.00	889.0
232.10	308.0
233.00	392.0
234.00	709.0
235.00	1283.0
236.00	532.0
237.00	803.0
238.10	244.0
239.00	316.0
239.90	178.0
241.10	654.0
241.90	1122.0
243.00	1576.0
244.00	17568.0
245.00	2187.0
246.00	4120.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

247.00	1376.0
248.80	529.0
249.90	216.0
252.00	195.0
252.80	331.0
255.00	103944.0
256.00	14821.0
257.00	1659.0
258.00	4992.0
259.00	629.0
261.10	220.0
261.80	254.0
264.10	261.0
264.90	2649.0
265.80	267.0
266.10	231.0
270.90	190.0
271.30	224.0
272.00	511.0
273.00	3148.0
274.00	7233.0
275.00	42704.0
276.00	6147.0
277.00	4275.0
278.20	486.0
283.10	300.0
283.90	357.0
285.00	793.0
285.80	240.0
289.00	263.0
290.00	278.0
290.90	164.0
292.10	223.0
292.90	600.0
293.80	326.0
294.80	261.0
296.00	13887.0
297.00	2243.0
300.90	183.0
303.00	972.0
304.10	459.0
305.90	161.0
308.00	190.0
314.10	369.0
314.90	1780.0
316.00	957.0
320.90	485.0
322.00	297.0
323.00	3740.0
324.10	578.0
325.90	314.0
327.00	854.0
327.90	392.0
329.00	174.0
332.00	675.0
332.90	204.0
334.00	2002.0
335.00	1234.0
338.80	269.0
341.00	668.0
345.90	692.0
346.90	162.0
352.00	829.0
353.00	704.0
354.00	966.0
355.10	244.0
358.80	224.0
365.00	6397.0
365.90	689.0
371.10	422.0
372.00	1434.0
372.90	217.0
374.10	197.0
383.00	539.0
383.80	226.0
389.90	578.0
402.00	514.0
403.00	826.0
403.90	182.0
420.90	1020.0
421.90	285.0
423.00	4848.0
423.90	705.0
435.80	171.0
438.50	218.0
441.00	15519.0
442.00	92936.0
443.00	19440.0
444.00	1709.0
445.10	155.0
479.20	186.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

m/z	Abundance
36.60	168.0
37.10	166.0
38.10	566.0
39.10	3668.0
40.10	419.0
44.10	1103.0
49.10	365.0
50.10	14166.0
51.10	56712.0
52.10	3175.0
55.00	442.0
56.10	1686.0
57.10	2972.0
58.10	379.0
60.20	158.0
61.00	908.0
62.10	635.0
63.10	2009.0
64.10	539.0
65.00	980.0
68.10	1340.0
69.00	60112.0
70.10	605.0
73.00	467.0
74.10	6323.0
75.10	11492.0
76.10	3132.0
77.10	63576.0
78.10	3538.0
79.10	4656.0
80.00	3746.0
81.00	5916.0
82.00	1195.0
83.00	1083.0
83.90	362.0
85.20	454.0
86.00	1569.0
87.00	632.0
88.00	245.0
91.00	1031.0
92.00	1189.0
93.00	10658.0
94.00	542.0
96.00	249.0
98.00	5837.0
99.10	4607.0
100.10	300.0
101.00	3113.0
102.30	189.0
103.00	668.0
104.00	2187.0
105.00	2625.0
107.00	23952.0
108.10	2806.0
109.10	1052.0
110.00	36432.0
111.00	6326.0
112.00	491.0
113.10	458.0
115.00	186.0
116.10	2160.0
117.00	17680.0
118.10	1440.0
119.90	540.0
122.00	2787.0
123.00	2517.0
123.90	1257.0
124.90	1126.0
127.00	66432.0
128.00	4663.0
129.00	30256.0
130.00	2006.0
131.00	339.0
132.10	349.0
132.60	292.0
132.80	305.0
134.00	981.0
135.00	2753.0
136.00	979.0
137.10	1060.0
137.90	288.0
138.90	380.0
140.00	223.0
141.10	4435.0
142.10	1282.0
143.00	868.0
144.00	383.0
146.10	937.0
147.00	2039.0
148.00	4940.0
149.10	1034.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

150.90	484.0
151.80	200.0
152.90	1098.0
153.90	1356.0
155.00	2399.0
156.00	3036.0
157.20	470.0
158.00	721.0
159.00	1050.0
160.00	1526.0
161.00	1727.0
162.00	671.0
163.60	282.0
165.00	1262.0
166.00	1720.0
167.00	5577.0
168.00	3751.0
169.00	765.0
169.90	279.0
171.10	232.0
171.90	1229.0
173.00	1022.0
174.00	1393.0
175.10	3742.0
175.90	951.0
177.00	1156.0
178.00	363.0
179.00	5504.0
180.00	4814.0
181.10	2178.0
182.10	437.0
184.10	745.0
185.10	2467.0
186.00	25288.0
187.00	7593.0
188.10	244.0
189.00	2058.0
189.90	349.0
191.00	772.0
192.10	2149.0
193.10	2542.0
194.10	791.0
196.00	3465.0
198.00	143424.0
199.00	9751.0
200.00	1256.0
200.80	353.0
201.20	447.0
201.40	402.0
202.00	193.0
202.90	917.0
204.00	7205.0
205.10	11854.0
206.00	43656.0
207.00	6430.0
208.00	2401.0
208.90	287.0
209.90	437.0
211.10	1991.0
212.00	482.0
213.00	396.0
214.90	716.0
216.00	1792.0
217.00	14067.0
217.90	2069.0
219.90	330.0
221.00	6020.0
222.00	1352.0
223.00	3770.0
224.00	27944.0
225.00	6551.0
226.10	603.0
227.00	15780.0
228.00	1783.0
229.00	2471.0
229.90	234.0
231.00	979.0
233.10	615.0
233.90	624.0
235.00	1097.0
235.90	355.0
237.00	1404.0
239.10	653.0
240.00	437.0
240.90	725.0
241.90	1139.0
243.00	1979.0
244.00	19456.0
245.00	2050.0
246.00	3921.0
246.90	989.0
249.00	780.0
250.80	219.0
251.90	416.0
253.10	591.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

253.40	535.0
255.00	104496.0
256.00	16952.0
256.70	681.0
258.00	5463.0
259.00	858.0
260.00	171.0
261.70	180.0
263.90	193.0
265.00	2686.0
265.90	758.0
266.90	173.0
269.60	232.0
271.00	286.0
272.10	204.0
273.00	3298.0
274.00	7262.0
275.00	41520.0
276.00	4686.0
277.00	4864.0
277.90	874.0
283.00	559.0
284.00	559.0
285.00	1292.0
289.20	188.0
289.90	234.0
293.00	344.0
294.10	347.0
294.60	190.0
296.00	15206.0
297.00	1925.0
297.80	271.0
300.90	171.0
301.90	192.0
302.90	1531.0
304.00	667.0
308.00	225.0
310.10	156.0
314.00	679.0
315.00	987.0
316.00	1229.0
321.10	445.0
322.00	907.0
323.00	4945.0
324.10	727.0
326.00	317.0
326.90	942.0
328.00	165.0
328.80	154.0
332.20	336.0
333.10	561.0
334.00	3158.0
335.00	927.0
338.90	218.0
341.00	534.0
342.00	535.0
346.00	726.0
352.00	1712.0
353.00	1001.0
354.10	1523.0
360.10	213.0
365.00	4993.0
366.00	982.0
370.80	181.0
372.00	2638.0
372.90	342.0
374.20	265.0
382.90	510.0
383.50	267.0
389.90	226.0
392.00	195.0
401.20	174.0
402.00	859.0
403.00	742.0
404.00	340.0
421.00	957.0
422.00	653.0
423.00	6081.0
424.00	1188.0
424.70	217.0
436.70	219.0
437.20	223.0
439.10	172.0
441.00	16584.0
442.00	110744.0
443.00	21040.0
444.00	1652.0
508.10	248.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02

Scan 2480 (17.858 min): BG036304.D
DFTPP02

m/z	Abundance
40.00	163.0
44.00	1281.0
59.20	198.0
73.00	155.0
206.90	299.0
210.80	181.0
218.40	158.0
311.80	221.0
380.50	256.0
461.90	214.0
518.90	233.0

Instrument :
BNA_G
ClientSampleId :
DFTPP02