

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG081719\  
 Data File : BG042304.D  
 Acq On : 18 Aug 2019 2:22  
 Operator : HP/JU  
 Sample : K4238-06  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 TP-18

Quant Time: Aug 19 07:57:34 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 19 06:42:36 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.01  | 152  | 61830    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.83 | 136  | 231026   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.66 | 164  | 167191   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.42 | 188  | 409536   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.69 | 240  | 466335   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 24.94 | 264  | 526054   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.56  | 112  | 400301   | 125.96 | ng    | 0.00     |
| 7) Phenol-d6                | 7.17  | 99   | 483396   | 112.83 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.17  | 82   | 353592   | 105.33 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.16 | 330  | 368807   | 194.20 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.28 | 172  | 1098091  | 115.84 | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.04 | 244  | 2135723  | 104.80 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG081719\  
Data File : BG042304.D  
Acq On : 18 Aug 2019 2:22  
Operator : HP/JU  
Sample : K4238-06  
Misc :  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
TP-18

Quant Time: Aug 19 07:57:34 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG072719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Aug 19 06:42:36 2019  
Response via : Initial Calibration

