


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG081719\  
 Data File : BG042261.D  
 Acq On : 16 Aug 2019 16:01  
 Operator : HP/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG072719.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Aug 19 06:42:36 2019



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2507

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 20.7      | 9675    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.5       | 50      | PASS             |
| 69          | 198          | 0.00         | 100          | 22.1      | 10310   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 36.2      | 16913   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 46720   | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 3240    | PASS             |
| 275         | 198          | 10           | 60           | 29.6      | 13838   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 1603    | PASS             |
| 441         | 198          | 0.01         | 100          | 13.7      | 6411    | PASS             |
| 442         | 198          | 50           | 100          | 85.5      | 39933   | PASS             |
| 443         | 442          | 15           | 24           | 19.9      | 7931    | PASS             |