


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG081719\
 Data File : BG042280.D
 Acq On : 17 Aug 2019 7:46
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG072719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 19 06:42:36 2019



AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2507

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.5	13270	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	23.3	14407	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	36.5	22573	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	61765	PASS
199	198	5	9	6.9	4282	PASS
275	198	10	60	28.5	17627	PASS
365	198	1	100	3.3	2066	PASS
441	198	0.01	100	12.0	7402	PASS
442	198	50	100	78.0	48154	PASS
443	442	15	24	18.5	8921	PASS