

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081815\
 Data File : BG018463.D
 Acq On : 18 Aug 2015 14:16
 Operator : UM/MA
 Sample : MDL-07-S-ML
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MDL-07-S-ML

Manual Integrations
 APPROVED

MMDadoda
 8/19/2015 2:13:04 PM

Quant Time: Aug 19 01:51:29 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 19 00:53:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	100292	20.00	ng/ul	0.00
18) Naphthalene-d8	10.82	136	483023	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.64	164	313052	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.38	188	794341	20.00	ng/ul	0.00
78) Chrysene-d12	21.64	240	809454	20.00	ng/ul	0.00
86) Perylene-d12	24.84	264	799740	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.45	96	15292	6.74	ng/uL	0.00
5) Phenol-d5	7.17	99	323834	35.58	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.34	67	208761	37.01	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	244453	35.06	ng/ul	0.00
13) 4-Methylphenol-d8	8.71	113	277083	36.23	ng/ul	0.00
19) Nitrobenzene-d5	9.17	128	136211	36.18	ng/ul	0.00
22) 2-Nitrophenol-d4	9.90	143	138980	36.24	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.43	165	260541	32.01	ng/ul	0.00
29) 4-Chloroaniline-d4	10.95	131	375233	40.80	ng/ul	0.00
44) Dimethylphthalate-d6	14.04	166	921605	34.99	ng/ul	0.00
47) Acenaphthylene-d8	14.34	160	1124384	33.90	ng/ul	0.00
52) 4-Nitrophenol-d4	14.83	143	185479	39.15	ng/ul	0.00
58) Fluorene-d10	15.63	176	813547	35.46	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.75	200	136673	34.98	ng/ul	0.00
71) Anthracene-d10	17.48	188	1325335	34.89	ng/ul	0.00
79) Pyrene-d10	19.76	212	1449384	34.51	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.63	264	1508269	35.52	ng/ul	0.00

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.48	88	1678	0.69 ng/uL#	83
4) Benzaldehyde	7.15	77	18998	3.56 ng/ul	96
6) Phenol	7.20	94	28172	3.03 ng/ul	93
8) Bis(2-Chloroethyl)ether	7.43	93	23203	3.25 ng/ul#	89
10) 2-Chlorophenol	7.58	128	19844	2.79 ng/ul	91
11) 2-Methylphenol	8.45	108	20455	2.84 ng/ul	83
12) 2,2'-oxybis(1-Chloropropan	8.55	45	30242	2.64 ng/ul#	90
14) Acetophenone	8.83	105	37965	3.44 ng/ul#	93
15) N-Nitroso-di-n-propylamine	8.81	70	21060	3.32 ng/ul	90
16) 4-Methylphenol	8.78	108	23066	2.94 ng/ul	80
17) Hexachloroethane	9.11	117	8102	2.65 ng/ul#	87
20) Nitrobenzene	9.21	77	28161	2.96 ng/ul#	96
21) Isophorone	9.73	82	58396	3.20 ng/ul#	97
23) 2-Nitrophenol	9.93	139	12256	2.88 ng/ul	93
24) 2,4-Dimethylphenol	9.99	107	25673	2.69 ng/ul	93
25) Bis(2-Chloroethoxy)methane	10.22	93	33831	2.97 ng/ul	98
27) 2,4-Dichlorophenol	10.47	162	20456	2.68 ng/ul#	87
28) Naphthalene	10.87	128	76293	2.99 ng/ul	97
30) 4-Chloroaniline	10.97	127	33755	3.61 ng/ul	100
31) Hexachlorobutadiene	11.16	225	13074	2.72 ng/ul	95
32) Caprolactam	11.75	113	9750m	3.17 ng/ul	
33) 4-Chloro-3-methylphenol	12.10	107	24565	2.78 ng/ul#	83
34) 2-Methylnaphthalene	12.47	142	54626	2.95 ng/ul	97

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081815\
 Data File : BG018463.D
 Acq On : 18 Aug 2015 14:16
 Operator : UM/MA
 Sample : MDL-07-S-ML
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 MDL-07-S-ML


Manual Integrations
 APPROVED

MMDadoda
 8/19/2015 2:13:04 PM

Quant Time: Aug 19 01:51:29 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 19 00:53:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1,2,4,5-Tetrachlorobenzene	12.84	216	29852	2.97	ng/ul#	88
38) Hexachlorocyclopentadiene	12.83	237	14489	2.42	ng/ul	88
39) 2,4,6-Trichlorophenol	13.08	196	16941	2.48	ng/ul#	82
40) 2,4,5-Trichlorophenol	13.15	196	18673	2.54	ng/ul	88
41) 1,1'-Biphenyl	13.48	154	76353	2.92	ng/ul#	94
42) 2-Chloronaphthalene	13.52	162	58590	2.89	ng/ul	96
43) 2-Nitroaniline	13.71	65	18514	2.83	ng/ul#	90
45) Dimethylphthalate	14.09	163	79407	3.06	ng/ul#	97
46) 2,6-Dinitrotoluene	14.21	165	14765	2.85	ng/ul	95
48) Acenaphthylene	14.36	152	101892	3.06	ng/ul	95
49) 3-Nitroaniline	14.54	138	18206	3.45	ng/ul	97
50) Acenaphthene	14.71	153	68501	2.94	ng/ul	97
51) 2,4-Dinitrophenol	14.75	184	4649	1.84	ng/ul#	68
53) 4-Nitrophenol	14.84	109	12459	3.02	ng/ul	87
54) Dibenzofuran	15.04	168	92501	3.03	ng/ul	92
55) 2,4-Dinitrotoluene	14.99	165	22048	2.99	ng/ul#	89
56) 2,3,4,6-Tetrachlorophenol	15.26	232	17501	2.73	ng/ul#	95
57) Diethylphthalate	15.45	149	88618	3.21	ng/ul	99
59) Fluorene	15.69	166	80007	3.05	ng/ul	94
60) 4-Chlorophenyl-phenylether	15.67	204	38669	3.11	ng/ul	93
61) 4-Nitroaniline	15.69	138	19250	3.30	ng/ul	96
64) 4,6-Dinitro-2-methylphenol	15.76	198	12958	3.11	ng/ul#	98
65) N-Nitrosodiphenylamine	15.89	169	70716	2.96	ng/ul	87
66) 4-Bromophenyl-phenylether	16.57	248	22712	2.64	ng/ul	92
67) Hexachlorobenzene	16.69	284	26933	2.96	ng/ul#	89
68) Atrazine	16.83	200	28113	3.14	ng/ul#	94
69) Pentachlorophenol	17.03	266	11968	1.97	ng/ul	93
70) Phenanthrene	17.43	178	138400	3.21	ng/ul	96
72) Anthracene	17.52	178	143597	3.27	ng/ul	97
73) 1,2,3,4-Tetrachlorobenzene	13.44	216	27852	2.53	ng/uL	97
74) Pentachlorobenzene	14.96	250	29167	2.78	ng/uL	96
75) Carbazole	17.78	167	133486	3.47	ng/ul	96
76) Di-n-butylphthalate	18.34	149	169804	3.38	ng/ul	98
77) Fluoranthene	19.43	202	168992	3.67	ng/ul	97
80) Pyrene	19.79	202	173196	3.16	ng/ul	97
81) Butylbenzylphthalate	20.68	149	68679	2.97	ng/ul	95
82) 3,3'-Dichlorobenzidine	21.54	252	56119	3.34	ng/ul#	92
83) Benzo(a)anthracene	21.62	228	156345	3.12	ng/ul	97
84) Bis(2-ethylhexyl)phthalate	21.54	149	102449	3.13	ng/ul	99
85) Chrysene	21.69	228	143714	3.06	ng/ul	99
87) Di-n-octyl phthalate	22.74	149	173431	3.36	ng/ul	100
88) Benzo(b)fluoranthene	23.82	252	156177	3.11	ng/ul	96
89) Benzo(k)fluoranthene	23.88	252	145333	3.00	ng/ul	97
91) Benzo(a)pyrene	24.69	252	146376	3.06	ng/ul#	96
92) Indeno(1,2,3-cd)pyrene	28.48	276	161779	2.93	ng/ul	99
93) Dibenzo(a,h)anthracene	28.54	278	133426	2.91	ng/ul	98
94) Benzo(g,h,i)perylene	29.60	276	134934	2.91	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 0.69 ng/uL

RT: 3.48 min Scan# 109

Delta R.T. -0.01 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	88	Resp:	1678
Ion Ratio	Lower	Upper	
88	100		
43	50.6	33.0	49.4#
58	108.2	73.4	110.0

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#4

Benzaldehyde

Concen: 3.56 ng/uL




RT: 7.15 min Scan# 734


Delta R.T. 0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	77	Resp:	18998
Ion Ratio	Lower	Upper	
77	100		
105	101.0	78.2	117.2
106	90.8	76.8	115.2





#6

Phenol

Concen: 3.03 ng/ul

RT: 7.20 min Scan# 742

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	94	Resp:	28172
Ion Ratio	Lower	Upper	
94	100		
65	28.6	23.0	34.6
66	31.2	31.0	46.4

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#8

Bis(2-Chloroethyl)ether

Concen: 3.25 ng/ul




RT: 7.43 min Scan# 782


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	93	Resp:	23203
Ion Ratio	Lower	Upper	
93	100		
63	70.7	61.4	92.2
95	44.3	25.9	38.9#









#10
2-Chlorophenol
Concen: 2.79 ng/ul
RT: 7.58 min Scan# 807
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	19844		
64	43.7	40.9	61.3	
130	34.6	24.5	36.7	




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#11
2-Methylphenol
Concen: 2.84 ng/ul
RT: 8.45 min Scan# 956
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	20455		
107	106.2	72.1	108.1	









#12
2,2'-oxybis(1-Chloropropane)
Concen: 2.64 ng/ul
RT: 8.55 min Scan# 972
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp Lower	Upper
45	100		
77	16.2	11.3	16.9
79	16.5	7.8	11.8#




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#14
Acetophenone
Concen: 3.44 ng/ul
RT: 8.83 min Scan# 1020
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp Lower	Upper
105	100		
77	76.5	61.9	92.9
51	20.0	25.5	38.3#









#15
N-Nitroso-di-n-propylamine
Concen: 3.32 ng/ul
RT: 8.81 min Scan# 1017
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
70	100		
42	53.5	50.7	76.1
101	8.9	7.4	11.2
130	17.2	15.8	23.6




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#16
4-Methylphenol
Concen: 2.94 ng/ul
RT: 8.78 min Scan# 1011
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
108	100		
107	96.3	94.6	142.0





#17

Hexachloroethane

Concen: 2.65 ng/ul

RT: 9.11 min Scan# 1067

Delta R.T. 0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G


ClientSampleId :

MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	8102		
201	110.8	73.6	110.4#	
199	62.9	48.6	72.8	

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM






Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 2.96 ng/ul


RT: 9.21 min Scan# 1085

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	28161		
123	40.8	32.8	49.2	
65	10.1	12.2	18.2#	






Abundance

Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BGC





#21

Isophorone

Concen: 3.20 ng/ul

RT: 9.73 min Scan# 1174

Delta R.T. -0.00 min

Lab File: BG018463.D


Acq: 18 Aug 2015 14:16

Instrument :

BNA_G

ClientSampleId :




MDL-07-S-ML



Tgt Ion	Ratio	Resp	Lower	Upper
82	100	58396		
95	8.0	5.2	7.8#	
138	16.2	13.8	20.8	

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#23

2-Nitrophenol

Concen: 2.88 ng/ul




RT: 9.93 min Scan# 1207


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
139	100	12256		
65	61.7	45.6	68.4	
109	26.8	25.4	38.2	





#24

2,4-Dimethylphenol

Concen: 2.69 ng/ul

RT: 9.99 min Scan# 1217

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	107	Resp:	25673
Ion Ratio	Lower	Upper	
107	100		
121	49.8	41.3	61.9
122	81.3	71.9	107.9

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#25

Bis(2-Chloroethoxy)methane

Concen: 2.97 ng/ul




RT: 10.22 min Scan# 1257


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	93	Resp:	33831
Ion Ratio	Lower	Upper	
93	100		
95	30.8	24.1	36.1
123	13.1	8.9	13.3









#27
 2,4-Dichlorophenol
 Concen: 2.68 ng/ul
 RT: 10.47 min Scan# 1299
 Delta R.T. 0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Instrument :
 BNA_G
 ClientSampleId :
 MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	20456		
164	79.0	51.7	77.5#	
98	35.8	30.4	45.6	




Manual Integrations
 APPROVED


MMDadoda
 8/19/2015 2:13:04 PM



#28
 Naphthalene
 Concen: 2.99 ng/ul
 RT: 10.87 min Scan# 1367
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	76293		
129	12.1	9.6	14.4	
127	15.1	10.5	15.7	









#30
 4-Chloroaniline
 Concen: 3.61 ng/ul
 RT: 10.97 min Scan# 1385
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Instrument :
 BNA_G
 ClientSampleId :
 MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
127	100		
129	33.6	26.7	40.1




Manual Integrations
 APPROVED


MMDadoda
 8/19/2015 2:13:04 PM



#31
 Hexachlorobutadiene
 Concen: 2.72 ng/ul
 RT: 11.16 min Scan# 1417
 Delta R.T. 0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.2	45.8	68.8
227	63.9	54.0	81.0





#32

Caprolactam

Concen: 3.17 ng/ul m

RT: 11.75 min Scan# 1517

Delta R.T. 0.03 min

Lab File: BG018463.D


Acq: 18 Aug 2015 14:16

Instrument :

BNA_G

ClientSampleId :




MDL-07-S-ML



Tgt Ion:	113	Resp:	9750
Ion	Ratio	Lower	Upper
113	100		
55	153.8	124.2	186.4
56	115.1	98.2	147.4

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#33

4-Chloro-3-methylphenol

Concen: 2.78 ng/ul




RT: 12.10 min Scan# 1576


Delta R.T. 0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	107	Resp:	24565
Ion	Ratio	Lower	Upper
107	100		
144	24.5	21.4	32.2
142	63.4	66.0	99.0#





#34

2-Methylnaphthalene

Concen: 2.95 ng/ul

RT: 12.47 min Scan# 1640

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G

ClientSampleId :

MDL-07-S-ML

Tgt Ion:142 Resp: 54626


Ion Ratio Lower Upper

142 100

141 94.5 78.0 117.0




Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E



#37

1,2,4,5-Tetrachlorobenzene

Concen: 2.97 ng/ul

RT: 12.84 min Scan# 1702

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:216 Resp: 29852


Ion Ratio Lower Upper

216 100

214 92.4 62.5 93.7

179 22.4 18.2 27.2

108 13.8 14.6 22.0#






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#38
Hexachlorocyclopentadiene
Concen: 2.42 ng/ul
RT: 12.83 min Scan# 1700
Delta R.T. 0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
237	100		
235	66.5	45.6	68.4
272	13.4	9.0	13.4




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#39
2,4,6-Trichlorophenol
Concen: 2.48 ng/ul
RT: 13.08 min Scan# 1743
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
196	100		
198	77.3	78.5	117.7#
200	26.4	25.1	37.7









#40
 2,4,5-Trichlorophenol
 Concen: 2.54 ng/ul
 RT: 13.15 min Scan# 1755
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Instrument :
 BNA_G
 ClientSampleId :
 MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	18673		
198	108.1	74.2	111.4	
200	31.7	25.2	37.8	




Manual Integrations
 APPROVED


MMDadoda
 8/19/2015 2:13:04 PM



#41
 1,1'-Biphenyl
 Concen: 2.92 ng/ul
 RT: 13.48 min Scan# 1811
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	76353		
153	47.1	35.2	52.8	
76	16.8	10.6	16.0	









#42
 2-Chloronaphthalene
 Concen: 2.89 ng/ul
 RT: 13.52 min Scan# 1818
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Instrument :
 BNA_G
 ClientSampleId :
 MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
162	100		
164	34.1	24.8	37.2
127	36.5	30.3	45.5




Manual Integrations
 APPROVED


MMDadoda
 8/19/2015 2:13:04 PM



#43
 2-Nitroaniline
 Concen: 2.83 ng/ul
 RT: 13.71 min Scan# 1851
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
65	100		
92	78.2	51.1	76.7#
138	97.3	75.1	112.7









#45
Dimethylphthalate
Concen: 3.06 ng/ul
RT: 14.09 min Scan# 1915
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.6	2.9	4.3#
164	10.8	7.7	11.5




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#46
2,6-Dinitrotoluene
Concen: 2.85 ng/ul
RT: 14.21 min Scan# 1935
Delta R.T. -0.01 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
165	100		
89	67.3	50.2	75.2
121	26.1	19.4	29.2









#48
Acenaphthylene
Concen: 3.06 ng/ul
RT: 14.36 min Scan# 1962
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
152	100		
151	23.0	15.4	23.2
153	13.0	10.8	16.2




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#49
3-Nitroaniline
Concen: 3.45 ng/ul
RT: 14.54 min Scan# 1991
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
138	100		
108	13.8	9.4	14.0
92	131.6	102.8	154.2





#50

Acenaphthene

Concen: 2.94 ng/ul

RT: 14.71 min Scan# 2020

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	153	Resp:	68501
Ion Ratio	Lower	Upper	
153	100		
152	43.7	38.2	57.4
154	87.3	68.6	103.0

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#51

2,4-Dinitrophenol

Concen: 1.84 ng/ul




RT: 14.75 min Scan# 2027


Delta R.T. 0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	184	Resp:	4649
Ion Ratio	Lower	Upper	
184	100		
63	46.0	55.3	82.9#
154	42.5	58.2	87.2#









#53
4-Nitrophenol
Concen: 3.02 ng/ul
RT: 14.84 min Scan# 2043
Delta R.T. -0.01 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
109	100	12459		
139	147.5	111.3	166.9	
65	153.7	105.3	157.9	




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#54
Dibenzofuran
Concen: 3.03 ng/ul
RT: 15.04 min Scan# 2077
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	92501		
139	32.1	29.7	44.5	





#55

2,4-Dinitrotoluene

Concen: 2.99 ng/ul

RT: 14.99 min Scan# 2069

Delta R.T. -0.00 min

Lab File: BG018463.D


Acq: 18 Aug 2015 14:16

Instrument :

BNA_G

ClientSampleId :

MDL-07-S-ML



Tgt Ion:165 Resp: 22048

Ion Ratio Lower Upper




165 100

63 38.6 36.7 55.1

182 4.9 2.2 3.2#

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#56

2,3,4,6-Tetrachlorophenol

Concen: 2.73 ng/ul

RT: 15.26 min Scan# 2115

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:232 Resp: 17501




Ion Ratio Lower Upper


232 100

131 44.1 37.4 56.0

130 2.7 1.4 2.0#

166 36.4 25.8 38.8





#57

Diethylphthalate

Concen: 3.21 ng/ul

RT: 15.45 min Scan# 2147

Delta R.T. -0.01 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	149	Resp:	88618
Ion Ratio	Lower	Upper	
149	100		
177	22.4	18.2	27.4
150	11.9	9.7	14.5

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#59

Fluorene

Concen: 3.05 ng/ul




RT: 15.69 min Scan# 2187


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	166	Resp:	80007
Ion Ratio	Lower	Upper	
166	100		
165	109.7	82.6	124.0
167	13.1	11.4	17.0









#60
 4-Chlorophenyl-phenylether
 Concen: 3.11 ng/ul
 RT: 15.67 min Scan# 2185
 Delta R.T. -0.00 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Instrument :
 BNA_G
 ClientSampleId :
 MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
204	100	38669		
206	37.1	27.1	40.7	
141	52.8	47.1	70.7	




Manual Integrations
 APPROVED


MMDadoda
 8/19/2015 2:13:04 PM



#61
 4-Nitroaniline
 Concen: 3.30 ng/ul
 RT: 15.69 min Scan# 2188
 Delta R.T. -0.01 min
 Lab File: BG018463.D
 Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	19250		
92	53.5	44.0	66.0	
108	76.6	64.6	97.0	





#64

4,6-Dinitro-2-methylphenol

Concen: 3.11 ng/ul

RT: 15.76 min Scan# 2199

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G


ClientSampleId :

MDL-07-S-ML




Tgt Ion:	198	Resp:	12958
Ion Ratio	Lower	Upper	
198	100		
182	7.8	3.4	5.0#
77	27.9	22.4	33.6

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 182.00 (181.70 to 182.70): E
Ion 77.00 (76.70 to 77.70): BGC



#65

N-Nitrosodiphenylamine

Concen: 2.96 ng/ul


RT: 15.89 min Scan# 2221

Delta R.T. -0.00 min



Lab File: BG018463.D


Acq: 18 Aug 2015 14:16

Tgt Ion:	169	Resp:	70716
Ion Ratio	Lower	Upper	
169	100		
168	62.3	60.2	90.2
167	35.4	33.4	50.2



Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E









#66
4-Bromophenyl-phenylether
Concen: 2.64 ng/ul
RT: 16.57 min Scan# 2338
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
248	100	22712		
250	104.4	76.4	114.6	
141	67.6	58.6	87.8	




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#67
Hexachlorobenzene
Concen: 2.96 ng/ul
RT: 16.69 min Scan# 2358
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	26933		
142	25.5	28.6	42.8#	
249	30.2	26.2	39.4	









#68
Atrazine
Concen: 3.14 ng/ul
RT: 16.83 min Scan# 2382
Delta R.T. -0.01 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
200	100	28113		
173	32.6	19.0	28.4#	
215	46.8	37.4	56.2	




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#69
Pentachlorophenol
Concen: 1.97 ng/ul
RT: 17.03 min Scan# 2416
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	11968		
264	75.4	55.4	83.0	
268	62.2	53.6	80.4	





#70

Phenanthrene

Concen: 3.21 ng/ul

RT: 17.43 min Scan# 2483

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	178	Resp:	138400
Ion Ratio	Lower	Upper	
178	100		
179	14.5	13.7	20.5
176	19.8	16.6	24.8

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#72

Anthracene

Concen: 3.27 ng/ul




RT: 17.52 min Scan# 2499


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	178	Resp:	143597
Ion Ratio	Lower	Upper	
178	100		
179	15.5	13.0	19.4
176	18.0	16.0	24.0





#73

1,2,3,4-Tetrachlorobenzene

Concen: 2.53 ng/uL

RT: 13.44 min Scan# 1805

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	216	Resp:	27852
Ion Ratio	Lower	Upper	
216	100		
214	84.0	64.0	96.0
218	49.4	38.7	58.1

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#74

Pentachlorobenzene

Concen: 2.78 ng/uL




RT: 14.96 min Scan# 2063


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	250	Resp:	29167
Ion Ratio	Lower	Upper	
250	100		
248	58.0	45.8	68.8
252	57.1	50.4	75.6





#75

Carbazole

Concen: 3.47 ng/ul

RT: 17.78 min Scan# 2544

Delta R.T. 0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	167	Resp:	133486
Ion Ratio		Lower	Upper
167	100		
166	20.9	18.9	28.3
139	12.6	10.1	15.1

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#76

Di-n-butylphthalate

Concen: 3.38 ng/ul




RT: 18.34 min Scan# 2639


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	149	Resp:	169804
Ion Ratio		Lower	Upper
149	100		
150	8.7	7.5	11.3
104	5.2	4.3	6.5





#77

Fluoranthene

Concen: 3.67 ng/ul

RT: 19.43 min Scan# 2824

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G


ClientSampleId :

MDL-07-S-ML




Tgt	Ion	Ratio	Resp	Lower	Upper
202	100		168992		
101	10.1	9.7	14.5		
100	10.0	7.8	11.8		

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



Abundance Ion 202.00 (201.70 to 202.70): E



#80

Pyrene

Concen: 3.16 ng/ul


RT: 19.79 min Scan# 2886

Delta R.T. -0.00 min



Lab File: BG018463.D


Acq: 18 Aug 2015 14:16

Tgt	Ion	Ratio	Resp	Lower	Upper
202	100		173196		
101	14.5	11.0	16.4		
100	10.8	9.8	14.6		



Abundance Ion 202.00 (201.70 to 202.70): E









#81
Butylbenzylphthalate
Concen: 2.97 ng/ul
RT: 20.68 min Scan# 3037
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	78.3	57.9	86.9
206	22.5	17.9	26.9




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#82
3,3'-Dichlorobenzidine
Concen: 3.34 ng/ul
RT: 21.54 min Scan# 3183
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
254	69.7	50.9	76.3
126	7.9	9.6	14.4#





#83

Benzo(a)anthracene

Concen: 3.12 ng/ul

RT: 21.62 min Scan# 3197

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G





ClientSampleId :

MDL-07-S-ML

Tgt Ion:	228	Resp:	156345
Ion Ratio	Lower	Upper	
228	100		
229	19.5	15.5	23.3
226	24.8	22.1	33.1

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#84

Bis(2-ethylhexyl)phthalate

Concen: 3.13 ng/ul




RT: 21.54 min Scan# 3183


Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Tgt Ion:	149	Resp:	102449
Ion Ratio	Lower	Upper	
149	100		
167	27.3	22.2	33.2
279	4.1	2.8	4.2









#85
Chrysene
Concen: 3.06 ng/ul
RT: 21.69 min Scan# 3208
Delta R.T. -0.01 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
228	100		
226	33.9	26.5	39.7
229	20.5	16.9	25.3




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#87
Di-n-octyl phthalate
Concen: 3.36 ng/ul
RT: 22.74 min Scan# 3388
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion:149 Resp: 173431









#88
Benzo(b)fluoranthene
Concen: 3.11 ng/ul
RT: 23.82 min Scan# 3572
Delta R.T. -0.00 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	156177		
253	24.7	18.1	27.1	
125	11.5	8.3	12.5	




Manual Integrations
APPROVED


MMDadoda
8/19/2015 2:13:04 PM



#89
Benzo(k)fluoranthene
Concen: 3.00 ng/ul
RT: 23.88 min Scan# 3582
Delta R.T. -0.01 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	145333		
253	20.8	17.8	26.6	
125	12.0	9.0	13.4	





#91

Benzo(a)pyrene

Concen: 3.06 ng/ul

RT: 24.69 min Scan# 3720

Delta R.T. -0.00 min

Lab File: BG018463.D

Acq: 18 Aug 2015 14:16

Instrument :

BNA_G


ClientSampleId :

MDL-07-S-ML




Tgt	Ion	Ratio	Resp	Lower	Upper
252	100		146376		
253	21.3		17.7	26.5	
125	9.8		10.2	15.2	

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#92

Indeno(1,2,3-cd)pyrene

Concen: 2.93 ng/ul


RT: 28.48 min Scan# 4364

Delta R.T. -0.01 min



Lab File: BG018463.D


Acq: 18 Aug 2015 14:16

Tgt	Ion	Ratio	Resp	Lower	Upper
276	100		161779		
138	20.9		15.8	23.6	
277	23.6		18.9	28.3	



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#93
Dibenzo(a,h)anthracene
Concen: 2.91 ng/ul
RT: 28.54 min Scan# 4374
Delta R.T. -0.02 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Instrument :
BNA_G
ClientSampleId :
MDL-07-S-ML

Tgt Ion	Ratio	Lower	Upper
278	100		
139	15.4	13.8	20.6
279	22.0	17.8	26.8

Manual Integrations
APPROVED

MMDadoda
8/19/2015 2:13:04 PM



#94
Benzo(g,h,i)perylene
Concen: 2.91 ng/ul
RT: 29.60 min Scan# 4555
Delta R.T. -0.02 min
Lab File: BG018463.D
Acq: 18 Aug 2015 14:16

Tgt Ion	Ratio	Lower	Upper
276	100		
138	19.7	16.5	24.7
277	21.2	19.0	28.6

