

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081816\
 Data File : BG023774.D
 Acq On : 19 Aug 2016 1:27
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013

Quant Time: Aug 19 03:56:06 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG081816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 19 01:43:20 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.10	152	94760	20.00	ng/ul	-0.02
18) Naphthalene-d8	10.93	136	417476	20.00	ng/ul	-0.01
35) Acenaphthene-d10	14.77	164	276398	20.00	ng/ul	-0.01
61) Phenanthrene-d10	17.53	188	719802	20.00	ng/ul	-0.01
75) Chrysene-d12	21.84	240	871241	20.00	ng/ul	0.00
83) Perylene-d12	25.22	264	677790	20.00	ng/ul	-0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.47	96	17111	8.20	ng/uL	0.00
5) Phenol-d5	7.25	99	183608	20.66	ng/ul	-0.02
7) Bis-(2-Chloroethyl)ether-d	7.41	67	127609	20.96	ng/ul	-0.02
9) 2-Chlorophenol-d4	7.63	132	131677	20.88	ng/ul	-0.02
13) 4-Methylphenol-d8	8.81	113	141209	20.33	ng/ul	-0.02
19) Nitrobenzene-d5	9.27	128	67821	20.11	ng/ul	-0.02
22) 2-Nitrophenol-d4	10.01	143	80653	20.97	ng/ul	-0.01
26) 2,4-Dichlorophenol-d3	10.55	165	141936	20.42	ng/ul	-0.02
29) 4-Chloroaniline-d4	11.07	131	186767	21.95	ng/ul	-0.02
43) Dimethylphthalate-d6	14.16	166	495746	21.64	ng/ul	-0.01
46) Acenaphthylene-d8	14.46	160	551818	21.27	ng/ul	-0.01
51) 4-Nitrophenol-d4	14.99	143	78057	21.08	ng/ul	-0.01
57) Fluorene-d10	15.76	176	428982	21.44	ng/ul	-0.01
62) 4,6-Dinitro-2-methylphenol	15.90	200	93401	20.47	ng/ul	-0.01
70) Anthracene-d10	17.63	188	681919	20.55	ng/ul	-0.01
76) Pyrene-d10	19.92	212	854185	20.43	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.99	264	638739	20.58	ng/ul	-0.02

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.50	88	18514	8.46 ng/uL#	86
4) Benzaldehyde	7.23	77	126522	22.07 ng/ul#	83
6) Phenol	7.28	94	193609	21.27 ng/ul#	76
8) Bis(2-Chloroethyl)ether	7.50	93	150182	21.62 ng/ul	87
10) 2-Chlorophenol	7.66	128	132034	20.53 ng/ul#	79
11) 2-Methylphenol	8.54	108	143810	20.77 ng/ul	89
12) 2,2'-oxybis(1-Chloropropan	8.62	45	212560	21.41 ng/ul	97
14) Acetophenone	8.92	105	228968	20.27 ng/ul#	78
15) N-Nitroso-di-n-propylamine	8.90	70	147894	21.14 ng/ul#	86
16) 4-Methylphenol	8.88	108	156486	20.78 ng/ul	89
17) Hexachloroethane	9.18	117	57970	20.42 ng/ul#	81
20) Nitrobenzene	9.32	77	204537	20.72 ng/ul#	80
21) Isophorone	9.84	82	385312	21.09 ng/ul#	92
23) 2-Nitrophenol	10.03	139	82728	19.97 ng/ul#	47
24) 2,4-Dimethylphenol	10.09	107	182838	20.45 ng/ul	87
25) Bis(2-Chloroethoxy)methane	10.32	93	204168	19.88 ng/ul	96
27) 2,4-Dichlorophenol	10.58	162	142675	20.56 ng/ul	96
28) Naphthalene	10.98	128	436944	20.46 ng/ul	97
30) 4-Chloroaniline	11.09	127	184989	22.17 ng/ul	95
31) Hexachlorobutadiene	11.26	225	100651	20.16 ng/ul	95
33) 4-Chloro-3-methylphenol	12.22	107	179451	20.92 ng/ul#	76
34) 2-Methylnaphthalene	12.59	142	341728	21.01 ng/ul	95
36) 1,2,4,5-Tetrachlorobenzene	12.96	216	187290	20.40 ng/ul	98


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081816\
 Data File : BG023774.D
 Acq On : 19 Aug 2016 1:27
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013

Quant Time: Aug 19 03:56:06 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG081816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 19 01:43:20 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) Hexachlorocyclopentadiene	12.92	237	61263	15.28	ng/ul	99
38) 2,4,6-Trichlorophenol	13.20	196	123000	20.37	ng/ul	96
39) 2,4,5-Trichlorophenol	13.28	196	126813	20.31	ng/ul	96
40) 1,1'-Biphenyl	13.59	154	447096	20.68	ng/ul	98
41) 2-Chloronaphthalene	13.64	162	341521	20.78	ng/ul	96
42) 2-Nitroaniline	13.85	65	147092	21.58	ng/ul#	62
44) Dimethylphthalate	14.21	163	490711	21.42	ng/ul	99
45) 2,6-Dinitrotoluene	14.34	165	101907	20.83	ng/ul#	80
47) Acenaphthylene	14.49	152	577413	21.36	ng/ul	98
48) 3-Nitroaniline	14.68	138	102545	22.25	ng/ul#	58
49) Acenaphthene	14.83	153	377628	20.86	ng/ul	99
50) 2,4-Dinitrophenol	14.90	184	45609	17.57	ng/ul#	83
52) 4-Nitrophenol	15.00	109	74278	19.97	ng/ul#	75
53) Dibenzofuran	15.17	168	558003	21.20	ng/ul#	87
54) 2,4-Dinitrotoluene	15.14	165	159818	22.36	ng/ul#	67
55) 2,3,4,6-Tetrachlorophenol	15.40	232	124151	21.56	ng/ul	99
56) Diethylphthalate	15.57	149	525837	22.47	ng/ul	97
58) Fluorene	15.82	166	466765	21.89	ng/ul	97
59) 4-Chlorophenyl-phenylether	15.80	204	231707	21.51	ng/ul	95
63) 4,6-Dinitro-2-methylphenol	15.91	198	99239	20.83	ng/ul#	81
64) N-Nitrosodiphenylamine	16.02	169	426945	20.03	ng/ul	97
65) 4-Bromophenyl-phenylether	16.71	248	166519	20.09	ng/ul	99
66) Hexachlorobenzene	16.83	284	179971	20.15	ng/ul	97
67) Atrazine	16.97	200	192748	22.08	ng/ul	96
68) Pentachlorophenol	17.19	266	72549	18.75	ng/ul	91
69) Phenanthrene	17.57	178	797464	20.93	ng/ul	99
71) Anthracene	17.67	178	822471	20.99	ng/ul	97
72) Carbazole	17.94	167	757199	23.68	ng/ul	99
73) Di-n-butylphthalate	18.48	149	982375	22.71	ng/ul#	94
74) Fluoranthene	19.59	202	1051309	25.90	ng/ul#	91
77) Pyrene	19.95	202	1070895	20.80	ng/ul#	92
78) Butylbenzylphthalate	20.82	149	473377	20.33	ng/ul#	83
79) 3,3'-Dichlorobenzidine	21.73	252	346855	20.03	ng/ul#	97
80) Benzo(a)anthracene	21.82	228	1046734	20.78	ng/ul	99
81) Bis(2-ethylhexyl)phthalate	21.70	149	554629	17.70	ng/ul	97
82) Chrysene	21.90	228	936782	20.67	ng/ul	99
84) Di-n-octyl phthalate	22.96	149	869067	22.31	ng/ul	100
85) Benzo(b)fluoranthene	24.14	252	833561	21.15	ng/ul#	94
86) Benzo(k)fluoranthene	24.21	252	808237	21.23	ng/ul#	92
88) Benzo(a)pyrene	25.06	252	783952	20.73	ng/ul#	93
89) Indeno(1,2,3-cd)pyrene	29.10	276	910219	20.47	ng/ul#	86
90) Dibenzo(a,h)anthracene	29.14	278	778328	20.41	ng/ul#	89
91) Benzo(g,h,i)perylene	30.31	276	758740	20.63	ng/ul#	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 8.46 ng/uL

RT: 3.50 min Scan# 113

Delta R.T. 0.00 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion: 88 Resp: 18514

Ion Ratio Lower Upper

88 100

43 31.8 16.6 24.8#




58 67.2 47.2 70.8



Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 43.00 (42.70 to 43.70): BGC

Ion 58.00 (57.70 to 58.70): BGC



#4

Benzaldehyde

Concen: 22.07 ng/uL

RT: 7.23 min Scan# 747

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion: 77 Resp: 126522

Ion Ratio Lower Upper

77 100

105 83.6 83.8 125.8#



106 83.7 77.9 116.9




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC





#6

Phenol

Concen: 21.27 ng/ul

RT: 7.28 min Scan# 756

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion: 94 Resp: 193609

Ion Ratio Lower Upper

94 100

65 31.8 16.8 25.2#




66 41.6 22.6 33.8#



Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 21.62 ng/ul

RT: 7.50 min Scan# 794

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion: 93 Resp: 150182

Ion Ratio Lower Upper

93 100

63 83.4 56.0 84.0



95 30.9 27.7 41.5




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 63.00 (62.70 to 63.70): BGC

Ion 95.00 (94.70 to 95.70): BGC





#10

2-Chlorophenol

Concen: 20.53 ng/ul

RT: 7.66 min Scan# 820

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:128 Resp: 132034

Ion Ratio Lower Upper

128 100

64 66.0 34.2 51.4#




130 34.9 27.0 40.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BG023774.D (-924) (-)

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 20.77 ng/ul

RT: 8.54 min Scan# 971

Delta R.T. -0.02 min

Lab File: BG023774.D


Acq: 19 Aug 2016 1:27

Tgt Ion:108 Resp: 143810

Ion Ratio Lower Upper



108 100


107 87.0 78.6 117.8



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
2,2'-oxybis(1-chloropropane)
Concen: 21.41 ng/ul
RT: 8.62 min Scan# 985
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27




Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion	Ratio	Lower	Upper
45	100		
77	15.6	13.4	20.2
79	11.2	10.2	15.4



#14
Acetophenone
Concen: 20.27 ng/ul
RT: 8.92 min Scan# 1036
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion	Ratio	Lower	Upper
105	100		
77	87.9	56.1	84.1#
51	29.3	14.6	22.0#









#15
N-Nitroso-di-n-propylamine
Concen: 21.14 ng/ul
RT: 8.90 min Scan# 1032
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27




Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion: 70 Resp: 147894
Ion Ratio Lower Upper
70 100
42 49.0 31.4 47.2#
101 10.0 8.6 13.0
130 15.6 19.0 28.4#



#16
4-Methylphenol
Concen: 20.78 ng/ul
RT: 8.88 min Scan# 1028
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion: 108 Resp: 156486
Ion Ratio Lower Upper
108 100
107 98.9 88.3 132.5





#17

Hexachloroethane

Concen: 20.42 ng/ul

RT: 9.18 min Scan# 1080

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:117 Resp: 57970

Ion Ratio Lower Upper

117 100

201 111.7 76.8 115.2




199 71.5 43.5 65.3#



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 20.72 ng/ul

RT: 9.32 min Scan# 1103

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion: 77 Resp: 204537

Ion Ratio Lower Upper

77 100

123 33.3 39.5 59.3#



65 13.3 9.0 13.4




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): BGC

Ion 65.00 (64.70 to 65.70): BGC









#21
Isophorone
Concen: 21.09 ng/ul
RT: 9.84 min Scan# 1191
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27




Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion: 82 Resp: 385312
Ion Ratio Lower Upper
82 100
95 8.2 5.1 7.7#
138 14.8 15.0 22.6#



#23
2-Nitrophenol
Concen: 19.97 ng/ul
RT: 10.03 min Scan# 1225
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion: 139 Resp: 82728
Ion Ratio Lower Upper
139 100
65 77.5 31.8 47.8#
109 39.7 17.1 25.7#









#24
 2,4-Dimethylphenol
 Concen: 20.45 ng/ul
 RT: 10.09 min Scan# 1235
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013


Tgt Ion: 107 Resp: 182838
 Ion Ratio Lower Upper
 107 100
 121 43.9 43.7 65.5
 122 76.9 70.3 105.5



#25
 Bis(2-Chloroethoxy)methane
 Concen: 19.88 ng/ul
 RT: 10.32 min Scan# 1274
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion: 93 Resp: 204168
 Ion Ratio Lower Upper
 93 100
 95 34.4 26.0 39.0
 123 10.4 10.1 15.1





#27

2,4-Dichlorophenol

Concen: 20.56 ng/ul

RT: 10.58 min Scan# 1318

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:162 Resp: 142675

Ion Ratio Lower Upper

162 100

164 64.4 51.6 77.4




98 41.9 28.3 42.5



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): B



#28

Naphthalene

Concen: 20.46 ng/ul

RT: 10.98 min Scan# 1386

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion:128 Resp: 436944

Ion Ratio Lower Upper

128 100

129 10.4 9.4 14.2



127 13.4 11.2 16.8




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#30
4-Chloroaniline
Concen: 22.17 ng/ul
RT: 11.09 min Scan# 1405
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27




Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion:127 Resp: 184989
Ion Ratio Lower Upper
127 100
129 30.5 26.9 40.3



#31
Hexachlorobutadiene
Concen: 20.16 ng/ul
RT: 11.26 min Scan# 1433
Delta R.T. -0.02 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion:225 Resp: 100651
Ion Ratio Lower Upper
225 100
223 60.9 54.7 82.1
227 64.4 52.2 78.4









#33
 4-Chloro-3-methylphenol
 Concen: 20.92 ng/ul
 RT: 12.22 min Scan# 1597
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013


Tgt Ion:107 Resp: 179451
 Ion Ratio Lower Upper
 107 100
 144 21.2 22.7 34.1#
 142 64.2 71.6 107.4#



#34
 2-Methylnaphthalene
 Concen: 21.01 ng/ul
 RT: 12.59 min Scan# 1660
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion:142 Resp: 341728
 Ion Ratio Lower Upper
 142 100
 141 89.5 75.2 112.8









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 20.40 ng/ul
 RT: 12.96 min Scan# 1723
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013

Tgt Ion:	216	Resp:	187290
Ion	Ratio	Lower	Upper
216	100		
214	78.9	61.7	92.5
179	21.7	17.0	25.4
108	15.7	14.2	21.2




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#37
 Hexachlorocyclopentadiene
 Concen: 15.28 ng/ul
 RT: 12.92 min Scan# 1717
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion:	237	Resp:	61263
Ion	Ratio	Lower	Upper
237	100		
235	62.4	49.6	74.4
272	9.5	9.3	13.9



Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E









#38
 2,4,6-Trichlorophenol
 Concen: 20.37 ng/ul
 RT: 13.20 min Scan# 1764
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013


Tgt Ion	Ratio	Lower	Upper
196	100		
198	100.2	77.4	116.0
200	30.1	26.6	39.8



#39
 2,4,5-Trichlorophenol
 Concen: 20.31 ng/ul
 RT: 13.28 min Scan# 1778
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion	Ratio	Lower	Upper
196	100		
198	103.3	79.8	119.8
200	33.4	24.0	36.0









#40
 1,1'-Biphenyl
 Concen: 20.68 ng/ul
 RT: 13.59 min Scan# 1831
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013

Tgt Ion	Ratio	Lower	Upper
154	100		
153	42.1	34.6	52.0
76	14.4	11.9	17.9




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): B






#41
 2-Chloronaphthalene
 Concen: 20.78 ng/ul
 RT: 13.64 min Scan# 1839
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion	Ratio	Lower	Upper
162	100		
164	32.6	25.4	38.2
127	41.7	30.2	45.4



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 127.00 (126.70 to 127.70): E









#42
2-Nitroaniline
Concen: 21.58 ng/ul
RT: 13.85 min Scan# 1875
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27




Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion: 65 Resp: 147092
Ion Ratio Lower Upper
65 100
92 62.5 67.1 100.7#
138 81.1 109.6 164.4#



#44
Dimethylphthalate
Concen: 21.42 ng/ul
RT: 14.21 min Scan# 1936
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion: 163 Resp: 490711
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 10.6 8.8 13.2









#45
 2,6-Dinitrotoluene
 Concen: 20.83 ng/ul
 RT: 14.34 min Scan# 1958
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SST02013


Tgt Ion	Ratio	Lower	Upper
165	100		
89	71.3	41.5	62.3
121	21.0	17.0	25.4



#47
 Acenaphthylene
 Concen: 21.36 ng/ul
 RT: 14.49 min Scan# 1983
 Delta R.T. -0.02 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.6	17.5	26.3
153	13.9	11.1	16.7









#48
 3-Nitroaniline
 Concen: 22.25 ng/ul
 RT: 14.68 min Scan# 2016
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013


Tgt Ion:138 Resp: 102545
 Ion Ratio Lower Upper
 138 100
 108 13.0 6.2 9.2#
 92 143.1 79.3 118.9#



#49
 Acenaphthene
 Concen: 20.86 ng/ul
 RT: 14.83 min Scan# 2042
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion:153 Resp: 377628
 Ion Ratio Lower Upper
 153 100
 152 46.0 37.3 55.9
 154 86.1 68.2 102.2









#50
2,4-Dinitrophenol
Concen: 17.57 ng/ul
RT: 14.90 min Scan# 2054
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Instrument :
BNA_G
ClientSampleId :
SSTD02013

Tgt Ion:184 Resp: 45609
Ion Ratio Lower Upper
184 100
63 68.0 40.5 60.7#
154 68.4 48.5 72.7




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BGC
Ion 154.00 (153.70 to 154.70): E






#52
4-Nitrophenol
Concen: 19.97 ng/ul
RT: 15.00 min Scan# 2071
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion:109 Resp: 74278
Ion Ratio Lower Upper
109 100
139 111.1 136.1 204.1#
65 135.5 106.0 159.0



Abundance Ion 109.00 (108.70 to 109.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 65.00 (64.70 to 65.70): BGC





#53

Dibenzenofuran

Concen: 21.20 ng/ul

RT: 15.17 min Scan# 2099

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :





SSTD02013

Tgt Ion:168 Resp: 558003

Ion Ratio Lower Upper

168 100

139 41.8 27.4 41.0#



#54

2,4-Dinitrotoluene

Concen: 22.36 ng/ul

RT: 15.14 min Scan# 2094

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27




Tgt Ion:165 Resp: 159818




Ion Ratio Lower Upper

165 100

63 51.9 25.7 38.5#

182 2.1 2.1 3.1









#55
2,3,4,6-Tetrachlorophenol
Concen: 21.56 ng/ul
RT: 15.40 min Scan# 2139
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27


Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion:	232	Resp:	124151
Ion	Ratio	Lower	Upper
232	100		
131	43.4	34.8	52.2
130	3.1	2.2	3.2
166	30.9	23.8	35.6



#56
Diethylphthalate
Concen: 22.47 ng/ul
RT: 15.57 min Scan# 2168
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion:	149	Resp:	525837
Ion	Ratio	Lower	Upper
149	100		
177	23.4	17.4	26.2
150	13.0	9.6	14.4





#58

Fluorene

Concen: 21.89 ng/ul

RT: 15.82 min Scan# 2210

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:166 Resp: 466765

Ion Ratio Lower Upper

166 100

165 97.4 80.3 120.5




167 13.0 10.3 15.5



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#59

4-Chlorophenyl-phenylether

Concen: 21.51 ng/ul

RT: 15.80 min Scan# 2207

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion:204 Resp: 231707

Ion Ratio Lower Upper

204 100

206 34.0 27.4 41.2



141 64.3 47.1 70.7




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#63
 4,6-Dinitro-2-methylphenol
 Concen: 20.83 ng/ul
 RT: 15.91 min Scan# 2225
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SST02013


Tgt Ion:198 Resp: 99239
 Ion Ratio Lower Upper
 198 100
 182 3.9 4.2 6.2#
 77 31.4 16.7 25.1#



#64
 N-Nitrosodiphenylamine
 Concen: 20.03 ng/ul
 RT: 16.02 min Scan# 2244
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion:169 Resp: 426945
 Ion Ratio Lower Upper
 169 100
 168 72.4 59.0 88.6
 167 34.8 30.6 46.0









#65
4-Bromophenyl-phenylether
Concen: 20.09 ng/ul
RT: 16.71 min Scan# 2361
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Instrument :
BNA_G
ClientSampleId :
SSTD02013

Tgt Ion:248 Resp: 166519
Ion Ratio Lower Upper
248 100
250 96.8 76.6 114.8
141 77.3 61.8 92.6




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#66
Hexachlorobenzene
Concen: 20.15 ng/ul
RT: 16.83 min Scan# 2382
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion:284 Resp: 179971
Ion Ratio Lower Upper
284 100
142 37.3 27.8 41.8
249 31.9 24.4 36.6



Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#67
Atrazine
Concen: 22.08 ng/ul
RT: 16.97 min Scan# 2406
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27




Instrument :
BNA_G
ClientSampleId :
SSTD02013


Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.6	21.0	31.6
215	44.6	37.9	56.9



#68
Pentachlorophenol
Concen: 18.75 ng/ul
RT: 17.19 min Scan# 2443
Delta R.T. -0.01 min
Lab File: BG023774.D
Acq: 19 Aug 2016 1:27

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.8	54.7	82.1
268	62.4	57.8	86.8





#69

Phenanthrene

Concen: 20.93 ng/ul

RT: 17.57 min Scan# 2508

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:178 Resp: 797464

Ion Ratio Lower Upper

178 100

179 16.5 12.9 19.3




176 19.4 15.8 23.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 20.99 ng/ul

RT: 17.67 min Scan# 2524

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion:178 Resp: 822471

Ion Ratio Lower Upper

178 100

179 17.5 12.3 18.5



176 19.5 16.3 24.5




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 23.68 ng/ul

RT: 17.94 min Scan# 2571

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:167 Resp: 757199

Ion Ratio Lower Upper

167 100

166 24.2 19.2 28.8




139 14.1 10.2 15.4



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#73

Di-n-butylphthalate

Concen: 22.71 ng/ul

RT: 18.48 min Scan# 2662

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion:149 Resp: 982375

Ion Ratio Lower Upper

149 100

150 10.4 7.1 10.7



104 7.2 3.3 4.9#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 25.90 ng/ul

RT: 19.59 min Scan# 2851

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:202 Resp: 1051309

Ion Ratio Lower Upper

202 100

101 9.1 10.6 16.0#




100 8.3 8.7 13.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 20.80 ng/ul

RT: 19.95 min Scan# 2913

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion:202 Resp: 1070895

Ion Ratio Lower Upper

202 100

101 11.3 12.5 18.7#



100 10.4 10.2 15.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B









#78
 Butylbenzylphthalate
 Concen: 20.33 ng/ul
 RT: 20.82 min Scan# 3061
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02013


Tgt Ion	Ratio	Lower	Upper
149	100		
91	82.1	54.2	81.4#
206	22.8	14.1	21.1#



#79
 3,3'-Dichlorobenzidine
 Concen: 20.03 ng/ul
 RT: 21.73 min Scan# 3216
 Delta R.T. -0.01 min
 Lab File: BG023774.D
 Acq: 19 Aug 2016 1:27

Tgt Ion	Ratio	Lower	Upper
252	100		
254	65.4	51.4	77.2
126	9.9	10.7	16.1#





#80

Benzo(a)anthracene

Concen: 20.78 ng/ul

RT: 21.82 min Scan# 3232

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013





Tgt Ion:228 Resp: 1046734

Ion Ratio Lower Upper

228 100

229 20.6 15.4 23.2

226 28.7 22.8 34.2



#81

Bis(2-ethylhexyl)phthalate

Concen: 17.70 ng/ul

RT: 21.70 min Scan# 3210

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27




Tgt Ion:149 Resp: 554629


Ion Ratio Lower Upper

149 100

167 28.9 21.9 32.9

279 4.2 2.9 4.3





#82

Chrysene

Concen: 20.67 ng/ul

RT: 21.90 min Scan# 3244

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:228 Resp: 936782

Ion Ratio Lower Upper

228 100

226 30.3 24.2 36.4




229 19.8 15.0 22.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Di-n-octyl phthalate

Concen: 22.31 ng/ul


RT: 22.96 min Scan# 3425

Delta R.T. -0.01 min



Lab File: BG023774.D


Acq: 19 Aug 2016 1:27

Tgt Ion:149 Resp: 869067



Abundance Ion 149.00 (148.70 to 149.70): E





#85

Benzo(b)fluoranthene

Concen: 21.15 ng/ul

RT: 24.14 min Scan# 3626

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27





Instrument :

BNA_G

ClientSampleId :

SSTD02013

Tgt Ion:	252	Resp:	833561
Ion Ratio	Lower	Upper	
252	100		
253	21.5	18.2	27.2
125	8.2	10.8	16.2#



#86

Benzo(k)fluoranthene

Concen: 21.23 ng/ul




RT: 24.21 min Scan# 3638


Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Tgt Ion:	252	Resp:	808237
Ion Ratio	Lower	Upper	
252	100		
253	22.3	15.8	23.6
125	8.7	10.6	16.0#





#88

Benzo(a)pyrene

Concen: 20.73 ng/ul

RT: 25.06 min Scan# 3783

Delta R.T. -0.01 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27





Instrument :

BNA_G

ClientSampleId :

SSTD02013

Tgt Ion:	252	Resp:	783952
Ion Ratio	Lower	Upper	
252	100		
253	22.7	19.0	28.4
125	9.2	12.0	18.0#



#89

Indeno(1,2,3-cd)pyrene

Concen: 20.47 ng/ul




RT: 29.10 min Scan# 4470


Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Tgt Ion:	276	Resp:	910219
Ion Ratio	Lower	Upper	
276	100		
138	15.5	20.8	31.2#
277	25.0	17.6	26.4





#90

Dibenzo(a,h)anthracene

Concen: 20.41 ng/ul

RT: 29.14 min Scan# 4478

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27

Instrument :

BNA_G

ClientSampleId :

SSTD02013


Tgt Ion:278 Resp: 778328

Ion Ratio Lower Upper

278 100

139 10.6 16.3 24.5#




279 22.8 19.1 28.7



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 20.63 ng/ul

RT: 30.31 min Scan# 4676

Delta R.T. -0.02 min

Lab File: BG023774.D

Acq: 19 Aug 2016 1:27


Tgt Ion:276 Resp: 758740

Ion Ratio Lower Upper

276 100

138 18.8 22.1 33.1#

277 22.8 17.0 25.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

