


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG081822\
 Data File : BG054648.D
 Acq On : 18 Aug 2022 9:24
 Operator : CG/JU
 Sample : DFTPP483
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP483

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 18 12:03:10 2022



AutoFind: Scans 2496, 2497, 2498; Background Corrected with Scan 2490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.6	29184	PASS
68	69	0.00	2	1.3	342	PASS
69	198	0.00	100	34.3	25915	PASS
70	69	0.00	2	0.2	58	PASS
127	198	10	80	45.5	34387	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	75632	PASS
199	198	5	9	6.8	5108	PASS
275	198	10	60	27.1	20516	PASS
365	198	1	100	3.7	2770	PASS
441	198	0.01	100	12.0	9062	PASS
442	442	50	100	100.0	55181	PASS
443	442	15	24	20.0	11041	PASS

m/z	Abundance
44.00	402.0
73.10	166.0
95.90	170.0
191.00	180.0
207.10	1586.0
208.00	283.0
281.00	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

m/z	Abundance
38.00	243.0
39.10	1833.0
40.00	411.0
44.00	391.0
49.20	176.0
50.10	6904.0
51.10	26048.0
52.10	1411.0
55.10	153.0
56.00	684.0
57.10	1785.0
61.10	305.0
62.00	394.0
63.10	1036.0
65.00	521.0
68.10	491.0
69.00	22624.0
73.00	320.0
74.10	2275.0
75.00	4025.0
76.10	1281.0
77.10	28256.0
78.10	2088.0
79.00	1845.0
80.10	1364.0
81.00	2100.0
82.00	549.0
83.00	392.0
85.00	381.0
86.00	435.0
87.00	289.0
91.00	378.0
92.00	396.0
93.00	2743.0
94.00	247.0
96.10	176.0
98.10	2396.0
99.00	1911.0
100.00	162.0
101.00	917.0
103.10	337.0
104.00	728.0
105.00	610.0
106.10	223.0
107.00	7875.0
108.00	1296.0
109.00	276.0
110.00	14902.0
111.00	2471.0
111.90	209.0
116.00	558.0
117.00	6344.0
118.00	562.0
120.00	151.0
122.10	602.0
123.00	784.0
124.00	393.0
125.00	386.0
127.00	29128.0
128.00	2232.0
129.00	11669.0
130.00	962.0
131.10	210.0
133.00	157.0
134.00	371.0
135.00	943.0
135.90	281.0
136.90	415.0
141.10	1384.0
142.00	453.0
143.00	468.0
146.00	240.0
147.10	694.0
148.00	1811.0
149.00	451.0
153.00	572.0
154.00	341.0
155.10	742.0
156.00	1147.0
157.10	177.0
157.90	339.0
158.90	214.0
160.00	563.0
161.00	611.0
161.90	200.0
165.00	630.0
166.00	352.0
167.00	3048.0
168.00	1926.0
168.90	267.0
172.00	291.0
173.10	430.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

174.00	598.0
175.00	1217.0
176.00	307.0
176.90	589.0
178.00	151.0
179.00	1966.0
180.00	1414.0
181.00	675.0
185.00	1041.0
186.00	7823.0
187.00	2199.0
188.00	346.0
189.00	511.0
190.90	363.0
192.00	704.0
193.00	839.0
194.00	161.0
196.10	1729.0
198.00	58328.0
199.00	4074.0
199.90	253.0
201.40	254.0
203.00	516.0
204.10	2330.0
205.00	3871.0
206.10	15901.0
207.00	3422.0
208.00	874.0
209.00	360.0
210.00	327.0
211.00	632.0
215.10	252.0
216.00	445.0
217.00	4206.0
218.00	575.0
221.00	1739.0
223.00	776.0
224.10	8712.0
225.00	2240.0
226.00	328.0
227.00	3940.0
228.00	517.0
229.00	817.0
230.00	155.0
231.10	338.0
234.00	239.0
235.00	352.0
235.90	160.0
236.90	252.0
239.00	189.0
241.00	247.0
242.00	464.0
243.00	473.0
244.10	6986.0
245.00	988.0
246.00	1461.0
247.00	271.0
249.00	233.0
252.90	153.0
255.00	34592.0
256.00	5142.0
257.00	409.0
258.00	1837.0
258.90	293.0
265.00	780.0
273.00	1008.0
274.00	2642.0
275.00	14140.0
276.00	2039.0
277.00	1406.0
278.00	227.0
281.00	229.0
283.10	273.0
284.90	251.0
293.00	276.0
296.00	4273.0
297.00	546.0
303.00	464.0
314.00	241.0
315.00	426.0
316.00	270.0
323.10	1554.0
324.00	234.0
326.90	231.0
333.00	153.0
334.00	775.0
335.00	207.0
346.00	221.0
352.00	497.0
353.00	325.0
354.00	370.0
354.20	374.0
365.00	1810.0
365.90	255.0
372.00	623.0
373.00	221.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

382.90	186.0
402.00	345.0
403.00	397.0
404.00	177.0
420.90	252.0
422.10	265.0
423.00	1791.0
424.10	473.0
441.00	5531.0
442.00	32792.0
443.00	6880.0
444.00	735.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

m/z	Abundance
37.10	171.0
38.10	510.0
39.10	2796.0
40.10	361.0
44.00	442.0
49.00	199.0
50.10	8292.0
51.10	34040.0
52.10	1770.0
55.10	213.0
56.00	933.0
57.00	2201.0
61.00	355.0
62.00	527.0
63.10	1376.0
64.00	176.0
65.00	629.0
68.10	534.0
69.00	30360.0
70.00	173.0
73.10	372.0
74.00	2854.0
75.00	5179.0
76.10	1700.0
77.10	38456.0
78.10	2561.0
79.10	2323.0
80.10	1727.0
81.00	2558.0
82.00	656.0
83.00	592.0
85.00	403.0
86.00	668.0
87.00	396.0
87.90	151.0
91.00	620.0
92.00	699.0
93.00	4127.0
94.10	395.0
96.10	255.0
98.10	2758.0
99.10	2380.0
99.90	197.0
101.00	1408.0
101.90	150.0
103.00	522.0
104.00	759.0
105.00	1157.0
106.10	368.0
107.00	11226.0
108.00	1653.0
110.00	20040.0
111.00	3405.0
112.00	327.0
112.90	153.0
116.00	679.0
117.10	9398.0
118.00	859.0
119.00	176.0
120.10	158.0
122.00	791.0
123.00	994.0
124.00	539.0
125.00	421.0
127.00	40384.0
128.00	3233.0
129.00	14998.0
130.00	1262.0
131.00	403.0
133.10	167.0
133.90	448.0
135.00	1370.0
135.90	554.0
137.00	637.0
140.00	409.0
141.00	2081.0
141.90	648.0
143.00	445.0
144.00	170.0
146.00	462.0
147.00	1254.0
148.00	2675.0
148.90	589.0
151.00	249.0
151.90	201.0
153.00	720.0
154.00	469.0
155.00	1069.0
156.10	1764.0
157.10	275.0
158.00	456.0
159.00	368.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

160.00	487.0
161.00	997.0
162.10	372.0
165.00	963.0
166.00	656.0
167.00	4410.0
168.10	2028.0
169.00	337.0
170.20	166.0
171.00	199.0
171.90	443.0
173.10	554.0
174.00	695.0
175.00	1644.0
176.10	508.0
177.00	715.0
177.90	282.0
179.00	3345.0
180.00	2210.0
181.00	1228.0
182.00	193.0
183.90	273.0
185.00	1603.0
186.00	12042.0
187.00	3458.0
188.10	353.0
188.90	901.0
191.00	537.0
192.00	1179.0
193.10	1358.0
194.10	311.0
196.00	2415.0
198.00	88912.0
199.00	6015.0
199.90	469.0
201.60	309.0
203.00	683.0
204.00	3581.0
205.10	5947.0
206.00	23440.0
207.00	4668.0
208.00	1309.0
209.00	433.0
210.10	401.0
211.00	778.0
215.00	349.0
216.10	603.0
217.00	6335.0
217.90	827.0
218.90	157.0
221.00	3385.0
221.90	1034.0
223.00	1519.0
224.10	14109.0
225.00	3353.0
226.00	348.0
227.00	5902.0
228.00	736.0
229.00	1346.0
231.00	523.0
234.00	432.0
235.00	464.0
235.90	229.0
237.00	550.0
238.90	185.0
240.10	204.0
241.00	347.0
242.00	721.0
243.00	753.0
244.00	10275.0
245.10	1495.0
246.00	2172.0
246.90	498.0
249.00	499.0
251.90	166.0
253.00	275.0
255.00	56448.0
256.00	8454.0
257.00	628.0
258.00	3260.0
258.90	531.0
265.00	1335.0
266.00	226.0
271.00	158.0
272.00	176.0
273.00	1584.0
274.00	4436.0
275.00	23760.0
276.00	3307.0
277.00	2246.0
278.00	389.0
281.10	197.0
282.90	259.0
284.00	194.0
285.00	279.0
292.90	611.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

m/z	Abundance
38.10	312.0
39.10	1941.0
40.00	362.0
44.00	434.0
48.90	201.0
50.10	6923.0
51.10	27464.0
52.10	1409.0
56.00	682.0
57.00	1958.0
61.00	363.0
62.00	361.0
63.10	1122.0
65.10	483.0
69.00	24760.0
73.10	232.0
74.10	2684.0
75.10	4276.0
76.10	1618.0
77.10	32152.0
78.10	2335.0
79.00	1809.0
80.00	1468.0
81.00	2001.0
82.00	463.0
83.00	475.0
85.10	412.0
86.00	531.0
87.00	340.0
91.00	549.0
92.00	489.0
93.00	3310.0
94.00	284.0
96.00	293.0
98.10	2341.0
99.00	2202.0
99.90	189.0
101.00	1136.0
103.00	403.0
104.00	807.0
105.00	706.0
105.90	204.0
107.00	9766.0
108.00	1584.0
110.00	17624.0
111.00	2531.0
112.10	338.0
116.00	536.0
117.00	8249.0
118.00	573.0
119.00	185.0
120.00	151.0
122.00	561.0
123.00	1072.0
124.00	518.0
125.00	455.0
127.00	33648.0
128.00	2831.0
129.00	13461.0
130.00	971.0
131.00	228.0
132.10	155.0
133.00	226.0
134.00	299.0
135.00	1052.0
136.00	454.0
137.00	482.0
138.10	168.0
140.10	218.0
141.00	1862.0
141.90	582.0
143.00	485.0
144.00	169.0
146.00	355.0
147.10	952.0
148.00	2327.0
149.10	558.0
150.90	297.0
153.00	689.0
154.00	420.0
155.00	1150.0
156.10	1473.0
157.00	372.0
158.00	401.0
158.90	244.0
160.00	652.0
161.10	802.0
161.90	209.0
162.90	175.0
165.00	678.0
165.90	671.0
167.00	4035.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

168.00	2034.0
169.00	355.0
172.00	305.0
173.00	573.0
174.00	863.0
175.10	1396.0
176.10	457.0
177.00	875.0
177.90	223.0
179.00	3029.0
180.00	2185.0
181.00	978.0
183.90	291.0
185.00	1461.0
186.00	11163.0
187.00	2895.0
188.00	277.0
189.00	852.0
190.00	175.0
191.00	543.0
192.00	1092.0
193.00	1076.0
194.00	297.0
195.10	179.0
196.00	2269.0
198.00	79656.0
199.00	5236.0
200.00	452.0
201.50	424.0
203.00	781.0
204.00	3083.0
205.00	5110.0
206.10	22096.0
207.00	4487.0
208.00	1139.0
209.00	502.0
210.00	355.0
211.00	870.0
215.00	205.0
216.00	530.0
217.00	6278.0
218.00	700.0
220.20	220.0
221.00	3546.0
221.80	923.0
223.00	1367.0
224.00	13288.0
225.00	3413.0
226.00	493.0
227.00	5406.0
228.00	650.0
229.00	1171.0
230.00	204.0
231.00	408.0
234.00	325.0
235.00	365.0
236.10	251.0
237.00	408.0
238.90	157.0
240.00	162.0
241.00	266.0
242.10	823.0
243.00	811.0
244.00	9969.0
245.10	1195.0
246.00	1789.0
247.00	518.0
248.90	365.0
253.00	240.0
255.00	55344.0
256.00	7557.0
257.00	598.0
258.00	3103.0
258.90	370.0
265.00	1258.0
266.00	231.0
271.00	183.0
273.00	1620.0
274.00	4347.0
275.00	23648.0
276.00	3078.0
277.00	1997.0
278.00	343.0
281.10	241.0
283.00	267.0
284.10	182.0
285.00	524.0
292.00	160.0
293.00	387.0
295.00	183.0
296.00	7572.0
297.00	1128.0
303.00	826.0
304.10	214.0
307.90	150.0
314.00	280.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

314.90	852.0
316.00	498.0
321.00	264.0
322.00	176.0
323.00	2189.0
324.00	433.0
327.00	439.0
328.00	323.0
333.10	325.0
334.00	1716.0
335.00	487.0
341.00	294.0
346.00	491.0
352.00	639.0
353.00	470.0
354.00	726.0
365.00	3460.0
365.90	464.0
370.90	200.0
372.00	1116.0
373.10	381.0
383.10	317.0
401.90	506.0
403.00	695.0
404.00	189.0
421.00	716.0
422.00	711.0
423.00	3980.0
424.00	975.0
441.10	11070.0
442.00	69872.0
443.00	13734.0
444.00	1108.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483

295.10	214.0
296.00	6591.0
297.00	1151.0
301.00	188.0
303.00	791.0
304.10	219.0
314.00	371.0
315.00	688.0
316.00	520.0
320.90	192.0
323.00	2203.0
324.00	477.0
327.00	491.0
328.10	223.0
332.00	151.0
333.00	257.0
334.00	1350.0
335.00	403.0
340.90	241.0
345.90	467.0
352.00	640.0
353.00	549.0
354.10	646.0
355.00	174.0
365.00	3040.0
365.90	500.0
371.10	219.0
372.00	1381.0
373.00	268.0
382.90	305.0
390.10	155.0
402.00	516.0
403.00	646.0
404.10	238.0
421.00	547.0
422.10	493.0
423.00	3529.0
424.00	901.0
441.00	10585.0
442.00	62880.0
443.00	12509.0
444.00	1219.0

Instrument :
BNA_G
ClientSampleId :
DFTPP483