


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG081915\
Data File : BG018465.D
Acq On : 19 Aug 2015 8:50
Operator : UM/MA
Sample : DFTPP37
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP37

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 20 04:10:50 2015



AutoFind: Scans 2549, 2550, 2551; Background Corrected with Scan 2544

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.4	44258	PASS
68	69	0.00	2	1.1	573	PASS
69	198	0.00	100	44.1	50861	PASS
70	69	0.00	2	0.6	299	PASS
127	198	10	80	52.7	60770	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	115373	PASS
199	198	5	9	7.7	8931	PASS
275	198	10	60	25.5	29469	PASS
365	198	1	100	3.8	4414	PASS
441	442	0.01	24	18.1	12929	PASS
442	198	50	100	62.0	71578	PASS
443	442	15	24	19.5	13983	PASS

m/z	Abundance
38.00	836.0
39.00	2582.0
39.90	1207.0
40.90	431.0
42.00	238.0
42.90	435.0
44.00	1060.0
44.80	191.0
48.90	284.0
50.00	10200.0
51.00	39592.0
52.00	2072.0
54.20	258.0
55.00	484.0
55.90	865.0
56.90	3032.0
58.00	407.0
58.90	415.0
61.00	516.0
61.90	789.0
63.00	1376.0
63.90	246.0
65.00	1322.0
67.90	982.0
68.90	45296.0
69.90	213.0
72.90	473.0
74.00	3520.0
75.00	7385.0
76.00	2872.0
77.00	50616.0
77.90	3259.0
78.90	2380.0
79.90	2018.0
80.90	3283.0
81.90	696.0
82.90	619.0
83.90	446.0
84.80	451.0
85.90	627.0
87.00	491.0
88.00	336.0
88.90	236.0
91.00	904.0
91.90	1183.0
92.90	6065.0
93.80	785.0
95.10	241.0
95.90	233.0
97.00	361.0
97.90	4641.0
98.90	3440.0
100.90	1668.0
101.70	303.0
102.90	663.0
103.90	1274.0
104.90	1172.0
106.90	15021.0
107.90	2528.0
109.00	363.0
109.90	26880.0
110.90	3994.0
111.90	1025.0
112.90	371.0
116.90	12353.0
117.90	895.0
119.90	264.0
122.00	940.0
122.90	1625.0
123.80	750.0
124.80	635.0
126.90	48984.0
127.90	3853.0
128.90	19136.0
129.90	1786.0
131.00	475.0
132.00	186.0
133.00	322.0
133.80	878.0
134.90	1280.0
136.00	304.0
136.80	1036.0
137.80	348.0
138.90	502.0
139.80	374.0
140.90	2720.0
141.90	799.0
142.90	547.0
144.80	265.0
145.80	572.0
146.90	945.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

147.90	2844.0
148.90	686.0
152.00	370.0
152.90	1029.0
153.90	444.0
154.90	1357.0
156.00	2363.0
157.00	384.0
157.50	239.0
158.00	527.0
158.90	419.0
159.90	932.0
160.90	1191.0
161.80	463.0
163.00	186.0
163.80	392.0
164.90	648.0
165.90	1098.0
166.90	5768.0
167.90	3181.0
168.90	165.0
171.00	190.0
171.80	751.0
172.90	457.0
173.90	1126.0
174.90	1938.0
175.90	644.0
176.90	1424.0
177.90	443.0
178.90	3875.0
180.00	2986.0
180.80	1226.0
182.00	642.0
183.00	217.0
183.90	338.0
184.90	2228.0
185.90	12824.0
186.90	4634.0
187.90	403.0
188.90	998.0
190.10	264.0
190.80	752.0
191.80	1426.0
193.00	1123.0
193.90	413.0
195.00	473.0
195.90	2334.0
197.90	92760.0
198.90	7431.0
199.90	474.0
201.20	912.0
203.00	1054.0
203.90	3156.0
204.90	7387.0
205.90	27512.0
206.90	3434.0
208.00	635.0
209.00	231.0
209.90	412.0
210.90	1319.0
212.00	229.0
214.70	247.0
216.10	336.0
216.90	7083.0
217.90	985.0
219.00	172.0
220.10	282.0
220.90	3692.0
222.00	1013.0
223.00	1876.0
223.90	12839.0
224.90	3233.0
225.90	382.0
226.90	6008.0
227.80	1052.0
228.80	1299.0
230.90	793.0
232.10	154.0
233.80	392.0
235.00	378.0
235.90	279.0
236.90	382.0
238.80	425.0
239.90	192.0
240.90	397.0
241.90	1042.0
243.00	934.0
244.00	9307.0
244.90	1465.0
245.90	2064.0
246.90	529.0
247.90	426.0
248.90	368.0
250.00	165.0
251.90	180.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

253.10	391.0
253.70	345.0
254.90	54600.0
255.90	8138.0
256.90	603.0
257.90	4063.0
259.00	820.0
260.70	207.0
264.90	1145.0
265.80	271.0
267.00	203.0
270.80	299.0
272.90	1765.0
273.90	4413.0
274.90	20840.0
275.90	2930.0
276.90	1768.0
278.00	225.0
282.00	226.0
282.90	302.0
284.00	160.0
285.00	454.0
292.90	455.0
293.50	240.0
295.90	6286.0
296.90	988.0
300.90	299.0
302.00	164.0
302.80	529.0
303.90	297.0
307.80	164.0
313.80	446.0
315.00	338.0
315.70	489.0
318.70	208.0
320.70	299.0
320.90	315.0
322.90	1869.0
324.00	621.0
326.70	407.0
332.10	153.0
333.00	547.0
333.90	1129.0
334.90	620.0
342.10	161.0
345.80	415.0
351.90	637.0
353.00	590.0
354.00	458.0
356.90	183.0
358.00	165.0
364.90	2152.0
365.90	254.0
371.80	659.0
372.00	668.0
373.00	217.0
384.00	154.0
390.00	273.0
402.00	255.0
402.90	385.0
404.10	390.0
406.40	160.0
412.20	171.0
420.70	339.0
422.00	173.0
422.90	1805.0
424.00	606.0
440.90	7727.0
442.00	40424.0
443.00	7820.0
443.80	738.0
444.80	209.0
495.60	167.0
509.60	203.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

m/z	Abundance
37.90	454.0
39.00	3895.0
39.90	1222.0
41.00	472.0
41.90	219.0
43.00	306.0
43.90	1375.0
44.90	321.0
47.90	154.0
49.00	350.0
50.00	13118.0
51.00	52192.0
52.00	2283.0
55.00	645.0
55.90	1108.0
56.90	3913.0
57.90	427.0
60.90	819.0
62.00	1012.0
63.00	2465.0
63.80	337.0
64.90	1416.0
68.90	59000.0
69.90	468.0
70.90	182.0
72.90	560.0
74.00	5225.0
75.00	9396.0
76.00	4276.0
77.00	67632.0
78.00	5003.0
78.90	3453.0
79.90	2765.0
80.90	4853.0
82.00	1519.0
82.90	1311.0
84.00	261.0
84.90	653.0
85.90	854.0
86.90	385.0
87.90	419.0
90.80	1162.0
91.90	1361.0
92.90	7433.0
93.90	591.0
95.00	341.0
96.00	663.0
97.00	194.0
97.90	5519.0
98.90	4822.0
99.90	356.0
100.90	2935.0
102.80	801.0
103.90	1819.0
105.00	990.0
105.90	542.0
106.90	22432.0
107.90	3281.0
108.90	698.0
109.90	34728.0
110.90	5228.0
111.90	703.0
114.00	272.0
114.60	171.0
115.90	1344.0
116.90	16186.0
117.90	894.0
118.90	231.0
119.80	376.0
121.90	849.0
122.90	1743.0
124.00	963.0
124.90	1246.0
125.70	672.0
126.90	72880.0
127.90	4416.0
128.90	25816.0
129.90	2705.0
130.90	346.0
131.90	332.0
133.90	1417.0
134.90	2523.0
136.00	763.0
137.00	821.0
137.90	269.0
138.60	199.0
140.00	435.0
140.90	3503.0
141.90	1360.0
142.90	1257.0
143.80	484.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

144.90	458.0
146.00	697.0
146.90	1996.0
147.90	4906.0
148.90	1158.0
149.90	409.0
150.90	529.0
151.90	263.0
152.90	1596.0
154.00	614.0
154.90	2477.0
155.90	2406.0
157.00	902.0
157.80	725.0
158.90	518.0
159.90	1301.0
160.90	1657.0
161.90	362.0
163.80	239.0
164.90	1797.0
165.90	1386.0
166.90	7693.0
168.00	3173.0
169.00	677.0
169.70	226.0
171.90	432.0
173.00	1259.0
174.00	1092.0
174.90	2773.0
176.00	759.0
176.90	780.0
177.90	753.0
178.90	5053.0
179.90	4052.0
180.90	1082.0
181.80	281.0
182.80	260.0
183.20	179.0
183.90	665.0
184.90	3163.0
185.90	19536.0
186.90	5868.0
187.80	702.0
188.90	1144.0
190.80	510.0
191.90	2323.0
192.90	2373.0
193.90	454.0
194.80	536.0
195.80	3522.0
197.90	131200.0
198.90	10674.0
199.90	813.0
201.50	482.0
202.90	1252.0
203.90	5020.0
204.90	8503.0
205.90	37056.0
207.00	5229.0
207.90	1062.0
208.90	558.0
210.00	754.0
210.90	1553.0
215.00	708.0
216.00	1237.0
216.90	10298.0
217.90	1694.0
219.00	179.0
219.90	176.0
220.90	5810.0
222.00	1051.0
222.90	2240.0
223.90	21984.0
224.90	5846.0
225.90	540.0
226.90	9476.0
227.90	1125.0
228.90	1721.0
229.90	397.0
230.90	601.0
231.60	172.0
233.80	412.0
234.90	709.0
236.00	480.0
236.90	555.0
238.10	184.0
238.90	392.0
239.90	613.0
240.90	478.0
241.90	1210.0
243.00	1201.0
243.90	13587.0
244.90	1807.0
245.90	3650.0
246.90	360.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

247.90	161.0
248.90	354.0
250.10	298.0
250.80	196.0
251.20	231.0
252.00	299.0
253.10	355.0
254.90	82576.0
255.90	11286.0
256.80	1430.0
257.90	4879.0
258.90	1038.0
259.90	380.0
264.90	2573.0
265.80	731.0
267.00	276.0
267.80	163.0
271.00	329.0
271.80	485.0
272.90	2653.0
273.90	5470.0
274.90	34168.0
275.90	4372.0
276.90	3551.0
277.80	479.0
279.00	173.0
281.90	285.0
283.00	597.0
283.90	313.0
284.90	1207.0
286.30	235.0
288.90	201.0
292.20	152.0
292.80	746.0
293.90	224.0
295.90	12297.0
296.90	1810.0
300.90	240.0
302.90	1944.0
303.90	256.0
306.80	167.0
307.90	155.0
309.70	202.0
313.90	403.0
314.90	1250.0
316.00	750.0
320.80	402.0
322.00	240.0
323.00	3386.0
323.90	795.0
326.00	177.0
326.80	666.0
333.00	289.0
334.00	1934.0
334.90	528.0
340.00	212.0
340.80	393.0
345.90	494.0
346.80	169.0
348.90	297.0
351.90	1249.0
352.90	599.0
353.90	1059.0
354.80	205.0
364.90	5288.0
365.80	858.0
370.10	153.0
370.90	192.0
371.80	1545.0
372.90	510.0
377.60	187.0
382.90	615.0
389.80	167.0
400.90	235.0
401.90	577.0
403.00	1260.0
403.70	312.0
410.00	259.0
420.90	529.0
421.90	495.0
422.90	5804.0
423.90	1094.0
430.10	222.0
438.70	316.0
439.20	160.0
440.90	14761.0
441.90	84016.0
443.00	15401.0
443.90	1924.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

m/z	Abundance
37.10	190.0
38.00	590.0
39.00	2734.0
40.00	1498.0
41.00	730.0
42.50	188.0
42.90	212.0
43.90	808.0
46.90	193.0
49.10	364.0
50.00	10401.0
51.00	42216.0
52.00	1475.0
53.00	155.0
54.90	399.0
56.00	1112.0
56.90	3366.0
58.00	329.0
59.00	335.0
61.00	723.0
61.90	744.0
63.00	2074.0
64.00	380.0
65.00	1346.0
66.00	171.0
66.70	229.0
67.90	739.0
68.90	48288.0
69.80	216.0
73.00	342.0
73.90	5168.0
74.90	6546.0
76.00	2695.0
77.00	54552.0
78.00	4167.0
78.90	3197.0
79.90	2630.0
80.90	3679.0
81.90	1141.0
82.90	971.0
84.10	210.0
85.00	951.0
85.90	1028.0
87.00	572.0
88.00	334.0
90.90	942.0
92.00	970.0
92.90	6167.0
93.90	367.0
94.70	223.0
95.80	743.0
96.90	438.0
97.90	5033.0
98.90	4064.0
99.90	395.0
100.90	1755.0
102.90	490.0
103.90	1380.0
104.90	1575.0
105.90	367.0
106.90	15799.0
108.00	2242.0
108.90	497.0
109.90	31136.0
110.90	4301.0
111.90	921.0
113.00	374.0
114.90	228.0
115.90	1061.0
116.90	12700.0
117.90	1214.0
119.90	285.0
120.80	182.0
121.90	1635.0
122.90	1913.0
123.90	1017.0
124.90	659.0
125.80	362.0
126.90	60448.0
127.90	4266.0
128.90	20456.0
129.90	2017.0
130.90	689.0
132.80	166.0
133.90	966.0
134.90	1670.0
136.20	311.0
136.80	876.0
137.80	362.0
139.80	367.0
140.90	2395.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

142.00	908.0
142.90	888.0
146.00	723.0
146.90	1460.0
147.90	4194.0
148.90	571.0
151.00	571.0
151.40	613.0
152.90	1221.0
153.90	698.0
154.90	2437.0
155.90	2637.0
156.80	464.0
158.00	549.0
159.00	386.0
159.90	1134.0
160.80	1342.0
162.00	292.0
163.00	173.0
164.90	998.0
166.00	949.0
166.90	7752.0
167.90	3467.0
168.90	595.0
170.00	296.0
170.70	167.0
171.90	1011.0
172.80	831.0
173.90	1612.0
174.90	1946.0
176.00	1026.0
176.90	1854.0
178.80	5238.0
179.90	3130.0
181.00	1549.0
183.80	561.0
184.90	2255.0
185.90	16968.0
186.90	5240.0
188.00	505.0
188.90	843.0
190.90	705.0
192.00	1839.0
193.00	1387.0
193.80	848.0
194.70	323.0
195.90	3865.0
197.90	122160.0
198.90	8689.0
199.80	582.0
200.70	166.0
201.50	433.0
203.00	1555.0
203.90	4734.0
204.90	8161.0
205.90	35744.0
206.90	5344.0
207.90	1493.0
208.80	1152.0
209.90	617.0
210.90	1175.0
211.60	724.0
214.90	444.0
215.90	1069.0
216.80	9334.0
217.90	1449.0
219.30	173.0
220.90	5247.0
221.90	1449.0
222.90	2184.0
223.90	21040.0
224.90	4957.0
225.90	544.0
226.90	8867.0
228.00	1070.0
228.90	1991.0
229.90	282.0
230.80	233.0
232.90	288.0
233.70	412.0
234.90	947.0
235.80	338.0
237.00	705.0
238.70	527.0
240.00	304.0
240.70	713.0
241.90	799.0
242.90	1148.0
243.90	15484.0
245.00	1892.0
245.90	3719.0
246.80	517.0
247.70	282.0
248.90	443.0
249.60	200.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

249.80	203.0
251.80	287.0
252.90	441.0
254.90	77576.0
255.90	11278.0
256.80	956.0
257.90	4638.0
259.00	698.0
264.00	152.0
264.90	2005.0
265.90	679.0
267.80	215.0
268.30	352.0
270.00	249.0
270.90	349.0
271.90	278.0
272.90	2268.0
273.90	6666.0
274.90	33400.0
275.90	3639.0
276.90	2525.0
277.80	527.0
282.90	416.0
283.90	221.0
284.80	458.0
289.80	224.0
291.00	231.0
291.90	300.0
292.80	807.0
293.90	167.0
294.70	188.0
295.90	11845.0
296.90	1637.0
297.90	156.0
301.90	390.0
302.80	1318.0
304.00	272.0
309.80	153.0
313.00	183.0
314.10	697.0
314.90	797.0
315.90	934.0
316.80	165.0
320.90	217.0
321.90	163.0
322.90	3782.0
323.90	604.0
326.80	838.0
328.00	293.0
331.90	424.0
332.90	419.0
334.00	1636.0
335.00	727.0
341.00	359.0
341.80	232.0
345.80	807.0
350.70	185.0
351.90	1227.0
353.00	313.0
354.00	986.0
354.70	287.0
364.90	5803.0
365.90	438.0
366.60	157.0
370.00	223.0
370.90	353.0
371.90	2572.0
372.70	488.0
378.80	368.0
382.90	728.0
384.80	158.0
389.90	516.0
401.90	646.0
402.80	1238.0
403.80	335.0
421.00	1053.0
421.90	562.0
422.90	4985.0
423.90	1553.0
439.20	171.0
440.90	16301.0
441.90	90296.0
442.90	18728.0
444.00	2100.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37

m/z	Abundance
38.60	241.0
39.90	1819.0
41.00	176.0
41.90	332.0
43.00	210.0
43.90	1233.0
48.90	249.0
50.90	408.0
54.90	195.0
59.90	185.0
76.80	273.0
81.00	192.0
82.90	159.0
90.90	391.0
95.10	198.0
95.90	183.0
96.90	186.0
116.00	160.0
119.10	185.0
121.10	164.0
133.00	263.0
133.90	181.0
158.70	219.0
176.70	418.0
193.10	193.0
206.80	370.0
209.10	206.0
280.90	393.0
282.00	206.0
302.70	196.0
461.40	200.0
469.90	159.0

Instrument :
BNA_G
ClientSampleId :
DFTPP37