


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082021\
 Data File : BG049904.D
 Acq On : 20 Aug 2021 7:38
 Operator : CG/JU
 Sample : DFTPP415
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP415

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG081921.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 20 10:07:50 2021



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2509

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	27957	PASS
68	69	0.00	2	0.5	163	PASS
69	198	0.00	100	44.4	32507	PASS
70	69	0.00	2	0.3	109	PASS
127	198	10	80	51.6	37811	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	73216	PASS
199	198	5	9	6.6	4810	PASS
275	198	10	60	33.7	24661	PASS
365	198	1	100	6.2	4507	PASS
441	198	0.01	100	15.8	11546	PASS
442	198	50	100	99.3	72680	PASS
443	442	15	24	19.9	14479	PASS

m/z	Abundance
44.15	314.0
49.25	154.0
73.15	163.0
107.05	158.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

m/z	Abundance
38.15	411.0
39.25	2013.0
40.15	281.0
44.05	173.0
50.15	7015.0
51.15	25816.0
52.15	1637.0
55.15	228.0
56.05	579.0
57.15	1674.0
61.15	240.0
62.25	366.0
63.15	1311.0
65.15	505.0
68.15	490.0
69.15	31448.0
73.25	232.0
74.15	3681.0
75.15	5395.0
76.25	1248.0
77.15	32128.0
78.15	2289.0
79.15	2024.0
80.15	2242.0
81.15	2629.0
82.15	884.0
83.15	727.0
85.15	433.0
86.15	840.0
87.15	271.0
91.05	576.0
92.05	535.0
93.15	3950.0
94.05	314.0
96.15	218.0
98.15	3027.0
99.15	2659.0
100.15	635.0
101.15	1540.0
103.05	577.0
104.15	1109.0
105.15	1091.0
106.15	443.0
107.15	11285.0
108.15	1702.0
110.15	15870.0
111.15	2940.0
112.05	353.0
113.15	176.0
116.25	919.0
117.15	10212.0
118.15	846.0
120.25	203.0
122.05	683.0
123.15	1082.0
124.05	582.0
125.15	896.0
127.15	34136.0
128.15	2968.0
129.15	15651.0
130.15	1286.0
131.25	235.0
134.15	719.0
135.15	1451.0
136.15	456.0
137.05	277.0
140.35	255.0
141.05	1883.0
142.15	424.0
143.15	438.0
144.15	226.0
144.95	174.0
146.25	486.0
147.15	980.0
148.15	1761.0
149.15	237.0
151.15	397.0
152.15	236.0
153.05	555.0
154.05	574.0
155.15	1248.0
156.25	1407.0
157.15	352.0
158.15	368.0
160.05	453.0
161.15	1136.0
162.05	204.0
162.35	205.0
163.95	173.0
165.05	828.0
166.35	493.0
167.15	3285.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

168.15	1970.0
169.15	310.0
170.05	167.0
171.25	172.0
172.15	431.0
173.05	590.0
174.15	875.0
175.15	1907.0
176.15	386.0
177.05	778.0
178.35	411.0
179.05	3462.0
180.15	2218.0
181.15	1060.0
182.25	182.0
184.05	345.0
185.15	1746.0
186.15	9733.0
187.15	2845.0
188.15	372.0
189.15	645.0
191.05	386.0
191.45	388.0
192.05	960.0
193.15	1481.0
195.25	299.0
196.15	1664.0
198.15	62584.0
199.15	4815.0
200.05	564.0
201.75	198.0
203.15	810.0
204.15	3268.0
205.15	4980.0
206.15	19696.0
207.15	2549.0
208.15	858.0
209.05	197.0
210.05	385.0
211.05	859.0
212.05	182.0
213.15	151.0
215.05	187.0
216.25	358.0
217.05	5582.0
218.15	675.0
221.15	2058.0
221.75	842.0
222.05	843.0
223.15	1644.0
224.15	12330.0
225.15	3129.0
226.25	345.0
227.15	5370.0
228.15	643.0
229.05	928.0
230.25	200.0
231.15	373.0
233.15	201.0
234.15	431.0
235.05	349.0
236.05	200.0
237.25	467.0
238.25	169.0
239.05	181.0
240.15	194.0
241.05	348.0
242.05	981.0
243.25	681.0
244.15	7085.0
245.25	755.0
246.15	2129.0
247.25	364.0
249.25	202.0
253.35	320.0
255.15	48144.0
256.15	6137.0
257.25	611.0
258.15	3204.0
258.95	593.0
265.05	1033.0
266.15	301.0
269.95	174.0
270.95	268.0
272.25	175.0
273.15	1388.0
274.15	3208.0
275.15	17448.0
276.15	2448.0
277.15	1813.0
278.05	407.0
279.05	203.0
283.25	273.0
284.05	219.0
285.15	415.0
292.15	187.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

293.25	423.0
294.05	186.0
296.15	7723.0
297.15	988.0
303.25	693.0
304.25	235.0
309.25	159.0
314.05	205.0
315.05	509.0
315.95	450.0
322.05	175.0
323.15	1668.0
324.15	325.0
327.15	314.0
328.25	456.0
333.15	161.0
334.25	1280.0
335.25	214.0
341.25	221.0
346.05	327.0
352.15	775.0
353.15	404.0
354.25	663.0
365.15	3260.0
366.25	456.0
371.05	161.0
372.15	1322.0
373.25	371.0
382.95	156.0
402.15	555.0
403.05	587.0
409.95	156.0
421.05	503.0
422.05	421.0
423.15	2441.0
424.25	586.0
441.15	7550.0
442.15	47064.0
443.15	9144.0
444.05	958.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

m/z	Abundance
38.25	491.0
39.25	2839.0
40.05	252.0
44.15	335.0
50.15	9129.0
51.15	34640.0
52.15	1517.0
53.25	159.0
55.05	219.0
56.15	798.0
57.15	2738.0
61.05	709.0
62.05	600.0
63.15	1488.0
65.15	595.0
69.15	38016.0
70.25	169.0
73.25	299.0
74.15	4265.0
75.15	6622.0
76.15	2193.0
77.15	41216.0
78.15	2880.0
79.15	2943.0
80.15	2663.0
81.15	2887.0
82.15	698.0
83.15	556.0
84.15	221.0
85.05	594.0
86.15	970.0
87.15	488.0
88.05	189.0
91.05	594.0
92.15	964.0
93.15	6097.0
94.05	332.0
95.05	234.0
96.15	318.0
97.05	419.0
98.15	4642.0
99.15	2960.0
100.15	458.0
101.15	1742.0
103.25	660.0
104.15	1133.0
105.15	997.0
106.25	511.0
107.15	15160.0
108.15	2233.0
110.15	22592.0
111.05	3657.0
112.05	375.0
113.15	175.0
116.15	930.0
117.15	13557.0
118.15	1004.0
119.05	243.0
120.15	371.0
122.15	1161.0
123.15	1401.0
124.05	554.0
125.15	544.0
127.15	43984.0
128.15	3678.0
129.15	19936.0
130.05	1308.0
131.05	248.0
134.05	453.0
135.15	1503.0
136.05	933.0
137.05	835.0
137.95	269.0
139.05	231.0
139.95	232.0
141.15	2897.0
142.15	791.0
143.15	455.0
144.15	339.0
146.15	183.0
147.15	1608.0
148.15	3520.0
149.15	806.0
150.35	232.0
151.05	245.0
152.05	268.0
153.15	841.0
154.15	907.0
155.15	1379.0
156.15	1648.0
157.25	527.0
158.05	865.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

159.15	291.0
160.15	1100.0
161.15	1620.0
162.05	241.0
163.35	164.0
165.15	1117.0
166.15	1165.0
167.15	4271.0
168.25	1972.0
169.15	399.0
171.25	290.0
172.05	460.0
173.25	638.0
174.15	1307.0
175.25	1972.0
176.15	665.0
177.15	1227.0
178.15	335.0
179.05	4035.0
180.15	2981.0
181.15	1129.0
182.15	244.0
182.95	181.0
183.95	155.0
185.15	2253.0
186.15	13683.0
187.15	4496.0
188.25	341.0
189.15	1044.0
189.95	322.0
191.15	746.0
192.15	1417.0
193.25	1554.0
194.15	244.0
195.05	172.0
196.15	2045.0
198.15	85112.0
199.15	5210.0
200.05	539.0
201.65	510.0
202.35	305.0
203.15	1133.0
204.25	4386.0
205.15	6979.0
206.15	27432.0
207.15	4219.0
208.15	1354.0
209.05	333.0
210.35	481.0
211.25	1283.0
212.15	204.0
215.25	281.0
216.05	648.0
217.15	8782.0
218.15	1250.0
219.05	259.0
219.95	168.0
221.15	2353.0
222.05	1345.0
223.15	2263.0
224.15	17432.0
225.15	3337.0
226.15	530.0
227.15	8472.0
228.15	652.0
229.15	1563.0
230.15	601.0
231.25	680.0
231.85	165.0
233.05	168.0
234.15	1025.0
235.05	647.0
236.15	269.0
237.15	532.0
239.15	406.0
239.95	234.0
240.85	228.0
241.25	280.0
242.15	1088.0
243.15	582.0
244.15	12477.0
245.15	1601.0
246.15	3063.0
247.15	551.0
249.05	306.0
252.25	337.0
253.15	408.0
255.15	72584.0
256.15	9374.0
257.25	604.0
258.15	4560.0
258.95	663.0
265.15	2752.0
266.25	285.0
271.25	242.0
272.15	154.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

273.15	1845.0
274.15	5252.0
275.15	28712.0
276.15	3886.0
277.15	2855.0
278.05	402.0
279.15	214.0
283.25	310.0
284.25	346.0
285.15	644.0
289.05	156.0
293.05	540.0
294.25	190.0
295.05	153.0
296.15	10830.0
297.05	1045.0
302.15	431.0
303.25	1303.0
304.15	238.0
308.25	176.0
314.15	559.0
315.05	1225.0
316.15	742.0
321.15	411.0
322.25	242.0
323.15	2706.0
324.15	646.0
327.15	533.0
328.15	409.0
332.05	356.0
332.45	231.0
333.05	438.0
334.15	2148.0
335.15	382.0
341.15	334.0
346.15	849.0
347.05	233.0
351.25	181.0
352.25	775.0
353.15	781.0
354.15	1000.0
355.35	163.0
365.15	5521.0
366.15	724.0
371.15	178.0
372.15	1796.0
373.05	566.0
383.05	264.0
390.15	371.0
391.15	209.0
402.15	643.0
403.05	834.0
404.15	359.0
421.05	517.0
423.15	4634.0
424.15	821.0
432.65	200.0
440.15	192.0
441.15	12401.0
442.15	83296.0
443.15	16221.0
444.15	1265.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

m/z	Abundance
38.15	438.0
39.15	1524.0
40.15	267.0
44.15	252.0
50.15	6789.0
51.15	23416.0
52.15	979.0
56.15	628.0
57.15	1696.0
61.05	532.0
62.05	428.0
63.25	1445.0
64.15	309.0
65.15	617.0
69.15	28056.0
70.05	157.0
73.15	183.0
74.15	3249.0
75.15	4506.0
76.15	1617.0
77.15	28608.0
78.15	1827.0
79.15	2351.0
80.15	2078.0
81.15	2772.0
82.25	508.0
83.15	379.0
85.15	376.0
86.15	454.0
87.15	528.0
91.15	801.0
92.15	815.0
93.15	4249.0
94.15	264.0
98.15	3164.0
99.15	2477.0
100.15	217.0
101.15	1893.0
102.95	287.0
104.15	982.0
105.05	817.0
106.05	290.0
107.15	10914.0
108.15	1335.0
109.15	592.0
110.15	18016.0
111.15	2516.0
112.15	350.0
116.15	904.0
117.15	10584.0
118.05	604.0
120.25	163.0
122.15	870.0
123.05	1104.0
124.15	428.0
125.15	404.0
127.15	35312.0
128.15	2920.0
129.15	15257.0
130.15	1201.0
131.15	285.0
134.15	694.0
135.25	1392.0
136.05	629.0
137.15	463.0
140.25	359.0
141.15	2098.0
142.15	532.0
143.15	568.0
144.15	215.0
146.25	370.0
147.05	1280.0
148.15	2484.0
149.05	298.0
149.95	287.0
151.15	348.0
152.15	170.0
153.15	762.0
154.05	712.0
155.15	1617.0
156.15	1430.0
157.05	235.0
158.15	519.0
159.15	378.0
160.15	798.0
161.15	883.0
162.15	347.0
165.05	698.0
166.15	992.0
167.15	3992.0
168.15	1337.0
169.15	528.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

170.05	213.0
171.15	200.0
172.05	368.0
173.15	431.0
174.25	930.0
175.25	2072.0
176.15	559.0
177.15	849.0
178.15	456.0
179.15	3878.0
180.15	2074.0
181.05	1185.0
182.05	253.0
184.15	379.0
185.15	1325.0
186.15	11660.0
187.15	3401.0
188.15	317.0
189.15	1091.0
191.05	295.0
192.15	1393.0
193.15	1135.0
194.15	404.0
196.15	2292.0
198.15	71952.0
199.15	4405.0
200.05	433.0
201.65	313.0
203.15	856.0
204.15	3320.0
205.15	5966.0
206.15	22576.0
207.15	2803.0
208.15	679.0
209.05	203.0
210.25	320.0
211.15	1069.0
213.15	255.0
215.05	299.0
216.15	558.0
217.15	7025.0
218.05	1020.0
220.15	207.0
221.05	2673.0
222.15	957.0
223.15	1463.0
224.15	13664.0
225.15	3422.0
226.25	415.0
227.15	6811.0
228.05	809.0
229.15	1532.0
231.15	500.0
234.15	408.0
235.15	516.0
236.25	350.0
237.25	743.0
239.15	482.0
240.05	226.0
241.25	322.0
242.05	930.0
243.25	1091.0
244.15	10002.0
245.15	1469.0
246.15	2271.0
247.15	685.0
248.05	166.0
249.15	341.0
249.95	202.0
251.15	170.0
252.35	300.0
253.45	368.0
255.15	62520.0
256.15	8883.0
256.85	451.0
257.15	486.0
258.15	4080.0
259.15	438.0
265.15	1999.0
266.15	394.0
271.15	221.0
272.15	271.0
273.25	1989.0
274.15	4481.0
275.15	27824.0
276.15	3641.0
277.15	2820.0
278.25	270.0
282.05	205.0
283.05	330.0
284.15	272.0
285.15	528.0
286.15	154.0
291.85	180.0
293.25	578.0
294.05	169.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415

295.05	209.0
296.15	9040.0
297.15	1362.0
301.15	281.0
301.95	185.0
303.15	1038.0
304.25	437.0
308.05	157.0
310.15	295.0
314.15	683.0
315.05	1158.0
316.15	834.0
317.35	177.0
321.25	330.0
323.15	2671.0
323.95	385.0
327.05	894.0
328.15	608.0
331.95	222.0
332.25	184.0
333.15	289.0
334.25	1677.0
335.05	580.0
340.05	154.0
341.15	439.0
342.15	289.0
346.25	498.0
347.15	196.0
352.05	908.0
353.25	642.0
354.15	456.0
359.05	210.0
365.05	4739.0
366.05	679.0
372.15	1679.0
373.15	413.0
383.05	320.0
390.15	269.0
390.95	244.0
392.05	150.0
402.25	653.0
403.15	858.0
404.15	420.0
421.05	820.0
422.15	874.0
423.15	4610.0
424.15	913.0
441.15	14686.0
442.15	87680.0
443.15	18072.0
444.35	889.0
475.55	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP415