

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082024\
 Data File : BG062482.D
 Acq On : 20 Aug 2024 18:38
 Operator : MA/JU
 Sample : PB162773BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK773

Quant Time: Aug 20 23:12:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 20 23:01:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.944	152	60422	20.000	ng/u1	0.00
20) Naphthalene-d8	10.734	136	273541	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.558	164	203221	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.302	188	438412	20.000	ng/u1	0.00
79) Chrysene-d12	21.543	240	426607	20.000	ng/u1	0.00
88) Perylene-d12	24.621	264	476984	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.409	96	6942	4.362	ng/uL	0.00
4) Pyridine-d5	3.808	84	106111	21.881	ng/u1	0.00
7) Phenol-d5	7.104	99	168039	27.956	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.268	67	98985	26.143	ng/u1	0.00
11) 2-Chlorophenol-d4	7.474	132	120220	29.279	ng/u1	0.00
15) 4-Methylphenol-d8	8.637	113	137372	27.265	ng/u1	0.00
21) Nitrobenzene-d5	9.089	128	59609	33.980	ng/u1	0.00
24) 2-Nitrophenol-d4	9.812	143	65005	37.251	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.352	165	130058	29.261	ng/u1	0.00
31) 4-Chloroaniline-d4	10.863	131	170692	24.216	ng/u1	0.00
46) Dimethylphthalate-d6	13.965	166	420944	26.629	ng/u1	0.00
49) Acenaphthylene-d8	14.253	160	472711	27.008	ng/u1	0.00
54) 4-Nitrophenol-d4	14.746	143	66410	26.335	ng/u1	0.00
60) Fluorene-d10	15.551	176	383759	26.917	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	55132	27.677	ng/u1	0.00
73) Anthracene-d10	17.402	188	582576	27.662	ng/u1	0.00
81) Pyrene-d10	19.681	212	731623	29.284	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.410	264	708464	27.988	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082024\
Data File : BG062482.D
Acq On : 20 Aug 2024 18:38
Operator : MA/JU
Sample : PB162773BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK773

Quant Time: Aug 20 23:12:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 20 23:01:24 2024
Response via : Initial Calibration

