

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082124\
 Data File : BG062504.D
 Acq On : 21 Aug 2024 11:51
 Operator : MA/JU
 Sample : PB162794TB 10X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLEB794

Quant Time: Aug 21 13:44:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 20 23:01:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	66026	20.000	ng/u1	0.00
20) Naphthalene-d8	10.734	136	301453	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.558	164	230436	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.301	188	494200	20.000	ng/u1	0.00
79) Chrysene-d12	21.537	240	468083	20.000	ng/u1	0.00
88) Perylene-d12	24.609	264	526290	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.402	96	8481	4.877	ng/uL	0.00
4) Pyridine-d5	3.813	84	121127	22.857	ng/u1	0.00
7) Phenol-d5	7.103	99	173186	26.367	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.268	67	101043	24.422	ng/u1	0.00
11) 2-Chlorophenol-d4	7.473	132	122060	27.204	ng/u1	0.00
15) 4-Methylphenol-d8	8.636	113	137864	25.040	ng/u1	0.00
21) Nitrobenzene-d5	9.095	128	61337	31.728	ng/u1	0.00
24) 2-Nitrophenol-d4	9.811	143	65646	34.135	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.352	165	133396	27.233	ng/u1	0.00
31) 4-Chloroaniline-d4	10.863	131	170770	21.984	ng/u1	0.00
46) Dimethylphthalate-d6	13.964	166	433450	24.182	ng/u1	0.00
49) Acenaphthylene-d8	14.252	160	496369	25.011	ng/u1	0.00
54) 4-Nitrophenol-d4	14.746	143	65806	23.014	ng/u1	0.00
60) Fluorene-d10	15.551	176	395591	24.470	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.662	200	53475	23.815	ng/u1	0.00
73) Anthracene-d10	17.395	188	633121	26.668	ng/u1	0.00
81) Pyrene-d10	19.680	212	757047	27.616	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.403	264	734710	26.305	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082124\
Data File : BG062504.D
Acq On : 21 Aug 2024 11:51
Operator : MA/JU
Sample : PB162794TB 10X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SLEB794

Quant Time: Aug 21 13:44:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 20 23:01:24 2024
Response via : Initial Calibration

