


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082217\
Data File : BG028411.D
Acq On : 22 Aug 2017 9:58
Operator : SJ/JU
Sample : DFTPP68
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP68

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 23 14:59:16 2017



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.1	16707	PASS
68	69	0.00	2	0.8	137	PASS
69	198	0.00	100	38.0	17557	PASS
70	69	0.00	2	0.3	53	PASS
127	198	10	80	42.3	19565	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	46256	PASS
199	198	5	9	6.7	3102	PASS
275	198	10	60	29.4	13584	PASS
365	198	1	100	3.7	1697	PASS
441	442	0.01	100	15.6	6085	PASS
442	198	50	100	84.4	39058	PASS
443	442	15	24	19.4	7585	PASS

m/z	Abundance
38.10	169.0
39.10	1310.0
40.10	2917.0
41.10	268.0
43.10	331.0
44.10	531.0
50.10	4117.0
51.10	16002.0
52.10	642.0
53.30	156.0
54.90	200.0
56.10	745.0
57.10	1264.0
61.00	344.0
62.10	297.0
63.00	723.0
65.00	324.0
69.00	17008.0
73.10	248.0
74.10	1516.0
75.10	3574.0
76.10	876.0
77.10	17264.0
78.10	1059.0
79.00	1408.0
80.00	999.0
81.10	1461.0
81.90	296.0
83.00	335.0
86.90	369.0
91.00	325.0
92.10	310.0
93.10	1546.0
94.00	391.0
95.00	330.0
95.80	159.0
98.00	1787.0
99.00	1327.0
101.00	850.0
103.00	213.0
104.00	621.0
105.00	279.0
106.10	227.0
107.00	4791.0
108.20	498.0
109.20	245.0
110.10	9260.0
111.10	1586.0
112.10	434.0
116.00	343.0
117.00	4069.0
122.10	330.0
123.00	568.0
125.00	273.0
127.00	16624.0
128.00	1613.0
129.00	7396.0
129.90	800.0
132.00	161.0
134.10	272.0
135.00	760.0
136.10	457.0
137.00	348.0
139.80	305.0
141.20	583.0
142.90	436.0
146.90	495.0
148.10	826.0
148.80	396.0
151.90	151.0
153.00	258.0
154.10	442.0
155.00	488.0
156.10	798.0
156.90	333.0
157.90	301.0
159.00	311.0
159.90	424.0
161.00	543.0
165.00	429.0
166.00	269.0
167.00	1610.0
167.90	850.0
169.00	450.0
171.10	186.0
173.20	252.0
174.00	331.0
175.00	577.0
176.00	184.0
176.90	598.0
177.60	263.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

178.10	186.0
179.00	1888.0
180.10	873.0
180.90	534.0
183.50	259.0
185.10	871.0
186.00	5085.0
187.10	1799.0
188.10	362.0
189.00	840.0
191.10	163.0
192.10	561.0
193.10	672.0
196.00	816.0
196.30	824.0
198.00	36704.0
199.00	2387.0
201.50	311.0
203.10	289.0
204.00	1383.0
205.00	2978.0
206.00	9266.0
207.00	2113.0
208.00	386.0
208.90	221.0
209.90	377.0
211.20	439.0
216.10	326.0
217.00	2877.0
218.10	369.0
221.00	1612.0
221.60	484.0
223.10	999.0
224.00	4982.0
225.00	1773.0
227.00	2305.0
227.90	427.0
229.10	652.0
234.00	255.0
235.00	174.0
236.00	186.0
236.90	152.0
241.00	259.0
242.00	444.0
243.10	378.0
244.00	4388.0
245.00	587.0
246.00	1069.0
247.10	256.0
252.70	210.0
255.00	23000.0
256.00	4062.0
257.00	170.0
258.00	1412.0
258.90	205.0
265.10	462.0
265.90	287.0
273.00	401.0
274.00	1899.0
275.00	11192.0
276.00	1164.0
276.90	807.0
278.10	239.0
279.00	209.0
283.00	158.0
284.90	221.0
292.20	157.0
292.80	352.0
296.00	2537.0
296.90	320.0
303.10	443.0
314.00	163.0
315.10	418.0
316.00	369.0
323.10	678.0
327.10	206.0
333.90	469.0
334.90	231.0
340.80	187.0
346.00	232.0
351.90	157.0
354.00	401.0
355.10	277.0
365.00	1452.0
365.80	154.0
372.10	879.0
402.00	392.0
403.00	277.0
422.20	193.0
423.00	1394.0
424.00	399.0
441.00	3741.0
442.00	22880.0
443.00	4398.0
444.10	521.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

m/z	Abundance
37.20	170.0
38.20	275.0
39.10	1888.0
40.10	2431.0
44.00	454.0
48.90	225.0
50.10	4707.0
51.10	19728.0
52.10	1105.0
55.30	150.0
56.00	628.0
57.10	1483.0
62.00	245.0
63.10	602.0
65.10	537.0
69.00	21376.0
70.10	161.0
73.10	235.0
74.10	1874.0
75.00	3279.0
76.10	1597.0
77.10	24288.0
78.10	1434.0
79.10	1930.0
80.00	1161.0
81.10	1696.0
82.00	444.0
83.00	372.0
85.00	407.0
86.00	645.0
87.10	410.0
87.90	264.0
91.10	368.0
92.00	361.0
93.00	2176.0
93.90	165.0
96.10	165.0
97.20	237.0
98.00	1965.0
99.00	1374.0
100.00	236.0
101.00	516.0
103.00	478.0
103.90	545.0
105.00	1009.0
106.10	400.0
107.00	7490.0
108.10	924.0
108.80	242.0
110.00	11704.0
111.10	1785.0
112.10	223.0
113.10	155.0
115.90	271.0
117.00	4649.0
118.00	173.0
122.00	648.0
123.00	815.0
123.90	299.0
125.00	260.0
126.00	207.0
127.00	22480.0
128.10	1731.0
129.00	10729.0
130.00	711.0
134.00	484.0
135.10	1039.0
135.90	384.0
137.10	602.0
138.10	308.0
139.70	164.0
140.90	870.0
141.90	408.0
146.00	387.0
147.00	734.0
148.00	1494.0
149.00	248.0
151.00	179.0
152.00	228.0
152.80	361.0
154.30	277.0
155.00	734.0
156.00	1007.0
157.10	350.0
158.00	207.0
159.10	225.0
160.00	472.0
161.00	960.0
162.10	185.0
165.10	490.0
166.00	648.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

167.00	2480.0
168.10	1251.0
169.00	618.0
169.80	212.0
171.00	255.0
171.90	162.0
173.00	518.0
174.00	744.0
175.10	1272.0
176.00	398.0
177.10	742.0
178.00	236.0
179.00	2200.0
180.00	1742.0
181.00	551.0
185.00	1753.0
186.10	8205.0
187.00	1804.0
188.00	254.0
189.00	598.0
191.00	267.0
191.90	463.0
193.00	687.0
196.00	1328.0
198.00	54168.0
198.90	3860.0
200.20	452.0
201.30	423.0
203.10	426.0
204.00	2183.0
205.00	4043.0
206.10	16576.0
207.10	2250.0
208.00	867.0
209.00	347.0
210.00	317.0
211.10	668.0
212.10	295.0
213.20	255.0
216.00	775.0
217.00	4941.0
218.00	785.0
219.90	267.0
221.10	2617.0
222.00	556.0
223.00	913.0
224.10	9342.0
225.10	2243.0
226.20	251.0
227.00	3586.0
228.00	611.0
229.10	474.0
230.00	230.0
230.80	200.0
233.80	211.0
235.10	266.0
237.10	280.0
239.00	155.0
239.80	166.0
241.00	277.0
242.00	545.0
242.90	537.0
244.00	6727.0
245.00	950.0
245.90	1470.0
246.90	664.0
248.80	342.0
249.20	250.0
251.10	222.0
252.90	250.0
253.70	540.0
254.00	547.0
255.00	39936.0
256.00	6511.0
257.10	257.0
258.00	2704.0
259.00	480.0
265.00	716.0
271.80	312.0
273.00	1264.0
274.00	3161.0
275.00	15217.0
276.00	2566.0
277.00	1032.0
278.00	162.0
279.20	180.0
280.90	202.0
284.10	159.0
285.20	246.0
289.00	215.0
292.70	208.0
295.00	308.0
296.00	4783.0
297.00	804.0
302.00	185.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

303.00	601.0
304.10	156.0
308.00	249.0
314.10	220.0
315.00	409.0
316.10	242.0
316.50	176.0
317.20	167.0
323.00	1402.0
323.90	153.0
324.10	157.0
325.10	285.0
326.90	464.0
332.10	153.0
334.10	742.0
334.90	394.0
345.90	333.0
346.90	207.0
352.10	773.0
353.10	646.0
354.10	551.0
355.10	241.0
359.00	160.0
365.00	1896.0
372.10	879.0
373.10	341.0
383.10	219.0
389.90	183.0
390.20	150.0
402.00	529.0
403.00	178.0
404.20	183.0
421.00	406.0
421.80	241.0
423.00	2432.0
424.10	498.0
429.40	152.0
435.50	204.0
441.00	7179.0
442.00	44712.0
443.10	8372.0
444.10	921.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

m/z	Abundance
38.20	243.0
39.10	900.0
40.00	2436.0
41.10	251.0
44.00	825.0
49.10	361.0
50.10	3887.0
51.10	14391.0
52.10	837.0
54.90	237.0
55.90	396.0
57.10	1084.0
61.00	427.0
62.10	296.0
63.10	505.0
65.10	642.0
67.10	243.0
68.20	412.0
69.00	14287.0
73.00	173.0
74.00	1532.0
75.10	2597.0
76.10	1230.0
77.10	18688.0
78.10	1131.0
79.00	1114.0
80.10	1251.0
81.00	1274.0
81.90	674.0
83.00	288.0
85.00	481.0
86.00	324.0
91.10	756.0
92.00	353.0
93.00	1811.0
98.10	1651.0
99.00	1468.0
101.00	987.0
103.00	239.0
104.00	658.0
104.90	453.0
107.00	5438.0
108.00	832.0
109.20	183.0
110.00	9511.0
111.00	1144.0
115.90	347.0
117.00	3973.0
118.00	273.0
122.10	552.0
123.00	631.0
124.80	178.0
127.00	19592.0
128.10	1144.0
129.00	6967.0
130.00	689.0
133.90	260.0
135.00	573.0
137.00	331.0
141.00	1295.0
142.00	229.0
143.00	337.0
144.90	188.0
145.80	175.0
147.10	620.0
148.00	996.0
148.90	318.0
151.20	241.0
153.00	510.0
154.20	156.0
155.00	338.0
156.00	1016.0
157.10	264.0
158.10	283.0
158.90	175.0
160.10	505.0
161.00	671.0
163.20	212.0
165.00	413.0
165.90	224.0
166.20	200.0
166.90	1715.0
167.90	1304.0
169.10	242.0
170.00	163.0
172.90	177.0
174.10	357.0
175.10	1019.0
176.10	450.0
177.00	566.0
178.90	1935.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

180.10	1046.0
181.00	419.0
184.00	341.0
185.10	1160.0
186.00	5745.0
187.00	2264.0
188.10	301.0
189.00	634.0
189.90	177.0
191.00	183.0
192.10	420.0
193.10	799.0
194.20	347.0
196.00	1230.0
198.00	47896.0
199.00	3059.0
199.70	188.0
200.00	164.0
201.30	395.0
201.70	298.0
202.50	205.0
203.00	524.0
204.00	2377.0
205.00	3486.0
206.00	14222.0
207.10	2106.0
208.00	665.0
208.90	177.0
210.10	209.0
211.00	848.0
216.00	211.0
217.00	3363.0
217.80	667.0
220.00	244.0
221.00	1464.0
222.00	483.0
223.00	655.0
224.00	7273.0
225.00	1533.0
226.20	320.0
227.00	3231.0
228.00	573.0
229.00	809.0
230.10	276.0
230.90	304.0
232.00	259.0
233.90	168.0
235.00	340.0
237.10	334.0
240.00	188.0
241.00	216.0
242.10	346.0
243.00	393.0
244.10	5412.0
245.00	603.0
246.00	1145.0
248.90	219.0
253.10	313.0
255.00	33872.0
256.00	5175.0
257.10	227.0
258.00	2523.0
258.90	372.0
259.80	151.0
265.00	450.0
266.00	233.0
272.00	153.0
273.20	816.0
274.00	2366.0
275.00	14345.0
276.10	2251.0
277.00	1237.0
278.00	180.0
283.00	241.0
292.90	368.0
294.00	190.0
295.00	170.0
296.00	3021.0
297.00	704.0
301.40	165.0
303.00	422.0
313.90	261.0
315.00	690.0
315.80	290.0
323.00	1529.0
323.90	389.0
326.90	235.0
332.90	272.0
334.10	1172.0
335.00	269.0
346.00	467.0
351.90	260.0
353.20	299.0
354.10	659.0
365.00	1743.0

Instrument :

BNA_G

ClientSampleId :

DFTPP68

365.90	156.0
371.00	206.0
372.00	838.0
373.10	171.0
389.90	187.0
391.90	187.0
402.00	208.0
403.10	710.0
403.90	158.0
421.10	570.0
422.00	336.0
423.00	2998.0
424.00	761.0
441.00	7335.0
442.00	49584.0
443.00	9985.0
444.10	803.0
524.40	156.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68

Scan 2463 (18.059 min): BG028411.D
DFTPP68

m/z	Abundance
40.10	1982.0
44.00	399.0
49.10	151.0
57.10	163.0
72.50	168.0
73.10	297.0
193.20	161.0
206.90	164.0

Instrument :
BNA_G
ClientSampleId :
DFTPP68