



2-Nitrophenol




$R = 2.416e-002 A^2 + 4.580e-002 A + 8.229e-003$
 Coef of Det (r^2) = 0.999208 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M
 Calibration Table Last Updated: Tue Aug 23 03:09:46 2022

Benzoic acid



$R = 2.167e-002 A^2 + 8.456e-002 A - 1.281e-002$
 Coef of Det (r^2) = 0.998568 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M
 Calibration Table Last Updated: Tue Aug 23 03:09:46 2022


2-Nitroaniline



Response = $3.081\text{e-}001 * \text{Amt} - 5.496\text{e-}002$
Coef of Det (r^2) = 0.994269 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M
Calibration Table Last Updated: Tue Aug 23 03:09:46 2022

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 4.021\text{e-}001 * \text{Amt} - 8.641\text{e-}002$$

Coef of Det (r^2) = 0.997498 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M

Calibration Table Last Updated: Tue Aug 23 03:09:46 2022


Pentachlorophenol



$\text{Response} = 1.482\text{e-}001 * \text{Amt} - 5.279\text{e-}002$
 Coef of Det (r^2) = 0.995529 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M
 Calibration Table Last Updated: Tue Aug 23 03:09:46 2022

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 7.724\text{e-}001 * \text{Amt} - 1.623\text{e-}001$$


Coef of Det (r^2) = 0.996734 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M

Calibration Table Last Updated: Tue Aug 23 03:09:46 2022

Butylbenzylphthalate

Response Ratio



$$\text{Response} = 5.196\text{e-}001 * \text{Amt} - 1.338\text{e-}001$$

Coef of Det (r^2) = 0.994548 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG082222.M

Calibration Table Last Updated: Tue Aug 23 03:09:46 2022