



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082224\
Data File : BG062535.D
Acq On : 22 Aug 2024 11:15
Operator : MA/JU
Sample : DFTPP417
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP417

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG081424.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Aug 20 23:01:24 2024



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.6	74589	PASS
68	69	0.00	2	0.5	278	PASS
69	198	0.00	100	33.2	59453	PASS
70	69	0.00	2	0.7	395	PASS
127	198	10	80	38.9	69664	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	179093	PASS
199	198	5	9	6.8	12155	PASS
275	198	10	60	27.1	48509	PASS
365	198	1	100	3.9	6913	PASS
441	198	0.01	100	11.2	20072	PASS
442	198	50	100	66.4	118893	PASS
443	442	15	24	19.1	22709	PASS

m/z	Abundance
41.15	173.0
43.15	339.0
44.05	955.0
83.95	200.0
167.15	188.0
207.15	295.0

Instrument :

BNA_G

ClientSampleId :

DFTPP417

m/z	Abundance
37.15	247.0
38.15	888.0
39.15	4688.0
40.05	298.0
41.25	315.0
44.15	920.0
45.05	211.0
49.25	578.0
50.15	16680.0
51.15	67008.0
52.15	2882.0
53.15	206.0
55.25	548.0
56.05	1629.0
57.15	4483.0
58.05	339.0
61.05	905.0
62.15	750.0
63.15	2484.0
63.95	361.0
65.15	1287.0
68.05	834.0
69.05	51200.0
70.15	469.0
73.05	554.0
74.15	4735.0
75.15	8560.0
76.15	2840.0
77.15	62880.0
78.15	4265.0
79.05	3866.0
80.05	2862.0
81.05	4302.0
82.05	926.0
83.15	814.0
84.05	205.0
85.05	899.0
86.05	1017.0
87.05	652.0
87.95	283.0
89.15	179.0
91.05	953.0
92.05	1109.0
93.05	7098.0
94.05	550.0
95.05	159.0
96.15	316.0
98.15	5492.0
99.05	3782.0
100.05	400.0
101.05	2750.0
101.95	190.0
103.15	664.0
104.05	1901.0
105.15	1416.0
106.15	450.0
107.05	18240.0
108.15	3029.0
110.05	30944.0
111.15	5120.0
112.05	866.0
112.95	200.0
115.05	163.0
116.15	784.0
117.05	15904.0
118.05	1017.0
119.15	247.0
120.15	171.0
121.95	1164.0
123.05	1767.0
124.05	764.0
125.05	756.0
127.15	60760.0
128.15	4896.0
129.05	23144.0
130.05	2284.0
131.15	603.0
132.25	317.0
134.15	813.0
135.05	2044.0
136.15	1039.0
137.15	666.0
138.25	181.0
140.15	433.0
141.05	3046.0
142.05	761.0
143.05	641.0
144.05	168.0
146.15	476.0
147.05	1923.0
148.05	3228.0
149.05	949.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

150.05	237.0
151.15	388.0
152.05	273.0
153.05	1147.0
154.05	736.0
155.15	1670.0
156.15	2731.0
157.05	493.0
158.05	589.0
159.05	614.0
160.05	982.0
161.15	1628.0
161.95	373.0
164.05	154.0
164.95	1118.0
166.05	808.0
167.15	6863.0
168.05	3680.0
169.05	737.0
170.05	331.0
171.05	302.0
172.05	458.0
173.05	895.0
174.15	1446.0
175.15	3043.0
176.05	677.0
177.05	1438.0
178.05	415.0
179.05	5246.0
180.15	3787.0
181.05	1648.0
182.05	291.0
185.15	2775.0
186.05	20152.0
187.15	5113.0
188.15	672.0
188.95	1076.0
190.05	381.0
191.05	622.0
192.05	1485.0
193.05	2439.0
194.05	551.0
195.15	347.0
196.05	3541.0
198.05	140864.0
199.05	8911.0
200.05	961.0
201.35	635.0
203.05	1512.0
204.15	4999.0
205.05	9732.0
206.15	37696.0
207.15	5786.0
208.05	1393.0
209.05	751.0
210.15	767.0
211.05	1592.0
212.95	156.0
215.05	474.0
216.15	925.0
217.05	9764.0
218.05	1441.0
219.05	169.0
221.05	5311.0
223.05	2181.0
224.15	21384.0
225.05	5547.0
226.15	647.0
227.05	10245.0
228.05	1415.0
229.05	2066.0
230.05	353.0
230.95	599.0
232.15	219.0
234.05	617.0
235.05	588.0
236.05	452.0
237.05	609.0
237.95	233.0
239.05	402.0
240.15	390.0
240.95	685.0
242.05	1088.0
243.15	1261.0
244.05	16052.0
245.15	2204.0
246.05	3218.0
247.05	827.0
247.65	204.0
249.05	573.0
253.05	308.0
255.05	84176.0
256.05	12980.0
257.05	1013.0
258.05	5750.0
259.05	897.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

265.05	2163.0
266.05	273.0
270.05	171.0
271.15	204.0
272.15	225.0
273.05	2685.0
274.05	6820.0
275.05	34048.0
276.05	4735.0
277.05	3345.0
278.05	470.0
278.95	267.0
282.95	335.0
284.15	345.0
285.15	446.0
289.15	187.0
291.95	154.0
293.05	706.0
293.95	275.0
296.05	10159.0
297.05	1730.0
297.95	159.0
300.95	176.0
301.95	262.0
303.05	1272.0
304.05	517.0
310.05	229.0
313.75	351.0
315.05	1199.0
316.05	815.0
321.05	406.0
321.95	197.0
323.05	3127.0
324.05	793.0
326.85	609.0
327.85	390.0
331.95	328.0
332.95	270.0
334.05	2297.0
335.25	514.0
341.15	409.0
345.85	710.0
352.05	1167.0
353.05	808.0
354.15	1038.0
355.05	210.0
365.05	4619.0
366.05	557.0
370.75	232.0
371.95	1541.0
372.95	542.0
383.05	704.0
383.95	158.0
401.95	810.0
403.15	798.0
404.05	362.0
420.95	833.0
421.95	850.0
423.05	4556.0
424.05	1133.0
441.05	13143.0
442.05	73224.0
443.05	13999.0
444.05	1288.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

m/z	Abundance
37.15	280.0
38.15	1134.0
39.15	6148.0
40.15	377.0
41.15	377.0
43.15	253.0
44.05	1169.0
45.15	307.0
49.15	532.0
50.15	22104.0
51.15	88064.0
52.15	4068.0
53.05	190.0
55.15	549.0
56.05	2316.0
57.15	5711.0
58.15	274.0
61.15	850.0
62.15	1090.0
63.15	3385.0
63.95	399.0
65.15	1526.0
69.05	73760.0
70.15	350.0
73.05	655.0
74.15	7272.0
75.05	10963.0
76.15	3754.0
77.15	81720.0
78.15	5723.0
79.15	5645.0
80.05	4190.0
81.15	5818.0
82.05	1672.0
83.05	1618.0
85.05	933.0
86.05	1535.0
87.05	798.0
89.15	155.0
91.05	1314.0
92.05	1413.0
93.05	9392.0
94.05	810.0
96.05	309.0
98.05	7084.0
99.15	5418.0
100.05	469.0
101.05	3465.0
103.15	1227.0
104.15	2240.0
105.05	2153.0
107.05	24176.0
108.15	3535.0
110.05	43096.0
111.15	6155.0
112.15	657.0
113.05	285.0
115.05	176.0
116.15	1293.0
117.05	21264.0
118.15	1497.0
120.05	280.0
121.15	178.0
122.15	1580.0
123.15	2539.0
124.05	944.0
125.05	1115.0
127.15	79080.0
128.15	6034.0
129.05	30520.0
130.05	2644.0
131.15	570.0
132.25	328.0
133.15	242.0
134.05	957.0
135.05	2755.0
136.05	852.0
137.15	1270.0
137.95	376.0
140.05	311.0
141.05	4009.0
142.05	1219.0
143.05	1144.0
144.15	226.0
145.05	217.0
146.05	699.0
147.15	2575.0
148.15	5452.0
149.15	1017.0
150.15	399.0
151.15	737.0
151.95	327.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

153.05	1636.0
154.05	1122.0
155.05	2352.0
156.15	3695.0
157.05	614.0
157.25	606.0
158.15	909.0
159.15	741.0
160.05	1148.0
161.15	1924.0
162.05	574.0
163.05	254.0
164.15	255.0
165.05	1980.0
166.05	1117.0
167.15	9110.0
168.05	5338.0
169.15	935.0
170.15	279.0
170.95	487.0
172.05	862.0
173.15	1384.0
174.05	2136.0
175.15	3437.0
176.05	923.0
177.05	1855.0
178.05	750.0
179.05	7390.0
180.05	4480.0
181.15	2331.0
182.05	480.0
184.05	660.0
185.05	4122.0
186.05	27928.0
187.05	8345.0
188.05	735.0
189.05	1508.0
190.05	468.0
191.05	1016.0
192.05	2404.0
193.05	3298.0
194.05	822.0
195.15	562.0
196.05	6141.0
198.05	208832.0
199.05	14127.0
200.15	1373.0
201.55	775.0
203.05	1906.0
204.05	8475.0
205.05	15824.0
206.15	58320.0
207.15	7760.0
208.05	1970.0
209.05	773.0
210.15	1023.0
211.05	1931.0
211.85	624.0
212.95	182.0
214.05	205.0
214.95	794.0
216.05	1478.0
217.05	16332.0
218.05	1981.0
219.05	223.0
221.05	7266.0
222.05	2124.0
223.05	3332.0
224.05	32968.0
225.05	8564.0
226.05	1098.0
227.05	14857.0
228.15	1978.0
229.05	3152.0
230.15	227.0
231.05	1261.0
232.05	153.0
232.95	190.0
233.95	1042.0
235.05	1221.0
236.15	588.0
237.05	1261.0
237.95	239.0
238.95	513.0
240.05	549.0
241.05	1065.0
242.05	1779.0
243.05	2034.0
244.15	25064.0
245.05	2949.0
246.05	5212.0
247.05	1089.0
248.15	190.0
249.05	1030.0
250.05	157.0
251.05	257.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

251.95	323.0
253.05	701.0
255.05	129408.0
256.05	19536.0
257.05	1946.0
258.05	8024.0
259.05	1320.0
259.95	241.0
260.95	289.0
264.05	220.0
265.05	3012.0
266.05	541.0
270.15	156.0
270.95	306.0
272.15	432.0
273.05	3907.0
274.05	10573.0
275.05	56096.0
276.05	7531.0
277.05	5259.0
278.05	890.0
282.25	155.0
283.05	694.0
283.95	323.0
284.95	962.0
285.95	198.0
286.15	195.0
289.15	182.0
289.85	163.0
291.15	240.0
292.05	299.0
293.05	1368.0
294.15	319.0
296.05	19120.0
297.05	2507.0
297.95	166.0
301.05	285.0
301.95	251.0
303.15	1944.0
304.05	573.0
307.85	204.0
309.05	212.0
310.05	253.0
314.05	752.0
315.05	1990.0
316.05	1276.0
321.15	516.0
321.95	326.0
323.15	5651.0
323.95	882.0
326.05	183.0
326.95	1095.0
328.05	627.0
329.05	161.0
331.95	358.0
333.15	464.0
334.05	3528.0
334.95	956.0
341.05	627.0
346.05	1280.0
346.95	295.0
352.15	1826.0
353.05	1165.0
354.05	1653.0
354.95	260.0
364.95	8236.0
365.95	1230.0
370.85	300.0
372.05	2442.0
373.15	579.0
382.95	864.0
390.15	434.0
391.05	226.0
391.95	224.0
401.05	208.0
401.95	1277.0
403.05	1603.0
404.05	599.0
420.95	1334.0
422.05	1222.0
423.05	7158.0
424.05	1896.0
424.85	173.0
441.05	21496.0
442.05	129280.0
443.05	25440.0
444.05	2631.0
445.05	163.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

m/z	Abundance
37.05	261.0
38.15	728.0
39.15	5192.0
40.15	367.0
41.05	256.0
43.15	318.0
44.05	1237.0
49.05	651.0
50.15	16912.0
51.15	68696.0
52.15	3143.0
53.05	275.0
55.05	522.0
56.05	2004.0
57.15	4272.0
58.25	345.0
61.15	1066.0
62.15	908.0
63.15	2245.0
64.15	398.0
65.05	1144.0
66.95	162.0
69.05	53400.0
70.15	366.0
73.15	513.0
74.15	5704.0
75.15	8727.0
76.25	2937.0
77.15	69000.0
78.15	4520.0
79.05	4135.0
80.05	3708.0
81.15	4976.0
82.15	1275.0
83.15	1156.0
84.05	272.0
85.05	825.0
86.05	1364.0
87.15	655.0
88.15	196.0
91.05	1062.0
92.05	1133.0
93.15	7660.0
94.15	378.0
95.25	184.0
95.95	356.0
97.15	159.0
98.15	5937.0
99.15	4667.0
100.05	373.0
101.05	2537.0
102.25	160.0
103.05	854.0
104.05	1712.0
105.15	1496.0
106.15	802.0
107.05	19864.0
108.05	3289.0
110.05	36208.0
111.15	5563.0
112.05	648.0
113.15	430.0
116.05	1280.0
117.05	17576.0
118.05	1295.0
118.95	151.0
120.05	229.0
122.05	1428.0
123.05	2154.0
124.05	810.0
125.05	990.0
127.15	69152.0
128.15	5146.0
129.05	29384.0
130.15	2347.0
131.05	699.0
133.15	171.0
133.95	770.0
135.15	2317.0
136.15	846.0
137.05	1186.0
138.05	342.0
140.05	373.0
141.15	4187.0
142.15	1320.0
142.95	735.0
144.05	363.0
144.95	175.0
146.05	894.0
147.15	2129.0
148.05	4778.0
149.05	982.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

149.95	258.0
151.25	731.0
151.95	406.0
153.05	1343.0
154.05	908.0
155.15	2139.0
156.15	3025.0
157.15	708.0
158.05	1110.0
159.05	564.0
160.15	1198.0
161.05	1830.0
162.15	487.0
163.15	165.0
163.95	225.0
165.05	1544.0
166.15	1328.0
167.05	8311.0
168.05	4157.0
169.15	791.0
169.95	298.0
171.05	348.0
172.05	821.0
173.05	959.0
174.05	1480.0
175.15	3191.0
176.15	1212.0
177.05	1566.0
178.05	752.0
179.05	6855.0
180.05	4596.0
181.05	1873.0
182.15	210.0
183.05	190.0
183.95	610.0
185.05	3458.0
186.15	24520.0
187.05	7392.0
188.05	568.0
189.15	1642.0
190.15	153.0
191.15	811.0
192.15	1953.0
193.15	2481.0
194.05	764.0
195.25	444.0
196.05	4925.0
198.05	187584.0
199.05	13427.0
200.05	926.0
201.55	763.0
202.25	291.0
203.05	1579.0
204.05	7052.0
205.05	12305.0
206.15	51592.0
207.15	7220.0
208.15	1830.0
208.95	822.0
210.15	980.0
211.15	2031.0
212.15	294.0
213.15	194.0
214.25	159.0
215.05	699.0
216.05	1373.0
217.05	14175.0
218.05	2086.0
219.15	221.0
220.15	329.0
221.15	6719.0
223.05	3462.0
224.15	29960.0
225.05	7540.0
225.95	665.0
227.05	13757.0
228.05	1887.0
229.05	2886.0
229.95	281.0
231.05	1124.0
231.95	231.0
232.85	257.0
234.05	673.0
235.05	1014.0
235.95	594.0
237.15	1029.0
238.95	513.0
240.05	588.0
241.05	987.0
242.15	1506.0
243.15	1691.0
244.05	24360.0
245.15	3404.0
246.05	5506.0
247.05	1170.0
247.85	263.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417

249.05	774.0
249.95	288.0
250.95	285.0
252.15	294.0
252.95	491.0
255.05	125016.0
256.05	17280.0
257.05	1428.0
258.05	7537.0
259.05	1451.0
259.95	286.0
261.15	223.0
264.15	304.0
265.05	3737.0
266.05	607.0
270.95	417.0
271.95	455.0
273.05	3498.0
274.05	9087.0
275.05	55384.0
276.05	8073.0
277.05	5155.0
278.05	984.0
281.85	230.0
283.15	573.0
283.95	454.0
285.05	1081.0
285.95	178.0
289.05	237.0
290.25	171.0
292.15	224.0
293.05	1177.0
294.05	243.0
296.05	17656.0
297.05	2665.0
297.95	257.0
300.95	277.0
301.95	363.0
303.05	2433.0
304.15	595.0
308.05	211.0
309.15	179.0
309.95	315.0
314.05	851.0
315.05	2332.0
316.05	1223.0
316.95	165.0
320.95	621.0
321.85	203.0
323.05	5575.0
324.05	1049.0
327.05	1118.0
328.05	510.0
329.15	252.0
332.05	469.0
333.05	568.0
334.05	3478.0
335.05	808.0
339.25	162.0
341.05	518.0
342.15	164.0
345.95	1309.0
346.95	276.0
352.05	2050.0
353.05	1032.0
354.05	2000.0
355.15	436.0
359.95	153.0
365.05	7885.0
365.95	1289.0
371.05	398.0
372.05	2623.0
373.05	885.0
382.95	739.0
383.95	357.0
389.95	355.0
391.15	264.0
400.95	178.0
402.05	1179.0
403.05	1880.0
403.85	709.0
404.95	200.0
421.05	1613.0
421.95	1336.0
423.05	9142.0
424.05	1788.0
425.25	215.0
441.05	25576.0
442.05	154176.0
443.05	28688.0
444.05	2891.0
445.15	178.0

Instrument :
BNA_G
ClientSampleId :
DFTPP417