


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082317\
Data File : BG028443.D
Acq On : 23 Aug 2017 9:44
Operator : SJ/JU
Sample : DFTPP69
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP69

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 24 14:06:18 2017



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.4	27426	PASS
68	69	0.00	2	1.2	330	PASS
69	198	0.00	100	35.2	26480	PASS
70	69	0.00	2	0.5	143	PASS
127	198	10	80	41.3	31133	PASS
197	198	0.00	2	0.4	330	PASS
198	198	100	100	100.0	75325	PASS
199	198	5	9	6.2	4699	PASS
275	198	10	60	30.4	22885	PASS
365	198	1	100	3.9	2958	PASS
441	442	0.01	100	16.7	11710	PASS
442	198	50	100	93.3	70253	PASS
443	442	15	24	19.3	13533	PASS

m/z	Abundance
38.10	250.0
39.10	1548.0
40.00	2095.0
41.10	166.0
43.20	236.0
44.00	658.0
49.00	267.0
50.10	7153.0
51.10	30624.0
52.10	1432.0
55.10	219.0
56.00	774.0
57.10	1843.0
58.20	242.0
59.20	229.0
60.90	300.0
61.10	301.0
62.00	363.0
63.00	1545.0
65.10	660.0
67.10	300.0
68.00	478.0
69.00	29416.0
73.10	401.0
74.10	2356.0
75.10	3607.0
76.10	1906.0
77.10	29600.0
78.10	2519.0
79.10	1857.0
80.00	1574.0
81.00	2606.0
82.00	504.0
83.10	572.0
84.00	264.0
85.00	553.0
86.10	921.0
87.00	241.0
90.90	353.0
92.00	959.0
93.10	3215.0
94.00	347.0
95.10	230.0
98.00	2683.0
99.00	1387.0
101.00	1306.0
103.00	451.0
104.10	644.0
105.00	692.0
107.10	9510.0
108.00	1530.0
109.00	478.0
110.00	17656.0
111.00	2501.0
112.00	612.0
113.00	198.0
116.10	618.0
117.00	6423.0
118.00	590.0
119.00	177.0
120.00	226.0
122.00	600.0
123.00	1045.0
124.00	513.0
125.10	428.0
127.00	33136.0
128.00	2526.0
129.00	12969.0
130.00	1096.0
131.10	358.0
134.10	551.0
135.00	1051.0
136.00	577.0
137.00	686.0
138.20	230.0
140.00	166.0
141.00	1925.0
142.10	486.0
143.00	712.0
146.00	400.0
147.10	951.0
148.00	1614.0
149.00	567.0
151.10	267.0
152.00	185.0
152.90	376.0
154.00	356.0
155.00	779.0
156.00	1714.0
157.00	412.0
158.00	288.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

160.00	846.0
161.00	945.0
162.10	336.0
164.00	255.0
165.10	859.0
166.00	782.0
167.00	2937.0
168.00	1575.0
168.90	243.0
171.30	292.0
172.00	229.0
173.00	840.0
174.00	756.0
175.10	1529.0
176.00	489.0
177.00	511.0
178.00	401.0
179.00	2526.0
180.00	2488.0
181.00	1011.0
182.20	205.0
183.20	291.0
184.00	374.0
185.00	1468.0
186.00	9671.0
187.00	3580.0
187.90	321.0
188.90	715.0
191.00	335.0
192.00	1093.0
193.00	961.0
194.00	319.0
194.60	163.0
196.00	1895.0
196.90	992.0
198.00	70584.0
199.00	4759.0
200.00	346.0
201.70	269.0
203.10	673.0
204.00	3045.0
205.00	4968.0
206.00	21160.0
207.10	2881.0
208.00	1310.0
209.00	245.0
210.60	466.0
211.00	365.0
212.10	355.0
212.90	153.0
214.90	331.0
215.90	572.0
217.00	5808.0
217.90	524.0
219.80	247.0
220.10	208.0
221.00	2149.0
222.00	1069.0
223.00	1280.0
224.00	12596.0
225.00	3223.0
226.00	349.0
227.00	4386.0
228.00	570.0
229.00	1179.0
231.00	400.0
233.00	271.0
234.10	298.0
234.90	347.0
237.00	180.0
238.10	302.0
242.10	558.0
243.10	542.0
244.00	7537.0
245.10	1518.0
246.00	1957.0
247.00	333.0
248.90	366.0
252.70	372.0
255.00	49928.0
256.00	7494.0
257.00	559.0
258.00	2875.0
259.00	325.0
264.90	1161.0
266.00	259.0
271.00	253.0
271.90	218.0
273.10	1480.0
274.00	2780.0
275.00	20704.0
276.00	3355.0
277.00	2171.0
277.90	363.0
280.00	153.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

282.90	254.0
284.10	192.0
285.00	408.0
289.10	152.0
291.90	209.0
293.00	232.0
294.00	507.0
295.10	183.0
296.00	5094.0
297.00	601.0
301.10	171.0
303.00	676.0
304.20	376.0
313.00	162.0
314.20	199.0
314.90	391.0
316.10	174.0
323.00	2079.0
324.10	551.0
326.80	408.0
327.80	295.0
332.90	172.0
334.00	1104.0
334.80	380.0
340.90	187.0
346.00	400.0
351.90	322.0
353.10	383.0
354.00	727.0
355.00	178.0
358.90	177.0
362.40	209.0
365.00	2487.0
365.90	274.0
372.00	791.0
373.10	302.0
373.90	199.0
383.00	383.0
389.90	161.0
391.00	263.0
402.10	497.0
403.00	495.0
403.90	246.0
421.00	540.0
422.10	547.0
423.00	2787.0
423.90	383.0
441.00	9216.0
442.00	53432.0
443.00	10961.0
444.00	1017.0
445.00	350.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

m/z	Abundance
38.00	336.0
39.10	1607.0
40.10	2357.0
41.00	174.0
44.10	323.0
49.10	299.0
50.10	7633.0
51.10	32280.0
52.10	1338.0
55.00	385.0
56.10	796.0
57.00	2445.0
58.10	260.0
61.00	538.0
62.10	596.0
63.10	1139.0
64.10	164.0
65.10	391.0
69.00	30880.0
70.20	239.0
72.80	360.0
74.10	3348.0
75.10	4479.0
76.10	1851.0
77.10	36504.0
78.10	2280.0
79.00	1922.0
80.10	1440.0
81.00	2968.0
82.00	899.0
83.10	1013.0
83.80	213.0
85.10	305.0
86.00	686.0
87.00	402.0
88.10	268.0
91.00	973.0
91.90	743.0
93.00	3602.0
93.90	222.0
95.00	161.0
96.10	248.0
97.00	151.0
98.10	3250.0
99.00	2695.0
100.00	497.0
101.00	1810.0
103.10	640.0
104.00	1220.0
105.10	1149.0
107.00	11257.0
108.00	1905.0
109.10	328.0
110.00	17768.0
111.00	2877.0
112.00	411.0
113.10	213.0
116.10	461.0
117.00	7479.0
118.10	875.0
119.10	223.0
120.00	173.0
122.00	644.0
123.10	812.0
124.10	696.0
125.00	533.0
127.00	35528.0
128.10	3045.0
129.00	14969.0
130.00	1332.0
131.00	191.0
133.00	153.0
134.00	760.0
135.00	1274.0
136.00	351.0
137.10	433.0
137.90	238.0
140.00	373.0
141.10	2085.0
142.00	755.0
143.00	480.0
144.00	262.0
146.00	261.0
147.00	1074.0
148.00	2909.0
149.00	384.0
150.00	162.0
151.10	514.0
152.00	207.0
153.00	631.0
154.00	312.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

155.10	1354.0
156.10	1566.0
156.90	232.0
157.90	645.0
158.90	315.0
160.00	822.0
161.10	833.0
161.80	222.0
162.10	223.0
165.00	915.0
166.00	672.0
167.10	3941.0
168.00	1477.0
169.00	453.0
170.00	193.0
170.90	247.0
172.00	287.0
173.00	481.0
174.00	1211.0
175.00	1761.0
176.00	435.0
177.00	1210.0
177.90	264.0
179.00	3707.0
180.10	2391.0
181.00	831.0
182.00	310.0
184.10	366.0
185.10	1581.0
186.00	13794.0
187.00	3538.0
188.00	477.0
189.10	832.0
190.00	220.0
190.90	301.0
192.00	1102.0
193.00	1620.0
194.00	375.0
195.10	233.0
196.00	2444.0
198.00	91664.0
198.90	5663.0
200.00	413.0
201.50	741.0
202.20	190.0
203.00	858.0
204.00	3612.0
205.00	6456.0
206.00	26576.0
207.10	3948.0
208.00	810.0
209.00	532.0
210.10	495.0
211.00	1167.0
215.00	322.0
216.00	693.0
217.00	6647.0
217.90	837.0
220.00	180.0
221.00	3840.0
221.90	890.0
222.90	1063.0
224.00	14694.0
225.00	4000.0
226.00	566.0
227.00	6704.0
228.00	966.0
229.00	1101.0
230.00	176.0
231.10	413.0
232.00	154.0
234.00	244.0
235.00	264.0
235.90	248.0
237.10	369.0
237.90	164.0
239.00	446.0
240.00	261.0
241.00	330.0
242.00	776.0
243.10	908.0
244.00	10440.0
245.10	1608.0
245.90	2260.0
247.00	502.0
247.80	173.0
248.90	657.0
251.10	393.0
252.30	197.0
253.20	537.0
255.00	61256.0
256.00	9382.0
257.00	585.0
258.00	3160.0
258.90	419.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

264.10	203.0
265.00	1559.0
265.90	246.0
269.80	151.0
271.00	209.0
271.90	251.0
273.10	1651.0
274.00	4233.0
275.00	27608.0
275.90	3909.0
277.00	2680.0
277.90	272.0
280.90	245.0
283.00	225.0
285.00	385.0
293.10	479.0
294.00	309.0
294.90	290.0
296.00	8622.0
297.00	1023.0
302.20	162.0
303.00	1034.0
304.10	326.0
307.00	171.0
314.00	372.0
315.00	1066.0
316.00	755.0
319.80	178.0
321.10	232.0
323.00	2716.0
324.10	680.0
327.00	268.0
328.00	548.0
331.70	221.0
333.00	337.0
334.00	1099.0
334.90	295.0
341.00	433.0
346.00	514.0
346.80	157.0
352.00	968.0
353.00	481.0
354.10	743.0
354.80	234.0
365.00	3484.0
366.00	464.0
370.00	168.0
370.70	233.0
372.00	1370.0
372.90	328.0
382.80	222.0
391.10	226.0
401.90	310.0
403.00	911.0
404.20	327.0
421.00	509.0
421.90	398.0
423.00	4241.0
424.00	1180.0
425.00	164.0
432.90	168.0
437.20	160.0
441.00	13738.0
442.00	82408.0
443.00	14804.0
444.00	1528.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

m/z	Abundance
39.10	1382.0
40.10	2484.0
41.10	242.0
43.10	330.0
44.00	620.0
49.30	390.0
50.10	4122.0
51.10	19376.0
52.10	953.0
55.00	405.0
56.10	663.0
57.00	2018.0
61.00	354.0
62.00	391.0
63.10	1007.0
65.20	517.0
68.00	512.0
69.00	19144.0
70.20	191.0
72.10	158.0
73.10	359.0
74.10	1810.0
75.10	2963.0
77.10	21848.0
78.10	1067.0
79.10	2067.0
80.00	1740.0
81.00	1442.0
82.00	535.0
83.10	642.0
84.10	265.0
85.00	612.0
86.10	536.0
87.20	165.0
88.10	243.0
91.00	407.0
92.00	364.0
93.00	2630.0
95.00	277.0
95.90	270.0
98.00	2071.0
99.00	1658.0
99.90	365.0
101.00	919.0
103.10	407.0
103.90	779.0
105.10	549.0
107.00	6408.0
108.10	1444.0
110.00	13849.0
111.00	1789.0
111.80	350.0
115.00	165.0
116.00	420.0
117.00	5595.0
117.80	221.0
122.00	487.0
123.00	881.0
124.00	197.0
125.10	502.0
127.00	24736.0
128.00	2511.0
129.00	9997.0
130.00	1136.0
131.10	460.0
133.00	191.0
134.00	361.0
135.00	775.0
135.90	275.0
140.00	323.0
141.00	1814.0
142.00	401.0
142.80	182.0
143.10	159.0
145.00	220.0
145.90	337.0
147.00	579.0
148.00	1575.0
149.00	686.0
151.70	177.0
153.00	417.0
154.10	253.0
155.00	616.0
156.00	1156.0
157.80	241.0
159.00	273.0
160.00	402.0
161.10	697.0
165.00	647.0
166.00	528.0
167.00	2865.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

167.90	1077.0
169.00	204.0
170.00	287.0
171.10	453.0
172.00	509.0
172.90	219.0
173.20	254.0
174.00	945.0
175.10	1169.0
176.00	249.0
177.00	694.0
178.00	433.0
179.00	1542.0
180.00	1495.0
180.90	977.0
183.00	244.0
183.80	248.0
184.30	160.0
185.10	730.0
186.00	9356.0
187.10	2625.0
188.00	455.0
188.90	791.0
190.80	251.0
192.10	677.0
193.00	1088.0
194.10	309.0
194.80	197.0
196.10	2055.0
198.00	63728.0
199.00	3675.0
200.00	396.0
201.00	437.0
201.50	295.0
203.00	630.0
204.00	2281.0
205.10	5408.0
206.10	18832.0
207.00	2781.0
208.00	945.0
209.00	248.0
210.10	344.0
210.90	539.0
211.40	483.0
212.00	198.0
213.10	254.0
215.00	263.0
217.00	4788.0
218.00	435.0
221.10	2572.0
223.00	1647.0
224.00	11500.0
225.00	2407.0
226.00	450.0
227.00	5510.0
228.00	594.0
228.90	758.0
231.00	236.0
233.80	314.0
235.00	299.0
235.80	164.0
237.00	297.0
239.00	244.0
239.70	165.0
240.10	192.0
241.00	298.0
241.90	852.0
243.00	1221.0
244.00	7801.0
245.00	970.0
246.00	1494.0
246.90	455.0
248.00	186.0
249.00	340.0
250.80	162.0
252.10	183.0
252.90	372.0
255.00	47976.0
256.00	6122.0
257.10	378.0
258.00	2449.0
258.90	398.0
261.10	260.0
265.00	1405.0
271.20	168.0
271.90	221.0
273.00	1119.0
274.00	4054.0
275.00	20344.0
276.00	2452.0
277.00	1990.0
277.90	365.0
282.90	162.0
284.10	175.0
285.00	763.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

293.00	642.0
294.00	177.0
296.00	5902.0
297.10	827.0
301.10	150.0
302.10	207.0
303.00	494.0
303.80	311.0
309.20	150.0
313.90	320.0
315.10	388.0
316.10	356.0
320.80	215.0
321.20	206.0
323.00	2315.0
324.10	397.0
326.20	430.0
326.90	360.0
332.90	381.0
334.00	1494.0
335.00	325.0
335.90	160.0
339.00	224.0
341.00	258.0
342.00	210.0
345.90	585.0
346.90	184.0
351.20	167.0
352.10	749.0
353.00	349.0
354.00	869.0
359.00	152.0
365.00	2903.0
365.90	302.0
372.00	938.0
373.00	265.0
383.00	333.0
391.00	196.0
402.00	222.0
403.00	805.0
403.90	252.0
421.00	610.0
422.00	370.0
423.00	3970.0
424.00	1021.0
425.10	249.0
441.00	12178.0
442.00	74920.0
443.00	14836.0
444.00	1625.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69

m/z	Abundance
40.10	2362.0
44.10	504.0
49.00	190.0
55.10	252.0
57.10	281.0
63.20	175.0
73.10	171.0
77.00	153.0
97.30	169.0
147.90	171.0
149.30	171.0
192.90	192.0
207.00	541.0
261.90	204.0
280.90	157.0
459.30	192.0

Instrument :
BNA_G
ClientSampleId :
DFTPP69