


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082417\
 Data File : BG028455.D
 Acq On : 24 Aug 2017 10:48
 Operator : SJ/JU
 Sample : DFTPP70
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP70

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 25 03:52:53 2017



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	15858	PASS
68	69	0.00	2	1.6	259	PASS
69	198	0.00	100	34.8	15838	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	38.7	17634	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	45536	PASS
199	198	5	9	7.0	3177	PASS
275	198	10	60	29.3	13344	PASS
365	198	1	100	4.1	1879	PASS
441	442	0.01	100	15.3	6206	PASS
442	198	50	100	88.8	40434	PASS
443	442	15	24	21.0	8509	PASS

m/z	Abundance
39.10	1274.0
40.00	3527.0
41.10	191.0
43.10	709.0
44.10	979.0
45.00	190.0
49.00	161.0
50.10	4200.0
51.10	17304.0
52.10	986.0
55.20	404.0
56.10	793.0
57.10	1174.0
61.00	266.0
62.10	199.0
63.10	590.0
65.00	437.0
66.90	201.0
69.00	15682.0
70.10	343.0
71.20	365.0
73.00	222.0
74.10	1717.0
75.10	3273.0
76.10	657.0
77.10	17880.0
78.10	1087.0
79.10	1406.0
80.00	1037.0
81.00	1583.0
82.00	425.0
83.10	637.0
84.00	168.0
85.00	447.0
85.90	305.0
88.10	247.0
91.10	616.0
92.10	222.0
93.10	2375.0
94.10	181.0
95.20	288.0
96.10	225.0
98.00	1599.0
99.00	1088.0
100.90	576.0
103.00	185.0
104.00	277.0
105.10	631.0
106.00	202.0
107.00	4780.0
108.00	920.0
109.00	254.0
110.00	9982.0
111.10	2422.0
112.00	232.0
116.20	319.0
117.00	4168.0
118.00	422.0
118.60	186.0
119.20	192.0
122.00	425.0
122.20	416.0
123.10	514.0
124.00	194.0
125.10	385.0
125.70	164.0
126.20	386.0
127.10	16888.0
128.10	1597.0
129.00	7168.0
130.00	1032.0
132.20	198.0
134.00	298.0
135.00	586.0
136.00	271.0
136.80	345.0
139.20	179.0
140.10	151.0
141.00	1088.0
142.20	650.0
143.10	330.0
147.00	699.0
148.00	1108.0
149.00	308.0
150.00	253.0
150.40	217.0
151.20	206.0
153.00	298.0
153.90	228.0
155.00	440.0
156.00	878.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

158.00	305.0
159.10	356.0
160.00	345.0
161.00	561.0
163.00	171.0
165.10	459.0
166.10	405.0
167.00	1862.0
168.00	749.0
169.00	397.0
170.60	239.0
171.10	262.0
171.90	224.0
173.10	355.0
174.00	486.0
175.00	665.0
176.00	435.0
177.00	620.0
177.90	189.0
179.00	1645.0
180.10	779.0
181.00	749.0
182.00	208.0
183.80	182.0
185.10	903.0
186.00	5684.0
187.10	1874.0
188.20	156.0
189.10	661.0
191.10	475.0
192.00	380.0
192.90	477.0
194.00	255.0
194.80	225.0
195.30	238.0
196.00	1126.0
198.00	40816.0
199.00	2904.0
199.80	170.0
202.90	307.0
204.10	2009.0
205.00	3185.0
206.00	11965.0
207.00	2287.0
208.00	347.0
210.10	253.0
210.70	335.0
211.20	561.0
216.00	320.0
217.00	3815.0
218.00	316.0
221.00	1968.0
221.80	388.0
223.00	847.0
224.00	7746.0
225.00	1801.0
226.10	158.0
227.00	3313.0
227.80	570.0
228.80	565.0
231.10	399.0
232.90	193.0
234.10	353.0
235.00	400.0
236.20	169.0
237.10	245.0
239.10	245.0
241.80	219.0
243.20	339.0
244.00	5463.0
245.00	804.0
246.00	886.0
247.00	279.0
249.10	321.0
253.10	262.0
255.00	27472.0
256.00	4067.0
256.90	307.0
258.00	1703.0
259.00	264.0
265.00	526.0
266.00	222.0
271.10	198.0
272.90	944.0
274.00	1688.0
275.00	11336.0
276.00	1982.0
277.00	1275.0
278.00	157.0
284.10	163.0
284.90	177.0
293.00	269.0
296.00	3777.0
296.90	397.0
303.10	416.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

304.00	283.0
307.70	171.0
314.20	162.0
314.90	269.0
316.00	468.0
323.00	978.0
327.90	350.0
333.10	153.0
334.00	854.0
335.00	154.0
350.20	158.0
351.70	157.0
352.90	501.0
354.00	337.0
365.00	1374.0
366.10	170.0
372.00	525.0
392.20	170.0
401.90	394.0
403.10	533.0
415.10	176.0
420.80	296.0
421.80	193.0
423.00	1772.0
424.10	591.0
435.30	184.0
441.10	4706.0
442.00	30744.0
443.00	6866.0
444.00	1083.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

m/z	Abundance
38.10	159.0
39.10	1684.0
40.00	3098.0
41.10	373.0
42.20	159.0
43.10	590.0
44.10	1142.0
45.00	235.0
50.00	4509.0
51.10	18544.0
52.00	1379.0
53.10	237.0
55.00	472.0
56.00	584.0
57.10	2121.0
61.00	314.0
62.00	376.0
63.00	822.0
65.00	356.0
68.10	513.0
69.00	19800.0
70.10	249.0
71.10	271.0
73.20	234.0
74.00	2358.0
75.10	3588.0
76.10	1027.0
77.10	20200.0
78.10	933.0
79.00	1706.0
80.00	1098.0
81.00	1567.0
81.90	384.0
83.10	667.0
84.20	180.0
85.00	482.0
85.90	443.0
87.00	201.0
91.10	280.0
92.10	612.0
93.00	2227.0
96.00	227.0
97.10	244.0
98.00	1806.0
99.00	1096.0
100.10	209.0
100.90	805.0
102.70	208.0
102.90	205.0
104.10	599.0
105.10	406.0
105.80	175.0
107.10	7017.0
108.10	912.0
109.10	469.0
110.00	10827.0
111.00	1275.0
112.00	336.0
115.20	181.0
116.00	630.0
117.00	4036.0
118.00	385.0
119.00	169.0
122.10	445.0
123.00	985.0
124.20	224.0
125.00	384.0
127.00	19872.0
128.00	2438.0
129.00	8373.0
130.00	934.0
134.00	362.0
135.00	911.0
136.10	435.0
137.00	534.0
139.90	199.0
141.10	1100.0
141.90	361.0
143.00	428.0
144.20	273.0
144.80	222.0
146.10	274.0
147.00	540.0
148.00	1201.0
149.10	564.0
149.90	184.0
151.10	226.0
153.20	585.0
154.10	365.0
155.10	1007.0
156.10	1507.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

156.80	257.0
157.80	192.0
158.10	229.0
159.00	212.0
160.00	547.0
161.00	761.0
162.00	184.0
164.20	208.0
165.10	673.0
165.90	493.0
167.10	2662.0
168.00	1297.0
170.20	201.0
171.10	311.0
172.00	359.0
173.00	478.0
174.10	705.0
175.10	1259.0
176.00	285.0
177.00	253.0
178.00	222.0
179.00	1832.0
180.00	1249.0
181.00	344.0
182.00	264.0
183.10	296.0
183.90	341.0
185.00	822.0
186.00	7447.0
187.00	1832.0
188.80	393.0
189.10	485.0
191.00	179.0
192.10	671.0
193.00	914.0
193.80	306.0
194.90	171.0
196.00	1426.0
196.60	516.0
198.00	52312.0
199.00	3841.0
200.00	231.0
201.40	222.0
201.90	277.0
203.00	705.0
204.10	2773.0
205.00	3993.0
206.10	13545.0
207.00	2118.0
208.10	590.0
209.00	262.0
209.90	293.0
211.00	679.0
214.80	331.0
216.00	491.0
217.00	4439.0
217.90	473.0
220.00	235.0
221.00	2718.0
222.00	542.0
223.10	906.0
224.10	7971.0
225.00	2360.0
226.10	154.0
227.00	3862.0
228.10	418.0
228.90	765.0
230.90	467.0
233.90	227.0
235.10	383.0
236.00	305.0
236.90	458.0
241.10	267.0
241.40	240.0
242.00	293.0
243.00	377.0
244.00	6706.0
245.00	940.0
246.00	1493.0
247.00	387.0
249.10	150.0
253.10	284.0
255.00	35424.0
256.00	5530.0
257.00	445.0
258.00	1765.0
259.00	408.0
265.00	698.0
266.00	191.0
273.00	1140.0
274.00	2908.0
275.00	14100.0
276.00	2376.0
277.00	1164.0
278.30	190.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

282.30	206.0
283.10	355.0
285.00	353.0
290.10	152.0
291.90	221.0
293.10	302.0
294.00	158.0
296.00	4558.0
297.10	882.0
303.00	819.0
303.90	398.0
314.00	362.0
315.00	336.0
316.10	235.0
323.00	1063.0
333.00	173.0
334.10	1126.0
335.00	356.0
341.00	190.0
346.10	254.0
347.10	207.0
351.90	250.0
353.10	276.0
354.00	456.0
355.00	171.0
365.00	2140.0
365.80	232.0
371.10	179.0
372.10	951.0
372.90	182.0
383.00	381.0
391.00	261.0
401.00	245.0
401.70	167.0
402.10	160.0
402.90	381.0
403.90	159.0
421.00	431.0
422.10	242.0
423.00	2820.0
424.00	480.0
441.00	7209.0
442.00	45656.0
443.10	8929.0
444.10	961.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

m/z	Abundance
39.10	845.0
40.10	3591.0
41.10	426.0
42.10	150.0
43.10	686.0
44.10	446.0
49.00	305.0
50.10	3073.0
51.10	12179.0
52.10	731.0
55.20	470.0
56.00	681.0
57.00	1041.0
58.70	200.0
59.10	167.0
60.10	217.0
61.10	325.0
62.20	209.0
63.10	611.0
65.10	192.0
67.10	216.0
68.00	264.0
69.00	12903.0
71.10	228.0
73.20	468.0
74.00	1184.0
75.10	2181.0
76.00	920.0
77.10	13986.0
78.10	933.0
79.00	951.0
80.10	716.0
81.10	1592.0
82.10	432.0
83.10	559.0
84.20	156.0
85.10	443.0
86.00	246.0
88.10	165.0
91.10	518.0
92.10	462.0
93.10	1983.0
94.00	385.0
95.00	168.0
96.10	167.0
97.10	340.0
98.10	1499.0
99.00	1012.0
100.20	151.0
101.10	701.0
103.50	189.0
104.00	296.0
104.90	400.0
107.00	4158.0
108.10	889.0
110.00	8821.0
111.00	1369.0
115.80	231.0
117.00	3529.0
118.00	401.0
119.00	206.0
120.20	192.0
121.10	164.0
121.80	172.0
123.00	974.0
124.10	342.0
125.00	431.0
127.00	16144.0
128.10	1731.0
129.00	6042.0
130.00	806.0
131.10	297.0
133.00	215.0
135.00	741.0
136.10	298.0
137.10	500.0
139.90	258.0
141.00	939.0
142.00	369.0
143.20	359.0
144.00	190.0
145.90	353.0
147.00	618.0
147.90	1001.0
148.90	256.0
151.10	169.0
152.00	170.0
152.90	482.0
154.10	211.0
155.00	451.0
156.00	1277.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

157.10	151.0
157.60	247.0
158.10	190.0
160.10	291.0
161.10	759.0
162.00	165.0
165.00	658.0
166.00	363.0
166.90	1321.0
168.00	665.0
168.90	159.0
172.00	362.0
172.90	219.0
173.90	652.0
175.10	870.0
175.90	179.0
176.90	440.0
177.90	400.0
179.00	1491.0
180.00	1069.0
180.90	510.0
184.10	224.0
185.00	689.0
186.00	6462.0
187.00	1946.0
189.10	275.0
191.00	164.0
191.90	629.0
193.10	916.0
194.00	179.0
196.00	943.0
198.00	43480.0
198.90	2786.0
200.70	169.0
201.50	441.0
202.90	255.0
204.00	1821.0
205.00	3552.0
206.00	13097.0
207.10	2453.0
209.10	418.0
210.20	345.0
211.20	395.0
215.00	216.0
216.00	187.0
217.00	3036.0
218.00	433.0
220.90	1080.0
221.90	241.0
222.90	700.0
224.00	7388.0
225.00	1653.0
227.00	2397.0
227.90	561.0
229.00	749.0
230.90	434.0
234.90	329.0
236.90	281.0
238.90	213.0
240.00	330.0
241.10	230.0
242.00	365.0
243.00	373.0
244.10	4487.0
245.10	581.0
246.00	852.0
247.00	243.0
253.30	422.0
255.00	33120.0
256.00	4381.0
257.10	600.0
258.00	2184.0
259.00	366.0
261.10	182.0
265.00	610.0
272.00	154.0
272.90	1308.0
274.00	1620.0
275.00	14598.0
276.00	1494.0
277.10	1260.0
281.30	294.0
283.90	194.0
284.90	150.0
286.30	196.0
289.20	153.0
292.80	255.0
295.00	171.0
296.00	4292.0
297.00	309.0
301.10	180.0
303.00	311.0
314.10	160.0
314.80	473.0
316.00	248.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

321.10	164.0
322.00	156.0
323.00	1420.0
323.90	229.0
325.10	222.0
327.10	358.0
328.00	212.0
334.10	807.0
335.00	359.0
339.00	163.0
341.10	294.0
345.90	386.0
352.10	213.0
353.10	235.0
353.90	538.0
357.20	163.0
365.00	2124.0
366.10	184.0
372.10	799.0
402.10	433.0
403.00	633.0
422.00	390.0
423.10	2491.0
424.00	985.0
425.10	161.0
440.00	194.0
441.00	6705.0
442.00	44904.0
443.00	9733.0
443.90	1479.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70

m/z	Abundance
40.00	3310.0
41.20	345.0
42.00	325.0
43.10	505.0
44.00	805.0
49.10	152.0
50.10	162.0
51.00	151.0
54.10	252.0
55.10	278.0
57.10	470.0
60.10	395.0
61.10	379.0
69.10	290.0
70.20	282.0
71.10	306.0
73.10	285.0
81.10	408.0
83.10	161.0
86.00	171.0
93.20	213.0
95.10	168.0
107.00	197.0
135.10	293.0
149.10	152.0
155.10	153.0
168.00	173.0
185.30	231.0
189.00	231.0
191.10	268.0
192.10	463.0
193.20	156.0
195.30	239.0
207.00	418.0
209.10	151.0
227.20	166.0
249.10	172.0
479.40	150.0

Instrument :
BNA_G
ClientSampleId :
DFTPP70