

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082516\
 Data File : BG023906.D
 Acq On : 25 Aug 2016 22:58
 Operator : UM/SJ
 Sample : H4588-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-04

Quant Time: Aug 26 01:45:08 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG080116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 24 17:42:04 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	123854	20.00	ng	0.00
21) Naphthalene-d8	10.93	136	542247	20.00	ng	0.00
38) Acenaphthene-d10	14.76	164	343463	20.00	ng	0.00
63) Phenanthrene-d10	17.53	188	759780	20.00	ng	0.00
75) Chrysene-d12	21.85	240	795028	20.00	ng	0.00
86) Perylene-d12	25.24	264	793561	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	386368	52.51	ng	0.00
7) Phenol-d6	7.25	99	368765	32.18	ng	0.00
23) Nitrobenzene-d5	9.27	82	839618	76.98	ng	0.00
41) 2,4,6-Tribromophenol	16.27	330	441041	87.79	ng	0.00
44) 2-Fluorobiphenyl	13.38	172	1565524	76.88	ng	0.00
78) Terphenyl-d14	20.14	244	1955359	74.63	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.20	163	211218	8.02	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082516\
Data File : BG023906.D
Acq On : 25 Aug 2016 22:58
Operator : UM/SJ
Sample : H4588-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MW-04

Quant Time: Aug 26 01:45:08 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG080116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 24 17:42:04 2016
Response via : Initial Calibration

