


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082517\
 Data File : BG028483.D
 Acq On : 25 Aug 2017 8:32
 Operator : SJ/JU
 Sample : DFTPP71
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP71

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 28 14:09:46 2017



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2461

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	21608	PASS
68	69	0.00	2	1.6	327	PASS
69	198	0.00	100	32.5	20393	PASS
70	69	0.00	2	0.6	130	PASS
127	198	10	80	38.6	24242	PASS
197	198	0.00	2	0.5	283	PASS
198	198	100	100	100.0	62800	PASS
199	198	5	9	6.8	4240	PASS
275	198	10	60	27.7	17381	PASS
365	198	1	100	4.0	2524	PASS
441	442	0.01	100	15.8	8451	PASS
442	198	50	100	84.9	53333	PASS
443	442	15	24	19.3	10267	PASS

m/z	Abundance
38.00	296.0
39.10	1703.0
40.10	1897.0
41.00	221.0
43.00	180.0
44.10	435.0
46.80	210.0
49.00	238.0
50.10	5035.0
51.10	22352.0
52.20	1357.0
55.20	517.0
56.10	849.0
57.00	1188.0
61.00	437.0
62.00	265.0
63.20	1009.0
65.10	449.0
66.90	187.0
68.10	583.0
69.00	20768.0
73.10	475.0
74.10	2425.0
75.10	3851.0
76.00	1379.0
77.10	23120.0
78.10	1342.0
79.00	1599.0
80.10	1261.0
81.00	2351.0
82.00	466.0
83.10	334.0
84.00	211.0
85.00	214.0
86.00	626.0
87.00	283.0
91.10	693.0
92.00	436.0
93.10	2495.0
93.90	220.0
98.10	1957.0
99.10	1065.0
100.10	174.0
101.00	695.0
103.00	377.0
104.00	1011.0
105.10	467.0
106.10	268.0
107.00	6827.0
108.10	1085.0
110.00	11453.0
111.00	2105.0
112.00	176.0
114.40	198.0
116.10	222.0
117.10	4664.0
118.10	493.0
121.90	385.0
123.00	922.0
124.00	349.0
125.00	671.0
127.00	21896.0
128.10	1409.0
129.00	8557.0
130.00	927.0
134.10	331.0
135.00	802.0
136.00	384.0
137.10	408.0
141.00	1651.0
142.10	712.0
143.10	441.0
144.80	159.0
147.00	690.0
148.00	1571.0
149.10	277.0
151.10	252.0
153.10	426.0
154.20	157.0
155.10	952.0
156.00	1367.0
156.90	284.0
157.80	396.0
158.90	218.0
159.90	629.0
161.00	533.0
162.00	267.0
163.20	185.0
165.10	402.0
166.10	547.0
167.10	2224.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

168.10	1124.0
168.90	340.0
171.00	192.0
172.00	263.0
173.00	476.0
174.00	760.0
175.00	944.0
177.00	690.0
179.00	2395.0
180.00	1605.0
181.20	464.0
182.10	261.0
185.00	1335.0
186.00	6806.0
187.10	2659.0
188.00	253.0
189.00	504.0
190.10	153.0
191.00	421.0
192.00	754.0
193.00	634.0
194.00	272.0
196.10	1640.0
196.60	715.0
198.00	54032.0
199.00	4481.0
200.10	471.0
201.80	226.0
203.00	348.0
204.10	1546.0
205.00	4368.0
206.00	15221.0
207.10	2539.0
207.90	495.0
209.00	381.0
210.30	236.0
211.10	875.0
213.00	281.0
216.10	502.0
217.00	3873.0
218.00	468.0
221.10	2044.0
222.00	816.0
223.00	725.0
224.00	8574.0
225.00	2021.0
226.00	269.0
227.00	3883.0
227.90	294.0
228.90	709.0
230.00	185.0
231.10	605.0
234.00	168.0
234.90	244.0
236.00	380.0
237.00	284.0
239.00	267.0
240.10	165.0
241.00	369.0
241.90	485.0
243.20	518.0
244.00	5606.0
245.00	560.0
246.10	1114.0
246.90	325.0
248.90	583.0
251.00	158.0
253.20	304.0
254.00	523.0
255.00	36760.0
256.00	5673.0
256.90	511.0
258.00	1435.0
258.90	475.0
260.90	155.0
264.10	153.0
265.00	903.0
271.00	270.0
273.00	833.0
274.00	2865.0
275.00	14255.0
276.00	1628.0
277.00	1294.0
277.90	241.0
280.90	151.0
283.20	253.0
283.80	160.0
285.00	315.0
286.90	212.0
288.00	203.0
292.10	152.0
292.90	331.0
296.00	4157.0
297.00	839.0
303.00	552.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

314.20	404.0
314.90	218.0
315.90	173.0
323.10	2007.0
324.20	273.0
327.00	273.0
327.90	278.0
332.10	296.0
333.20	307.0
334.10	903.0
335.10	225.0
341.10	255.0
345.90	310.0
352.00	280.0
353.10	226.0
354.00	319.0
365.00	1550.0
365.90	255.0
372.10	887.0
373.20	151.0
383.10	215.0
401.90	162.0
403.10	492.0
404.00	291.0
420.90	538.0
422.10	374.0
423.00	2625.0
424.00	322.0
441.00	6145.0
442.00	36408.0
443.00	7029.0
444.00	553.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

m/z	Abundance
38.10	171.0
39.10	1126.0
40.00	2080.0
41.10	412.0
44.00	469.0
49.00	220.0
50.10	6130.0
51.10	25344.0
52.10	1244.0
55.10	321.0
56.00	603.0
57.10	2004.0
61.00	405.0
62.10	425.0
63.00	661.0
65.10	463.0
67.20	235.0
69.00	24432.0
70.10	156.0
73.00	301.0
74.00	2463.0
75.10	3527.0
76.10	1160.0
77.10	24672.0
78.10	1376.0
79.00	1713.0
80.10	1285.0
81.00	1880.0
82.10	595.0
83.10	499.0
84.10	258.0
84.90	187.0
85.90	424.0
87.00	378.0
88.00	168.0
91.00	672.0
92.20	437.0
93.00	3203.0
94.10	323.0
96.10	164.0
98.10	2849.0
99.00	2095.0
100.00	181.0
101.00	1537.0
102.90	439.0
104.10	779.0
105.00	649.0
106.00	416.0
107.10	9337.0
108.10	1113.0
109.10	810.0
110.00	14656.0
111.10	2277.0
112.10	283.0
116.10	603.0
117.10	5243.0
117.90	458.0
119.10	179.0
120.00	220.0
121.20	187.0
121.90	468.0
123.10	1053.0
124.00	335.0
125.00	459.0
127.00	29008.0
128.10	3389.0
129.00	11055.0
130.00	1178.0
134.00	623.0
135.00	927.0
136.00	451.0
137.00	359.0
140.10	170.0
141.00	1416.0
142.10	705.0
143.00	664.0
146.10	303.0
147.00	676.0
148.00	1913.0
148.90	335.0
151.40	322.0
153.20	443.0
154.00	440.0
155.00	888.0
156.00	1584.0
157.00	423.0
158.10	465.0
160.20	599.0
161.00	1079.0
161.70	193.0
165.00	936.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

166.00	493.0
167.00	2672.0
167.90	1332.0
169.00	222.0
170.90	161.0
172.10	430.0
173.00	390.0
174.10	774.0
175.10	1371.0
175.90	237.0
177.00	769.0
178.00	228.0
179.00	2787.0
180.00	2288.0
181.00	961.0
182.10	265.0
184.20	223.0
185.10	1580.0
186.00	9668.0
187.00	2361.0
188.10	417.0
188.90	690.0
191.10	219.0
192.00	895.0
193.00	1257.0
193.60	255.0
194.10	242.0
195.00	240.0
196.00	1756.0
198.00	73464.0
198.90	4309.0
200.10	513.0
201.40	171.0
203.10	674.0
204.00	3130.0
205.00	4833.0
206.10	20432.0
207.00	2627.0
207.90	510.0
209.00	424.0
210.10	391.0
210.70	605.0
212.20	161.0
215.00	447.0
216.00	451.0
217.00	6225.0
218.00	579.0
219.10	173.0
221.00	2465.0
222.00	815.0
223.00	1046.0
224.00	11021.0
225.00	3168.0
226.10	329.0
227.00	5616.0
227.90	848.0
229.00	1212.0
230.00	234.0
230.90	191.0
234.00	337.0
235.10	438.0
235.90	158.0
236.90	439.0
240.10	200.0
241.00	338.0
242.00	844.0
243.10	528.0
244.00	7547.0
245.10	1207.0
246.10	1443.0
247.00	399.0
247.80	166.0
248.90	330.0
251.90	192.0
252.90	592.0
255.00	49624.0
256.00	6937.0
257.10	610.0
258.00	3204.0
259.00	485.0
265.00	1204.0
270.80	269.0
271.90	155.0
273.00	1602.0
274.00	3946.0
275.00	20840.0
276.10	3162.0
276.90	1634.0
278.10	510.0
280.90	162.0
283.10	317.0
284.10	297.0
285.00	370.0
289.00	219.0
293.00	466.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

294.10	232.0
296.00	5068.0
296.90	982.0
301.10	167.0
301.70	162.0
303.00	586.0
304.10	375.0
307.80	170.0
311.90	215.0
314.00	258.0
314.90	877.0
316.00	366.0
323.10	1509.0
324.00	465.0
327.00	450.0
332.00	325.0
334.10	1414.0
334.90	423.0
341.10	440.0
346.00	272.0
352.10	750.0
353.10	472.0
354.00	645.0
355.20	177.0
365.00	3130.0
366.00	558.0
371.10	278.0
372.00	1272.0
372.90	158.0
383.00	353.0
390.00	185.0
392.00	190.0
401.80	303.0
403.00	498.0
404.00	343.0
421.00	389.0
422.00	646.0
423.00	4001.0
424.00	539.0
425.10	218.0
433.60	207.0
441.00	10447.0
442.00	61304.0
443.00	11941.0
444.00	1199.0
544.50	183.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

m/z	Abundance
37.10	213.0
38.10	194.0
39.10	995.0
40.00	1949.0
44.00	513.0
49.20	199.0
50.10	4238.0
51.10	17128.0
52.10	1544.0
55.00	334.0
56.10	798.0
57.00	1363.0
61.20	288.0
62.10	242.0
63.10	897.0
64.10	173.0
65.10	442.0
67.20	175.0
67.90	399.0
69.00	16600.0
70.10	236.0
74.10	2555.0
75.00	2555.0
76.10	1110.0
77.10	20856.0
78.10	1199.0
79.10	1040.0
80.00	1463.0
81.10	1991.0
82.00	304.0
83.10	638.0
85.00	315.0
85.90	474.0
87.00	236.0
91.00	364.0
92.10	456.0
93.10	2161.0
94.20	202.0
95.10	200.0
96.10	186.0
98.00	1785.0
99.00	1253.0
100.10	241.0
101.00	606.0
103.00	313.0
104.00	560.0
105.10	553.0
105.90	240.0
107.10	6759.0
108.10	1024.0
108.90	204.0
110.00	11291.0
111.00	1408.0
112.10	332.0
116.00	635.0
117.00	5006.0
117.90	277.0
121.50	171.0
122.10	394.0
123.10	532.0
124.20	338.0
125.00	469.0
127.10	21824.0
128.10	1935.0
129.00	7233.0
130.00	724.0
134.00	258.0
135.00	1237.0
136.20	494.0
137.00	242.0
139.90	165.0
141.00	1281.0
142.00	566.0
142.80	484.0
145.00	162.0
146.10	293.0
147.10	634.0
148.00	1356.0
149.20	324.0
150.00	154.0
153.10	579.0
154.00	434.0
155.00	560.0
156.00	1382.0
157.10	329.0
158.10	150.0
159.90	348.0
160.90	580.0
164.00	254.0
165.00	397.0
166.00	659.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

167.10	2114.0
168.00	1018.0
169.00	242.0
169.90	206.0
171.30	165.0
171.80	309.0
172.90	411.0
173.90	593.0
175.00	1098.0
175.90	295.0
176.90	469.0
178.10	220.0
179.00	1741.0
180.00	1582.0
181.00	823.0
183.00	228.0
184.00	335.0
185.10	1023.0
186.10	7576.0
187.10	1938.0
188.10	350.0
188.90	506.0
191.00	360.0
192.00	867.0
192.90	929.0
194.10	215.0
195.10	175.0
196.10	1706.0
196.80	851.0
198.00	60904.0
199.00	3931.0
200.00	472.0
200.90	261.0
201.10	250.0
201.40	272.0
202.90	602.0
204.00	2484.0
205.00	3378.0
206.10	16472.0
207.00	2414.0
208.00	511.0
210.00	306.0
211.00	390.0
213.10	208.0
215.10	173.0
216.10	395.0
217.00	4774.0
218.00	750.0
221.00	2107.0
222.00	592.0
223.00	954.0
224.00	9063.0
225.00	2681.0
226.10	170.0
227.00	4225.0
227.90	878.0
229.00	1109.0
231.00	375.0
232.90	155.0
233.80	189.0
235.00	296.0
236.20	223.0
236.90	452.0
240.10	162.0
241.00	365.0
242.00	463.0
242.60	300.0
243.20	551.0
244.00	7201.0
245.00	728.0
246.00	1753.0
247.00	396.0
249.10	325.0
251.00	191.0
253.10	249.0
255.00	42304.0
256.00	5615.0
256.90	317.0
258.00	2237.0
259.10	859.0
264.00	185.0
265.00	862.0
272.30	167.0
273.00	801.0
274.00	2874.0
275.00	17048.0
276.00	3272.0
277.00	1565.0
278.10	315.0
285.00	311.0
290.00	210.0
293.00	308.0
296.00	5639.0
296.90	668.0
303.10	727.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

304.00	220.0
310.00	228.0
314.00	315.0
315.10	523.0
316.10	240.0
320.90	208.0
323.00	1858.0
324.10	377.0
327.00	586.0
328.10	285.0
333.20	210.0
334.00	1144.0
335.10	378.0
340.90	333.0
346.10	333.0
350.20	200.0
352.10	386.0
353.00	708.0
354.00	641.0
365.00	2892.0
366.00	319.0
371.00	342.0
372.00	967.0
382.90	602.0
388.60	250.0
399.30	168.0
402.10	253.0
403.00	528.0
421.00	503.0
422.20	506.0
423.10	3529.0
424.00	776.0
441.00	8763.0
442.00	62288.0
443.00	11832.0
444.10	818.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71

m/z	Abundance
40.00	2129.0
41.10	267.0
44.00	427.0
48.90	194.0
57.20	316.0
69.00	207.0
73.10	160.0
81.10	339.0
84.90	160.0
95.00	181.0
97.00	190.0
98.00	186.0
103.20	150.0
105.00	167.0
143.00	206.0
207.00	228.0
207.90	153.0
311.40	171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP71