



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082621\
 Data File : BG049975.D
 Acq On : 26 Aug 2021 10:38
 Operator : CG/JU
 Sample : DFTPP416
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP416

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG081921.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 30 14:59:55 2021



AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2508

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	29485	PASS
68	69	0.00	2	1.2	421	PASS
69	198	0.00	100	42.3	34182	PASS
70	69	0.00	2	0.7	242	PASS
127	198	10	80	50.8	41061	PASS
197	198	0.00	2	0.5	429	PASS
198	198	100	100	100.0	80816	PASS
199	198	5	9	6.6	5373	PASS
275	198	10	60	33.6	27136	PASS
365	198	1	100	5.6	4509	PASS
441	198	0.01	100	15.0	12099	PASS
442	198	50	100	93.2	75315	PASS
443	442	15	24	19.1	14371	PASS

m/z	Abundance
39.95	172.0
44.35	199.0
69.15	215.0
73.25	320.0
77.25	178.0
91.15	163.0
129.15	176.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

m/z	Abundance
37.05	156.0
38.15	242.0
39.15	2357.0
40.15	206.0
44.15	250.0
50.15	7035.0
51.15	26920.0
52.15	1344.0
55.15	519.0
56.05	958.0
57.15	2037.0
61.05	211.0
62.15	667.0
63.15	1128.0
64.15	329.0
65.15	787.0
68.15	673.0
69.15	29640.0
70.25	180.0
73.05	441.0
74.15	3302.0
75.15	4785.0
76.15	2147.0
77.15	32216.0
78.15	2270.0
79.15	2866.0
80.05	2134.0
81.15	3011.0
82.15	771.0
83.05	560.0
85.25	492.0
86.05	529.0
87.05	421.0
91.15	775.0
92.15	680.0
93.15	4124.0
94.15	225.0
95.95	163.0
98.15	3838.0
99.15	2779.0
100.05	209.0
101.15	1513.0
103.15	512.0
104.15	974.0
105.15	994.0
107.15	11828.0
108.15	1932.0
108.95	242.0
110.15	18744.0
111.15	3287.0
112.05	509.0
117.15	9689.0
118.15	801.0
119.85	189.0
122.05	518.0
123.05	1319.0
124.15	410.0
125.15	675.0
125.85	194.0
127.15	32832.0
128.15	2708.0
129.15	15002.0
130.15	1498.0
131.15	257.0
132.15	165.0
133.05	169.0
134.05	892.0
135.05	1352.0
136.15	630.0
137.25	768.0
138.15	250.0
141.15	1982.0
142.15	322.0
143.15	449.0
144.05	152.0
146.05	233.0
147.05	956.0
148.05	2479.0
149.05	448.0
151.05	249.0
152.15	231.0
153.15	689.0
154.15	324.0
155.15	1210.0
156.25	1824.0
157.15	370.0
157.95	249.0
159.05	610.0
160.15	480.0
161.15	923.0
162.25	192.0
165.05	1070.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

166.15	599.0
167.15	3560.0
168.05	1600.0
169.25	151.0
169.95	158.0
171.25	209.0
172.05	441.0
173.15	590.0
174.15	946.0
175.25	1997.0
176.15	237.0
177.15	495.0
178.15	414.0
179.15	2539.0
180.15	2412.0
181.15	815.0
184.25	193.0
185.25	1565.0
186.15	10712.0
187.15	3138.0
188.15	205.0
189.05	724.0
191.05	307.0
192.15	1116.0
193.15	850.0
194.25	368.0
196.15	2194.0
196.85	452.0
198.15	61592.0
199.15	4874.0
200.15	604.0
201.55	275.0
203.05	478.0
204.15	2907.0
205.15	5064.0
206.15	19496.0
207.25	2584.0
208.15	1016.0
210.25	538.0
210.95	491.0
215.25	454.0
216.15	713.0
217.15	6205.0
218.15	828.0
221.15	2169.0
221.95	638.0
223.05	1011.0
224.15	12169.0
225.25	3276.0
226.25	269.0
227.15	5040.0
228.05	753.0
229.05	1170.0
231.25	425.0
234.25	285.0
235.15	380.0
236.05	231.0
237.05	398.0
239.25	163.0
240.15	349.0
240.95	273.0
242.15	299.0
243.05	854.0
244.15	7963.0
245.15	1092.0
246.15	2007.0
247.15	309.0
248.15	243.0
249.25	360.0
252.15	154.0
252.65	150.0
253.55	367.0
255.15	47168.0
256.15	7087.0
257.15	549.0
258.15	3260.0
259.15	352.0
265.15	1421.0
266.05	194.0
272.15	239.0
273.15	1477.0
274.25	3737.0
275.15	18408.0
276.25	2276.0
277.15	2204.0
278.05	472.0
281.15	213.0
283.05	312.0
285.25	477.0
290.95	172.0
293.25	327.0
296.15	6582.0
297.15	868.0
303.25	505.0
304.05	220.0
313.85	162.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

314.25	187.0
315.25	516.0
316.15	648.0
323.15	2128.0
324.15	298.0
327.25	337.0
327.95	275.0
334.25	1050.0
335.25	408.0
346.25	538.0
347.15	187.0
352.05	433.0
353.25	517.0
354.15	405.0
365.15	2834.0
366.15	308.0
372.05	1021.0
383.25	367.0
390.25	177.0
391.05	158.0
402.05	333.0
403.25	598.0
421.25	366.0
422.05	546.0
423.15	2604.0
424.25	583.0
440.05	176.0
441.25	8040.0
442.25	44688.0
443.15	9536.0
444.05	641.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

m/z	Abundance
38.05	295.0
39.15	2761.0
40.15	375.0
41.25	183.0
50.15	8752.0
51.15	33560.0
52.15	1954.0
55.25	331.0
56.15	906.0
57.15	2676.0
58.25	168.0
61.25	441.0
62.15	514.0
63.15	1998.0
64.15	204.0
65.25	976.0
69.15	41784.0
70.15	286.0
71.15	199.0
73.05	444.0
74.15	3987.0
75.15	7837.0
76.15	2427.0
77.15	43712.0
78.25	3323.0
79.15	3985.0
80.15	2486.0
81.15	3089.0
82.15	1126.0
83.15	1055.0
84.05	206.0
85.15	659.0
86.15	980.0
87.25	640.0
91.15	486.0
92.15	1009.0
93.15	6994.0
94.05	523.0
95.05	244.0
96.05	261.0
98.15	5120.0
99.15	3871.0
100.05	321.0
101.05	1728.0
102.15	222.0
103.05	625.0
104.15	1202.0
105.15	1002.0
107.15	16174.0
108.15	2602.0
110.15	24888.0
111.15	3637.0
112.15	483.0
116.25	883.0
117.15	13052.0
118.05	980.0
118.95	217.0
121.15	213.0
122.15	1168.0
123.15	1568.0
124.15	1038.0
125.15	792.0
127.15	49688.0
128.15	3331.0
129.15	22256.0
130.15	1937.0
131.15	256.0
134.05	464.0
135.15	1675.0
136.15	561.0
137.15	918.0
138.35	184.0
140.05	364.0
141.15	2825.0
142.15	739.0
143.15	978.0
144.15	254.0
146.05	404.0
147.15	1525.0
148.15	3155.0
149.05	452.0
150.25	254.0
151.25	435.0
152.25	193.0
153.05	1163.0
154.15	647.0
155.15	1691.0
156.25	1958.0
157.15	805.0
158.05	475.0
159.15	535.0
160.15	918.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

161.15	1291.0
161.85	486.0
164.25	204.0
165.05	1206.0
166.05	738.0
167.15	4993.0
168.15	2320.0
169.15	438.0
170.25	246.0
171.15	360.0
172.25	292.0
173.15	558.0
174.05	1227.0
175.15	2258.0
176.15	406.0
177.05	1040.0
178.05	409.0
179.15	4769.0
180.25	3156.0
181.15	1416.0
182.25	347.0
184.15	435.0
185.15	2438.0
186.15	14868.0
187.25	4777.0
188.15	357.0
189.15	1385.0
189.95	200.0
191.15	452.0
192.15	1499.0
193.15	1409.0
194.15	329.0
194.75	191.0
196.15	2799.0
196.75	836.0
198.15	94400.0
199.15	5933.0
200.15	886.0
201.85	239.0
203.15	1126.0
204.15	4605.0
205.15	6923.0
206.15	29256.0
207.15	4532.0
208.25	1103.0
209.15	525.0
210.05	624.0
211.25	1061.0
211.85	285.0
213.05	270.0
215.25	501.0
216.25	772.0
217.15	8727.0
218.15	1537.0
219.25	219.0
221.05	3412.0
221.95	855.0
223.15	1763.0
224.25	19216.0
225.15	4323.0
226.25	429.0
227.15	8053.0
228.25	1132.0
229.15	1823.0
230.15	244.0
231.15	744.0
233.25	319.0
234.15	615.0
235.25	753.0
235.95	341.0
237.15	800.0
238.15	197.0
239.25	574.0
240.05	333.0
241.15	444.0
242.15	1013.0
243.25	1093.0
244.15	13278.0
245.25	1692.0
246.15	3185.0
247.05	624.0
247.95	178.0
249.15	543.0
250.05	174.0
250.95	253.0
252.05	264.0
253.35	632.0
255.15	72960.0
256.15	11048.0
257.25	1171.0
258.15	4645.0
259.05	695.0
263.95	220.0
265.05	2054.0
266.15	434.0
270.15	233.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

271.35	186.0
272.05	327.0
273.15	1928.0
274.15	5837.0
275.15	33088.0
276.15	4331.0
277.15	3316.0
278.15	565.0
283.15	343.0
284.15	326.0
285.05	643.0
290.15	162.0
292.05	158.0
293.15	578.0
294.15	378.0
295.05	257.0
296.15	11228.0
297.05	2030.0
302.15	229.0
303.15	1070.0
310.25	159.0
314.05	667.0
315.15	1520.0
316.15	536.0
320.95	269.0
321.35	201.0
323.15	2997.0
324.15	500.0
326.95	735.0
328.15	633.0
332.05	246.0
333.15	389.0
334.15	1837.0
335.15	819.0
336.25	226.0
341.25	436.0
346.05	949.0
352.15	921.0
353.05	455.0
353.35	476.0
354.15	1019.0
354.95	164.0
355.35	173.0
365.15	5351.0
366.15	571.0
370.05	296.0
371.15	253.0
372.15	1143.0
373.25	246.0
383.05	492.0
384.05	150.0
390.05	214.0
391.45	154.0
392.05	234.0
402.15	625.0
403.15	1137.0
404.15	246.0
421.05	751.0
422.35	534.0
423.15	4347.0
424.05	843.0
441.15	13217.0
442.15	85240.0
443.25	15794.0
444.25	1717.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

m/z	Abundance
37.05	206.0
38.25	319.0
39.15	2198.0
41.25	283.0
44.05	369.0
49.05	215.0
50.15	7595.0
51.15	27976.0
52.15	1510.0
55.05	265.0
56.05	1210.0
57.15	1612.0
61.15	469.0
62.25	629.0
63.15	1355.0
64.15	194.0
65.15	771.0
68.25	591.0
69.15	31768.0
70.15	259.0
73.05	327.0
74.15	3871.0
75.15	5604.0
76.15	2053.0
77.15	33368.0
78.15	2185.0
79.15	2711.0
80.15	1842.0
81.15	2866.0
82.15	991.0
83.15	454.0
84.95	482.0
86.05	770.0
87.15	374.0
88.05	158.0
91.15	831.0
92.05	915.0
93.15	5149.0
94.15	352.0
96.25	203.0
98.15	2955.0
99.15	2461.0
100.15	199.0
101.15	2126.0
102.15	207.0
103.15	613.0
104.15	1307.0
105.05	1055.0
106.15	363.0
107.15	13303.0
108.15	2293.0
109.15	553.0
110.15	19872.0
111.15	3277.0
112.05	297.0
113.25	254.0
116.15	704.0
117.15	10180.0
118.15	884.0
119.15	215.0
120.25	213.0
122.15	867.0
123.15	1451.0
124.15	799.0
125.15	886.0
127.15	40664.0
128.15	3328.0
129.15	18224.0
130.15	1601.0
130.95	247.0
132.15	165.0
134.25	459.0
135.05	1315.0
136.05	501.0
137.15	732.0
138.85	184.0
140.15	336.0
141.15	2263.0
142.15	758.0
143.05	789.0
146.15	511.0
147.15	1632.0
148.15	2719.0
149.15	658.0
150.15	187.0
151.35	242.0
153.05	1033.0
154.15	607.0
155.25	1916.0
156.15	1965.0
157.25	202.0
158.25	311.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

159.05	572.0
160.15	1033.0
161.15	1116.0
162.05	358.0
165.15	1148.0
166.15	849.0
167.15	4814.0
168.15	2089.0
169.15	426.0
171.05	217.0
172.15	477.0
173.15	705.0
174.15	888.0
175.15	2134.0
176.15	518.0
177.25	1127.0
178.25	357.0
179.15	3582.0
180.15	2918.0
181.05	1441.0
182.55	191.0
183.05	261.0
184.15	417.0
185.15	1868.0
186.15	13753.0
187.15	4187.0
188.05	412.0
189.05	1276.0
190.25	192.0
191.05	706.0
192.05	1423.0
193.15	1262.0
194.25	255.0
194.85	177.0
196.15	1898.0
198.15	86456.0
199.15	5311.0
200.15	249.0
201.55	363.0
202.15	200.0
203.15	1028.0
204.15	3731.0
205.15	5621.0
206.15	26800.0
207.15	3897.0
208.15	1013.0
209.15	434.0
210.25	424.0
211.15	965.0
213.05	202.0
215.25	404.0
216.15	539.0
217.15	8779.0
218.15	1069.0
220.15	296.0
221.15	2693.0
222.05	960.0
223.25	1706.0
224.25	15081.0
225.15	4436.0
226.05	385.0
227.15	8689.0
228.15	1480.0
229.15	1324.0
230.15	252.0
231.15	634.0
233.15	308.0
234.05	862.0
235.05	571.0
236.05	459.0
237.15	407.0
239.15	469.0
240.15	438.0
241.15	360.0
242.25	979.0
243.05	1025.0
244.25	11496.0
245.15	1743.0
246.25	2611.0
247.25	597.0
249.05	386.0
252.15	173.0
252.85	340.0
253.35	278.0
254.05	726.0
255.15	69696.0
256.15	10890.0
257.15	749.0
258.15	5142.0
259.15	835.0
263.85	152.0
265.15	1966.0
272.25	406.0
273.15	2197.0
274.15	5035.0
275.15	29912.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416

276.15	4813.0
277.15	2998.0
278.05	395.0
283.05	377.0
284.15	211.0
285.25	672.0
289.05	157.0
292.25	174.0
293.15	627.0
293.95	261.0
294.95	258.0
296.15	11088.0
297.25	1429.0
301.15	183.0
302.05	239.0
303.25	1398.0
304.25	264.0
305.15	158.0
314.25	639.0
315.15	1189.0
316.25	772.0
321.05	480.0
322.35	306.0
323.25	3475.0
324.25	612.0
327.05	719.0
328.05	444.0
332.25	426.0
333.05	433.0
334.15	1823.0
335.15	487.0
336.15	187.0
341.15	412.0
342.05	295.0
346.15	653.0
352.15	1265.0
353.15	708.0
354.05	1203.0
355.05	314.0
365.15	5341.0
366.15	730.0
371.15	412.0
372.15	1960.0
373.15	328.0
383.15	719.0
384.35	151.0
390.05	362.0
402.15	755.0
403.15	1179.0
404.15	357.0
421.15	1115.0
422.15	982.0
423.15	4592.0
424.15	1298.0
425.05	169.0
426.15	163.0
439.05	155.0
441.25	15040.0
442.15	96016.0
443.15	17784.0
444.15	1656.0
445.05	236.0

Instrument :
BNA_G
ClientSampleId :
DFTPP416