

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082622\
 Data File : BG054749.D
 Acq On : 26 Aug 2022 9:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 02:06:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 17:50:55 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.00
2	1,4-Dioxane	40.000	38.980	2.6	109	0.00
3	Pyridine	40.000	39.893	0.3	108	0.00
4	n-Nitrosodimethylamine	40.000	38.839	2.9	105	0.00
5 S	2-Fluorophenol	80.000	78.749	1.6	106	0.00
6	Aniline	40.000	39.571	1.1	107	0.00
7 S	Phenol-d6	80.000	79.309	0.9	107	0.00
8	2-Chlorophenol	40.000	39.110	2.2	106	0.00
9	Benzaldehyde	40.000	36.107	9.7	107	0.00
10 C	Phenol	40.000	39.502	1.2	106	0.00
11	bis(2-Chloroethyl)ether	40.000	39.647	0.9	108	0.00
12	1,3-Dichlorobenzene	40.000	39.106	2.2	108	0.00
13 C	1,4-Dichlorobenzene	40.000	39.090	2.3	107	0.00
14	1,2-Dichlorobenzene	40.000	38.575	3.6	106	0.00
15	Benzyl Alcohol	40.000	40.120	-0.3	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.463	3.8	105	0.00
17	2-Methylphenol	40.000	39.894	0.3	106	0.00
18	Hexachloroethane	40.000	38.791	3.0	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.631	3.4	104	0.00
20	3+4-Methylphenols	40.000	39.721	0.7	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	38.216	4.5	106	0.00
23 S	Nitrobenzene-d5	80.000	76.414	4.5	105	0.00
24	Nitrobenzene	40.000	38.152	4.6	105	0.00
25	Isophorone	40.000	37.994	5.0	103	0.00
26 C	2-Nitrophenol	40.000	39.527	1.2	106	0.00
27	2,4-Dimethylphenol	40.000	38.387	4.0	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.051	4.9	104	0.00
29 C	2,4-Dichlorophenol	40.000	39.141	2.1	105	0.00
30	1,2,4-Trichlorobenzene	40.000	38.400	4.0	106	0.00
31	Naphthalene	40.000	38.350	4.1	105	0.00
32	Benzoic acid	40.000	35.560	11.1	100	0.00
33	4-Chloroaniline	40.000	38.657	3.4	105	0.00
34 C	Hexachlorobutadiene	40.000	38.427	3.9	107	0.00
35	Caprolactam	40.000	38.114	4.7	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.255	1.9	106	0.00
37	2-Methylnaphthalene	40.000	38.634	3.4	105	0.00
38	1-Methylnaphthalene	40.000	38.188	4.5	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.458	3.9	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.465	3.8	102	0.00
42 S	2,4,6-Tribromophenol	80.000	77.790	2.8	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.818	3.0	102	0.00
44	2,4,5-Trichlorophenol	40.000	39.129	2.2	104	0.00
45 S	2-Fluorobiphenyl	80.000	76.955	3.8	105	0.00
46	1,1'-Biphenyl	40.000	37.993	5.0	103	0.00
47	2-Chloronaphthalene	40.000	38.015	5.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082622\
 Data File : BG054749.D
 Acq On : 26 Aug 2022 9:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 02:06:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 17:50:55 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.718	0.7	105	0.00
49	Acenaphthylene	40.000	38.651	3.4	104	0.00
50	Dimethylphthalate	40.000	37.843	5.4	103	0.00
51	2,6-Dinitrotoluene	40.000	38.739	3.2	101	0.00
52 C	Acenaphthene	40.000	38.443	3.9	103	0.00
53	3-Nitroaniline	40.000	39.342	1.6	103	0.00
54 P	2,4-Dinitrophenol	40.000	35.906	10.2	97	0.00
55	Dibenzofuran	40.000	38.446	3.9	104	0.00
56 P	4-Nitrophenol	40.000	39.186	2.0	99	0.00
57	2,4-Dinitrotoluene	40.000	40.107	-0.3	104	0.00
58	Fluorene	40.000	38.272	4.3	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.870	2.8	104	0.00
60	Diethylphthalate	40.000	37.902	5.2	102	0.00
61	4-Chlorophenyl-phenylether	40.000	38.263	4.3	104	0.00
62	4-Nitroaniline	40.000	39.207	2.0	102	0.00
63	Azobenzene	40.000	38.042	4.9	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.116	-0.3	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.821	2.9	104	0.00
67	4-Bromophenyl-phenylether	40.000	39.037	2.4	103	0.00
68	Hexachlorobenzene	40.000	39.000	2.5	103	0.00
69	Atrazine	40.000	39.259	1.9	102	0.00
70 C	Pentachlorophenol	40.000	39.114	2.2	99	0.00
71	Phenanthrene	40.000	38.472	3.8	103	0.00
72	Anthracene	40.000	38.605	3.5	102	0.00
73	Carbazole	40.000	38.717	3.2	103	0.00
74	Di-n-butylphthalate	40.000	38.518	3.7	101	0.00
75 C	Fluoranthene	40.000	38.784	3.0	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	45.910	-14.8	112	0.00
78	Pyrene	40.000	38.991	2.5	102	0.00
79 S	Terphenyl-d14	80.000	77.027	3.7	102	0.00
80	Butylbenzylphthalate	40.000	37.913	5.2	101	0.00
81	Benzo(a)anthracene	40.000	38.559	3.6	104	0.00
82	3,3'-Dichlorobenzidine	40.000	39.749	0.6	104	0.00
83	Chrysene	40.000	38.447	3.9	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.951	5.1	102	0.00
85 c	Di-n-octyl phthalate	40.000	38.067	4.8	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	38.739	3.2	104	-0.01
88	Benzo(b)fluoranthene	40.000	38.725	3.2	103	0.00
89	Benzo(k)fluoranthene	40.000	38.290	4.3	104	-0.01
90 C	Benzo(a)pyrene	40.000	38.524	3.7	103	-0.01
91	Dibenzo(a,h)anthracene	40.000	38.737	3.2	104	-0.02
92	Benzo(g,h,i)perylene	40.000	38.427	3.9	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082622\
Data File : BG054749.D
Acq On : 26 Aug 2022 9:52
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 27 02:06:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 17:50:55 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0