

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082815\
 Data File : BG018514.D
 Acq On : 28 Aug 2015 14:35
 Operator : UM/MA
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02015

Quant Time: Aug 31 01:52:29 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 31 01:23:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.485	0.450	7.2	108	0.00
3 S	1,4-Dioxane-d8	0.453	0.391	13.7	99	0.00
4	Benzaldehyde	1.065	1.224	-14.9	116	0.00
5 S	Phenol-d5	1.815	1.872	-3.1	118	0.00
6	Phenol	1.852	1.896	-2.4	117	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.125	1.132	-0.6	115	0.00
8	Bis(2-Chloroethyl)ether	1.424	1.450	-1.8	113	0.00
9 S	2-Chlorophenol-d4	1.390	1.335	4.0	112	0.00
10	2-Chlorophenol	1.417	1.383	2.4	110	0.00
11	2-Methylphenol	1.436	1.444	-0.6	113	0.00
12	2,2'-oxybis(1-Chloropropane	2.286	1.871	18.2	91	0.00
13 S	4-Methylphenol-d8	1.525	1.542	-1.1	116	0.00
14	Acetophenone	2.203	2.339	-6.2	117	0.00
15 P	N-Nitroso-di-n-propylamine	1.264	1.246	1.4	113	0.00
16	4-Methylphenol	1.567	1.629	-4.0	120	0.00
17	Hexachloroethane	0.611	0.554	9.3	106	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
19 S	Nitrobenzene-d5	0.156	0.162	-3.8	121	0.00
20	Nitrobenzene	0.393	0.388	1.3	113	0.00
21	Isophorone	0.757	0.757	0.0	115	0.00
22 S	2-Nitrophenol-d4	0.159	0.164	-3.1	116	0.00
23 C	2-Nitrophenol	0.176	0.180	-2.3	117	0.00
24	2,4-Dimethylphenol	0.395	0.380	3.8	113	0.00
25	Bis(2-Chloroethoxy)methane	0.472	0.468	0.8	113	0.00
26 S	2,4-Dichlorophenol-d3	0.337	0.349	-3.6	121	0.00
27 C	2,4-Dichlorophenol	0.316	0.325	-2.8	119	0.00
28	Naphthalene	1.057	1.077	-1.9	116	0.00
29 S	4-Chloroaniline-d4	0.381	0.466	-22.3#	119	0.00
30	4-Chloroaniline	0.387	0.454	-17.3	114	0.00
31 C	Hexachlorobutadiene	0.199	0.189	5.0	110	0.00
32	Caprolactam	0.127	0.139	-9.4	125	0.01
33 C	4-Chloro-3-methylphenol	0.366	0.366	0.0	117	0.00
34	2-Methylnaphthalene	0.766	0.785	-2.5	117	0.00
35	1-Methylnaphthalene	0.748	0.790	-5.6	119	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
37	1,2,4,5-Tetrachlorobenzene	0.641	0.643	-0.3	122	0.00
38	Hexachlorocyclopentadiene	0.382	0.295	22.8#	92	0.00
39 C	2,4,6-Trichlorophenol	0.437	0.438	-0.2	120	0.00
40	2,4,5-Trichlorophenol	0.470	0.468	0.4	120	0.00
41	1,1'-Biphenyl	1.669	1.673	-0.2	120	0.00
42	2-Chloronaphthalene	1.297	1.272	1.9	119	0.00
43	2-Nitroaniline	0.419	0.394	6.0	116	0.00
44 S	Dimethylphthalate-d6	1.683	1.682	0.1	122	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082815\
 Data File : BG018514.D
 Acq On : 28 Aug 2015 14:35
 Operator : UM/MA
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02015

Quant Time: Aug 31 01:52:29 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 31 01:23:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.660	1.633	1.6	119	0.00
46	2,6-Dinitrotoluene	0.331	0.348	-5.1	126	0.00
47 S	Acenaphthylene-d8	2.119	2.143	-1.1	121	0.00
48	Acenaphthylene	2.125	2.209	-4.0	124	0.00
49	3-Nitroaniline	0.337	0.399	-18.4	135	0.00
50 C	Acenaphthene	1.491	1.508	-1.1	123	0.00
51	2,4-Dinitrophenol	0.161	0.144	10.6	120	0.00
52 S	4-Nitrophenol-d4	0.303	0.329	-8.6	136	0.00
53	4-Nitrophenol	0.263	0.243	7.6	110	0.00
54	Dibenzofuran	1.948	2.021	-3.7	124	0.00
55	2,4-Dinitrotoluene	0.472	0.501	-6.1	128	0.00
56	2,3,4,6-Tetrachlorophenol	0.409	0.419	-2.4	122	0.00
57	Diethylphthalate	1.764	1.718	2.6	120	0.00
58 S	Fluorene-d10	1.466	1.512	-3.1	125	0.00
59	Fluorene	1.676	1.736	-3.6	127	0.00
60	4-Chlorophenyl-phenylether	0.793	0.809	-2.0	122	0.00
61	4-Nitroaniline	0.372	0.434	-16.7	141	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.098	0.089	9.2	118	0.00
64	4,6-Dinitro-2-methylphenol	0.105	0.096	8.6	121	0.00
65	N-Nitrosodiphenylamine	0.602	0.594	1.3	126	0.00
66	4-Bromophenyl-phenylether	0.216	0.209	3.2	124	0.00
67	Hexachlorobenzene	0.229	0.233	-1.7	130	0.00
68	Atrazine	0.225	0.241	-7.1	129	0.00
69 C	Pentachlorophenol	0.153	0.138	9.8	116	0.00
70	Phenanthrene	1.085	1.117	-2.9	132	0.00
71 S	Anthracene-d10	0.957	0.981	-2.5	129	0.00
72	Anthracene	1.106	1.153	-4.2	131	0.00
73	1,2,3,4-Tetrachlorobenzene	0.277	0.270	2.5	123	0.00
74	Pentachlorobenzene	0.264	0.250	5.3	122	0.00
75	Carbazole	0.970	1.113	-14.7	135	0.00
76	Di-n-butylphthalate	1.266	1.301	-2.8	129	0.00
77 C	Fluoranthene	1.159	1.366	-17.9	134	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	139	0.00
79 S	Pyrene-d10	1.038	1.047	-0.9	132	0.00
80	Pyrene	1.352	1.353	-0.1	131	0.00
81	Butylbenzylphthalate	0.572	0.561	1.9	131	0.00
82	3,3'-Dichlorobenzidine	0.415	0.439	-5.8	127	0.00
83	Benzo(a)anthracene	1.240	1.251	-0.9	134	0.00
84	Bis(2-ethylhexyl)phthalate	0.808	0.802	0.7	132	0.00
85	Chrysene	1.159	1.138	1.8	131	0.00
86 I	Perylene-d12	1.000	1.000	0.0	139	0.01
87	Di-n-octyl phthalate	1.290	1.359	-5.3	129	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG082815\
Data File : BG018514.D
Acq On : 28 Aug 2015 14:35
Operator : UM/MA
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02015

Quant Time: Aug 31 01:52:29 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG080915.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 31 01:23:00 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.257	1.239	1.4	133	0.00
89	Benzo(k)fluoranthene	1.210	1.224	-1.2	137	0.00
90 S	Benzo(a)pyrene-d12	1.062	1.079	-1.6	137	0.00
91 C	Benzo(a)pyrene	1.195	1.191	0.3	133	0.00
92	Indeno(1,2,3-cd)pyrene	1.383	1.377	0.4	135	0.02
93	Dibenzo(a,h)anthracene	1.148	1.122	2.3	134	0.00
94	Benzo(g,h,i)perylene	1.161	1.144	1.5	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0