


Data Path : Z:\HPCHEM1\BNA_G\Data\BG082917\
 Data File : BG028539.D
 Acq On : 29 Aug 2017 15:42
 Operator : SJ/JU
 Sample : DFTPP73
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP73

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Wed Aug 30 01:19:11 2017



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	14840	PASS
68	69	0.00	2	1.9	261	PASS
69	198	0.00	100	33.4	13957	PASS
70	69	0.00	2	0.6	80	PASS
127	198	10	80	41.4	17294	PASS
197	198	0.00	2	0.6	258	PASS
198	198	100	100	100.0	41749	PASS
199	198	5	9	6.9	2895	PASS
275	198	10	60	30.5	12715	PASS
365	198	1	100	4.1	1721	PASS
441	442	0.01	100	17.7	6559	PASS
442	198	50	100	88.6	37008	PASS
443	442	15	24	20.8	7700	PASS

m/z	Abundance
38.10	261.0
39.10	871.0
41.20	169.0
43.10	285.0
44.10	285.0
49.00	404.0
50.10	4031.0
51.10	16984.0
52.10	798.0
55.20	397.0
56.10	710.0
57.10	1039.0
61.10	170.0
62.10	332.0
63.10	691.0
65.10	652.0
68.10	324.0
69.00	15515.0
70.10	242.0
73.10	352.0
74.00	1436.0
75.10	3216.0
76.10	765.0
77.10	16448.0
78.10	1089.0
79.10	1084.0
80.00	1042.0
81.10	1241.0
82.10	548.0
83.10	870.0
85.10	161.0
85.90	420.0
87.10	254.0
91.10	457.0
91.90	264.0
93.00	1452.0
94.00	251.0
95.10	195.0
96.00	264.0
97.10	367.0
98.00	1346.0
99.10	1408.0
101.10	910.0
103.00	246.0
104.10	389.0
105.10	632.0
107.10	6360.0
108.10	1235.0
108.80	202.0
110.00	8822.0
111.00	1456.0
112.00	279.0
116.00	399.0
117.10	3265.0
118.10	264.0
122.10	327.0
123.10	426.0
124.00	263.0
124.90	268.0
127.00	17784.0
128.00	1245.0
129.00	7652.0
130.00	569.0
131.00	327.0
133.00	255.0
134.00	295.0
135.00	774.0
135.90	486.0
137.00	267.0
141.10	942.0
142.10	468.0
143.00	246.0
147.10	512.0
148.00	1247.0
149.10	309.0
152.90	293.0
154.00	278.0
155.10	926.0
156.10	629.0
156.90	329.0
157.90	207.0
160.00	678.0
161.10	628.0
165.00	396.0
166.10	821.0
167.00	1678.0
168.00	609.0
169.00	414.0
170.00	266.0
171.90	283.0
173.00	311.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

174.00	513.0
175.10	716.0
177.00	601.0
179.00	1157.0
180.00	870.0
180.90	553.0
181.80	156.0
184.00	269.0
185.20	742.0
186.10	5327.0
187.00	1861.0
188.00	210.0
189.00	377.0
189.90	187.0
192.00	543.0
193.00	638.0
193.80	162.0
195.90	1358.0
197.00	776.0
198.00	41560.0
199.00	2708.0
201.10	190.0
203.10	356.0
204.00	1330.0
205.10	2599.0
206.00	11397.0
207.00	1662.0
208.00	522.0
209.00	197.0
210.20	157.0
211.00	627.0
212.00	177.0
216.00	181.0
217.00	3526.0
218.00	471.0
221.00	1788.0
221.90	561.0
223.10	806.0
224.00	6717.0
225.10	1809.0
226.20	258.0
227.00	3056.0
228.00	500.0
229.10	707.0
230.00	197.0
230.70	177.0
231.20	232.0
233.90	414.0
235.00	186.0
239.00	210.0
240.10	269.0
241.10	343.0
242.00	626.0
243.10	617.0
244.00	4503.0
245.00	718.0
246.00	1555.0
247.20	152.0
249.20	254.0
253.80	519.0
255.00	27296.0
256.00	3355.0
257.00	466.0
258.00	2046.0
259.00	211.0
265.10	590.0
265.90	193.0
271.90	295.0
273.10	869.0
274.00	1923.0
275.00	12420.0
276.00	1626.0
277.00	997.0
279.20	173.0
283.00	151.0
285.10	166.0
292.00	297.0
293.10	309.0
294.90	187.0
296.00	3274.0
296.90	515.0
303.00	422.0
314.10	210.0
315.00	490.0
315.90	294.0
323.10	1061.0
324.10	279.0
334.10	749.0
334.90	218.0
341.00	235.0
352.10	471.0
353.00	312.0
354.10	314.0
356.30	160.0
365.00	1377.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

366.00	180.0
372.00	509.0
383.10	193.0
391.10	167.0
402.10	253.0
421.00	204.0
422.10	321.0
423.00	1210.0
424.00	623.0
438.10	188.0
441.10	5411.0
442.00	28688.0
443.00	6222.0
444.00	529.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

m/z	Abundance
38.00	218.0
39.20	1432.0
39.90	225.0
43.00	222.0
44.00	747.0
50.10	4262.0
51.10	16720.0
52.00	636.0
56.00	523.0
57.10	1359.0
61.00	230.0
62.10	178.0
63.00	728.0
64.30	269.0
65.20	214.0
68.00	459.0
69.00	16880.0
73.10	171.0
74.00	1958.0
75.00	2987.0
76.10	611.0
77.10	20040.0
78.10	1092.0
79.10	1386.0
80.00	615.0
81.00	2051.0
82.10	417.0
83.10	346.0
83.90	224.0
85.10	397.0
86.10	173.0
91.10	465.0
92.10	579.0
93.10	2213.0
95.10	208.0
95.50	165.0
97.10	225.0
98.00	1996.0
99.10	1462.0
101.00	1028.0
103.10	387.0
104.10	799.0
105.10	605.0
106.00	329.0
107.00	5849.0
108.00	810.0
109.00	156.0
110.00	10932.0
111.00	2054.0
111.80	185.0
115.90	162.0
117.00	4288.0
117.90	484.0
122.00	194.0
123.00	736.0
124.00	438.0
125.00	472.0
127.10	20832.0
128.00	1641.0
129.00	8140.0
130.10	1141.0
134.00	430.0
135.00	455.0
135.90	277.0
137.10	310.0
138.90	225.0
141.00	1270.0
141.80	271.0
142.00	268.0
142.90	417.0
143.90	287.0
146.10	253.0
146.90	673.0
148.00	955.0
148.90	187.0
150.90	398.0
153.00	299.0
154.00	311.0
155.10	470.0
156.00	1188.0
157.90	419.0
158.90	462.0
160.10	433.0
161.00	614.0
163.20	218.0
164.90	547.0
166.20	414.0
167.00	1517.0
168.00	876.0
168.90	197.0
171.00	195.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

172.00	481.0
173.00	552.0
174.00	861.0
175.00	1151.0
176.10	199.0
177.00	392.0
178.10	241.0
179.10	1755.0
180.00	1448.0
181.00	316.0
183.10	285.0
184.10	220.0
184.90	696.0
186.10	6895.0
187.00	1982.0
188.10	259.0
189.00	695.0
191.10	244.0
192.00	300.0
193.00	459.0
194.10	179.0
195.10	287.0
196.00	1221.0
198.00	49312.0
199.00	3790.0
200.00	230.0
201.00	282.0
201.90	388.0
203.00	384.0
204.10	1661.0
205.10	3594.0
206.10	14606.0
207.10	2509.0
208.10	619.0
209.20	309.0
209.80	268.0
210.30	275.0
211.00	570.0
211.80	238.0
214.90	490.0
215.90	332.0
217.00	3212.0
218.00	656.0
218.90	202.0
221.00	2045.0
223.10	1384.0
224.00	7491.0
225.00	2299.0
226.10	392.0
227.00	3545.0
228.00	691.0
229.00	1178.0
230.00	162.0
230.80	207.0
231.10	188.0
232.70	243.0
234.00	399.0
234.80	205.0
236.00	414.0
237.00	388.0
238.90	250.0
241.00	235.0
242.10	454.0
243.10	430.0
244.10	5871.0
245.10	603.0
246.10	1152.0
247.00	288.0
249.00	286.0
252.90	360.0
255.00	34528.0
256.00	4676.0
257.10	431.0
258.00	1669.0
259.10	361.0
261.10	176.0
264.20	170.0
265.00	800.0
265.90	330.0
273.00	991.0
274.00	2527.0
275.00	14939.0
276.00	1474.0
276.90	853.0
277.80	182.0
283.00	192.0
284.00	269.0
285.00	472.0
293.10	230.0
294.00	336.0
296.00	4339.0
297.00	321.0
303.10	377.0
303.80	258.0
314.00	519.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

315.00	536.0
316.00	498.0
317.00	159.0
320.90	177.0
323.10	1202.0
324.00	195.0
327.10	410.0
333.10	248.0
333.90	861.0
335.00	311.0
341.10	204.0
345.90	208.0
352.00	443.0
353.00	532.0
354.00	543.0
358.80	200.0
365.00	1937.0
366.00	224.0
370.00	193.0
372.00	891.0
373.00	168.0
382.30	184.0
382.90	154.0
390.00	159.0
402.00	257.0
403.00	451.0
404.00	351.0
421.00	252.0
422.00	528.0
423.10	2655.0
424.00	662.0
441.10	6739.0
442.00	44304.0
443.00	9582.0
444.10	1011.0
444.90	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

m/z	Abundance
38.10	245.0
39.10	1032.0
40.00	264.0
43.10	220.0
44.10	683.0
49.10	163.0
50.10	2538.0
51.10	10818.0
52.10	738.0
55.10	384.0
56.10	264.0
57.10	844.0
61.00	152.0
63.10	604.0
63.70	171.0
64.80	265.0
69.00	10096.0
73.00	387.0
74.10	1141.0
75.00	1817.0
76.10	701.0
77.10	11028.0
78.00	555.0
79.10	798.0
80.00	630.0
81.10	749.0
82.10	171.0
83.00	406.0
86.00	191.0
91.10	451.0
91.90	266.0
93.10	1646.0
95.10	209.0
96.20	156.0
97.20	201.0
98.10	1212.0
99.00	1532.0
101.00	636.0
103.00	161.0
104.10	315.0
105.00	312.0
107.00	4134.0
108.00	496.0
108.90	212.0
110.00	5723.0
111.00	1103.0
112.10	335.0
113.20	250.0
115.90	215.0
117.00	2387.0
121.10	224.0
122.00	281.0
123.00	358.0
124.20	170.0
127.10	13267.0
128.00	1050.0
129.00	5549.0
130.10	486.0
133.10	264.0
134.00	234.0
135.00	665.0
136.10	158.0
136.90	276.0
137.70	259.0
141.00	896.0
142.00	208.0
143.00	266.0
147.10	350.0
148.10	705.0
149.10	176.0
151.20	327.0
153.10	417.0
154.00	381.0
155.10	454.0
156.00	540.0
157.00	181.0
158.10	224.0
159.00	386.0
160.00	357.0
161.00	338.0
161.90	192.0
165.00	312.0
166.00	318.0
167.00	1499.0
167.90	437.0
169.10	168.0
172.00	173.0
172.90	243.0
174.10	380.0
175.10	614.0
175.70	316.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

176.90	566.0
179.00	1715.0
180.10	754.0
181.10	476.0
184.00	171.0
185.00	309.0
186.00	4929.0
187.00	990.0
188.20	198.0
188.80	469.0
189.10	423.0
191.00	380.0
192.00	320.0
193.10	608.0
194.00	303.0
194.90	161.0
196.00	859.0
198.00	34376.0
199.00	2188.0
199.80	258.0
200.50	178.0
203.00	342.0
204.00	1687.0
205.00	2045.0
206.10	10502.0
207.10	1350.0
207.90	350.0
209.10	242.0
210.30	367.0
211.60	200.0
216.00	491.0
217.00	3217.0
218.10	195.0
221.00	1518.0
223.00	835.0
224.10	7455.0
225.00	1935.0
227.10	2741.0
228.10	395.0
229.00	684.0
231.10	461.0
232.70	150.0
233.90	195.0
234.90	406.0
237.00	349.0
239.10	277.0
241.10	357.0
242.00	341.0
243.00	692.0
244.10	4319.0
245.10	812.0
246.10	940.0
246.90	215.0
249.10	396.0
251.20	166.0
252.10	163.0
253.30	190.0
255.00	23840.0
256.00	3340.0
256.80	311.0
258.00	954.0
259.00	292.0
265.10	474.0
272.30	185.0
273.10	796.0
274.00	2138.0
275.00	10787.0
276.00	1682.0
277.00	1114.0
279.10	250.0
285.10	162.0
286.00	166.0
293.10	417.0
295.10	354.0
296.00	2815.0
297.10	427.0
302.20	151.0
303.10	416.0
303.90	173.0
315.00	311.0
316.00	402.0
318.10	157.0
322.00	286.0
323.10	1403.0
324.00	234.0
332.00	201.0
332.90	203.0
334.00	633.0
335.20	275.0
340.10	187.0
341.00	401.0
342.00	217.0
351.00	161.0
351.90	674.0
352.90	167.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

354.00	399.0
365.00	1851.0
366.10	175.0
372.00	261.0
383.10	178.0
402.00	252.0
403.10	500.0
404.20	196.0
421.00	408.0
421.90	392.0
423.00	1588.0
424.00	275.0
441.10	7528.0
442.00	38032.0
443.00	7297.0
444.00	820.0
493.30	177.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73

m/z	Abundance
41.00	194.0
42.90	234.0
44.10	485.0
49.10	187.0
55.10	275.0
69.10	206.0
76.80	155.0
95.20	351.0
96.90	181.0
107.10	206.0
108.90	163.0
125.00	151.0
132.90	167.0
147.20	174.0
149.00	271.0
166.80	154.0
167.50	204.0
171.20	218.0
181.10	154.0
189.00	165.0
206.90	165.0
261.20	160.0
289.30	160.0
295.00	163.0
339.10	262.0

Instrument :
BNA_G
ClientSampleId :
DFTPP73