

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082922\
 Data File : BG054767.D
 Acq On : 29 Aug 2022 10:43
 Operator : CG/JU
 Sample : PB147277BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB147277BL

Quant Time: Aug 30 02:44:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 17:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.148	152	52662	20.000	ng	0.00
21) Naphthalene-d8	10.974	136	212477	20.000	ng	0.00
39) Acenaphthene-d10	14.782	164	142867	20.000	ng	0.00
64) Phenanthrene-d10	17.525	188	333306	20.000	ng	0.00
76) Chrysene-d12	21.814	240	341486	20.000	ng	-0.01
86) Perylene-d12	25.181	264	346157	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.686	112	453911	150.094	ng	0.00
7) Phenol-d6	7.290	99	591204	142.947	ng	0.00
23) Nitrobenzene-d5	9.317	82	359853	88.908	ng	-0.01
42) 2,4,6-Tribromophenol	16.268	330	251313	140.775	ng	0.00
45) 2-Fluorobiphenyl	13.407	172	859283	90.400	ng	0.00
79) Terphenyl-d14	20.128	244	1562709	90.672	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082922\
Data File : BG054767.D
Acq On : 29 Aug 2022 10:43
Operator : CG/JU
Sample : PB147277BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB147277BL

Quant Time: Aug 30 02:44:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 17:50:55 2022
Response via : Initial Calibration

